

# Lawrence Berkeley National Laboratory

## Recent Work

### **Title**

COIL CALIBRATION OF 10 COILS - TEST RESULTS DRAWING NO. 19M5696

### **Permalink**

<https://escholarship.org/uc/item/4rq521tb>

### **Author**

Nelson, Donald H.

### **Publication Date**

1981-06-15

LBID-511  
c.1



# Lawrence Berkeley Laboratory

UNIVERSITY OF CALIFORNIA

Engineering & Technical  
Services Division

RECEIVED  
LAWRENCE  
BERKELEY LABORATORY

MAY 24 1982

LIBRARY AND  
DOCUMENTS SECTION

**For Reference**

Not to be taken from this room



LBID-511  
c.1

## **DISCLAIMER**

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

**ENGINEERING NOTE**MME Book  
No. 630BFILE NO.  
MT 296PAGE  
1 of 7

SUBJECT

Coil Calibration of 10 Coils - Test Results  
Drawing No. 19M5696

NAME

Donald H. Nelson

DATE

June 15, 1981

Introduction

On May 20, 1981, Ed Cyr and I calibrated (i.e., determined the area of) 10 coils. Six of these coils are to be installed in a six coil array to be used for harmonic measurements of a superconducting magnet Isabelle prototype scheduled for measurements starting August 1, 1981. On May 22nd, we made more accurate determinations of the relative area of coils 2 through 10 (relative to coil 1 arbitrarily selected as the standard for comparison). On May 26th, we demonstrated the measurement technique to M.I. Green, W.V. Hassenzahl and T. Lauritzen.

Results

Table I summarizes the areas as determined by method 1, May 20th. See Test Plan - MT 295.

Table II shows the results of the comparative measurements (method 2 of MT 295) made on May 22nd and May 26th. For comparison I have included relative areas based on the absolute (but less accurate) measurements represented by Table I data in Table II.

Based on these data, we selected coil pairs and spares as shown in Table III. The choice of which coil of each pair would be selected as the outer coil was left to Lauritzen's discretion. Since the outer radius coils are most sensitive to errors in fabrication, we recommended that coils with loose turns be positioned to the inside position. Engineering drawing no. 19M5696 shows the coil locations in the array.

Method of Measurements

Method 1: We followed the procedures described in the test plan. For method 1, we measured magnetic induction over the 45 cm occupied by the coil centerline when in place. Measurements made both before and after the absolute calibration of the coil are represented by Figure 1. I numerically integrated these equally spaced data then

**ENGINEERING NOTE**MME Book  
No. 630BFILE NO.  
MT 296PAGE  
2 of 7

SUBJECT

Coil Calibration of 10 Coils - Test Results  
Drawing No. 19M5696NAME  
Donald H. NelsonDATE  
June 15, 1981

Coil	Area (m <sup>2</sup> )	Area (rpt.) (m <sup>2</sup> )	$\Delta$ Area (%)	Dev (m <sup>2</sup> )	from Average (%)	Dev (m <sup>2</sup> )	of (rpt.) (%)
1	16.722			0.026	0.156		
2	16.682			-0.014	-0.084		
3	16.671	16.658	(-0.08%)	-0.025	-0.150	-0.038	-0.0230
4	16.678			-0.018	-0.108		
5	16.705			0.009	0.054		
6	16.693			-0.003	-0.018		
7	16.677			-0.019	-0.114		
8	16.706	16.704	(-0.01%)	0.010	0.060	0.008	0.048
9	16.691			0.005	-0.030		
10	16.735	16.722	(-0.08%)	0.039	0.234	0.026	0.156
AVG	16.696						

TABLE I Coil Area Summary (Method 1)

**ENGINEERING NOTE**

MME Book  
No. 630B

FILE NO.  
MT 296

PAGE  
3 of 7

SUBJECT

Coil Calibration of 10 Coils - Test Results  
Drawing No. 19M5696

NAME  
Donald H. Nelson

DATE  
June 15, 1981

Coil No.	HIGH PRECISION (RELATIVE AREAS)		ABSOLUTE AREA DATA		
	Deviation* (%)	R E P E A T		DEVIATION* USING TABLE I DATA	
		No. 1 as Ref	No. 8 as Ref	Main Data (%)	Repeat Data (%)
2	-0.27			-0.24	
3	-0.35			-0.30	-0.38
4	-0.33		-0.29	-0.26	
5	-0.11			-0.10	
6	-0.20			-0.17	
7	-0.30			-0.27	
8	-0.11		(Ref)	-0.10	-0.11
9	-0.24	-0.21	-0.22	-0.19	
10	0.03			0.08	0.00

\*Deviation with respect to coil no. 1

TABLE II Comparative Data With Comparisons to Absolute (Table I) Data

**ENGINEERING NOTE**MME Book  
No. 630BFILE NO.  
MT 296

PAGE

4 of 7

SUBJECT

Coil Calibration of 10 Coils - Test Results  
Drawing No. 19M5696

NAME

Donald H. Nelson

DATE

June 15, 1981

<u>Coil Pair</u>	<u>S e l e c t e d C o i l s</u>		<u>Spare</u>
A	4	7	3
B	5	8	10
C	6	9	2

TABLE III Pair Selection

LBID-511

MT 296  
 UNIFORMITY OF MAGNETIC FIELD  
 AEC NO. 135652  
 BEV BEAM-LINE 30  
 MAGNET M3

Figure 1  
 8 X 16 X 36 C MAGNET  
 $E_{SHUNT} = 20 MV$

COIL WIDTHS  
 3.5 cm

5 X [cm]

Y, DISTANCE FROM Z

$B_z(0, Y, 0)$   
 MAGNETIC INDUCTION  
 ( $10^3$  GAUSS)

~ 3 G/cm


~ 3 G/cm

SYMD.  
 X ⊙ + ⊠  
 Y ⊙ ⊙ ⊙ ⊙  
 Z ⊙ ⊙ ⊙ ⊙

10.950

X Before Calibration

⊙ After Calibration (2 MV/100V)





LAWRENCE BERKELEY LABORATORY - UNIVERSITY OF CALIFORNIA <b>ENGINEERING NOTE</b>	MME Book No. 630B	FILE NO. MT 296	PAGE 6 of 7
SUBJECT Coil Calibration of 10 Coils - Test Results Drawing No. 19M5696		NAME Donald H. Nelson	DATE June 15, 1981

divided by 45 cm to determine the average field over the coil volume. The HP 97 program used for performing the integration is saved in MME Data Book No. 630B. The before and after (calibration) determinations of average magnetic induction agreed to 0.02%.

We calibrated the electronic integrator with a flux standard and determined each coil area based on 1) a known flux linkage from the flux standard, and 2) a determination of  $B_{AVG}$  from the NMR data by solving equation 1.

$$nA = \frac{\psi_{SLFS}^{1/2} E_{Flip\ Coil}}{B_{AVG} E_{SLFS}} \quad (1)$$

$nA$  = turns area of coil under test ( $m^2$ )

$\psi_{SLFS}$  = flux linkage provided by the SLFS (Wb)

$B_{AVG}$  = average magnetic induction over the coil length (T)

$E_{Flip\ Coil}$  = output potential of integrator generated by flipping the coil in known magnetic induction ( $B_{AVG}$ ) (V)

$E_{SLFS}$  = output potential of integrator generated by a known calibration signal from the SLFS, i.e., due to  $\psi_{SLFS}$  (V)

Method 2: The second method described in the test plan is inherently a more accurate determination of coil area based on a known area reference coil. In selecting coil pairs, it is more important that the relative areas be known than the absolute, so we did not require knowledge of the area of the reference coil to high accuracy.

Method 2 is more accurate than method 1 because 1) there are fewer variables involved in the calibration, e.g., a knowledge of the magnitude of the field is unimportant, only that it doesn't change over the relatively short time of comparison, 2) only short term reproducibility of both the magnetic field and the measurement system gain are needed for method 2; whereas stability over the entire measurement period is required for method 1.

**ENGINEERING NOTE**MME Book  
No. 630B

FILE NO.

MT 296

PAGE

7 of 7

SUBJECT

NAME

DATE

June 15, 1981

An HP 97 program using equations from the test plan (Method 2 - V-D) was used to determine the relative areas of the coils. This program is listed in MME Book No. 630B.

Discussion

Based on the few repeats, I estimate that the absolute calibration data is good to  $\pm 0.1\%$  as predicted. The high precision data is probably  $\pm 0.02\%$ .

The determination of areas by Method 1 was facilitated by the outstanding performance of the LBL/CERN Nuclear Magnetic Resonance Magnetometer. The success of these NMR systems is largely due to the effort of Dr. Michael I. Green

Distribution

W.S. Gilbert  
M.I. Green  
E.C. Hartwig/L.J. Wagner/W.H. Deuser  
W.V. Hassenzahl  
T. Lauritzen  
R.M. Main  
D.J. Rondeau  
C.E. Taylor

This work was supported by the U.S. Dept. of Energy under Contract DE-AC03-76SF00098.

This report was done with support from the Department of Energy. Any conclusions or opinions expressed in this report represent solely those of the author(s) and not necessarily those of The Regents of the University of California, the Lawrence Berkeley Laboratory or the Department of Energy.

Reference to a company or product name does not imply approval or recommendation of the product by the University of California or the U.S. Department of Energy to the exclusion of others that may be suitable.

TECHNICAL INFORMATION DEPARTMENT  
LAWRENCE BERKELEY LABORATORY  
UNIVERSITY OF CALIFORNIA  
BERKELEY, CALIFORNIA 94720