

# UCSF

## UC San Francisco Previously Published Works

### Title

Evaluation of Intra- and Interscanner Reliability of MRI Protocols for Spinal Cord Gray Matter and Total Cross-Sectional Area Measurements

### Permalink

<https://escholarship.org/uc/item/4z3003b2>

### Journal

Journal of Magnetic Resonance Imaging, 49(4)

### ISSN

1053-1807

### Authors

Papinutto, Nico  
Henry, Roland G

### Publication Date

2019-04-01

### DOI

10.1002/jmri.26269

Peer reviewed

# Evaluation of Intra- and Interscanner Reliability of MRI Protocols for Spinal Cord Gray Matter and Total Cross-Sectional Area Measurements

Nico Papinutto, PhD,\* and Roland G. Henry, PhD

**Background:** In vivo quantification of spinal cord atrophy in neurological diseases using MRI has attracted increasing attention.

**Purpose:** To compare across different platforms the most promising imaging techniques to assess human spinal cord atrophy.

**Study Type:** Test/retest multiscanner study.

**Subjects:** Twelve healthy volunteers.

**Field Strength/Sequence:** Three different 3T scanner platforms (Siemens, Philips, and GE) / optimized phase sensitive inversion recovery (PSIR), T<sub>1</sub>-weighted (T<sub>1</sub>-w), and T<sub>2</sub>\*-weighted (T<sub>2</sub>\*-w) protocols.

**Assessment:** On all images acquired, two operators assessed contrast-to-noise ratio (CNR) between gray matter (GM) and white matter (WM), and between WM and cerebrospinal fluid (CSF); one experienced operator measured total cross-sectional area (TCA) and GM area using JIM and the Spinal Cord Toolbox (SCT).

**Statistical Tests:** Coefficient of variation (COV); intraclass correlation coefficient (ICC); mixed effect models; analysis of variance (t-tests).

**Results:** For all the scanners, GM/WM CNR was higher for PSIR than T<sub>2</sub>\*-w ( $P < 0.0001$ ) and WM/CSF CNR for T<sub>1</sub>-w was the highest ( $P < 0.0001$ ). For TCA, using JIM, median COVs were smaller than 1.5% and ICC  $> 0.95$ , while using SCT, median COVs were in the range 2.2–2.75% and ICC 0.79–0.95. For GM, despite some failures of the automatic segmentation, median COVs using SCT on T<sub>2</sub>\*-w were smaller than using JIM manual PSIR segmentations. In the mixed effect models, the subject was always the main contributor to the variance of area measurements and scanner often contributed to TCA variance ( $P < 0.05$ ). Using JIM, TCA measurements on T<sub>2</sub>\*-w were different than on PSIR ( $P = 0.0021$ ) and T<sub>1</sub>-w ( $P = 0.0018$ ), while using SCT, no notable differences were found between T<sub>1</sub>-w and T<sub>2</sub>\*-w ( $P = 0.18$ ). JIM and SCT-derived TCA were not different on T<sub>1</sub>-w ( $P = 0.66$ ), while they were different for T<sub>2</sub>\*-w ( $P < 0.0001$ ). GM area derived using SCT/T<sub>2</sub>\*-w versus JIM/PSIR were different ( $P < 0.0001$ ).

**Data Conclusion:** The present work sets reference values for the magnitude of the contribution of different effects to cord area measurement intra- and interscanner variability.

**Level of Evidence:** 1

**Technical Efficacy:** Stage 4

J. MAGN. RESON. IMAGING 2019;49:1078–1090.

Quantifying spinal cord atrophy and the more recently described spinal cord gray matter atrophy in various neurologic conditions including trauma, inflammation, or neurodegeneration has gained increasing attention, particularly with the development of dedicated spinal cord imaging techniques.<sup>1–6</sup> Spinal cord dedicated volumetric 3D T<sub>1</sub>-weighted (T<sub>1</sub>-w) protocols, similar to the ones widely used for brain volume

estimation, are becoming a standard for total cross-sectional area (TCA) measurements.<sup>7</sup> However, on images acquired with this and other conventional T<sub>1</sub>-w and T<sub>2</sub>-weighted (T<sub>2</sub>-w) protocols, the gray matter (GM) / white matter (WM) contrast is suboptimal to allow separate assessment of these two tissues.

The most promising imaging techniques used so far to measure GM area/volume in the spinal cord are based on

View this article online at [wileyonlinelibrary.com](http://wileyonlinelibrary.com). DOI: 10.1002/jmri.26269

Received May 2, 2018, Accepted for publication Jul 5, 2018.

\*Address reprint requests to: N.P., Dept. of Neurology, University of California San Francisco, 94158, San Francisco, CA, USA. E-mail: [nico.papinutto@ucsf.edu](mailto:nico.papinutto@ucsf.edu)

From the Department of Neurology, University of California San Francisco, 94158, San Francisco, CA, USA

This is an open access article under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs License, which permits use and distribution in any medium, provided the original work is properly cited, the use is non-commercial and no modifications or adaptations are made.

$T_2^*$ -weighted ( $T_2^*$ -w) 3D or 2D gradient echo (GRE)<sup>8–10</sup> or 2D  $T_1$ -w phase sensitive inversion recovery (PSIR) protocols.<sup>11,12</sup> These protocols also have good WM/CSF (cerebrospinal fluid) contrast that allows assessment of the TCA. Direct comparison of the  $T_2^*$ -w and PSIR techniques in terms of intrascanner and interscanner reliability of TCA and GM area measurements is fundamental for effect-size and sample-size estimates in studies quantifying spinal cord and cord gray matter tissues atrophy.

The goal of this study was to perform direct comparisons of three selected optimized protocols for spinal cord GM and TCA segmentation on the same group of 12 healthy controls, on three 3T scanners produced by the three main vendors of human magnetic resonance imaging (MRI) scanners: Siemens, Philips, and General Electric (GE). The chosen protocols used product sequences available on all systems.

## Materials and Methods

### Research Participants

Twelve healthy subjects (five males, seven females, mean age/standard deviation [SD]: 33.5/9.7 years) with no history of neurological disorder were enrolled in the study. The Committee on Human Research at our institution approved the study protocols. Written informed consent was obtained from all participants.

### Image Acquisition

All participants were scanned twice with a 30-minute MRI protocol, with repositioning in between the scans (test/retest, 1 hour of total scan time per scanner) on three different scanners: a Siemens 3T Skyra (Siemens Healthineers, Erlangen, Germany), a Philips 3T Ingenia (Royal Philips, Amsterdam, The Netherlands), and a GE 3T Discovery MR750 (General Electric Healthcare, Chicago, IL). Between the test and retest scans, participants were asked to get off the scanner table and have a little walk in the MR room. The Siemens, Philips, and GE scanners were equipped, respectively, with a 64-channel head-neck coil, a neurovascular (NV) coil, and a HNS CTL 123 coil, all providing good signal-to-noise ratio in the upper cervical cord region. The 1-hour sessions on the three different scanners for each participant were performed within a month for all subjects (median, mean/SD: 20, 19.1/8.3 days). All study acquisitions were performed between October 5 2017 and November 27 2017.

The scanning protocol included a sagittal cervical cord localizer, an axial single-slice 2D PSIR acquisition at the spinal cord disc level C2-C3, a  $T_2^*$ -w axial 2D MEDIC/M-FFE/MERGE (nomenclature, respectively for Siemens, Philips, and GE) acquired covering the cervical portion of the spinal cord from vertebra C1 down to about vertebra C6, and a  $T_1$ -w sagittal 3D MPRAGE/ $T_1$ -TFE/BRAVO acquisition centered on the C3 vertebral body.

The PSIR protocol was optimized on the three scanners based on experience with the previously developed protocols for the Skyra Siemens scanner.<sup>11–13</sup> The specific sequence/parameters on each scanner were optimized conditional on the specific software/hardware available. The driving optimization rationale was to achieve similar contrast-to-noise ratio (CNR) across the scanners and

hardware configurations. Perfectly matching all the acquisition parameters and acquisition times was not considered a priority.

The  $T_2^*$ -w and  $T_1$ -w protocols used in the study were optimized in a worldwide collaborative initiative.<sup>14</sup> The protocols are freely available for download on the website <https://osf.io/tt4z9/> and are periodically updated to follow modifications suggested by the participants to the initiative. For the GE scanner, minor further tuning of some parameters was necessary because of hardware limitations (no anterior neck coils were available, therefore parallel imaging was not possible).

Parameters for the 2D PSIR, 3D  $T_1$ -w, and 2D  $T_2^*$ -w protocols are reported in Table 1. Note that the definitions of some of the parameters, in particular repetition time (TR), inversion time (TI), and echo time (TE), vary by vendor. For example, the TR is defined for Siemens as the time between inversion/preparation pulses, while for Philips and GE the TR is defined as time between excitation pulses.

### Qualitative Quality Assessment

Two operators (NP and RGH) visually assessed the quality of all the images and assigned a consensual score (0, bad; 0.5, average; 1, good) to each image for three image characteristics: overall quality, overall noise/excessive motion, and GM/WM delineation/contrast (when present). Values were summed to assign a total score for each scanner and protocol for each of the three characteristics (maximum score = 24).

### Data Processing

All data processing and analyses were performed by a single operator (NP) with more than 12 years of experience in brain and spinal cord MRI acquisition/analysis methods.

### CNR Evaluation

CNR between GM and WM ( $CNR_{GM/WM}$ ) and between WM and CSF ( $CNR_{WM/CSF}$ ) was calculated by two operators (NP and RGH) in regions of interest (ROIs) at the C2-C3 disc level on the test images for each scanner/protocol/subject.

The two operators manually drew a GM ROI on the anterior part of the GM (an area spanning the anterior horns). Three WM ROIs were symmetrically drawn in the region of the lateral and posterior columns. Two CSF ROIs were symmetrically placed in the right and left spinal canal. The same group of ROIs was used for all the acquisitions of a subject. Examples of ROIs are reported in Figs. 1, 2 and 3

CNR between tissues 1 and 2 was computed for each subject, scanner, and protocol as previously defined<sup>15–18</sup>:

$$CNR_{12} = |SI_1 - SI_2| / \sqrt{SD_1^2 + SD_2^2} \quad (1)$$

where  $SI_1$ ,  $SI_2$ ,  $SD_1$ , and  $SD_2$  respectively indicate the mean intensity value within the tissues 1 and 2 ROIs, and the corresponding standard deviations.

Average values and SDs between operators were computed and differences between scanners and protocols tested using two-tailed *t*-tests ( $P < 0.05$ ).

### Total Cross-Sectional Area and GM Area Reliabilities

To estimate intra- and interscanner reliability of area measurements, TCA and GM were computed for all the available images acquired with the different protocols.

In order to assess the potential impact of the segmentation method on the spinal cord metrics, we selected the two most widely used approaches based on recent literature. The first approach is JIM v6 (Xinapse Systems, <http://www.xinapse.com>) that was previously

used in different studies.<sup>1-3,11,12,19-25</sup> The second approach is the open source Spinal Cord Toolbox (SCT) (<https://sourceforge.net/p/spinalcordtoolbox/wiki/Home/>) that has seen recent utilization.<sup>14,23,26-33</sup>

While the methods based on JIM have been previously optimized and tested for GM/TCA segmentation of single-slice 2D PSIR images and for TCA extracted by 3D T<sub>1</sub>-w and T<sub>2</sub>/T<sub>2</sub>\*-w images, the SCT has been optimized and tested for TCA extraction from T<sub>1</sub>-w/T<sub>2</sub>-w acquisitions and for GM/TCA segmentation of T<sub>2</sub>\*-w

**TABLE 1. 2D Phase Sensitive Inversion Recovery (PSIR), 3D T1-w, and 2D T2\*-w Protocol Parameters**

<u>2D PSIR</u>	Siemens	Philips	GE
Sequence name	CV	T1-TFE	PSMDE
Dimension	2D	2D	2D
TR (msec)	4000	9.5	8.00
TE (msec)	3.22	4.7	3.76
TI (msec)	400	300	400
# averages	3	5	20
Shots	9	10	52
Segments	26	24	4/5
Flip angle (deg)	10	15	25
Voxel size (mm)	.78 × .78 × 5	.78 × .78 × 5	.78 × .78 × 5
Field of view (mm)	200 × 200 × 5	200 × 200 × 5	200 × 200 × 5
BW(Hz/Px)	250	151.7	113.6 (22.73 kHz tot)
Phase encoding dir.	R > > L	R > > L	A > > P
Parallel acc. factor	no	no	no
Acq. time (min:sec)	1:52	2:30	~3 (dep. on heart rate)
Cardiac gating	simulated	not needed	finger pulse
Orientation	axial	axial	axial

<u>3D T1-w</u>	Siemens	Philips	GE
Sequence name	MPRAGE	T1-TFE	BRAVO
Dimension	3D	3D	3D
TR (msec)	2000	7.77	8.71
TE (msec)	3.72	3.56	3.66
TI (msec)	1000	1000	450
# averages	1	1	2
Flip angle (deg)	9	8	12
Voxel size (mm)	1 × 1 × 1	1 × 1 × 1	1 × 1 × 1
Field of view (mm)	320 × 260 × 192	256 × 256 × 192	256 × 256 × 96

TABLE 1. Continued

<b>3D T1-w</b>	<b>Siemens</b>	<b>Philips</b>	<b>GE</b>
BW (Hz/Px)	150	191.5	97.65 (25.00 kHz tot)
Phase encoding dir.	A > > P	A > > P	R > > L
Parallel acc. factor	2	2	no
Acq. time (min:sec)	4:44	4:56	7:28
Orientation	sagittal	sagittal	sagittal

<b>2D T2*-w</b>	<b>Siemens</b>	<b>Philips</b>	<b>GE</b>
Sequence name	MEDIC	M-FFE	MERGE
Dimension	2D	2D	2D
TR (msec)	627	625	777
TE (msec)	15	2.52	13.77
# echoes	3	3	3
# averages	2	1	2
Flip angle (deg)	30	30	30
Voxel size (mm)	0.5 × .0.5 × 3	0.5 × .0.5 × 3	0.5 × .0.5 × 3
Field of view (mm)	160 × 160 × 45	160 × 160 × 45	128 × 128 × 45
BW (Hz/Px)	240	241.1	195.31 (25.00 kHz tot)
Phase encoding dir.	R > > L	R > > L	A > > P
Parallel acc. factor	2	no	no
Acq. time (min:sec)	3:52	4:02	6:44
Orientation	axial	axial	axial

acquisitions with multiple slices/volumetric coverage. Therefore, JIM was used to segment TCA on the T<sub>1</sub>-w, PSIR, and T<sub>2</sub>\*-w images, and to segment GM on PSIR images. The automatic SCT was used to segment TCA on T<sub>1</sub>-w and T<sub>2</sub>\*-w images, and to segment GM on T<sub>2</sub>\*-w images. We did not use the SCT on the single-slice 2D PSIR images, because the SCT is optimized for multiple slices/volumetric coverage, and JIM for the GM segmentation of T<sub>2</sub>\*-w images, because we believe there are no exhaustive published data regarding the reliability of manual segmentations with this combination of software/contrast.

**2D PSIR IMAGES.** TCA and GM areas for each participant and scanner were measured on the phase-sensitive reconstructed images. TCA estimates were obtained in a semiautomated way using an active surface model<sup>34</sup> available in JIM, with a method previously shown to have high intra- and inter-rater reliability.<sup>1,11</sup> Briefly, this was done using the *cord finder toolkit* with fixed settings (nominal cord diameter 8 mm, number of shape coefficients 24, order of longitudinal

variation 12). The marker requested by the toolkit was positioned by a single experienced operator (NP) on the mid-sagittal WM, directly posterior to the gray commissure.

GM areas were manually measured using JIM with a segmentation technique that has been shown to be highly reliable. GM area was segmented three times using JIM by NP for each participant and scanner. The average GM area obtained from the three segmentations was finally calculated.<sup>1,11</sup>

From previous experience<sup>1,3,11,35,36</sup> the interoperator variability of the segmentations performed with the JIM methods are expected to have coefficient of variation (COV) <0.5% and intraclass correlation coefficient (ICC) >0.99 for the TCA semiautomated measurements, and COV in the range 3–5% and ICC ~0.90 for the GM area manual measurements.

**T<sub>1</sub>-W AND T<sub>2</sub>\*-W IMAGES.** Two methods were used to calculate TCA on T<sub>1</sub>-w and T<sub>2</sub>\*-w acquisitions. The first

method, semiautomated, was used in previous publications.<sup>3,19,20</sup> TCA on  $T_1$ -w images was measured by reslicing the sagittal acquisitions and extracting five consecutive 1-mm-thick axial slices perpendicular to the long axis of the cord at the C2-C3 disc level, and measuring the average area of the cord using the semiautomated *cord finder toolkit* of JIM with the same fixed settings used for PSIRs. The markers requested by the toolkit were placed at the center of the spinal cord in each of the five slices. For  $T_2^*$ -w images a similar process was applied to a single axial slice at the C2-C3 disc level without any reslicing.

The second method used the fully automatized SCT. Original images were preprocessed in the native space and then registered to the PAM50 spinal cord template.<sup>31</sup> The TCA was extracted from  $T_1$ -w and  $T_2^*$ -w scans following automatic cord segmentation (using in order the commands “sct\_propseg,” “sct\_label\_vertebrae,” “sct\_label\_utils,” “sct\_register\_to\_template,” and “sct\_warp\_template”<sup>33</sup> with default parameters, following the documentation available at <https://sourceforge.net/p/spinalcordtoolbox/wiki/tools/>) and then averaged within the C3 vertebra automatically labeled by the software (command “sct\_process\_segmentation”). The GM area on  $T_2^*$ -w images was extracted with the SCT following automatic gray matter segmentation (command “sct\_segment\_graymatter”)<sup>32</sup> and extracting the averaged value within the C3 vertebra (command “sct\_process\_segmentation”).

### Statistical Analysis

All statistical analyses were performed using JMP Pro 13 (SAS Institute, Cary, NC).

- The coefficient of variation ( $COV = 100 \times (\text{absolute difference}) / \text{mean of measurements}$ ) for all the test/retest couples of TCA and GM area measured with the different segmentation methods/protocols was calculated and its median/mean (SD) on the group of subjects computed for each scanner.
- The ICC was calculated between all the test/retest couples of TCA and GM area measured, for each different segmentation method/protocol, and for each scanner.
- Bland–Altman plots were produced for each of the combinations segmentation method/protocol, representing each scanner with a different symbol. On the Bland–Altman plots, the difference of the retest and test measurements was reported on the y-axis and their mean value on the x-axis.
- Mixed models with scanner as fixed effect, and test–retest and subject as nested random effects were used to estimate the contribution of subject, test–retest acquisition, and scanner to the variance of obtained measures.
- To visualize interscanner differences in the calculated areas, the average values between test and retest acquisitions were

computed and graphed for each of the combinations method/protocol.

- To evaluate the effect of acquisition protocol when using the same segmentation method, analysis of variance (ANOVA) ( $t$ -tests) were used between couple of measurements for TCA (PSIR,  $T_1$ -w, and  $T_2^*$ -w images segmented with JIM and  $T_1$ -w and  $T_2^*$ -w images segmented with SCT). ANOVA was used also to evaluate the effect of the segmentation method on the same protocol ( $T_1$ -w images and  $T_2^*$ -w images segmented with JIM and SCT). Finally, ANOVA was used to see if there were statistically significant differences ( $P < 0.05$ ) in the GM area values obtained with the two couples segmentation method/protocol (SCT/ $T_2^*$ -w vs. JIM/PSIR).

## Results

### Qualitative Quality Assessment

All images acquired with PSIR and  $T_2^*$ -w protocols at the C2-C3 disc level for the 12 healthy controls on the three different scanners are reported in Figs. 1 and 2, demonstrating overall good quality of both the PSIR and  $T_2^*$ -w images. Overall quality consensus scores for PSIR were 24 for all the scanners, while for  $T_2^*$ -w images they were 23, 18, and 22, respectively for Siemens, Philips, and GE.

Consensus visual qualitative assessment scores suggest that the PSIR images on the GE scanner appeared slightly more noisier/affected by motion (22) compared with the other two scanners (24), while according to this score, the  $T_2^*$ -w images appeared to be of a slightly better quality on Siemens (23) and GE (24) scanners compared with Philips (22). With regard to the GM delineation/contrast, visual qualitative assessment indicated consistent good quality for the PSIR images (23, 24, 23 for Siemens, Philips, and GE), while the quality of some of the  $T_2^*$ -w images (12 over 72, which means about 17%) was suboptimal (scores 24, 19, 23 for Siemens, Philips, and GE).

Illustrative images acquired on a single subject with the  $T_1$ -w protocol are shown in Fig. 3. The quality of  $T_1$ -w images was in general consistent with the reported example and consistent across the different scanners (overall quality scores and noise/motion were 24, 24, and 23, respectively, for Siemens, Philips, and GE). We excluded from the following analyses two  $T_1$ -w acquisitions because the subject clearly moved during the acquisition (both were test scans on the GE scanner).

### CNR Evaluation

In Table 2 the GM/WM and WM/CSF CNR measured at the C2-C3 disc level for PSIR,  $T_2^*$ -w, and  $T_1$ -w protocols is reported.

The GM/WM CNR for the PSIR protocol was higher compared with the  $T_2^*$ -w protocol for all the scanners ( $P < 0.0001$ ). GM/WM CNR for the  $T_2^*$ -w protocol on the




FIGURE 1: PSIR images acquired at the C2-C3 disc level for the 12 subjects on the three different scanners. Each row is a subject and from left to right: Siemens scanner test acquisition, Siemens scanner retest acquisition, Philips scanner test acquisition, Philips scanner retest acquisition, GE scanner test acquisition, and GE scanner retest acquisition. In the bottom right image an example of ROIs used for the CNR evaluation is reported.


Philips scanner was lower if compared with Siemens ( $P < 0.0001$ ) that was lower if compared with GE ( $P < 0.0001$ ), confirming the visual qualitative impression of slightly worse quality on Philips. The visual impression that the PSIR images were noisier on GE, however, is not supported by the CNR evaluation.

The WM/CSF CNR for the T<sub>1</sub>-w protocol was consistently higher for all the scanners compared with the PSIR and T<sub>2</sub>\*-w protocol images ( $P < 0.0001$ ). The WM/CSF CNR for PSIR and T<sub>2</sub>\*-w protocols was comparable for GE ( $P = 0.06$ ) and Philips ( $P = 0.042$ ), while for Siemens for PSIR it was much higher ( $P < 0.0001$ ).

#### TCA and GM Area Measurements

Test-retest COV and ICC for measured TCA and GM area for all the combinations of segmentation methods and protocols are reported in Table 3.

For the TCA, with the JIM semiautomatic method, median COVs are very similar across the three protocols and smaller than 1.5% (with the only exception of PSIR on the GE scanner). ICC were always  $>0.95$ . The SCT performed very similarly when measuring TCA on T<sub>1</sub>-w and T<sub>2</sub>\*-w protocols for all the scanners, with median COV in the range 2.2–2.75% and ICC in the range 0.79–0.95.



**FIGURE 2:**  $T_2^*$ -w images acquired at the C2-C3 disc level for the 12 subjects on the three different scanners. Each row is a subject and from left to right: Siemens scanner test acquisition, Siemens scanner retest acquisition, Philips scanner test acquisition, Philips scanner retest acquisition, GE scanner test acquisition, and GE scanner retest acquisition. In the bottom right image an example of ROIs used for the CNR evaluation is reported.

With regard to GM, median COV of measurements obtained with the SCT on  $T_2^*$ -w images were consistently smaller compared to manual segmentation on PSIR images. Nevertheless, mean and SD of COV for the SCT/ $T_2^*$ -w combination are in general larger than on JIM/PSIR, because of outlier values due to failures of the automatic segmentation algorithm (Bland–Altman plots reported in Fig. 4). The lowest ICC were 0.3423 for SCT/ $T_2^*$ -w on GE and 0.6915 for JIM/PSIR on Philips.


Furthermore, despite similar COV median values, mean/SD are bigger (and ICC smaller) for SCT on  $T_2^*$ -w

images than for  $T_1$ -w images, due to fewer automatic segmentation errors of the SCT on  $T_1$ -w images.

**Statistical Analysis**

In the mixed effect models, subject was always the main contributor to the variance of the area measurements (always significant, Wald  $P < 0.05$ ). Among all the combinations of segmentation methods/protocols/tissues, session (test–retest) was statistically significant in explaining the variance of the area measurements only for TCA measured on  $T_1$ -w images with the SCT ( $P < 0.0001$ ). Scanner (fixed effect) was instead





**FIGURE 3:** Illustrative example of  $T_1$ -w images acquired for a single healthy subject. Top row: sagittal original acquisition. Bottom row: axial reslicing at the C2-C3 disc level. From left to right: Siemens scanner test acquisition, Siemens scanner retest acquisition, Philips scanner test acquisition, Philips scanner retest acquisition, GE scanner test acquisition, and GE scanner retest acquisition. In the bottom right image an example of ROIs used for the CNR evaluation is reported.

often statistically significant in the mixed effect models for TCA measurements. The value of the estimated intercept (that can be read as values for Siemens for the way the variables were ordered in the model), the biases of the measurements that the model attributes to the other scanners and the related  $P$ -values are reported in Table 4.

The mean value of the test and retest acquisitions for the 12 healthy subjects are reported in Fig. 5 for each segmentation method/protocol/area.

According to the ANOVA, there was a statistically significant difference for TCA measurements on the  $T_2^*$ -w protocol segmented with JIM, compared to both PSIR ( $P = 0.0021$ ) and  $T_1$ -w ( $P = 0.0018$ ). No difference was found between the  $T_1$ -w and  $T_2^*$ -w protocols when the segmentation was performed using the SCT ( $P = 0.18$ ).

Regarding the difference attributable to the segmentation method on the same protocol for TCA measurement, on  $T_1$ -w images JIM and SCT were not statistically different ( $P = 0.66$ ), while there was difference for the  $T_2^*$ -w contrast ( $P < 0.0001$ ).

Finally, there were very significant differences in the GM area obtained using the SCT/ $T_2^*$ -w vs. JIM/PSIR segmentation method/protocol combinations ( $P < 0.0001$ ). However, no scanner-related bias was detected for GM areas.

All results are graphed in Fig. 6, where  $P$ -values for the comparisons between different combinations of segmentation method/protocol are also reported.

## Discussion

In this work we present for the first time analysis of a rich MRI dataset, acquired on the same 12 healthy subjects on three 3T scanners produced by the main commercial brands with the most promising protocols for TCA and GM area assessments. A qualitative assessment and quantitative

evaluation of CNR and intra- and interscanner reliability of area measurements at the C2-C3 spinal cord level is presented.

For GM delineation, the quality of PSIR images was more consistent than  $T_2^*$ -w images. These qualitative visual impressions were confirmed by CNR evaluations that showed that GM/WM CNR for PSIR images was higher than for  $T_2^*$ -w images for all the scanners. The observed tendency for fuzzy appearance of GM on  $T_2^*$ -w compared to PSIR images may be the result from higher sensitivity of the  $T_2^*$ -w protocol to susceptibility artifacts and motion.

**TABLE 2. Between Operators Mean (Standard Deviation) Contrast-to-Noise Ratio (CNR) Between Gray Matter (GM) and White Matter (WM) Tissues, and Between WM and Cerebrospinal Fluid (CSF), for the Three Different Acquisition Protocols (PSIR,  $T_2^*$ -w and  $T_1$ -w) on the Three Different Scanners**

$CNR_{GM/WM}$	PSIR	$T_2^*$ -w	$T_1$ -w
Siemens	2.11 (0.15)	1.56 (0.06)	—
Philips	3.14 (0.16)	1.09 (0.05)	—
GE	2.39 (0.04)	1.67 (0.02)	—

$CNR_{WM/CSF}$	PSIR	$T_2^*$ -w	$T_1$ -w
Siemens	8.45 (0.22)	3.54 (1.38)	9.91 (0.14)
Philips	3.06 (0.11)	3.71 (1.05)	9.08 (0.04)
GE	3.40 (0.07)	4.64 (2.16)	7.26 (0.05)

**TABLE 3. Test-Retest COV (Median, Top Row, and Mean (SD), Middle Row) and ICC (Bottom Row) for TCA and GM Area Measurements on Images Acquired With the Three Protocols (PSIR, T1-w, T2\*-w) and Using the Different Segmentation Methods (JIM, SCT) on the Group of 12 Controls**

	JIM TCA T1-w	JIM TCA PSIR	JIM TCA T2*-w	SCT TCA T1-w	SCT TCA T2*-w	SCT GM T2*-w	JIM GM PSIR
Siemens	0.84 1.23(1.25) 0.9892	1.09 1.55(1.40) 0.9853	1.43 2.19(2.35) 0.9671	2.27 2.53(2.16) 0.9450	2.42 5.87(12.96) 0.8121	2.91 8.31(15.91) 0.6516	3.41 4.31(2.94) 0.8751
Philips	1.16 1.55(1.32) 0.9828	1.14 1.21(0.83) 0.9923	1.36 1.85(1.90) 0.9720	2.67 3.55(2.74) 0.8753	2.24 3.91(4.04) 0.8652	4.82 5.05(4.04) 0.8077	7.21 6.52(3.30) 0.6915
GE	1.02 1.34(0.96) 0.9893	2.38 2.56(1.53) 0.9515	0.82 1.15(0.96) 0.9903	2.36 3.37(3.75) 0.8770	2.76 5.60(7.14) 0.7937	3.64 8.30(10.46) 0.3423	6.52 6.25(4.64) 0.7443

Data for each scanner are reported.


3D T<sub>1</sub>-w images were consistently of good quality for all the scanners in terms of spinal cord/CSF delineation. This is not surprising, considering that the used protocols are optimizations of 3D inversion recovery spoiled gradient echo protocols that have become a standard for atrophy assessment on brain images over more than a decade.

The goal of having comparable CNR across different vendors was overall achieved with the chosen sequences/parameters/hardware, with in general 3D T<sub>1</sub>-w protocols giving higher WM/CSF CNR than the other two protocols and PSIR higher GM/WM CNR than T<sub>2</sub>\*-w protocols. It has to be mentioned that, since there were different hardware configuration/protocol choices, we preferred not to correct CNR for acquisition times/coverage when evaluating CNR at

the C2-C3 level. It has also to be mentioned that different resolutions can affect the CNR.

For TCA estimates, all protocols performed very similarly in terms of intra- and interscanner reliability, for a given segmentation method. The semiautomatic method based on JIM showed better test-retest COV and ICC on both 3D T<sub>1</sub>-w and T<sub>2</sub>\*-w protocols compared with the automatic SCT method.

We therefore think there is not an obvious choice of best protocol if the goal of a study is TCA evaluation. The choice has to be driven by a series of factors and considerations such as the acquisition time that can be spent on a protocol, the spinal cord levels to be covered, the specific hardware available, and the need of assessing TCA alone or



**FIGURE 4: Bland-Altman plots reporting the difference between the TCA and GM area retest and test measurements (y-axis) and their mean value (x-axis), for all the combinations of software and protocol.**




FIGURE 5: Plots reporting the mean of the values measured for the TCA and GM area in the test and retest acquisitions (y-axis) for the 12 healthy subjects (x-axis), for all the combinations of segmentation method and protocol (indicated above the plot).

GM as well. It is worth noting that  $T_2^*$ -w protocols gave significant biases in comparison to 3D  $T_1$ -w and PSIR protocols for TCA estimates when JIM segmentation was used.

When using JIM, biases across scanners tended to be statistically more significant than when using the SCT; the test-retest COV and biases across scanners with JIM measurements were all bigger than the interoperator variability reported in the previous literature. The statistical sensitivity to scanner model for JIM could be explained considering the lower interoperator and intrascanner variability of JIM-based segmentations, thereby providing statistical power to detect small biases across scanners.


GM segmentations were performed with the method/protocol couples JIM/PSIR and SCT/ $T_2^*$ -w. The latter combination gave lower median COVs. There was a statistically significant large bias between values obtained with the two segmentation method/protocol combinations.  $T_2^*$ -w derived GM areas were much smaller than PSIR derived ones. This could be due to the contrast difference or the higher resolution for the  $T_2^*$ -w protocols. It has been shown previously that higher resolution PSIR gives smaller partial volume effects on the WM/GM edge and therefore smaller GM area estimates.<sup>37</sup>

Automatic methods have obvious advantages if compared to manual or semiautomatic methods, but if they often need corrections they are essentially semiautomatic methods.

The SCT was not robust as an automatic method on these spinal cord data and had higher failure rates on  $T_2^*$ -w images than  $T_1$ -w images. This explains the low median values, but bigger mean and standard deviations in some cases. In a few cases, the SCT repeated the same error for both test and retest acquisitions of a particular subject, in particular for  $T_2^*$ -w images (for example, wrong vertebra assignment, or wrong total cord delineation and subsequently GM segmentation). These systematic errors gave a good intrascanner reproducibility of GM area, but it was evident with a visual check that the segmentations were not accurate in both the test and retest acquisitions. We also noticed that the SCT provided a GM segmentation result even when there was no WM/GM delineation (a suboptimal  $T_2^*$ -w image, but even on a 3D  $T_1$ -w images with no GM visible). The SCT can be very useful also in these situations to create a probabilistic GM mask to be used to calculate metrics on other acquisitions/contrasts, but it could give misleading information if used to quantitatively assess the GM area.

These observations could explain why in a published work that tested the SCT GM segmentation method and other methods, the Dice Similarity Coefficients and Jaccard Index indicated moderate overlap of segmentations obtained with SCT with gold-standard manual segmentations.<sup>15</sup>

Manually correcting SCT errors was beyond the scope of the present work, but there is clearly room for improving



**FIGURE 6:** ANOVA for TCA and GM area measured with the different combinations segmentation method/protocol. *P*-values for the different couples of comparisons are reported and highlighted in bold when differences were statistically significant ( $P < 0.05$ ).

this very useful tool, maybe tuning it to the different specific acquisitions. Manual and semiautomatic methods have the disadvantage of being time-consuming and can have high interrater variability. Nevertheless, the semiautomatic method based on JIM consistently gave better reliability for TCA estimates on all the tested protocols. For the GM manual segmentations performed with JIM, the observed test-retest COVs were of the same order of magnitude of the interoperator variability previously reported. The statistical power provided by the GM segmentation technique was therefore not sufficient to disentangle all the possible sources

of variability in the measurement (scanner, positioning at acquisition, test-retest, interoperator variability of segmentations).

A limitation of this study is that software/hardware varied across scanners. For example, the available GE scanner did not have anterior neck coils, which prohibited the use of parallel imaging. For this reason, the implementation of some protocols on GE was a little different than on Siemens and Philips, where we managed to set more similar protocols. The T<sub>1</sub>-w sequence on GE had a smaller field of view (FOV) compared to the other scanners, different bandwidth, but two averages were made to compensate for the CNR lost. Different FOVs can affect the quality of shimming and the CNR, and therefore affect the quality of segmentations. A different choice could have been made; for example, an IR-SPGR sequence could have been used instead of BRAVO, or increasing the FOV instead of making two averages. Analogous differences due to the lack of parallel acceleration capability were present in the T<sub>2</sub>\*-w protocol. Also for PSIR, the way the different vendors implement the cardiac gating in the protocol forced differences in the settings on the different scanners.

Another limitation of this study is that we performed analyses with only two segmentation methods. We also decided not to perform segmentation for every protocol/segmentation method combination but constrained our analyses to only those applications already shown to be appropriate for the given segmentation method. Further developments in segmentation methods may help to reduce the variability in the area estimates. These data could be used for testing (and possibly improving) other algorithms.

Other limitations are the absence of a T<sub>2</sub>-w protocol, the limited number of subjects, and the fact that only data on healthy subjects were acquired (all choices forced by the very demanding protocol that required an hour of scan on three different scanners in a very short time frame).

**TABLE 4. Mixed Effect Model Results**

	JIM TCA T1-w	JIM TCA PSIR	JIM TCA T2*-w	SCT TCA T1-w	SCT TCA T2*-w	SCT GM T2*-w	JIM GM PSIR
Siemens (intercept)	79.94	80.63	84.62	78.81	77.49	13.90	20.28
Philips	-0.62 <b><i>P</i> = 0.0012</b>	0.37 <i>P</i> = 0.25	-1.54 <b><i>P</i> &lt; 0.0001</b>	0.82 <i>P</i> = 0.13	0.26 <i>P</i> = 0.74	0.04 <i>P</i> = 0.81	0.09 <i>P</i> = 0.48
GE	-0.56 <b><i>P</i> = 0.0039</b>	-1.42 <b><i>P</i> &lt; 0.0001</b>	0.64 <b><i>P</i> = 0.016</b>	-1.59 <b><i>P</i> = 0.0047</b>	0.84 <i>P</i> = 0.28	0.24 <i>P</i> = 0.15	0.08 <i>P</i> = 0.53

Estimated scanner contribution to biases: Estimated intercept (corresponding to Siemens) and the bias and related *P* value for the other scanners are reported (in bold when  $< 0.05$ ). Values of areas are in square mm.

While there may be further room for optimization, we believe that our efforts reflect the expected biases and variability due to the choice of scanners, protocols, and segmentation methods.

The present work suggests that multiscanner/multicenter studies for TCA/GM segmentation are feasible with all the techniques explored in the study.

This study may set reference values for the magnitude of the contribution of different effects (scanner, protocol, segmentation method) to TCA/GM area measurement intra- and interscanner variability.

The data and results reported in the present study may help in making informed decisions when planning a specific study, depending on the different acquisition settings and study goals.

Further optimization of protocols and segmentation algorithms is warranted and this study can help in determining what are the directions in which the spinal cord MRI community should move along.

---

## Acknowledgments

First and foremost, the authors would like to thank the twelve subjects who participated in this study. Repeating the MRI acquisitions on three different scanners in such a short time frame required significant time and effort. The authors would also like to thank Julien Cohen-Adad of Polytechnique Montréal for his great work in leading the international effort that defined most of the T<sub>1</sub>-w and T<sub>2</sub>\*-w protocols used in the study, as well as Antonella Castellano, Letterio Salvatore Politi and Andrea Falini of San Raffaele Institute (Milan, Italy), who were fundamental in optimizing the PSIR protocol for Philips scanners. Additionally, the authors would like to express their gratitude for the indispensable support received from Siemens (Gerhard Laub, Sinyeob Ahn, Kevin J. Johnson), Philips (Marcello Cadioli) and General Electric (Suchandrima Banerjee, Patrick D. Koon). Finally, the authors would like to thank William A. Stern, and all the MRI technologists and staff members who with their professionalism, kindness and availability made performing this study possible. This project was supported by the Research Evaluation & Allocation Committee (REAC), School of Medicine, University of California, San Francisco.

## References

- Schlaeger R, Papinutto N, Panara V, et al. Spinal cord gray matter atrophy correlates with multiple sclerosis disability. *Ann Neurol* 2014;76:568–580.
- Schlaeger R, Papinutto N, Zhu AH, et al. Association between thoracic spinal cord gray matter atrophy and disability in multiple sclerosis. *JAMA Neurol* 2015;72:897–904.
- Castellano A, Papinutto N, Cadioli M, et al. Quantitative MRI of the spinal cord and brain in adrenomyeloneuropathy: in vivo assessment of structural changes. *Brain* 2016;139(Pt 6):1735–1746.
- Papinutto and Henry: Spinal Cord MRI Protocols for Atrophy
- Freund P, Curt A, Friston K, Thompson A. Tracking changes following spinal cord injury: insights from neuroimaging. *Neuroscientist* 2013;19:116–128.
- Cohen-Adad J, El Mendili MM, Morizot-Koutlidis R, et al. Involvement of spinal sensory pathway in ALS and specificity of cord atrophy to lower motor neuron degeneration. *Amyotroph Lateral Scler Frontotemporal Degener* 2013;14:30–38.
- El Mendili MM, Cohen-Adad J, Pelegrini-Issac M, et al. Multi-parametric spinal cord MRI as potential progression marker in amyotrophic lateral sclerosis. *PLoS One* 2014;9:e95516.
- Rocca MA, Comi G, Filippi M. The role of T1-weighted derived measures of neurodegeneration for assessing disability progression in multiple sclerosis. *Front Neurol* 2017;8:433.
- De Leener B, Taso M, Cohen-Adad J, Callot V. Segmentation of the human spinal cord. *Magma* 2016;29:125–153.
- Held P, Dorenbeck U, Seitz J, Frund R, Albrich H. MRI of the abnormal cervical spinal cord using 2D spoiled gradient echo multiecho sequence (MEDIC) with magnetization transfer saturation pulse. A T2\* weighted feasibility study. *J Neuroradiol* 2003;30:83–90.
- Yiannakas MC, Kearney H, Samson RS, et al. Feasibility of grey matter and white matter segmentation of the upper cervical cord in vivo: a pilot study with application to magnetisation transfer measurements. *NeuroImage* 2012;63:1054–1059.
- Papinutto N, Schlaeger R, Panara V, et al. 2D phase-sensitive inversion recovery imaging to measure in vivo spinal cord gray and white matter areas in clinically feasible acquisition times. *J Magn Reson Imaging* 2015;42:698–708.
- Papinutto N, Schlaeger R, Panara V, et al. Age, gender and normalization covariates for spinal cord gray matter and total cross-sectional areas at cervical and thoracic levels: A 2D phase sensitive inversion recovery imaging study. *PLoS One* 2015;10:e0118576.
- Papinutto N, Datta E, Zhu AH, et al. Multisite feasibility study of spinal cord gray matter and total cord areas measurements on 2D phase sensitive inversion recovery images. In: *Proc 24th Annual Meeting ISMRM, Singapore*; 2016.
- Alley S, Gilbert G, Gandini Wheeler-Kingshott CAM, et al. Consensus acquisition protocol for quantitative MRI of the cervical spinal cord at 3T. In: *Proc 26th Annual Meeting ISMRM, Paris*; 2018.
- Prados F, Ashburner J, Blaiotta C, et al. Spinal cord grey matter segmentation challenge. *NeuroImage* 2017;152:312–329.
- Grussu F, Schneider T, Zhang H, et al. Neurite orientation dispersion and density imaging of the healthy cervical spinal cord in vivo. *NeuroImage* 2015;111:590–601.
- Yiannakas MC, Grussu F, Louka P, et al. Reduced field-of-view diffusion-weighted imaging of the lumbosacral enlargement: a pilot in vivo study of the healthy spinal cord at 3T. *PLoS One* 2016;11:e0164890.
- Gringel T, Schulz-Schaeffer W, Eloff E, et al. Optimized high-resolution mapping of magnetization transfer (MT) at 3 Tesla for direct visualization of substructures of the human thalamus in clinically feasible measurement time. *J Magn Reson Imaging* 2009;29:1285–1292.
- Papinutto N, Bakshi R, Bischof A, et al. Gradient nonlinearity effects on upper cervical spinal cord area measurement from 3D T1-weighted brain MRI acquisitions. *Magn Reson Med* 2018;79:1595–1601.
- Liu Z, Yaldizli O, Pardini M, et al. Cervical cord area measurement using volumetric brain magnetic resonance imaging in multiple sclerosis. *Mult Scler Relat Disord* 2015;4:52–57.
- Aymerich FX, Auger C, Alonso J, et al. Cervical cord atrophy and long-term disease progression in patients with primary-progressive multiple sclerosis. *AJNR Am J Neuroradiol* 2018;39:399–404.
- Singhal T, Tauhid S, Hurwitz S, Neema M, Bakshi R. The effect of glatiramer acetate on spinal cord volume in relapsing-remitting multiple sclerosis. *J Neuroimaging* 2017;27:33–36.

23. Yiannakas MC, Mustafa AM, De Leener B, et al. Fully automated segmentation of the cervical cord from T1-weighted MRI using PropSeg: Application to multiple sclerosis. *NeuroImage Clin* 2016;10:71–77.
24. Kim G, Khalid F, Oommen VV, et al. T1- vs. T2-based MRI measures of spinal cord volume in healthy subjects and patients with multiple sclerosis. *BMC Neurol* 2015;15:124.
25. Valsasina P, Rocca MA, Horsfield MA, et al. Regional cervical cord atrophy and disability in multiple sclerosis: a voxel-based analysis. *Radiology* 2013;266:853–861.
26. Martin AR, De Leener B, Cohen-Adad J, et al. Monitoring for myelopathic progression with multiparametric quantitative MRI. *PLoS One* 2018;13:e0195733.
27. Martin AR, De Leener B, Cohen-Adad J, et al. Can microstructural MRI detect subclinical tissue injury in subjects with asymptomatic cervical spinal cord compression?. A prospective cohort study. *BMJ Open* 2018;8:e019809.
28. Paquin ME, El Mendili MM, Gros C, et al. Spinal cord gray matter atrophy in amyotrophic lateral sclerosis. *AJNR Am J Neuroradiol* 2018;39:184–192.
29. Rasoanandrianina H, Grapperon AM, Taso M, et al. Region-specific impairment of the cervical spinal cord (SC) in amyotrophic lateral sclerosis: A preliminary study using SC templates and quantitative MRI (diffusion tensor imaging/inhomogeneous magnetization transfer). *NMR Biomed* 2017;30.
30. Martin AR, De Leener B, Cohen-Adad J, et al. Clinically feasible microstructural MRI to quantify cervical spinal cord tissue injury using DTI, MT, and T2\*-weighted imaging: assessment of normative data and reliability. *AJNR Am J Neuroradiol* 2017;38:1257–1265.
31. De Leener B, Levy S, Dupont SM, et al. SCT: Spinal Cord Toolbox, an open-source software for processing spinal cord MRI data. *NeuroImage* 2017;145(Pt A):24–43.
32. Dupont SM, De Leener B, Taso M, et al. Fully-integrated framework for the segmentation and registration of the spinal cord white and gray matter. *NeuroImage* 2017;150:358–372.
33. De Leener B, Kadoury S, Cohen-Adad J. Robust, accurate and fast automatic segmentation of the spinal cord. *NeuroImage* 2014;98:528–536.
34. Horsfield MA, Sala S, Neema M, et al. Rapid semi-automatic segmentation of the spinal cord from magnetic resonance images: application in multiple sclerosis. *NeuroImage* 2010;50:446–455.
35. Kearney H, Yiannakas MC, Abdel-Aziz K, et al. Improved MRI quantification of spinal cord atrophy in multiple sclerosis. *J Magn Reson Imaging* 2014;39:617–623.
36. Bischof A, Olney NT, Rosen HJ, et al. Phase Sensitive Inversion Recovery (PSIR) spinal cord imaging as a potential biomarker for Motor Neuron Disease. In: *Proc 26th Annual Meeting ISMRM, Paris; 2018.*
37. Datta E, Papinutto N, Schlaeger R, et al. Gray matter segmentation of the spinal cord with active contours in MR images. *NeuroImage* 2017;147:788–799.