

UCSF

UC San Francisco Previously Published Works

Title

IPVS policy statement on HPV nucleic acid testing guidance for those utilising/considering HPV as primary precancer screening: Quality assurance and quality control issues.

Permalink

<https://escholarship.org/uc/item/5d50w51d>

Authors

Garland, Suzanne

Iftner, Thomas

Cuschieri, Kate

et al.

Publication Date

2023-02-01

DOI

10.1016/j.jcv.2022.105349

Peer reviewed



Published in final edited form as:

J Clin Virol. 2023 February ; 159: 105349. doi:10.1016/j.jcv.2022.105349.

IPVS POLICY STATEMENT ON HPV NUCLEIC ACID TESTING guidance for those utilising/considering HPV as primary precancer screening: Quality Assurance and Quality Control issues

Suzanne M. Garland^{a,b,c,*}, Thomas Iftner^d, Kate Cuschieri^{e,f}, Andreas M Kaufmann^g, Marc Arbyn^h, Silvia de Sanjoseⁱ, Mario Poljak^j, Joakim Dillner^k, Elizabeth R. Unger^l, IPVS Policy Committee

Suzanne M Garland^{a,b,c}, Margaret Stanley^m, Anna-Barbara Moscickiⁿ, Yin Ling Woo^o, Neerja Bhatla^p, Karen L. Chan^q, Joel Palefsky^r, Anna R Giuliano^s, Andreas Kaufmann^g, Silvia de Sanjoseⁱ, Julia M.L. Brotherton^{t,u}, Sarah Feldman^v

^aDepartment of Obstetrics and Gynaecology, University of Melbourne, Parkville 3052, Victoria, Australia

^bCentre for Women's Infectious Diseases Research, The Royal Women's Hospital, Parkville 3052, Victoria, Australia

^cInfection and Immunity, Murdoch Children's Research Institute, Parkville 3052, Victoria, Australia

^dInstitute Med. Virology and Epidemiology, of Viral Diseases Institut Med. Virologie und Epidemiologie der Viruskrankheiten Elfriede-Aulhorn-Str. 6 | 72076 Tübingen, Germany

^eScottish HPV Reference Laboratory, Royal Infirmary of Edinburgh, Edinburgh, Scotland

^fHPV Research Group, University of Edinburgh, Edinburgh, Scotland

^gCharité – Universitätsmedizin Berlin, corporate member of Freie Universität Berlin and Humboldt-Universität zu Berlin, Clinic for Gynecology, Augustenburger Platz 1, 13353 Berlin, Germany

^hUnit Cancer Epidemiology, Belgian Cancer Centre, Sciensano, 1050 Brussels, Belgium

ⁱISGlobal, Barcelona, Spain

^jInstitute of Microbiology and Immunology, Faculty of Medicine, University of Ljubljana, Slovenia

^kMedical Diagnostics Karolinska Karolinska University Hospital, Center for Cervical Cancer Prevention, 14186 Stockholm, Sweden

^lDivision of High-Consequence Pathogens and Pathology, National Center for Emerging and Zoonotic Infectious Diseases, Centers for Disease Control and Prevention, Atlanta GA, USA.

* **Corresponding author** Prof Suzanne M. Garland, suzanneg@unimelb.edu.au, Centre for Women's Infectious Diseases, The Royal Women's Hospital, Melbourne, Locked Bag 300 | Corner Grattan St & Flemington Rd, Parkville VIC 3052 Australia.

Declaration of Competing Interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

^mDepartment of Pathology, Tennis Court Rd, Cambridge, CB2 1QP, UK

ⁿDepartment of Pediatrics, University of California, Los Angeles, USA 90095

^oDepartment of Obstetrics and Gynaecology, Faculty of Medicine, University Malaya Kuala Lumpur, Malaysia

^pDepartment of Obstetrics & Gynaecology, All India Institute of Medical Sciences, New Delhi, 110029, India

^qDepartment of Obstetrics and Gynaecology Queen Mary Hospital, Hong Kong

^rUniversity of California, San Francisco. CA, USA

^sCenter for Immunization and Infection Research in Cancer (CIIRC), Moffitt Cancer Center, Tampa, FL, 33612, USA

^tAustralian Centre for the Prevention of Cervical Cancer, East Melbourne 3002, Victoria, Australia

^uMelbourne School of Population and Global Health, University of Melbourne, Parkville 3052, Victoria, Australia

^vBrigham and Women's Hospital, Gynecology and Reproductive Biology, Harvard Medical School, Boston, MA, USA.

Abstract

We advise that only clinically validated HPV assays which have fulfilled internationally accepted performance criteria be used for primary cervical screening. Further, assays should be demonstrated to be fit for purpose in the laboratory in which they will ultimately be performed, and quality materials manuals and frameworks will be helpful in this endeavor. Importantly, there is a fundamental shortage of well validated, low-cost, low complexity HPV tests that have demonstrated utility in a near-patient setting; representing a significant challenge and focus for future development in order to reach the WHO's goal of eliminating cervical cancer.

Keywords

HPV; primary; cervical; screening; validation

Call to action for global cervical cancer elimination strategy endorsed at 73rd World Health Assembly of the WHO.

On May 18th, 2018, the WHO Director General Dr Tedros Adhanom Ghebreyesus made a call to action for cervical cancer elimination as a public health problem [1]. At the closure of the 73rd World Health Assembly of the WHO, a virtual meeting was held on November 17th, 2020, at which the official announcement of the global elimination strategy to accomplish this objective was launched. Three pillars of action as a strategy to achieve this were described, including, utilising a high precision assay for cervical screening in at least 70% of women worldwide with at least two lifetime screenings at 35 and 45 years of age, 90% of young girls completing an HPV vaccinate series by 15 years of age and 90% of women

recognised with cervical disease receiving adequate care and treatment, including palliative care. Together these comprise the 2030 targets aimed at cervical cancer becoming a rare disease affecting <4/100,000 women per annum. Although we do have the tools to achieve this goal, the COVID-19 pandemic that began March 2020, has raised additional challenges.

One pillar of action: Increased coverage of cervical screening with high precision assays

Cervical cytology, as first described by George Papanicolaou in 1928 [2], performs well when adopted repeatedly (but has a poor negative predictive value when the frequency is greater than three yearly) for the target population and is implemented in a setting with high quality assurance (QA) and control (QC) and appropriate and timely follow up of cervical abnormalities. However, one-time cervical cytology has relatively low sensitivity and greatest effectiveness occurs due to the cumulative sensitivity associated with recurrent screening. The reality for those (few) women in Low and Middle Income Countries [LMICs] able to access cytology services is one to two cytology tests per lifetime. Often when LMICs have introduced cervical screening it has been with the adoption of VIA [Visual Inspection with Acetic Acid] as VIA requires few tools (dilute white vinegar and the naked eye) to observe cervical abnormalities, allowing for same day cryotherapy or thermoablation. However, VIA is subjective, has low specificity, and does require training and ongoing quality assurance.

Observational studies [3], followed by randomised controlled trials (RCTs) [4-8], (largely performed in higher income countries) have consistently shown the superior performance of molecular nucleic acid testing [NAT] in detecting HPV DNA or RNA for cervical cancer screening, as these tests are more sensitive in predicting the development or detection of underlying lesions, as well as being more objective with less inter-operator variation. In higher HPV prevalence populations, such as immunocompromised women including women living with HIV (WLWH), NAT screening is suboptimal and requires biomarkers for dysplasia specificity. (See detail below under HPV NAT primary screening in LMIC) In general, a negative HPV test is associated with a substantially lower risk of cervical precancer and cancer over time than a negative cytology or negative VIA result.

International performance guidelines

In the pivotal RCTs, designed to assess the performance of HPV testing as a primary screen for cervical disease, the index assays used were Hybrid Capture II (HC2, Qiagen, Gaithersburg, MD, USA) or GP5+6+ PCR-EIA, which both detect DNA of 13 or 14 oncogenic or high-risk (hr) HPV genotypes [4-8]. The International Agency for Research on Cancer [IARC] has classified 12 HPV types as oncogenic for cervical cancer (16,18,31,33,35,39,45,51,52,56,58 and 59). These 12 types are categorised by IARC as group I carcinogens. Often hrHPV assays target two more types: HPV68 (classified as belonging to IARC group IIa, = *probably carcinogenic*) and HPV66 (group IIb = *possibly carcinogenic*). In addition, certain HPV tests target more *possibly carcinogenic* types (26, 53, 67, 70, 73, 82). These types can be found in some rare cases of cervical cancer. However, for cervical cancer screening, it is sufficient to target only the 12 oncogenic types

(group I), since adding more types only marginally increases the sensitivity for precancerous lesions that may progress to cancer, but substantially decreases the clinical specificity [9]. Of note HPV16 and HPV18 carry the greatest oncogenic risk globally with HPV16 being the most oncogenic HPV. [9]. Based on international consensus, equivalency criteria were created through which other HPV NATs could be assessed in order to determine acceptance for use in cervical population screening [10]. These criteria, known as the Meijer criteria [10], incorporated metrics for non-inferior accuracy (clinical sensitivity and specificity) and intra- & inter-laboratory reproducibility to detect cervical intraepithelial neoplasia grade 2 and worse [CIN2+] lesions in screening samples compared to either of these two standard comparator tests. However, these criteria do not include the cumulative CIN3+ incidence rate after a negative test result (screening efficiency) which is critical for screening programs with extended intervals to avoid interval (pre) cancers.

It is noteworthy that a plethora of screening NAT tests exists. A 2012 review, identified 125 distinct assays and at least 84 variants, but with very few validated in clinical practice [11] or in adequately designed studies. Subsequently in 2015, a systematic review of screening and validation studies was performed and: this yielded a short subset of assays fulfilling the international criteria [12]. To address this paucity of validated assays, including those with genotyping capabilities, the VALGENT (VALidation of HPV GENotyping Tests) framework was created by Arbyn et al to help expedite the evaluation of HPV tests according to the Meijer 2009 criteria [13]. VALGENT is a robust protocol for validation and includes capacity for the assessment of HPV assays [13]. These validation frameworks have utilised cross-sectional and longitudinal design studies of clinician-collected cervical specimens.

A recently published updated review in 2020 revealed an almost doubling of distinct HPV assays, with 254 assays with 425 variants available on the global market: yet most had no analytical or clinical evaluation [14]. Furthermore, more than 90% had no regulatory evaluation, nor were evaluated following a stringent clinical validation protocol [14]. In April 2021, an updated list of HPV assays suitable for primary cervical cancer screening was published for guidance for countries choosing to use an HPV NAT assay [15]. (See Table 1 for sufficiently validated assays to use.) In addition, at the time of preparation of this document, three HPV assays have been through the WHO prequalification process [PQ] and it is likely this number will increase. WHO PQ is a process created to provide assurance as new *in vitro* diagnostic medical devices/ products enter the global market, with respect to their quality, safety and performance. For example, for products such as HPV NAT tests, PQ serves as a quality assurance mark for WHO Member States, UN agencies, and international procurers. The process for PQ includes the following: a review of the product dossier provided by the manufacturer (product description, risk analysis, analytical and clinical validation and verification studies, stability studies, etc.); an inspection of the manufacturing site(s); and an independent performance evaluation and labeling review. PQ assessment specifically focuses on relevant aspects for resource-limited settings [16].

A recent comprehensive overview by Cuschieri et al details quality-considerations and challenges, including assay and platform validation, internal quality control selection, infrastructure to prevent sample contamination and strengths and weaknesses of external quality assurance schemes [17]. In addition, generic guidance for quality monitoring and

assurance of nucleic acid amplification tests (irrespective of target) can be helpful and some recent useful guidelines from 2010-2018 are referenced here [18-21].

Also strongly recommended is continuous monitoring of testing and participation in a real-time QC program, similar to programs used to monitor testing for other infectious agents. This is accomplished by laboratories testing an externally-sourced QC sample, which has reactivity close to the assay's lower cut-off or limit of detection with each test run and monitoring the test results of that sample over time to ensure consistency of performance. The results from a participating laboratory can be compared with those from other laboratories using the same assay and QC sample. This process provides real-time monitoring of assay analytical performance by the QC provider, and results that fall short of expected performance can be immediately investigated [22].

Apart from the use of validated assays, it is also important that all laboratory services themselves are accredited. Accredited requires that the laboratory needs to show evidence of its ability to correctly assess the content of blinded samples (proficiency testing panels). Another example of an EQA, is the international HPV DNA proficiency panel, traceable to International Standards, was assembled by the WHO HPV LabNet in 2007 and has been issued regularly ever since [23, 24]. (See table 2 for necessary measures of quality control when HPV NAT tests are performed in cervical cancer primary screening).

The future

With time and additional research, the role of extended genotyping over partial genotyping for HPV16/18 of those women found positive on screening will become evident [25]. In the future, addition of biomarkers of progression (for example host and/or viral methylation markers) will predict those with HPV 16/18 CIN2+ requiring treatment and reduce unnecessary treatment [26]. Clinical algorithms of care will therefore need continual review for modification, together with evaluation and endorsement by such bodies for example in the EU in updating the CE -IVD-R" mark. Moreover, based on data from South Africa, WLWH may benefit from a restricted genotype detection (to 8 genotypes) for cervical screening and potentially an altered threshold for positivity [27, 28].

Sampling through self-collection

Most cervical screening programs do not reach the aim of 70% coverage and for some women and other persons with a cervix who do not identify as women, the lack of culturally appropriate facilities and/or providers often results in a decision to forego a reluctance to have a pelvic examination and screening. [29] [30 Distance to reach a health care professional can also be an important barrier. Furthermore, many countries do not have the infrastructure or adequate personnel to have organized cytology screening programs. Validated HPV DNA assays based on NAT have similarly accuracy for CIN2/3+ on vaginal self-samples compared to cervical clinician-taken specimens [31]. As more cervical- screening programs are offering self-collected sampling in conjunction with NAT, particularly those who previously used VIA, this should reduce cultural, socio-economic and gender barriers to screening, and increase equity. The approach to self-sampling, whether

as a universal offer or as way to engage a “hard to reach population” is likely to differ according to country-specific infrastructure and priorities. Nevertheless, before embedding self-sampling into any screening program, appropriate validation of the each component of the screening needs assessment and includes the collection and transport devices as well as the HPV assay combination, in the laboratory(s) that will be performing the testing routinely - is key, particularly as few test manufacturers – including many of those listed in Table 1 - have a formal claim for self-sampling at time of preparation of this document. To support this, a new validation protocol for emerging HPV tests is being developed which incorporates the contemporaneous collection of a self-sample as well as a clinician taken sample. The VALHUDES protocol generates evidence on tests with similar clinical accuracy on self-versus clinician-taken samples [32]. The principle is that if the HPV test validated on clinician collected samples has similar performance on self- collected samples, then the test can be used for screening self-collected samples. Moreover, for alternative self-collection devices for an already clinically validated combination test or device, analytical test concordance (self vs clinician collected) could be accepted as sufficient evidence [33].

HPV NAT primary screening in LMIC

The highest burden of cervical cancer is seen in countries with low financial resources, lack of screening activities, non-organized and non-quality-controlled programs, high prevalence of risk factors such as endemic HIV, and with high endemic HPV prevalence also at older ages. In the context of high HPV prevalence and insufficient resources for triage, HPV NAT testing can become problematic and needs adequate modification to ensure specificity. Reduction in the spectrum of types tested for to the eight most carcinogenic HPV types has been proposed [24]. The addition of biomarkers, such as cellular HPV-induced proteins or gene expression modifying promoter methylation tests, to the screening program may be necessary to identify women in true need of therapy and to avoid overwhelming the health care system [25, 26].

Impact of COVID-19

The SARS-CoV-2 pandemic has overwhelmed hospitals limiting their capacity to perform routine gynaecological procedures. As patients with serious medical conditions that should undergo elective interventions generally do not have the same entitlement to acute care as COVID-19 patients, this one-sided prioritisation system may have life-threatening consequences, especially for affected patients diagnosed with CIN3 or with cervical cancer [34]. Lockdown measures affecting clinics, public transportation, and schools have severely compromised screening, vaccination and treatment efforts: catch-up programs should be planned and offered.

Although oncological operations or interventions in symptomatic patients have suffered during COVID-19 [35] office-based gynecologists and cervical cancer screening centers that carry the major burden of cervical screening have also reported that women attend at lower rates, as they fear acquiring COVID-19 through contact with other patients.

The SARS-CoV-2 pandemic thereby counteracts the worldwide efforts to reduce cases and mortality of cervical cancer. However, a silver lining of the COVID-19 pandemic has been the increased availability of molecular platforms and appropriate staff qualified in molecular technology in most countries due to the rapid scale up of COVID-19 testing: these can now also be deployed for HPV testing. Most countries have also built up a re-call system (for those who are COVID-19 positive) which can be utilized for managing those with positive HPV screening tests. It is noteworthy that in many countries, the idea of recalling patients was thought to be nearly impossible. Yet, COVID-19 has forced the strengthening of public health infrastructure with its effective recall system, the infrastructure which must be incorporated into cervical screening in the long term.

Education and Addressing stigma

Switching from cytology to HPV testing is a major paradigm change. While it is important to consider the technical requirements along with QA and QC issues, professionals and programs must place equal emphasis on education and addressing the potential stigma associated with HPV testing. The majority of healthcare professionals and the public still focus on HPV as a sexually transmitted DISEASE (rather than an usually asymptomatic, transient and near universal infection for which we have effective prophylactic vaccines to prevent associated cancers) and this needs to be addressed if we are to achieve the 70% uptake among eligible women. To emphasise this point, the mucosal HPVs are largely (although not exclusively) transmitted by sexual contact (even intimate genital skin to genital skin) and are a very common infection, whereas the resultant diseases (clinical manifestations eg warts, precancerous lesions, cancer) are less so. Accordingly, with a reduction in infection from widespread vaccination, reductions in HPV-related diseases will occur. Scientific progress made in terms of technology must be matched with appropriate messaging/education for it to be accepted widely. Therefore, it is imperative that the narrative for primary HPV testing be standardized and agreed locally with tailoring to the local community as required.

Conclusion

Guidelines for cervical screening using NAT assays have recently been published by WHO [36] and in mid-2022 assessed the role for use of HPV mRNA detection as a primary screening test [37]. However, these guidelines, which support the use of HPV NAT detection in a screen, triage and treat or a screen and treat approach starting at the age of 30 years with regular screening every 5 to 10 years, recommend regulatory approved HPV assays. We await the next phase of the development of implementation guidelines for laboratories which will focus on choice of assays.

We emphasise that assays validated for clinical screening must be able to detect those with prevalent or incipient CIN3+ and are different to those adopted for surveillance (vaccine impact/effectiveness) which are of higher analytical sensitivity, detecting a lower quantity of HPV, not necessarily of clinical relevance. The cost of clinical screening assays needs to be affordable for all. Continuous and rigorous QA and QC checks are essential. In processing self-collected samples, amplification assays should be used. Although the

COVID-19 pandemic has created several challenges including limited reagent availability, more laboratories are now equipped with the knowledge and rapid large-throughput instruments to ultimately use for HPV assays as a result of the pandemic. Although the pandemic has taught us how screening and vaccination programs can be compromised, it has also highlighted the importance of catch up and returning to routines.

Acknowledgement

The findings and conclusions in this report are those of the authors and do not necessarily represent the official position of the Centers for Disease Control and Prevention.

Funding

This work is supported by National Health and Medical Research Council Investigator Grant APP1197951 (SMG).

Appendix

Members of the IPVS Policy Committee are Garland SM (Chair), Stanley M, Moscicki AB, Woo YL, Bhatla N, Chan, K, Palefsky J, Giuliano AR, Kaufmann A, de Sanjose S, Brotherton JML, Feldman S.

References

- [1]. WHO Director-General calls for all countries to take action to help end the suffering caused by cervical cancer [cited 2019. Available from: <https://www.who.int/news/item/18-05-2018-who-dg-calls-for-all-countries-to-take-action-to-help-end-the-suffering-caused-by-cervical-cancer>.
- [2]. Tan SY, Tatsumura Y. George Papanicolaou (1883-1962): Discoverer of the Pap smear. *Singapore Med J.* 2015;56(10):586–7. [PubMed: 26512152]
- [3]. Khan MJ, Castle PE, Lorincz AT, Wacholder S, Sherman M, Scott DR, et al. The elevated 10-year risk of cervical precancer and cancer in women with human papillomavirus (HPV) type 16 or 18 and the possible utility of type-specific HPV testing in clinical practice. *J Natl Cancer Inst.* 2005;97(14):1072–9. [PubMed: 16030305]
- [4]. Ronco G, Giorgi-Rossi P, Carozzi F, Confortini M, Dalla Palma P, Del Mistro A, et al. Efficacy of human papillomavirus testing for the detection of invasive cervical cancers and cervical intraepithelial neoplasia: a randomised controlled trial. *Lancet Oncol.* 2010;11(3):249–57. [PubMed: 20089449]
- [5]. Arbyn M, Ronco G, Anttila A, Meijer CJ, Poljak M, Ogilvie G, et al. Evidence regarding human papillomavirus testing in secondary prevention of cervical cancer. *Vaccine.* 2012;30 Suppl 5:F88–99. [PubMed: 23199969]
- [6]. Wright TC, Stoler MH, Behrens CM, Sharma A, Zhang G, Wright TL. Primary cervical cancer screening with human papillomavirus: end of study results from the ATHENA study using HPV as the first-line screening test. *Gynecol Oncol.* 2015;136(2):189–97. [PubMed: 25579108]
- [7]. Haedicke J, Iftner T. A review of the clinical performance of the Aptima HPV assay. *J Clin Virol.* 2016;76 Suppl 1:S40–s8. [PubMed: 26614686]
- [8]. Ogilvie GS, van Niekerk D, Krajden M, et al. Effect of Screening With Primary Cervical HPV Testing vs Cytology Testing on High-grade Cervical Intraepithelial Neoplasia at 48 Months: The HPV FOCAL Randomized Clinical Trial. *JAMA.* 2018;320(1):43–52. doi:10.1001/jama.2018.7464 [PubMed: 29971397]
- [9]. Bouvard V, Baan R, Straif K, Grosse Y, Secretan B, El Ghissassi F, et al. A review of human carcinogens—Part B: biological agents. *Lancet Oncol.* 2009;10(4):321–2. [PubMed: 19350698]
- [10]. Meijer CJ, Berkhof J, Castle PE, Hesselink AT, Franco EL, Ronco G, et al. Guidelines for human papillomavirus DNA test requirements for primary cervical cancer screening in women 30 years and older. *Int J Cancer.* 2009;124(3):516–20. [PubMed: 18973271]

- [11]. Poljak M, Cuzick J, Kocjan BJ, Iftner T, Dillner J, Arbyn M. Nucleic acid tests for the detection of alpha human papillomaviruses. *Vaccine*. 2012;30 Suppl 5:F100–6. [PubMed: 23199952]
- [12]. Arbyn M, Snijders PJ, Meijer CJ, Berkhof J, Cuschieri K, Kocjan BJ, et al. Which high-risk HPV assays fulfil criteria for use in primary cervical cancer screening? *Clin Microbiol Infect*. 2015;21(9):817–26. [PubMed: 25936581]
- [13]. Arbyn M, Depuydt C, Benoy I, Bogers J, Cuschieri K, Schmitt M, et al. VALGENT: A protocol for clinical validation of human papillomavirus assays. *J Clin Virol*. 2016;76 Suppl 1:S14–s21. [PubMed: 26522865]
- [14]. Poljak M, Oštrbenk Valen ak A, Gimpelj Domjani G, Xu L, Arbyn M. Commercially available molecular tests for human papillomaviruses: a global overview. *Clin Microbiol Infect*. 2020;26(9):1144–50. [PubMed: 32247892]
- [15]. Arbyn M, Simon M, Peeters E, Xu L, Meijer C, Berkhof J, et al. 2020 list of human papillomavirus assays suitable for primary cervical cancer screening. *Clin Microbiol Infect*. 2021;27(8):1083–95. [PubMed: 33975008]
- [16]. Overview of the WHO prequalification of in vitro diagnostics assessment: prequalification of in vitro diagnostics. Version 9. Geneva: World Health Organization; 2021. Licence: CC BY-NC-SA 3.0 IGO [Available from: https://extranet.who.int/pqweb/sites/default/files/documents/21-01-27-Overview-DX-Prequalification-Requirements-PQDx_007-v9.pdf].
- [17]. Cuschieri K, Schuurman R, Coughlan S. Ensuring quality in cervical screening programmes based on molecular human papillomavirus testing. *Cytopathology*. 2019;30(3):273–80. [PubMed: 30657615]
- [18]. Miller JM, Binnicker MJ, Campbell S, Carroll KC, Chapin KC, Gilligan PH, et al. A Guide to Utilization of the Microbiology Laboratory for Diagnosis of Infectious Diseases: 2018 Update by the Infectious Diseases Society of America and the American Society for Microbiology. *Clin Infect Dis*. 2018;67(6):e1–e94. [PubMed: 29955859]
- [19]. Members of the 2013 and 2014 AMP Clinical Practice. Molecular Diagnostic Assay Validation: Update to the 2009 AMP Molecular Diagnostic Assay Validation White Paper. Association for Molecular Pathology. September, 2014. Accessed 6/15/20.
- [20]. CLSI. Validation and Verification of Multiplex Nucleic Acid Assays. 2nd ed. CLSI guideline MM17. Wayne, PA: Clinical and Laboratory Standards Institute; 2018.
- [21]. Burd EM. Validation of laboratory-developed molecular assays for infectious diseases. *Clin Microbiol Rev*. 2010;23(3):550–76. [PubMed: 20610823]
- [22]. Garland SM, Dimech W, Collignon P, Cooley L, Nimmo GR, Smith DW, et al. The new screening program to prevent cervical cancer using HPV DNA: getting the balance right in maintaining quality. *J Pathol Clin Res*. 2018;4(4):207–12. [PubMed: 30058126]
- [23]. Eklund C, Zhou T, Dillner J. Global proficiency study of human papillomavirus genotyping. *J Clin Microbiol*. 2010;48(11):4147–55. [PubMed: 20844222]
- [24]. Eklund C, Mühr LSA, Lagheden C, Forslund O, Robertsson KD, Dillner J. The 2019 HPV Labnet international proficiency study: Need of global Human Papillomavirus Proficiency Testing. *J Clin Virol*. 2021;141:104902. [PubMed: 34186414]
- [25]. Demarco M, Hyun N, Carter-Pokras O, Raine-Bennett TR, Cheung L, Chen X, et al. A study of type-specific HPV natural history and implications for contemporary cervical cancer screening programs. *EClinicalMedicine*. 2020;22:100293. [PubMed: 32510043]
- [26]. Kremer WW, van Zummeren M, Breytenbach E, Richter KL, Steenbergen RDM, Meijer C, et al. The use of molecular markers for cervical screening of women living with HIV in South Africa. *Aids*. 2019;33(13):2035–42. [PubMed: 31385866]
- [27]. Johnson LG, Saidu R, Mbulawa Z, Williamson AL, Boa R, Tergas A, et al. Selecting human papillomavirus genotypes to optimize the performance of screening tests among South African women. *Cancer Med*. 2020;9(18):6813–24. [PubMed: 32706163]
- [28]. Kuhn L, Saidu R, Boa R, Tergas A, Moodley J, Persing D, et al. Clinical evaluation of modifications to a human papillomavirus assay to optimise its utility for cervical cancer screening in low-resource settings: a diagnostic accuracy study. *Lancet Glob Health*. 2020;8(2):e296–e304. [PubMed: 31981559]

- [29]. Arbyn M, Smith SB, Temin S, Sultana F, Castle P. Detecting cervical precancer and reaching underscreened women by using HPV testing on self samples: updated meta-analyses. *Bmj*. 2018;363:k4823. [PubMed: 30518635]
- [30]. Arbyn M, Peeters E, Benoy I, Vanden Broeck D, Bogers J, De Sutter P, et al. VALHUDES: A protocol for validation of human papillomavirus assays and collection devices for HPV testing on self-samples and urine samples. *J Clin Virol*. 2018;107:52–6. [PubMed: 30195193]
- [31]. Saville M, Hawkes D, Keung MHT, Ip ELO, Silvers J, Sultana F, Malloy MJ, Velentzis LS, Canfel KI, Wrede CD, Brotherton JML, Analytical performance of HPV assays on vaginal self-collected vs practitioner-collected cervical samples: the SCoPE study, *J of Clin Virol*. 2020:127, 1–7.
- [32]. Giorgi Rossi P, Fortunato C, Barbarino P et al. Self-sampling to increase participation in cervical cancer screening: an RCT comparing home mailing, distribution in pharmacies, and recall letter. *Br J Cancer* 112, 667–675 (2015). 10.1038/bjc.2015.11 [PubMed: 25633037]
- [33]. Arbyn M, Castle PE, Schiffman M, Wentzensen N, Heckman-Stoddard B, Sahasrabudhe VV. Meta-analysis of agreement/concordance statistics in studies comparing self- vs clinician-collected samples for HPV testing in cervical cancer screening. *Int J Cancer*. 2022.
- [34]. Matsuo K, Huang Y, Matsuzaki S, Klar M, Wright JD. Effect of delay in surgical therapy for early-stage cervical cancer: An implication in the coronavirus pandemic. *Eur J Cancer*. 2020 Nov;139:173–176. doi: 10.1016/j.ejca.2020.08.025 [PubMed: 32992156]
- [35]. Glasbey J, Ademuyiwa A, Adisa A, AlAmeer E, Arnaud AP, Ayasra F, et al. Effect of COVID-19 pandemic lockdowns on planned cancer surgery for 15 tumour types in 61 countries: an international, prospective, cohort study. *The Lancet Oncology*. 2021;22(11):1507–17. [PubMed: 34624250]
- [36]. Human papillomavirus (HPV) nucleic acid amplification tests (NAATs) to screen for cervical pre-cancer lesions and prevent cervical cancer: World Health Organization; [updated 16 March 2022. Policy brief. Available from: <https://www.who.int/publications/i/item/9789240045248>.
- [37]. Arbyn M, Simon M, de Sanjosé S, Clarke MA, Poljak M, Rezhake R, et al. Accuracy and effectiveness of HPV mRNA testing in cervical cancer screening: a systematic review and meta-analysis. *The Lancet Oncology*. 2022;23(7):950–60. [PubMed: 35709810]

Table 1.

List of HPV assays validated for cervical cancer screening as stand-alone test or in combination with cytology. Adapted from Arbyn et al CMI 2021 [15].

•	Hybrid Capture 2 HPV DNA Test (Qiagen, Gaithersburg, MD, USA)
•	GP5+/6+ PCR-EIA (Diassay, Rijkswijk, the Netherlands)
•	Abbott RealTime High Risk HPV Test *** (Abbott Molecular, Des Plaines, IL, USA)
•	Anyplex II HPV HR Detection (Seegene, Seoul, South Korea)
•	BD Onclarity HPV Assay* (BD Diagnostics, Sparks, MD, USA)
•	Cobas 4800 HPV Test* (Roche Molecular Diagnostics, Pleasanton, CA, USA)
•	HPV-Risk Assay (Self-Screen BV, Amsterdam, The Netherlands)
•	PapilloCheck HPV-Screening Test (Greiner Bio-One, Frickenhausen, Germany)
•	Xpert HPV *** (Cepheid, Sunnyvale, CA, USA)
•	Alinity m HR HPV Assay (Abbott Molecular, Des Plaines, IL, USA)
•	Cobas 6800/8800 HPV Test* (Roche Molecular Diagnostics, Pleasanton, CA, USA)
•	APTIMA HPV Assay** (Hologic, Bedford, MA, USA)

* FDA approved for HPV alone primary screening

** FDA approved only with co-testing with cytology

*** WHO Prequalification of In Vitro Diagnostics

NB CareHPV™ *** Test is prequalified (2018), but not formally validated according to Meijer's criteria and/or VALGENT

Table 2.

Recommended QC measures for HPV NAT cervical cancer screening tests.

<p><u>Assay-specific</u></p> <ul style="list-style-type: none">• Including the 12 HR-HPV genotypes classified as carcinogenic (IARC group I)• Not including more than 13 to 14 genotypes in total, if not a fully genotyping assay• No additional HR-HPV that reduce the clinical specificity• No low-risk HPV genotypes included• Clinically validated assay or validated in comparison to standard comparator tests HCII or GP5+/6+ PCR• Internal control for sample cellularity• Positive control for assay performance• Negative control for excluding contamination• Approved as IVD by relevant approval bodies (e.g., FDA, CE, WHO) <p><u>Laboratory-specific</u></p> <ul style="list-style-type: none">• Infrastructure to prevent sample contamination with post PCR products• Participation in regular proficiency testing of external samples• Monitoring of processes• QC/QA management
--

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript