

UCSF

UC San Francisco Previously Published Works

Title

The fourth generation Alere™ HIV Combo rapid test improves detection of acute infection in MTN-003 (VOICE) samples

Permalink

<https://escholarship.org/uc/item/6xr217v9>

Authors

Livant, Edward
Heaps, Amy
Kelly, Cliff
[et al.](#)

Publication Date

2017-09-01

DOI

10.1016/j.jcv.2017.06.006

Copyright Information

This work is made available under the terms of a Creative Commons Attribution License, available at <https://creativecommons.org/licenses/by/4.0/>

Peer reviewed



Published in final edited form as:

J Clin Virol. 2017 September ; 94: 15–21. doi:10.1016/j.jcv.2017.06.006.

The fourth generation Alere™ HIV Combo rapid test improves detection of acute infection in MTN-003 (VOICE) samples

Edward Livant^{a,1}, Amy Heaps^{b,1}, Cliff Kelly^c, Rashika Maharaj^d, Natasha Samsunder^e, Lindiwe Nhlangulela^f, Patrick Karugaba^g, Ravindre Panchia^h, Jeanne Mrazzazzo^{i,2}, Zvavahera Mike Chirenje^j, Urvi M. Parikh^{b,*}, and on behalf of the VOICE Study Team

^aMagee-Womens Research Institute, Pittsburgh, PA, United States ^bUniversity of Pittsburgh, Pittsburgh, PA, United States ^cSCHARP, Seattle, WA, United States ^dSouth Africa Medical Research Council HIV Prevention Research Unit, Durban, South Africa ^eCAPRISA, Durban, South Africa ^fAurum Institute, Klerksdorp, South Africa ^gMU-JHU Research Collaboration, Kampala, Uganda ^hPerinatal HIV Prevention Unit, University of the Witwatersrand, Johannesburg, South Africa ⁱUniversity of Washington, Seattle, WA, United States ^jUZ-UCSF, Harare, Zimbabwe

Abstract

Background—Early and accurate detection of HIV is crucial when using pre-exposure prophylaxis (PrEP) for HIV prevention to avoid PrEP initiation in acutely infected individuals and to minimize the risk of drug resistance in individuals with breakthrough infection.

Objective—To determine if fourth-generation antigen/antibody (Ag/Ab) rapid diagnostic tests (RDT) would have detected HIV infection earlier than the third-generation RDT used in MTN-003 (VOICE).

Study design—5029 VOICE participants were evaluated with third-generation Alere Determine™ HIV-1/2, OraQuick ADVANCE® Rapid HIV-1/2, Uni-Gold™ Recombigen® HIV-1/2 and Bio-Rad GS HIV-1/2 + O EIA; and fourth-generation Alere Determine™ HIV-1/2 Ag/Ab Combo, Conformité Européenne (CE)-Marked Alere™ HIV Combo and Bio-Rad HIV

*Corresponding Author: Urvi M. Parikh, Ph.D S804 Scaife Hall, 3550 Terrace St Pittsburgh, PA 15261, United States. ump3@pitt.edu (U.M. Parikh).

¹These authors contributed equally to this work.

²Current affiliation is University of Alabama, Birmingham, AL, United States.

Author contributions

EL, AH and UMP designed the study, performed biological assays, interpreted data, and wrote the manuscript. CK performed statistical analyses. RM, NS, LN, PK and RP oversaw local HIV testing; collection, processing and shipment of plasma. JM and ZMC designed and implemented the VOICE study. All authors revised and approved the manuscript.

Ethical approval

The study was conducted in accordance with the Declaration of Helsinki and patients gave their written consent for the study. The local research ethic committee approved the study (Comité de Protection des Personnes Sud Méditerranée IV, Q2011.12.01). The MTN Laboratory Center operates under University of Pittsburgh IRB approval REN16030130. Additional ethical approval was obtained for this testing at select sites: CAPRISA eThekweni (BREC BE184/5); MU-JHU (UNCST HS 1760); WRHI (University of Witwatersrand 090408); MRC (MRC EC08-011) and CAPRISA Aurum (BREC BE060/15).

Competing interests

The authors declare that they have no competing interests.

Appendix A. Supplementary data

Supplementary data associated with this article can be found, in the online version, at <http://dx.doi.org/10.1016/j.jcv.2017.06.006>.

Combo Ag/Ab EIA. Multispot[®], GS HIV-1 Western Blot (WB) and Geenius[™] (Bio-Rad) were also evaluated.

Results—Of 57 antibody-negative pre-seroconversion plasma samples with HIV RNA > 20 copies/mL identified, 16 (28%) were reactive by CE-Marked Alere[™] HIV Combo (1 Ab; 9 Ag; 6 Ag/Ab reactive) and 4 (7%) by Alere Determine[™] HIV-1/2 Ag/Ab Combo (2 Ab; 2 Ag; 0 Ag/Ab reactive) ($p = 0.0005$). Multispot[®] confirmed only 1 of 16 acute infections while WB and Geenius[™] confirmed none. GS HIV Combo Ag/Ab EIA identified 27 of 57 (47%) pre-seroconversion RNA-positive samples.

Conclusion—In VOICE, 28% of infections missed by current third-generation RDT would have been identified with the use of CE-Marked Alere[™] HIV Combo. Geenius[™], Multispot[®] and WB were all insensitive (< 10%) in confirming infections detected by fourth-generation assays. An improved diagnostic algorithm that includes a fourth-generation RDT with HIV RNA testing will be essential for efficiently identifying seroconverters on PrEP.

Keywords

HIV-1; Acute infection; 4th generation rapid test; Pre-exposure prophylaxis; HIV confirmatory test; HIV diagnostics

1. Background

Early and accurate detection of HIV is crucial when using pre-exposure prophylaxis (PrEP) for HIV prevention to avoid PrEP initiation in acutely infected individuals and to minimize the risk of drug resistance in individuals with breakthrough infection [1]. The World Health Organization recommends two sequential rapid tests for HIV diagnosis [2] however third generation rapid tests detect only HIV antibodies and may miss up to 75% of early acute HIV infection cases [3].

Fourth generation HIV rapid tests use a lateral flow cassette to separately assay for both anti-HIV antibodies and p24 antigen, however numerous studies have demonstrated the United States Food and Drug Administration (FDA)-approved Alere Determine[™] HIV-1/2 Ag/Ab Combo to be insensitive for detection of acute infection. In a total of 54 acute seroconversion samples from 3 studies of acute infection from Malawi, Swaziland, Rwanda and Zambia, only 1 sample (1.9%) was positive for the p24 antigen component [4–6]. Reported sensitivity for antigen detection was higher in samples from the United Kingdom (50%), the United States (45.5% and 75.8%), Italy (88.2%) and in mixed-subtype seroconversion panels (86.6%) suggesting clade differences [3,7–10]. Evaluation of Gag-expressing virus-like particles (VLP) showed that 6 of 7 subtype B VLPs were detected by Alere Determine[™] HIV-1/2 Ag/Ab Combo, compared to only 1 of 5 subtype C, and 3 of 31 non B/non-C-subtypes [11].

In February 2015, Alere released a re-formulated fourth generation rapid test kit, the Conformité Européenne (CE)-Marked Alere[™] HIV Combo. Only one study to date has evaluated this new assay, reporting an 88% sensitivity compared to the Abbott Architect HIV Ag/Ab Combo assay using stored plasma or serum samples from the United Kingdom [12]. Also unknown is the performance of HIV confirmatory tests such as Bio-Rad

Multispot[®] HIV-1/HIV-2 Rapid Test, Geenius[™] HIV-1/2 Supplemental Assay and GS HIV-1 Western Blot in the context of an algorithm that includes fourth generation rapid tests. This is the first evaluation of the new CE-Marked Alere[™] HIV Combo in pre-seroconversion samples collected in South Africa, Uganda and Zimbabwe.

2. Objectives

The objective of this study was to evaluate the performance of third and fourth generation HIV rapid diagnostic tests and confirmatory assays to detect acute infection in a large cohort of participants from the MTN-003 (VOICE) HIV prevention trial.

3. Study design

3.1. Study population and specimen collection

VOICE ([ClinicalTrials.gov](https://clinicaltrials.gov/ct2/show/study/NCT00705679) NCT00705679) was a randomized, Phase 2 B placebo-controlled trial to evaluate the safety and effectiveness of oral tenofovir disoproxil fumarate (TDF), oral TDF-emtricitabine (FTC) and vaginal tenofovir 1% gel for the prevention of HIV infection in 5029 HIV-uninfected women from 15 clinical sites in South Africa, Uganda and Zimbabwe. Population characteristics and trial results have been described previously [13].

In VOICE, participants were monitored monthly for seroconversion with one (Uganda) or two (South Africa and Zimbabwe) third generation HIV rapid diagnostic tests (RDT) performed point-of-care using venous or fingerstick-drawn whole blood at clinical research sites. RDTs used included Alere Determine[™] HIV-1/2 (Alere Medical Co. Ltd., Matsudo, Japan), OraQuick ADVANCE[®] Rapid HIV-1/2 Antibody Test (OraSure Technologies, Inc., Bethlehem, PA) and/or Uni-Gold[™] Recombigen[®] HIV-1/2 rapid test (Trinity Biotech[™] Wicklow, Ireland). To diagnose HIV infection, clinical research sites followed an HIV algorithm during the study (Fig. 1). Positive or discordant RDT results were confirmed by GS HIV-1 Western Blot (Bio-Rad Laboratories, Redmond, WA) using plasma from a separate draw and HIV-1 RNA PCR using Abbott RealTime HIV-1 (Abbott Molecular, Inc., Des Plaines, IL) was performed on plasma following an indeterminate or negative Western blot result to assess acute infection (Fig. 1). Assay interpretation was defined by the manufacturer's package insert for each test. In the VOICE study, plasma from the visit at which seroconversion was detected and stored plasma from the seroconverting participant's enrollment visit were shipped to the University of Pittsburgh and re-tested to verify HIV infection using GS HIV-1/2 + O EIA (Bio-Rad), GS HIV-1 Western Blot and HIV-1 RNA PCR (Abbott).

3.2. Identification of specimens from acutely infected VOICE participants

After completion of the VOICE study, stored plasma specimens collected up to 91 days prior to detection of seroconversion were shipped to the University of Pittsburgh. Pre-seroconversion plasma specimens were tested for the presence of HIV-1 RNA by the COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HIV-1 Test, v2.0 with a limit of quantitation of 20 copies/mL (Roche Diagnostics, Indianapolis, IN). Specimens with detectable HIV-1 RNA (including results of < 20 copies/mL, detected) were re-tested with OraQuick ADVANCE[®] and Uni-Gold[™] Recombigen[®] and if antibody negative, were considered to be from an

acutely infected participant. Only those specimens identified as “pre-seroconversion” defined as having detectable HIV-1 RNA and negative antibody reactivity were used for the evaluation of fourth generation RDTs.

3.3. Third and fourth generation diagnostic test evaluations

Third generation RDT-negative samples were tested with the fourth generation FDA-approved Alere™ Determine™ HIV-1/2 Ag/Ab Combo rapid test (Alere North America, Waltham, MA) and the CE-Marked Alere™ HIV Combo Rapid Test (Alere Medical Co. Ltd., Matsudo-cho, Japan) (Supplementary Table S1). The Determine™ Combo and Alere™ HIV Combo RDTs are named similarly but have different formulations, different places of manufacture, and different performance specifications. Samples were also tested in duplicate by the third generation Bio-Rad GS HIV-1/2 + O Enzyme Immunoassay (EIA) and the fourth generation Bio-Rad GS HIV Combo Ag/Ab EIA, both FDA-Approved (Bio-Rad Laboratories).

3.4. Confirmatory test evaluations

Pre-seroconversion samples found to be positive by at least one of the above fourth generation tests were further tested for the presence of antibodies to HIV-1 and/or HIV-2 by Multispot® HIV-1/HIV-2 Rapid Test, Geenius™ HIV-1/2 Supplemental Assay and GS HIV-1 Western Blot (Bio-Rad).

3.5. Statistical analysis

McNemar’s test was used to compare the proportion of positive/reactive test results between assays. For calculating sensitivity, specificity, positive and negative predictive value of the results of the third generation HIV RDTs performed in VOICE, the gold standard was defined by the results of the study algorithm (Fig. 1) which included additional WB and/or HIV-1 RNA testing by a local and/or central laboratory during the study. Samples that tested negative or indeterminate for RDT and/or WB but had detectable HIV-1 RNA were considered endpoints in the VOICE study. The gold standard for the same calculations on the results of the HIV tests evaluated in this study was detectable or undetectable HIV-1 RNA by Roche TaqMan. All confidence intervals are exact (Clopper-Pearson) 95% confidence intervals.

4. Results

4.1. Third generation rapid test performance in VOICE

Between 2009 and 2011, 5029 enrolled VOICE participants had 77,915 follow-up visits to screen for HIV infection. Of the 72,169 visits where two rapid tests were used, 55,346 (77%) were conducted with Uni-Gold™ Recombigen® and Determine™ HIV-1/2 while 16,823 (23%) were conducted using OraQuick ADVANCE® and Determine™ HIV-1/2. The rate of discordancy was similar for both rapid test combinations (0.15%). Ultimately, rapid testing correctly identified HIV infection status for 312 seroconverters and 4717 non-seroconverters at 72,055 of 72,169 (99.84%) follow-up visits in VOICE, however final HIV status differed from rapid test results at 114 visits (50 visits had a false positive rapid result while 64 visits had a false negative rapid result) (Fig. 2). Actual sensitivity for all 3 tests was

markedly lower than manufacturer's claims; Determine™ HIV-1/2 had the highest sensitivity at 88.78%, followed by Uni-Gold™ Recombigen® at 85.86% and OraQuick ADVANCE® at 61.54% (Table 1).

4.2. Pre-Seroconversion HIV-1 RNA

The low sensitivity of third generation rapid tests for detecting seroconversion in VOICE was investigated using pre-seroconversion plasma. Of 312 participants who seroconverted in VOICE, stored plasma collected within 91 days prior to the date of the first positive point-of-care rapid test was available for 215 HIV-1 seroconverters from 246 visits. HIV-1 RNA was performed on 230 of 246 (93%) specimens (14 of 16 not tested had an undetectable HIV-1 RNA result at a closer visit; 2 of 16 had specimen integrity issues). Of 230 specimens, HIV-1 RNA was detected in 32 of 48 (67%) collected 12–29 days pre-seroconversion, 33 of 98 (34%) collected 30–59 days pre-seroconversion, and 7 of 84 (8%) collected 60–91 days pre-seroconversion (Table 2).

4.3. 4th generation rapid tests and immunoassays

Of 72 RNA-positive pre-seroconversion plasma specimens, 15 were positive for Uni-Gold® Recombigen® and/or OraQuick Advance® upon re-testing and were excluded from further analysis (Table 2). Of the remaining 57, 16 (28%) were positive by CE-HIV Combo RDT (1 Ab only, 9 Ag only, 6 dual Ag/Ab) and only 4 (7%) were positive by FDA-Determine™ Combo RDT (2 Ab only, 2 Ag only, 0 dual Ag/Ab) (Table 3). The CE-HIV Combo RDT detected 21% more infections than the FDA-Determine™ Combo RDT ($p = 0.0005$). 27 of 57 (47%) were positive by HIV Combo EIA and 8 of 57 (14%) were positive by HIV-1/2 + O EIA. HIV Combo EIA detected 33% more infections than HIV-1/2 + O EIA ($p < 0.0001$) (Fig. 3). Specificity was high ($> 98.7\%$) for all fourth generation tests, while sensitivity was highest for HIV Combo EIA (57.75%) followed by CE-HIV Combo RDT (42.25%) and Determine™ Combo RDT (23.94%) (Table 1).

4.4. Confirmatory tests

Because some HIV testing algorithms require confirmatory testing, we evaluated the concordance of fourth generation rapid test results with three Bio-Rad HIV confirmatory tests, Western blot, Multispot® and Geenius™. Of 16 reactive CE-HIV Combo RDT samples, 9 were indeterminate by Western blot. Only one of 16 was reactive by Multispot®, and no samples were reactive by Geenius™, with the exception of two specimens that were false reactive with a gp140 band yielding an HIV-2 indeterminate result (Table 3).

5. Discussion

Using pre-seroconversion plasma specimens from the VOICE study, we evaluated fourth generation RDT to determine if acute HIV infection would have been detected earlier compared to the third generation RDT used in the trial. The high incidence of acute infections in VOICE resulted in much lower sensitivities of third generation RDT (61.5% to 88.8%) relative to manufacturers' claims (99.6% to 100%) and confirmatory tests also performed poorly in detecting acute infection. In our study, 28% of infections from participants with detectable HIV-1 RNA and negative third generation RDT were detected

by CE-HIV Combo. These data support the use of CE-HIV Combo in HIV diagnostic algorithms in high incidence settings.

Numerous studies have found the antigen component in fourth generation RDT to be insensitive and have reported that these tests do not provide any advantage for the diagnosis of acute infection [4–6,9,10,14–17]. Notably, we found major differences in test performance using two different versions of the same rapid test from the same company; CE-HIV Combo detected 21% more acute infections than FDA-Determine™ Combo. This is the first study to directly compare both versions of the Alere™ HIV Combo tests, and it highlights the importance of using the appropriate kit in diagnostic algorithms. The 4th Generation HIV Combo EIA had the highest improvement in detection of any test evaluated (47%) (Fig. 3), however, cannot be used in algorithms requiring point-of-care testing.

Acute HIV detection is paramount for both PrEP clinical trials and PrEP rollout. In trials, delayed detection of infection can increase the number of person-years a participant is on an investigational product before reaching an endpoint, and more importantly, how long a participant continues taking an investigational product while already HIV infected. In five trials of TDF/FTC PrEP, the greatest risk of drug resistance occurred in participants who enrolled with undetected acute infection [1]. In VOICE, the majority of seroconverters obtained their HIV status within 3 months of acquiring HIV infection due to monthly testing, but the interval between infection and diagnosis in PrEP users outside the trial setting will be longer due to quarterly or less frequent HIV testing. Modeling studies are needed to evaluate the cost-benefit of implementing the CE-HIV Combo in PrEP programs.

Our study had several limitations. Plasma from seroconverters was stored quarterly, limiting the precision for calculating the exact number of days that the CE-HIV Combo could detect infection earlier than third generation RDT. We found that 15 of 230 specimens that tested negative by Uni-Gold™ Recombigen® and/or OraQuick ADVANCE® at clinical sites tested positive upon re-testing with stored plasma using the same RDT, suggesting that actual test sensitivity may be lower in point-of-care settings using fingerstick-collected blood compared to research settings due to imprecision in execution of test procedures.

The updated CE-Marked Alere™ HIV Combo Rapid Test has potential for improving current HIV algorithms for detection of acute HIV infection. Better confirmatory tests are needed to avoid discrepant results that require nucleic acid testing to resolve. Shortening the window period with antigen/antibody HIV tests will benefit individuals on PrEP for timely HIV diagnosis, and could help lower population-based incidence by decreasing transmission of HIV during acute infection.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgments

We would like to acknowledge the VOICE participants for their time and effort in this study. We also thank the entire VOICE team, especially the site laboratory staff involved in administrating and performing the rapid tests at each of the MTN clinical sites: MRC, Durban South Africa; CAPRISA eThekweni, Durban South Africa; WRHI


Johannesburg, South Africa; CAPRISA Aurum, Klerksdorp, South Africa; UZ-UCSF Harare Zimbabwe; MU-JHU, Kampala Uganda.

Funding

This work was supported by the Microbicide Trials Network, which is funded by the National Institute of Allergy and Infectious Diseases (5U01AI068633), with co-funding from the National Institute of Child Health and Human Development and the National Institute of Mental Health, all components of the U.S. National Institutes of Health.

References

1. Parikh UM, Mellors JW. Should we fear resistance from tenofovir/emtricitabine preexposure prophylaxis? *Curr. Opin. HIV AIDS*. 2016; 11(1):49–55. [PubMed: 26633640]
2. WHO. Service Delivery Approaches to HIV Testing and Counselling (HTC): a Strategic HTC Programme Framework. World Health Organization; Geneva: 2012. http://apps.who.int/iris/bitstream/10665/75206/1/9789241593877_eng.pdf
3. Patel P, et al. Rapid HIV screening: missed opportunities for HIV diagnosis and prevention. *J. Clin. Virol.* 2012; 54(1):42–47. [PubMed: 22381919]
4. Rosenberg NE, et al. Detection of acute HIV infection: a field evaluation of the Determine[®] HIV-1/2 Ag/Ab combo test. *J. Infect. Dis.* 2012; 205(4):528–534. [PubMed: 22207651]
5. Duong YT, et al. Poor performance of the determine HIV-1/2 Ag/Ab combo fourth-generation rapid test for detection of acute infections in a National Household Survey in Swaziland. *J. Clin. Microbiol.* 2014; 52(10):3743–3748. [PubMed: 25122853]
6. Kilembe W, et al. Failure of a novel, rapid antigen and antibody combination test to detect antigen-positive HIV infection in African adults with early HIV infection. *PLoS One*. 2012; 7(6):e37154. [PubMed: 22715363]
7. Fox J, Dunn H, O'Shea S. Low rates of p24 antigen detection using a fourth-generation point of care HIV test. *Sex. Transm. Infect.* 2011; 87(2):178–179. [PubMed: 21084439]
8. Faraoni S, et al. Evaluation of a rapid antigen and antibody combination test in acute HIV infection. *J. Clin. Virol.* 2013; 57(1):84–87. [PubMed: 23380659]
9. Masciotra S, et al. Performance of the Alere Determine HIV-1/2 Ag/Ab Combo rapid test with specimens from HIV-1 seroconverters from the US and HIV-2 infected individuals from ivory coast. *J. Clin. Virol.* 2013; 58(Suppl 1):e54–8. [PubMed: 23911678]
10. Beelaert G, Fransens K. Evaluation of a rapid and simple fourth-generation HIV screening assay for qualitative detection of HIV p24 antigen and/or antibodies to HIV-1 and HIV-2. *J. Virol. Methods*. 2010; 168(1–2):218–222. [PubMed: 20561542]
11. Vetter BN, et al. Generation of a recombinant Gag virus-like-particle panel for the evaluation of p24 antigen detection by diagnostic HIV tests. *PLoS One*. 2014; 9(10):e111552. [PubMed: 25343245]
12. Fitzgerald N, et al. Diagnosing acute HIV infection at point of care: a retrospective analysis of the sensitivity and specificity of a fourth-generation point-of-care test for detection of HIV core protein p24. *Sex. Transm. Infect.* 2016; 93
13. Marrazzo JM, et al. Tenofovir-based preexposure prophylaxis for HIV infection among African women. *N. Engl. J. Med.* 2015; 372(6):509–518. [PubMed: 25651245]
14. Stekler JD, et al. Performance of Determine Combo and other point-of-care HIV tests among Seattle MSM. *J. Clin. Virol.* 2016; 76:8–13. [PubMed: 26774543]
15. Smallwood M, et al. Evaluation of a rapid point of care test for detecting acute and established HIV infection, and examining the role of study quality on diagnostic accuracy: a Bayesian meta-analysis. *PLoS One*. 2016; 11(2):e0149592. [PubMed: 26891218]
16. Lanzafame M, et al. Performance of Alere Determine HIV-1/2 Ag/Ab Combo rapid test for acute HIV infection: a case report. *Infez. Med.* 2015; 23(1):48–50. [PubMed: 25819051]
17. Chetty V, Moodley D, Chuturgoon A. Evaluation of a 4th generation rapid HIV test for earlier and reliable detection of HIV infection in pregnancy. *J. Clin. Virol.* 2012; 54(2):180–184. [PubMed: 22445263]

**Fig. 1.**

Endpoint Algorithm for Diagnosis of HIV-1 Infection used at Clinical Sites during the VOICE Study. Participants were monitored monthly for seroconversion with one (Uganda) or two (South Africa and Zimbabwe) third generation HIV rapid diagnostic tests including Alere Determine™ HIV-1/2, OraQuick ADVANCE® Rapid HIV-1/2 Antibody Test and/or Uni-Gold™ Recombigen® HIV-1/2. Positive or discordant results were confirmed by GS HIV-1 Western Blot (Western Blot 1) using plasma from a separate draw. HIV-1 RNA PCR (HIV viral load) was performed on plasma following an indeterminate or negative Western blot result to assess acute infection. Participants confirmed as HIV-uninfected resumed study product and continued in VOICE while participants whose Western Blot tested positive returned to the study site to have a second Western blot performed on plasma from a separate draw. A second positive Western blot confirmed HIV infection and the participant

was exited from the study. Discordant results underwent investigation and further testing by the Study Laboratory Management Group.

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript




Fig. 2. Third Generation Rapid Testing Results on Whole Blood Performed During Study Follow-Up Visits at Clinical Sites.




Fig. 3. Proportion of RNA Positive/3rd Generation Rapid Negative Samples Detected as HIV Positive by Third and Fourth Generation EIA and Fourth Generation Rapid Tests. 57 pre-seroconversion plasma samples with HIV RNA > 20 copies/ml that tested HIV negative using third generation antibody (Ab)-only rapid diagnostic tests were evaluated with Bio-Rad GS HIV-1/2 + O EIA, Bio-Rad HIV Combo Antigen (Ag)/Ab EIA, Alere Determine™ HIV-1/2 Ag/Ab Combo, and Conformité Européene (CE)-Marked Alere™ HIV Combo. 8 of 57 (14%) were positive by Bio-Rad GS HIV-1/2 + O EIA and 27 of 47 (47%) were positive by Bio-Rad HIV Combo Ag/Ab EIA ($p < 0.0001$). 4 of 57 (7%) were positive by Alere Determine™ HIV-1/2 Ag/Ab Combo (2 Ab only, 2 Ag only, 0 dual Ag/Ab) and 16 of 57 (28%) were positive by CE-Marked Alere™ HIV Combo (1 Ab only, 9 Ag only, 6 dual Ag/Ab).

Table 1

Sensitivity, Specificity, Positive and Negative Predictive Values of Rapid and Immunoassay Tests.

| | Tests Performed (N) | Sensitivity (95% CI) | Specificity (95% CI) | Positive Predictive Value (95% CI) | Negative Predictive Value (95% CI) |
|--|---------------------|-------------------------------------|---|-------------------------------------|--|
| Determine™ HIV-1/2 ^a | 72,196 | n = 401 88.78 (85.27, 91.70) | n = 71,795 99.96 (99.94, 99.97) | n = 388 91.75% (88.56, 94.29) | n = 71,808 99.94% (99.92, 99.95) |
| OrqQuick® ADVANCE ^a | 16,854 | n = 26 61.54% (40.57, 79.77) | n = 16,828 99.99% (99.97, 100.00) | n = 17 94.12% (71.31, 99.85) | n = 16,837 99.94% (99.89, 99.97) |
| Uni-Gold™ Recombigen® ^a | 61,075 | n = 389 85.86% (82.00, 89.17) | n = 60,686 99.97% (99.95, 99.98) | n = 352 94.89% (92.04, 96.94) | n = 60,723 99.91% (99.88, 99.93) |
| FDA-Approved Determine™ Combo ^b | 229 | n = 71 23.94% (14.61, 35.54) | n = 158 98.73% (95.50, 99.85) | n = 19 89.47% (66.86, 98.70) | n = 210 74.29% (67.82, 80.05) |
| CE-Marked HIV Combo ^b | 229 | n = 71 42.25% (30.61, 54.56) | n = 158 98.73% (95.50, 99.85) | n = 32 93.75% (79.19, 99.23) | n = 197 79.19% (72.84, 84.63) |
| GS HIV-1/2 + O EIA ^b | 229 | n = 71 30.99% (20.54, 43.08) | n = 158 99.37% (96.52, 99.98) | n = 23 95.65% (78.05, 99.89) | n = 206 76.21% (69.80, 81.85) |
| GS HIV Combo Ag/Ab EIA ^b | 229 | n = 71 57.75% (45.44, 69.39) | n = 158 99.37% (96.52, 99.98) | n = 42 97.62% (87.43, 99.94) | n = 187 83.96% (77.90, 88.91) |

^aTests performed at trial sites during VOICE using fingerstick or venous-collected whole blood.

^bTests performed as part of post-trial analysis using stored plasma.

Table 2

Proportion of Positive or Reactive Test Result from Pre-Seroconversion Plasma Specimens.

| Number of days that sample was collected prior to positive third generation rapid test result at the clinical site | Number (%) of Positive or Reactive Test Results | | | |
|--|---|------------|-------------------|----------------------------|
| | 12–29 days | 30–59 days | 60–91 days | TOTAL |
| | N=48 | N=98 | N=84 | N = 230 |
| Roche TaqMan, v2.0 HIV-1 RNA ≥ 20 copies/mL or < 20 copies/mL, detected | 32 (67%) | 33 (34%) | 7 (8%) | 72 (31%) |
| Uni-Gold™ Recombigen® | 8 (17%) | 5 (5%) | 1 (1%) | 14 (6%) |
| OraQuick ADVANCE® | 6 (13%) | 1 (1%) | 2 (2%) | 9 (4%) |
| | N=48 | N=98 | N=83 ^a | N = 229^a |
| CE-Marked Alere™ HIV Combo | 16 (33%) | 14 (14%) | 2 (2%) | 32 (14%) |
| FDA-Approved Determine™ Combo | 12 (25%) | 6 (6%) | 1 (1%) | 19 (8%) |
| GS HIV-1/2 + O EIA | 13 (27%) | 8 (8%) | 2 (2%) | 23 (10%) |
| GS HIV Combo Ag/Ab EIA | 23 (48%) | 16 (16%) | 3 (4%) | 42 (18%) |
| Multispot HIV-1/HIV-2 | 7 (15%) | 4 (4%) | 1 (1%) | 12 (5%) |
| GS HIV-1 Western Blot | 6 (13%) | 3 (3%) | 1 (1%) | 10 (4%) |
| Geenius HIV-1/2 Supplemental | 6 (13%) | 4 (4%) | 1 (1%) | 11 (5%) |

^aOne specimen was only tested by Roche TaqMan, Uni-Gold™ Recombigen® and OraQuick ADVANCE®.

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Table 3

Rapid, Immunoassay and Confirmatory Test Results for HIV-1 RNA-Positive, Third Generation Rapid Negative and HIV Combo-Positive Pre-Seroconverter Specimens.

| PID | Sample Collected # days Prior to Positive 3rd Generation Rapid Test | HIV-1 RNA (Copies/mL) | Rapid Test | | Immunoassay | | Confirmatory Tests | | |
|-----|---|-----------------------|---------------------|--------------------------------|----------------|--------------------|-----------------------|--------------------------------|--------------------------------------|
| | | | CE-Marked HIV Combo | Determine™ HIV-1/2 Ag/Ab Combo | GS HIV-1/2 + O | GS HIV Combo Ag/Ab | Multispot HIV-1/HIV-2 | GS HIV-1 Western Blot (bands) | Geenius HIV-1/2 Supplemental (bands) |
| 1 | 27 | 1,904,350 | Ag/Ab | Ab | R | R | NR | Ind. (gp41, p24) | HIV Neg. |
| 2 | 28 | 1,554,640 | Ag/Ab | Ab | NR | R | NR | Ind. (gp41) | HIV Neg. |
| 3 | 29 | 4,790,990 | Ag | Ag | NR | R | NR | Ind. (p55/51, gp41) | HIV-2 Ind. (gp140) |
| 4 | 32 | 17,546,710 | Ag | Ag | NR | R | NR | Neg. | HIV Neg. |
| 5 | 12 | 65,672,120 | Ag | Neg. | NR | R | NR | Neg. | HIV Neg. |
| 6 | 20 | 2,176,510 | Ag | Neg. | NR | R | NR | Ind. (*gp41, *p18) | HIV Neg. |
| 7 | 23 | 18,683,070 | Ag/Ab | Neg. | R | R | HIV-1 R | Neg. | HIV Neg. |
| 8 | 29 | 6,373,520 | Ag/Ab | Neg. | R | R | NR | Neg. | HIV Neg. |
| 9 | 30 | 20,594,620 | Ag/Ab | Neg. | NR | R | NR | Neg. | HIV Neg. |
| 10 | 30 | 3,426,650 | Ag | Neg. | NR | R | NR | Ind. (gp41) | HIV Neg. |
| 11 | 32 | 2660 | Ag | Neg. | NR | NR | NR | Ind. (*gp41, p55/51) | HIV-2 Ind. (gp140) |
| 12 | 35 | 2,180,160 | Ag | Neg. | R | R | NR | Ind. (*gp41) | HIV Neg. |
| 13 | 52 | 154 | Ab | Neg. | R | R | NR | Ind. (*p55/51, *p41, p40, p24) | HIV Neg. |
| 14 | 56 | 3,405,070 | Ag | Neg. | NR | R | NR | Ind. (*gp41) | HIV Neg. |
| 15 | 57 | 533,140 | Ag | Neg. | NR | R | NR | Neg. | HIV Neg. |
| 16 | 82 | 12,706,620 | Ag/Ab | Neg. | NR | R | NR | Neg. | HIV Neg. |

Abbreviations: Participant identifier (PID); Antigen (Ag); Antibody (Ab); Antigen/Antibody (Ag/Ab); Negative (Neg.); Reactive (R); Non-Reactive (NR); Indeterminate (Ind.).