

Lawrence Berkeley National Laboratory

Recent Work

Title

A SENSITIVE DIFFERENTIAL PRESSURE SYSTEMS FOR MEASURING CRYOGENIC LIQUID DEPTHS

Permalink

<https://escholarship.org/uc/item/7k70j8c9>

Authors

Pope, William L.
Mclaughlin, Edwin F.

Publication Date

1961-04-21

UCRL 9672

UNIVERSITY OF
CALIFORNIA

Ernest O. Lawrence

*Radiation
Laboratory*

TWO-WEEK LOAN COPY

*This is a Library Circulating Copy
which may be borrowed for two weeks.
For a personal retention copy, call
Tech. Info. Division, Ext. 5545*

BERKELEY, CALIFORNIA

DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

to be published in Proceedings of 1961 Cryogenic Engineering Conference, 8-15-61

UCRL-9672
Limited distribution

UNIVERSITY OF CALIFORNIA
Lawrence Radiation Laboratory
Berkeley, California

Contract No. W-7405-eng-48

**A SENSITIVE DIFFERENTIAL PRESSURE SYSTEM
FOR MEASURING CRYOGENIC LIQUID DEPTHS**

William L. Pope and Edwin F. McLaughlin

April 21, 1961

**A SENSITIVE DIFFERENTIAL PRESSURE SYSTEM
FOR MEASURING CRYOGENIC LIQUID DEPTHS**

William L. Pope and Edwin F. McLaughlin

**Lawrence Radiation Laboratory
University of California
Berkeley, California**

April 21, 1961

ABSTRACT

A system for measuring depths of cryogenic liquids is described. The indicating device is a modified differential pressure gage. The level sensing probes are of various types, either permanent or removable. The heat leak to cryogenic liquids may be made negligibly small.

A SENSITIVE DIFFERENTIAL PRESSURE SYSTEM FOR MEASURING CRYOGENIC LIQUID DEPTHS*

William L. Pope and Edwin F. McLaughlin

Lawrence Radiation Laboratory
University of California
Berkeley, California

April 21, 1961

THE PRESSURE GAGE

The gage used is a modification of the "Magnehelic" gage (F. W. Dwyer Co.), a differential pressure gage available with a minimum full scale range of 15 in. of liquid hydrogen and a corresponding least count of 0.3 in. of liquid hydrogen. (See Fig. 1). This sensitivity is made possible by a unique magnetic coupling between the force-balance system and the indicating mechanism.

The pressure-sensing element is a flexible synthetic-rubber diaphragm. In the modification, this diaphragm is protected against overtravel and overpressures up to 40 psi in either direction. The case is redesigned to be vacuum tight (essential for purging) and to withstand an internal pressure of 150 psi. After modification, the gages are calibrated with an inclined-tube manometer and show typical errors of 4% of full scale.

THE LEVEL SENSING PROBES

These probes can be either permanent or removable. The permanent probes are those commonly used to measure a hydrostatic differential pressure, with the low-pressure tap in the vapor above the cryogenic liquid and the high-pressure tap coming out of the bottom of the liquid container. This latter tap is a stainless steel tube wound in a spiral with a slight upward cast from the horizontal. The length of tube in the spiral minimizes the heat leak; the upward cast insures a stable boiling interface. The free end of the stainless steel spiral is connected to a copper tube of low thermal resistance which leads out the top flange of the vacuum jacket.

The removable probes are of the gas-bubbler type. The low-pressure tap again is exposed to the vapor above the cryogenic fluid. The high-pressure tap is exposed to the liquid at the bottom of the container by means of a tube inserted through the top of the vessel. The liquid-vapor interface in the tube is kept at

* This work was done under the auspices of the U. S. Atomic Energy Commission.

the bottom of the tube by slowly bleeding in a noncondensable gas. Continued flow of gas is not always necessary; these probes when charged with helium will measure liquid-hydrogen levels for many hours before helium must be added.

A more versatile type of probe is one that generates its own gas. This probe has two coaxial tubes leading to the bottom of the liquid. The outer tube furnishes vacuum insulation for the inner tube; the inner tube is a series combination of copper and stainless steel. The lower end of the inner tube is stainless steel with a large temperature gradient down its short length. This gradient keeps the boiling interface near the bottom of the stainless section and also keeps the vapor in the inner tube well above condensation temperature.

THE SYSTEM

Two or more pressure gages can be paralalled to read from the same probe. The gages have been installed with as much as 40 ft of 3/16-O. D. connecting tubing between the gages and the probes. Because of the high sensitivity of the gage, the system must be vacuum tight to avoid erroneous readings. Jitter of the gage during filling and emptying of the vessel is not extreme for either the permanent probe or the gas-generating probe. Small fluctuations present can be damped out with a pneumatic RC circuit using the volume of the gage itself and a short piece of capillary tubing in series with the high-pressure tap.

Figure Legend

Fig. 1. The modified Magnehelic gage.

