

Lawrence Berkeley National Laboratory

Recent Work

Title

COIL CALIBRATION OF 10 COILS - TEST PLAN

Permalink

<https://escholarship.org/uc/item/86s5j4v5>

Author

Nelson, Donald H.

Publication Date

1981-05-20



Lawrence Berkeley Laboratory

UNIVERSITY OF CALIFORNIA

Engineering & Technical
Services Division

RECEIVED
LAWRENCE
BERKELEY LABORATORY

MAY 24 1982

LIBRARY AND
DOCUMENTS SECTION

For Reference

Not to be taken from this room



DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

LAWRENCE BERKELEY LABORATORY - UNIVERSITY OF CALIFORNIA		CODE	SERIAL	PAGE
ENGINEERING NOTE		MT295	MME Book No. 630B	1 of 9
AUTHOR	DEPARTMENT	LOCATION	DATE	
Donald H. Nelson	Electronics Engineering	B25A-124	May 20, 1981	
PROGRAM - PROJECT - JOB				
Superconducting Dipole Isabelle Prototype Magnetic Measurements Project				
TITLE				
COIL CALIBRATION OF 10 COILS - TEST PLAN				
<p>PURPOSE: Six coils are to be installed in the multipole-array for testing superconducting dipole - Isabelle prototype.</p> <p>GOAL: Absolute determination of coils area to $\pm 0.1\%$ Relative sensitivity (area) to $\pm 0.01\%$ of average</p> <p>COIL DATA (Prior to calibration): Table I lists resistance measurements for the 10 coils (from pages 1 and 4 - fabrication log)</p> <p>TEST EQUIPMENT & COORDINATE SYSTEM: Figure 1* shows the test equipment and coordinate system. Table II lists the specific equipment.</p> <p>TEST PROCEDURE FOR ABSOLUTE AREA CALIBRATION ($\pm 0.1\%$ of average) (Method 1)*</p> <p>I. Measurement of BAVG** = average field over volume occupied by coils to be tested.</p> <p>A. Set and monitor magnet current</p> <ol style="list-style-type: none"> 1. NMR(ref) ~ 1.0 (T) record both magnet current and NMR 2. Check magnetic field stability <ol style="list-style-type: none"> a. Measure and record magnet current monitor, i.e., shunt potential (mV) b. Measure and record magnetic induction at NMR reference location (Tesla) <p>B. Measure BAVG</p> <ol style="list-style-type: none"> 1. 10 equally spaced measurements (5 cm centers) <ol style="list-style-type: none"> a. Check NMR(ref) periodically (as determined from stability studies - A2 above) 2. Numerically integrate to compute BAVG; $BAVG = \int_{-22.5}^{+22.5} B_z dx / 45$ 				
<p>*For method 1, one coil is tested at a time, and the reference coil and coil separator shown in Figure 1 are not used. The input from the reference coil to the flux standard is replaced by a "short-circuit".</p> <p>**Table III defines abbreviations.</p>				

LAWRENCE BERKELEY LABORATORY • UNIVERSITY OF CALIFORNIA		CODE	SERIAL	PAGE
ENGINEERING NOTE		MT295		2 OF 9
AUTHOR	DEPARTMENT	LOCATION	DATE	
Donald H. Nelson	Electronics Engineering	B25A-124	May 20, 1981	

COIL CHECK

DATE	TIME	COIL No.	METER CHECK	$R_{100'}$ ^{KΩ}	R_{COIL} ^{KΩ}	TEMP.	COMMENT.
5-18-81	9:12	1	✓	.1191	3.7094	22.5	
		2	✓	.1191	3.7435	22.5	
		3	✓	.1189	3.7262	22.4	
		4	✓	.1188	3.7424	22.6	
		5	✓	.1187	3.7165	22.6	
		6	✓	.1189	3.7125	22.5	
		7	✓	.1188	3.7413	22.4	
		8	✓	.1190	3.7122	22.4	
		9	✓	.1190	3.7421	22.5	
		10	✓	.1190	3.7436	22.5	


HIGH FIELD MAGNET DEVELOPMENT

SEARCH COILS

1. CHECK RESISTANCE OF WIRE USED TO WIND THE SEARCH COILS.

A. WOUND 100 FT. OF #40 (PHELPS-DODGE SY-BONDEZE) ON A WOODEN SPOOL. WAS WOUND UNDER SAME TENSION AS SEARCH COILS ARE TO BE WOUND.

2. HOW TO CHECK COIL RESISTANCE



A. CONNECT CURRENT TERMINALS FROM METER TO 1 & 3.


B. FOR METER CHECK - SHORT OUT RESISTANCE POINTS

C. FOR R_{100FT} - CONNECT RESISTANCE POINTS TO 1 & 2

D. FOR R_{COIL} - CONNECT RESISTANCE POINTS TO 3 & 4


TABLE I Resistance Measurements of 10 Coils

SUBJECT			SKETCH LAWRENCE BERKELEY LABORATORY UNIVERSITY OF CALIFORNIA		JOB ORDER INFORMATION	Job Tag
DRAWN BY ED CYR						No. No.
DATE 6/17/81	BLOG. NO. 25A	ROOM NO. 119	Serial No.	Date Reqd.		
APPROVED BY			Date Issued	Date Reqd.		
DATE			DATE		Deliver To	



NOTE: The two coils (one in the case of calibration method 1) are positioned within the magnet by 1/2 inch bolts through the coil fiducial holes. The bolt heads (or nuts) contact the fiducial holes in the positioning fixture. The coil(s), two at a time in method 2, are positioned by the snug fitting bolts to $\pm 1/32$ inch.

FIGURE 1 Schematic Diagram of Test Equipment Used for Coil Calibration



RL-112 (Rev. 8/75) .600-55589

SKETCH NUMBER

ENGINEERING NOTE

AUTHOR Donald H. Nelson	DEPARTMENT Electronics Engineering	LOCATION B25A-124	DATE May 20, 1981	CODE MT295	SERIAL	PAGE 4 of 9
-----------------------------------	--	-----------------------------	-----------------------------	----------------------	---------------	-----------------------------

EQUIPMENT	DESCRIPTION	IDENTIFICATION	
		No. 1	No. 2
Magnet	8" x 16" x 36" -C	Bev. AEC No. 135652	
Power Supply		Bev. S/N 707 & 702	
Current Mon. Shunt	2500 A/50 mV	Bev. S/N 25 R = 19.99 $\mu\Omega$	
Voltmeter #1	Dixson Model VT 200		
Flux Standard	MME SLFS $\emptyset = 1.0144$ V.S.	MME S/N 39	
Divider	Dekavider Model RV622A	MME AEC No. 107603	
Integrator	MME Model '71	S/N 2 R = 56.2 k C = 3 μ F Div = 1.000	
Voltmeter #2	H.P. Model 3455A	MME AEC No. 517459	
NMR Control Unit	LBL/CERN	S/N 023	S/N 025
Amplifier	LBL/CERN	S/N 26	S/N 27
Probe	LBL - Range 5	S/N 151	S/N 014
Scope #1	Tektronix Model 465B	S/N B045512	
Scope #2	NLS Model MS-215		S/N 1793

TABLE II EQUIPMENT LIST

ENGINEERING NOTE

CODE

MT295

SERIAL

PAGE

5
OF 9

AUTHOR

Donald H. Nelson

DEPARTMENT

Electronics Engineering

LOCATION

B25A-124

DATE

May 20, 1981

TEST PROCEDURE FOR ABSOLUTE AREA CALIBRATION ($\pm 0.1\%$ of average) (continued)

II. Calibrate 10 coils

A. Calibrate integrator/DVM combination

1. Connect coil to be tested

2. Measure and record E(SLFS)*

B. Measure flux linkage due to flipping coil (single coil)

1. Five measurements minimum

C. Record BNMR (reference location)

D. Calculate turns-area product, $nA(i)$.

1. Unadjusted (for NMR data)

$$nA(i) \text{ (m}^2\text{)} = (\text{SLFLUX}/\text{BAVG}) * 0.5 * \text{E(flip)}/\text{E(SLFS)}$$

2. Adjusted (for NMR data)

$$nA_{\text{adj}}(i) \text{ (m}^2\text{)} = nA(i) * \text{BNMR}(i)/\text{BNMR}(\text{avg})$$

*See Table III for abbreviations and units

ENGINEERING NOTE

MT295

6
OF
9

AUTHOR

Donald H. Nelson

DEPARTMENT

Electronics Engineering

LOCATION

B25A-124

DATE

May 20, 1981

ABBREVIATION (Symbol)	DEFINITION	UNITS
NMR(ref)	Nuclear Magnetic Resonance (at reference locations)	Teslas
BNMR	Magnetic induction measured with NMR magnetometer	Teslas
SLFLUX	Flux-linkage (generated when flux- standard primary current reversed)	Webers
BAVG	Average magnetic induction at location of coil under test	Teslas
E(flip)	Integrator output potential due to flipping coil under test, i.e., turning coil 180 degrees about its long axis	Volts
E(SLFS)	Integrator output potential due to reversing SLFS primary current	Volts
nA(i)	Turns-area product (of coil no. i)	m ²
nAadj(i)	nA(i) adjusted for change in BAVG as determined by NMR(ref)	m ²

TABLE III DEFINITIONS OF ABBREVIATIONS

LAWRENCE BERKELEY LABORATORY - UNIVERSITY OF CALIFORNIA.		CODE	SERIAL	PAGE
ENGINEERING NOTE		MT295		7 9
AUTHOR	DEPARTMENT	LOCATION	DATE	
Donald H. Nelson	Electronics Engineering	B25A-124	May 20, 1981	

TEST PROCEDURE FOR RELATIVE AREA CALIBRATION (+0.01% of Reference Coil)(Method 2)

- I. Select a reference coil
- II. Mount reference coil and test coil on a fixture which will allow the positions of the two coils to be interchanged. (See Figure 1)
- III. Connect coils in aiding arrangement, i.e., as coil positions are interchanged, the signals from the two coils are of the same sign.
 - A. Divide sum of signals to avoid overloading integrator
 - B. Calibrate to determine system sensitivity
 - C. Monitor field changes (throughout calibration procedure) with NMR #2
 - D. Record a series of integrator output potentials as coil positions are interchanged (7 measurements minimum)
- IV. Connect coils in opposing arrangement, i.e., reverse the sense of coil under test by the SLFS input #2 selector/reversing switch
 - A. Increase fraction of signal input to the integrator
 - B. Calibrate to determine system sensitivity
 - C. Monitor field changes (throughout calibration procedure) with NMR #2
 - D. Record a series of integrator output potentials as coil positions are interchanged (7 measurements minimum)
- V. Calculate relative area
 - A. Quantity of interest:

$$(\Delta nA)_i = 100 \frac{(nA)_{ref} - (nA)_i}{(nA)_{ref}} \quad (\%) \text{ of } (nA)_{ref}$$

1. $(\Delta nA)_i$ = relative area of i-th coil (% of ref area)
2. $(nA)_{ref}$ = area of reference coil (m^2)
3. $(nA)_i$ = area of i-th coil (m^2)

ENGINEERING NOTE

V. Calculate relative area (continued)

B. Measures quantities

1. $\Delta E_{i+} = \Delta(E_{ref} + E_i)$ (V)

2. $\Delta E_{i-} = \Delta(E_{ref} - E_i)$ (V)

3. E_{SLFS+} = Output potential of integrator due to flux linkage change when SLFS primary current reversed (while recording ΔE_{i+}) (V)

4. B_{NMR+} = Magnetic induction as measured by NMR #2 while recording ΔE_{i+} (T)

5. E_{SLFS-} = Output potential of integrator due to flux linkage change when SLFS primary current reversed (while recording ΔE_{i-}) (V)

6. B_{NMR-} = Magnetic induction as measured by NMR #2 while recording ΔE_{i-} (T)

7. \emptyset_{SLFS} = SLFS flux linkage (Previously determined) (Wb)

C. Requirements to obtain quantity of interest (A1) from measured quantities (B)

1. $B_{NMR+} = B_{NMR-} \pm 0.01\%$

2. Interchange of coils repeatable to $\pm 0.01\%$ 3. Coil mounting fixture must not introduce errors of $\geq 0.01\%$ 4. \emptyset_{SLFS} is constant to $\pm 0.01\%$ during measurements of 3 & 5 above

5. Integrator/DVM resolution adequate for resolving 0.01%

D. Equations used

1. $\Delta \emptyset_- = \frac{\Delta E_{i-}}{E_{SLFS-}} \emptyset_{SLFS} = (nA_{ref} - nA_i) \Delta B_{AVG}$

2. $\Delta \emptyset_+ = \frac{\Delta E_{i+}}{E_{SLFS+}} \emptyset_{SLFS} = (nA_{ref} + nA_i) \Delta B_{AVG}$

3. $nA_{ref} = \frac{\Delta \emptyset_+ + \Delta \emptyset_-}{2 \Delta B_{AVG}}$

Note: To arrive at the quantity of interest, divide 1 by $1/2(1 + 2)$ and multiply by 100. ΔB_{AVG} cancels, if requirements C 1, 2 are 3 are met.

LAWRENCE BERKELEY LABORATORY - UNIVERSITY OF CALIFORNIA		CODE	SERIAL	PAGE
ENGINEERING NOTE		MT295		9
AUTHOR	DEPARTMENT	LOCATION	DATE	
Donald H. Nelson	Electronics Engineering	B25A-124	May 20, 1981	

COIL PHASE DETERMINATION

A test plan for determining the phase of each coil (the angle between the coil's electrical axis and the outward normal of the coils top surface) is given in Appendix A - MT 299 which will be distributed on request.

DISTRIBUTION

W.S. Gilbert
M.I. Green
E.C. Hartwig/L.J. Wagner/W.H. Deuser
W.V. Hassenzahl
T. Lauritzen
R.M. Main
D.J. Rondeau
C.E. Taylor

This work was supported by the U.S. Dept. of Energy under Contract DE-AC03-76SF00098.

This report was done with support from the Department of Energy. Any conclusions or opinions expressed in this report represent solely those of the author(s) and not necessarily those of The Regents of the University of California, the Lawrence Berkeley Laboratory or the Department of Energy.

Reference to a company or product name does not imply approval or recommendation of the product by the University of California or the U.S. Department of Energy to the exclusion of others that may be suitable.

TECHNICAL INFORMATION DEPARTMENT
LAWRENCE BERKELEY LABORATORY
UNIVERSITY OF CALIFORNIA
BERKELEY, CALIFORNIA 94720