

UC Irvine

UC Irvine Previously Published Works

Title

Photothermal treatment of port-wine stains using erythrocyte-derived particles doped with indocyanine green: a theoretical study.

Permalink

<https://escholarship.org/uc/item/9vg9325c>

Journal

Journal of biomedical optics, 23(12)

ISSN

1083-3668

Authors

Burns, Joshua M
Jia, Wangcun
Nelson, J Stuart
[et al.](#)

Publication Date

2018-11-01

DOI

10.1117/1.jbo.23.12.121616

Peer reviewed

Photothermal treatment of port-wine stains using erythrocyte-derived particles doped with indocyanine green: a theoretical study

Joshua M. Burns
Wangcun Jia
J. Stuart Nelson
Boris Majaron
Bahman Anvari

Photothermal treatment of port-wine stains using erythrocyte-derived particles doped with indocyanine green: a theoretical study

Joshua M. Burns,^a Wangcun Jia,^b J. Stuart Nelson,^b Boris Majaron,^{c,d} and Bahman Anvari^{a,b,*}

^aUniversity of California, Riverside, Department of Bioengineering, Riverside, California, United States

^bUniversity of California, Irvine, Beckman Laser Institute and Medical Clinic, Irvine, California, United States

^cJožef Stefan Institute, Department of Complex Matter, Ljubljana, Slovenia

^dUniversity of Ljubljana, Faculty of Mathematics and Physics, Ljubljana, Slovenia

Abstract. Pulsed dye laser irradiation in the wavelength range of 585 to 600 nm is currently the gold standard for treatment of port-wine stains (PWSs). However, this treatment method is often ineffective for deeply seated blood vessels and in individuals with moderate to heavy pigmentation. Use of optical particles doped with the FDA-approved near-infrared (NIR) absorber, indocyanine green (ICG), can potentially provide an effective method to overcome these limitations. Herein, we theoretically investigate the effectiveness of particles derived from erythrocytes, which contain ICG, in mediating photothermal destruction of PWS blood vessels. We refer to these particles as NIR erythrocyte-derived transducers (NETs). Our theoretical model consists of a Monte Carlo algorithm to estimate the volumetric energy deposition, a finite elements approach to solve the heat diffusion equation, and a damage integral based on an Arrhenius relationship to quantify tissue damage. The model geometries include simulated PWS blood vessels as well as actual human PWS blood vessels plexus obtained by the optical coherence tomography. Our simulation results indicate that blood vessels containing micron- or nano-sized NETs and irradiated at 755 nm have higher levels of photothermal damage as compared to blood vessels without NETs irradiated at 585 nm. Blood vessels containing micron-sized NETs also showed higher photothermal damage than blood vessels containing nano-sized NETs. The theoretical model presented can be used in guiding the fabrication of NETs with patient-specific optical properties to allow for personalized treatment based on the depth and size of blood vessels as well as the pigmentation of the individual's skin. © The Authors. Published by SPIE under a Creative Commons Attribution 3.0 Unported License. Distribution or reproduction of this work in whole or in part requires full attribution of the original publication, including its DOI. [DOI: [10.1117/1.JBO.23.12.121616](https://doi.org/10.1117/1.JBO.23.12.121616)]

Keywords: laser therapy; mathematical modeling; near-infrared; optical materials; red blood cells; skin.

Paper 180485SSR received Aug. 6, 2018; accepted for publication Nov. 2, 2018; published online Nov. 29, 2018.

1 Introduction

Port-wine stains (PWSs) are congenital and progressive malformations of dermal capillaries that occur in approximately three children per 1000 live births.¹ Pulsed dye lasers (PDL) with wavelengths in the range of 585 to 600 nm have become the most prevalent treatment for PWSs.^{2–5} This treatment uses laser light to target the endogenous hemoglobin for photothermolysis of the blood vessels (BVs). Although clinical results can be obtained in the treatment of light red or red macular PWS, especially in children and light-skin individuals, clearing is inefficient, requiring multiple therapeutic sessions. It is reported that about 60% of patients receive reduction in redness and lesion size after 10 treatment sessions, but only 10% to 20% experience complete clearance of the stain.^{4,6} In addition, PDL treatments within safe therapeutic radiant dosages of ≈ 7 to 16 J/cm², when used in combination with cryogen spray cooling of skin,⁷ do not provide the clinically desirable photothermal effects to completely destroy BVs located at depths greater than about 500 μ m below skin surface.

Another challenge with the current approach is related to the treatment of patients with moderate to heavy skin pigmentation (Fitzpatrick types IV to VI). The skins of these patients

contain a large number of melanosome organelles within the melanocytes, located primarily in the basal epidermis layer. Melanosomes contain melanin, a polymeric pigment with broad absorption spectrum ranging from UV to visible, including the 585- to 600-nm treatment band. Therefore, melanin acts as a chromophore competing with hemoglobin to partially absorb the photons intended to reach the dermal vascular plexus. The result is nonspecific thermal injury to the epidermis and inefficient photodestruction of target BVs.

A potential phototherapeutic approach is to use an exogenous absorber that can be activated by near-infrared (NIR) light to generate heat. **The advantage of changing the therapeutic wavelength from 585 nm to NIR wavelength of 755 nm is associated with nearly threefold reduction in the absorption coefficient of a single melanosome.**⁸ One particularly promising exogenous absorber is indocyanine green (ICG). It is one of the least toxic agents administered to humans and the only FDA-approved NIR dye for specific imaging applications.^{9,10} ICG has also been investigated as a photothermal agent for treatment of PWSs^{11,12} and as a photosensitizer for photodynamic therapy of choroidal melanomas and breast adenocarcinomas.^{13,14}

Despite its current use in clinical settings, ICG's major disadvantage is its short half-life in blood (3 to 4 min).⁹ Due to its amphiphilic nature, ICG can nonspecifically bind to various biomolecules in the blood, including high- and low-density

*Address all correspondence to: Bahman Anvari, E-mail: anvarib@ucr.edu

lipoproteins and albumin. Hepatic parenchymal cells uptake ICG and/or the ICG-bound molecular complexes leading to biliary excretion into the intestine.¹⁵

To extend the circulation time and reduce nonspecific interactions with plasma proteins and other biomolecules, ICG has been encapsulated into various constructs including synthetic polymers,^{16–19} micelles,^{20–22} liposomes,^{23–25} and silica/silicate matrices.^{26,27} Encapsulation within constructs derived from erythrocytes are particularly advantageous due to their potential biocompatibility.

We have previously reported on fabrication of vesicles derived from erythrocytes and loaded with ICG, and their utility for fluorescence imaging and photodestruction of human cells.²⁸ We refer to these constructs as NIR erythrocyte-mimicking transducers (NETs). A particular feature of NETs is that their diameter can be tuned from micron- to nano-scale, and their ICG content can be adjusted independently during the fabrication process. We recently characterized the effects of NETs' diameter and ICG concentration on the resulting optical properties, absorption coefficient (μ_a), and reduced scattering coefficient (μ_s') of these constructs.²⁹ These results are particularly useful in development of mathematical models aimed at quantification of light and laser-induced temperature distributions in tissues containing NETs. As a first step toward potential application of NETs in phototherapy of PWSs, we analyze in dedicated numerical simulations the photothermal response of BVs containing NETs, and compare the results with those obtained with the current treatment approach of using PDL in 585- to 600-nm range. These results can be used to identify the optimum formulation of NETs and appropriate irradiation parameters for NIR photothermal treatment of PWSs.

2 Materials and Methods

2.1 Overview of Model

In Fig. 1, we present a block diagram of the mathematical model consisting of three components: a Monte Carlo light-transport model to estimate the energy deposition in response to laser irradiation,^{30,31} finite-element model to compute the spatiotemporal temperature distribution based on the heat diffusion equation, and an Arrhenius rate process integral to quantify the resulting thermal damage. Each component of the model is described in the following sections.




Fig. 1 Block diagram of the mathematical model components.




Fig. 2 Simulated PWS geometry.

2.2 Model Geometry

We used two geometries to model the human skin with PWS. The first geometry simulated the skin and consisted of a 60- μm -thick epidermis and 1940- μm -thick dermis containing BVs (Fig. 2). The epidermis was composed of a 45- μm melaninless layer and a 15- μm melanocyte-filled layer located at the basal epidermis. BVs were assumed to be cylinders with diameters of 200 μm , running parallel to the y -axis for the entire length of the geometry (1 mm). The top position of the vessel was located either at 500 or 800 μm below skin surface.

In addition, we also constructed a PWS skin geometry from an image obtained by optical coherence tomography (OCT) of a PWS patient (Fig. 3).³² We first removed motion artifact associated with the image cube, 5 mm (length, in x -direction) \times 5 mm (width, in y -direction) \times 2.8 mm (height, in z -direction), by taking the Fourier transform of each two-dimensional (2-D) section (5 mm \times 5 mm) (axial resolution = 8 μm) and applying a bandpass filter with cutoff normalized frequencies of 0.1 and 1.0, where 1.0 corresponds to half of the sampling frequency in the horizontal direction, and 0.0 and 0.01 in the vertical direction. We then took the inverse Fourier transform of each image and converted into binary values through thresholding. Finally, speckle noise was removed using a three-dimensional (3-D) median filter.

2.3 Optical Properties of Human Skin

The baseline absorption coefficient for melaninless epidermis $\mu_{a,\text{base}}$ (mm^{-1}) and bloodless dermis was approximated by³³




Fig. 3 Flowchart of processing an OCT image of PWS skin for use in mathematical model. With respect to the coordinate system shown in Fig. 2, the coordinate system here is rotated 90 deg counterclockwise around the x -axis. The images are horizontal cross-sections in the x - y plane at depth of $z = 680 \mu\text{m}$.

$$\mu_{a,\text{base}} = 0.0244 + 8.53 \cdot \exp\left(-\frac{\lambda - 154}{66.2}\right), \quad (1)$$

where λ is the wavelength in nm. The absorption coefficient for the basal layer of the epidermis $\mu_{a,\text{basal}}$ is a combination of the $\mu_{a,\text{base}}$ and the absorption coefficient of melanosomes $\mu_{a,\text{mel}}$:

$$\mu_{a,\text{basal}} = f_{\text{mel}} \cdot \mu_{a,\text{mel}} + (1 - f_{\text{mel}}) \cdot \mu_{a,\text{base}}, \quad (2)$$

where f_{mel} is the volume fraction of melanosomes. In this study, we present lightly, moderately, and heavily pigmented skin with $f_{\text{mel}} = 4\%$, 15% , and 50% , respectively,³⁴ and $\mu_{a,\text{mel}} (\text{mm}^{-1})$ defined as³³

$$\mu_{a,\text{mel}} = 6.6 \times 10^{10} \cdot \lambda^{-3.33}. \quad (3)$$

For simulated single vessels, we accounted for attenuation of the incident light by absorption in the remaining PWS BVs by adjusting the absorption coefficient of the dermis ($\mu_{a,\text{der}}$) (Table 1) as³⁰

Table 1 Adjusted absorption coefficient of dermis used in this study.

z (μm)	r_i (μm)	f_i (%)	$\mu_{a,\text{der}}$ (585 nm) (mm^{-1})	$\mu_{a,\text{der}}$ (755 nm) (mm^{-1})
60 to 260	25	5.2	0.071	0.025
260 to 460	37	8.0	0.090	0.024
460 to 660	34	3.8	0.062	0.025
660 to 860	27	2.5	0.053	0.025
860 to 1060	24	2.0	0.050	0.025
1060 to 2000	24	1.5	0.047	0.025
PWS blood vessels imaged by OCT all z (μm)			0.073	0.026

$$\mu_{a,\text{der}} = f_i C_i \mu_{a,\text{blood}} + (1 - f_i) \mu_{a,\text{der}}, \quad (4)$$

where C_i is the correction factor accounting for optical screening within a BV with radius r_i , and f_i is the fractional volume occupied by the PWS BVs. In this adjustment, we account for the average r_i of the PWS BVs within a range of subsurface depths (z) (see Table 1) and assume 45% hematocrit (hct) with oxygen saturation of 70%:³⁰

$$C_i \approx 0.039 + 0.486 \exp\left[-\frac{\mu_{a,\text{blood}}(45\% \text{hct}) r_i}{0.193}\right] + 0.468 \exp\left[-\frac{\mu_{a,\text{blood}}(45\% \text{hct}) r_i}{0.914}\right]. \quad (5)$$

The adjusted values of $\mu_{a,\text{der}}$ at 585 nm vary with depth and the corresponding values of r_i and f_i , whereas those for 755 nm are nearly independent of depth.

For OCT simulated vessels, $\mu_{a,\text{der}}$ was expressed as

$$\mu_{a,\text{der}} = f_{\text{blood}} \cdot \mu_{a,\text{blood}} + (1 - f_{\text{blood}}) \cdot \mu_{a,\text{bas}}, \quad (6)$$

where f_{blood} is the volume fraction of blood in the healthy dermis, chosen to be 0.2% consistent with literature reported value,³³ and $\mu_{a,\text{blood}} (\text{mm}^{-1})$ is the absorption coefficient of blood with oxygen saturation of 70%. In this case, values of $\mu_{a,\text{der}}$ were not adjusted since the actual distributions of BVs were directly obtained from the OCT images.

The scattering coefficients (μ_s) (mm^{-1}) of the melaninless epidermis, epidermal basal layer, and dermis were estimated as³³

$$\mu_s = \frac{2 \times 10^4 \cdot \lambda^{-1.5} + 2 \times 10^{11} \cdot \lambda^{-4}}{1 - g}, \quad (7)$$

where the scattering anisotropy factor (g) was assumed to be 0.8 for 585 nm³⁵ and 0.91 at 755 nm,³⁶ the irradiation wavelengths used in this study.

2.4 Optical Properties of NETs and Blood Vessels

Spectrally dependent values of μ_a and μ_s' for micron-sized NETs (μ_{NETs}) and nano-sized NETs (nNETs) with respective mean diameters (d_{mean}) of $\approx 4 \mu\text{m}$ and 92 nm were previously estimated using an integrating sphere in conjunction with an inverse adding-doubling algorithm.²⁹ Values of μ_a and μ_s for blood with 45% hct and oxygen saturation of 70% were obtained from the literature.³⁷ We then determined the effective optical properties of BV as

$$\mu_{a,\text{BV}} = (1 - f_{\text{NETs}}) \mu_{a,\text{blood}} + f_{\text{NETs}} \mu_{a,\text{NETs}}, \quad (8)$$

$$\mu_{s,\text{BV}} = (1 - f_{\text{NETs}}) \mu_{s,\text{blood}} + f_{\text{NETs}} \mu_{s,\text{NETs}}, \quad (9)$$

$$g_{\text{BV}} = (1 - f_{\text{NETs}}) g_{\text{blood}} + f_{\text{NETs}} g_{\text{NETs}}, \quad (10)$$

where $\mu_{a,x}$, $\mu_{s,x}$, and g_x are the respective effective absorption and scattering coefficients, and anisotropy factor, with the subscript x standing for either BVs containing blood and volume fraction of NETs (f_{NETs}) (assumed 10% or 25%), blood, or NETs. We varied $\mu_{a,\text{NETs}}$ so that $\mu_{a,\text{BV}}$ had values in the

Table 2 Optical properties used in this study. Optical properties for blood are based on 45% hematocrit.

Wavelength	Optical properties	Melaninless epidermis	Basal layer containing	Basal layer containing	Basal layer containing	Dermis		μ NETs	nNETs	$\mu_{a,BV}$
			melanosomes (light)	melanosomes (moderate)	melanosomes (heavy)	Blood	($d_{\text{mean}} = 4 \mu\text{m}$)	($d_{\text{mean}} = 92 \text{ nm}$)		
585 nm	μ_a (mm^{-1})	0.037	1.65	6.07	20.15	^a	17.91	—	—	
	μ_s (mm^{-1})	15.61	15.61	15.61	15.61	15.61	76.05	—	—	
	g	0.80	0.80	0.80	0.80	0.80	0.970	—	—	
755 nm	μ_a (mm^{-1})	0.025	0.71	2.58	8.61	^a	0.44	6.04 to 56.04	6.04 to 56.04	1 to 18
	μ_s (mm^{-1})	17.55	17.55	17.55	17.55	17.55	79.85	2.20	0.06	59.90 to 72.09
	g	0.91	0.91	0.91	0.91	0.91	0.983	0.99	0.55	0.87 to 0.98

^aSee Table 1.

range of 1 to 18 mm^{-1} . To determine the ICG concentration in NETs fabrication buffer [ICG] (μM) to produce a desired $\mu_{a,BV}$, we first used a linear regression to quantify the relationship between the previously reported estimates of $\mu_{a,NETs}$, and the utilized values of [ICG]

$$[\text{ICG}] = 17.8 \cdot \mu_{a,NETs}. \quad (11)$$

Next, by rearranging Eq. (8) for $\mu_{a,NETs}$, and substituting the expression into Eq. (11), we obtained [ICG]:

$$[\text{ICG}] = 17.8 \left[\frac{\mu_{a,BV} - (1 - f_{NETs})\mu_{a,blood}(45\%hct)}{f_{NETs}} \right]. \quad (12)$$

Using Eq. (12), we determined [ICG] values of $\approx 108, 286, 464, 642, 820,$ and $998 \mu\text{M}$ to produce the corresponding $\mu_{a,BV}$ values of 1, 2, 3, 4, 5, and 6 mm^{-1} . For example, values of $f_{NETs} = 25\%$ and $\mu_{a,BV} = 18 \text{ mm}^{-1}$ resulted in $[\text{ICG}] = 1258 \mu\text{M}$. Values of the optical properties used in this study are provided in Table 2.

2.5 Mathematical Model

The 3-D Monte Carlo model for light transport and energy deposition was developed earlier by Majaron et al.^{30,31} Our simulations involved one million photons for 585- or 755-nm flat-top laser beams with diameter of 6 mm for simulated PWS geometry, and 8 mm for the skin model based on the PWS patient OCT record. With a 2.3-GHz central processing unit, each Monte Carlo simulation took ≈ 2 to 11 h depending on the geometry and optical properties of the tissue. For example, the run time was ≈ 10.5 h for a simulated geometry containing nNETs ($f_{NETs} = 10\%$, $\mu_{a,BV} = 6 \text{ mm}^{-1}$), and lightly pigmented skin ($f_{mel} = 4\%$) irradiated at 755 nm. For an OCT geometry, an example run time was ≈ 2.0 h for a moderately pigmented skin ($f_{mel} = 15\%$) irradiated at 585 nm without NETs.

We used the heat diffusion equation to calculate the spatio-temporal temperature profiles in response to laser irradiation:

$$\kappa \left[\frac{\partial T^2(x, z, t)}{\partial^2 x} + \frac{\partial T^2(x, z, t)}{\partial^2 z} \right] + S(x, z) = \rho C \frac{\partial T(x, z, t)}{\partial t}, \quad (13)$$

where κ is the thermal conductivity, $T(x, z, t)$ is the skin temperature ($^{\circ}\text{C}$) (assuming symmetry about the y -axis), x is the lateral distance from the beam center (m), z is the depth into the skin, t is time (s), $S(x, z)$ is the heat source term resulting from absorption of photons (W m^{-3}), ρ is the density (kg m^{-3}), and C is the specific heat capacity ($\text{J kg}^{-1}\text{C}^{-1}$). Thermal properties of human skin used in the study were density, $\rho = 1200 \text{ kg m}^{-3}$, specific heat capacity, $C = 3600 \text{ J kg}^{-1}\text{C}^{-1}$, and $\kappa = 0.26, 0.53, 0.53 \text{ W m}^{-1}\text{C}^{-1}$ for epidermis, dermis, and BVs, respectively.³⁸ The baseline skin temperature was 35°C .

We assumed that a cryogen spurt was sprayed onto the skin surface immediately prior to the laser pulse.^{39–43} To account for this type of skin cooling, we implemented a convective surface boundary condition as follows:

$$-\kappa \frac{\partial T(x, z, t)}{\partial z} \Big|_{z=0} = h[T_{\text{med}} - T(x, z, t)|_{z=0}], \quad (14)$$

where T_{med} is the temperature of the cryogen film on the skin surface ($^{\circ}\text{C}$), and h is the convective heat transfer coefficient at the skin surface. We utilized the relevant thermal parameters associated with the boundary condition from our previous work^{43–45} and list them in Table 3. The heat-diffusion equation was solved numerically using the commercial finite-elements software package, FEMLAB[®]. The finite-elements run time for simulated geometries was 2 to 3 min using 1200×400

Table 3 Parameters for surface thermal boundary condition.

Time interval	h ($\text{W/m}^2\text{C}$)	T_{surface} ($^{\circ}\text{C}$)	Duration (ms)
Cryogen spurt application	4000	-50	100
Cryogen pool residence	3000	-26	200
Rewarming	10	25	500

nodes. For OCT geometries, run time was 12 to 18 h using 500×350 nodes.

Thermal damage to the skin was quantified using the Arrhenius rate damage integral

$$\Omega(x, z) = A \int_0^\tau \exp\left[-\frac{E}{RT(x, z, t)}\right] dt, \quad (15)$$

where $\Omega(x, z)$ is the damage index, A is the frequency factor ($A = 1.8 \times 10^{51} \text{ s}^{-1}$ for bulk skin and $A = 7.6 \times 10^{66} \text{ s}^{-1}$ for blood), E is the damage activation energy ($E = 327,000 \text{ J mol}^{-1}$ for bulk skin and $E = 455,000 \text{ J mol}^{-1}$ for blood), and R is the universal gas constant ($R = 8.314 \text{ J mol}^{-1} \text{ K}^{-1}$).^{46,47} Value of $\Omega(x, z) = 1$ is associated with irreversible damage to 63% of the tissue. Laser pulse duration was fixed at 3 ms, and it was applied immediately after termination of a 100-ms cryogen spurt. Values of $\Omega(x, z)$ were calculated following the laser pulse and through the duration of the rewarming interval. The percent photothermal damage to the BV (% damage) was defined as the area of the damaged vessel at this time divided by the total cross-sectional area.

3 Results and Discussion

3.1 Simulated PWS Geometry

3.1.1 Threshold incident dosage for epidermal damage in absence of NETs

We first determined the threshold incident dosage (D_{th}) for epidermal damage in response to 585 and 755 nm irradiation without BVs. For lightly, moderately, and heavily pigmented skin, values of D_{th} were determined as 8, 3, and 1 J/cm^2 for 585 nm, and 21, 6, and 3 J/cm^2 for 755-nm irradiation, respectively (Table 4). These simulation results confirm that use of 755 nm increases the threshold for epidermal injury, due to the lower absorption of melanin at this wavelength compared to 585 nm. These results are consistent with a previous clinical study where laser irradiation at 755 nm with incident dosages greater than 10 J/cm^2 effectively induced PWS lightening.⁴⁸

3.1.2 Photothermal response of PWS blood vessels in the absence of NETs

The percent photothermal damage profiles to BVs without NETs for lightly ($f_{mel} = 4\%$) pigmented skin at the $D_{th} = 8 \text{ J/cm}^2$ at 585 nm and 21 J/cm^2 at 755 nm are

Table 4 Estimated threshold incident dosages for epidermal damage (D_{th}) in absence of NETs for simulated PWS skin geometry.

Wavelength	Skin pigmentation	f_{mel} (%)	D_{th} (J cm^{-2})
585 nm	Light	4	8
	Moderate	15	3
	Heavily	50	1
755 nm	Light	4	21
	Moderate	15	6
	Heavily	50	3




Fig. 4 Damage profiles shown in red ($\Omega \geq 1$) for blood vessels located at (a, b) 500 and (c, d) 800 μm below skin surface for lightly pigmented skin ($f_{mel} = 4\%$) in response to (a, c) 585 and (b, d) 755-nm laser irradiation without NETs.

shown in Fig. 4. Complete damage to BVs without NETs was observed for BVs in response to 585-nm laser irradiation at $D_{th} = 8 \text{ J/cm}^2$ [Figs. 4(a) and 4(c)]. However, this vascular damage was also accompanied by damage to the surrounding dermis. For BVs without NETs irradiated at 755 nm, there was 79 and 50% damage to vessels at depths of 500 and 800 μm , respectively, without thermal injury to perivascular dermis [Figs. 4(b) and 4(d)].

For moderately pigmented skin ($f_{mel} = 15\%$) in response to 585-nm irradiation at $D_{th} = 3 \text{ J/cm}^2$, % damage to the vessels at 500 and 800 μm were 34% and 11%, respectively (Fig. 5). In response to 755-nm irradiation at $D_{th} = 6 \text{ J/cm}^2$, there was no damage to the BVs at either depth. Lastly, for heavily pigmented skin ($f_{mel} = 50\%$) in response to 585-nm irradiation at $D_{th} = 1 \text{ J/cm}^2$ and 755 nm irradiation at $D_{th} = 3 \text{ J/cm}^2$, there was no damage to the vessels at either depth (results not shown). The decrease in % damage for blood vessels with moderately and heavily pigmented skin can be attributed to the decrease in D_{th} , as well as increased light absorption within the basal epidermis.




Fig. 5 Damage profiles shown in red ($\Omega \geq 1$) for blood vessels located at (a, b) 500 and (c, d) 800 μm below skin surface for moderately pigmented skin ($f_{mel} = 15\%$) in response to (a, c) 585 and (b, d) 755-nm laser irradiation without NETs.

3.1.3 Photothermal response of PWS blood vessels containing NETs

The percent photothermal damage profiles to blood vessels containing μ NETs or nNETs ($f_{\text{NETs}} = 10\%$) for lightly pigmented skin ($f_{\text{mel}} = 4\%$) in response to 755 nm irradiation at $D = 21 \text{ J/cm}^2$ are shown in Fig. 6. There was 100% damage to blood vessels located at depths of 500 and 800 μm below skin surface containing μ NETs, or nNETs with $\mu_{a,\text{BV}}$ of 1 mm^{-1} (Fig. 6). As the depth of the blood vessel increased from 500 to 800 μm , there was a decrease in damage to the dermis, which was further reduced when using nNETs as compared to μ NETs. These results suggest PWS patients with light pigmentation and blood vessels as deep as 500 to 800 μm could potentially benefit from the delivery of μ NETs or nNETs that yield $\mu_{a,\text{BV}} = 1 \text{ mm}^{-1}$ to achieve photothermal injury to the blood vessels. This value of $\mu_{a,\text{BV}}$ corresponds to μ NETs or nNETs with $\mu_{a,\text{NETs}} \approx 6 \text{ mm}^{-1}$, which can be fabricated by using $\approx 108 \mu\text{M}$ ICG in the fabrication buffer [Eq. (12)].

For moderately pigmented skin ($f_{\text{mel}} = 15\%$), $f_{\text{NETs}} = 10\%$, and 755-nm laser irradiation at $D_{\text{th}} = 6 \text{ J/cm}^2$, there was no damage to the blood vessels when using μ NETs or nNETs that yielded $\mu_{a,\text{BV}} = 1 \text{ mm}^{-1}$ [Fig. 7(a)]. For $\mu_{a,\text{BV}} = 3 \text{ mm}^{-1}$, the blood vessel at depth of 500 μm showed 94 and 86% damage when containing μ NETs or nNETs, respectively [Figs. 7(a) and 7(b)]. For this value of $\mu_{a,\text{BV}}$, when the depth of the blood vessel was increased to 800 μm , respective % damage values were reduced to 89% and 74% in the presence of μ NETs or nNETs [Figs. 7(c) and 7(d)].

In general, for the case of the blood vessel containing μ NETs at depth of 500 μm , % damage to blood vessels followed an exponentially increasing behavior with $\mu_{a,\text{BV}}$, and approached an asymptotic value of $\approx 95\%$ for values of $\mu_{a,\text{BV}} \geq 3 \text{ mm}^{-1}$, corresponding to $\mu_{a,\text{NETs}} \geq 26 \text{ mm}^{-1}$ [Fig. 7(a)]. For the blood vessel containing nNETs at depth of 500 μm , and blood vessels containing μ NETs or nNETs at depth of 800 μm , the relationship between $\mu_{a,\text{BV}}$ and % damage to blood vessels followed a sigmoidal behavior where below a certain threshold value of $\mu_{a,\text{BV}}$ (and the corresponding $\mu_{a,\text{NETs}}$) there was no damage to the vessel [Fig. 7(a)]. These results indicate that once $\mu_{a,\text{BV}}$




Fig. 6 Damage profiles shown in red ($\Omega \geq 1$) for blood vessels containing μ NETs, or nNETs at $f_{\text{NETs}} = 10\%$ with $\mu_{a,\text{BV}} = 1 \text{ mm}^{-1}$, and located at depths of (a, b) 500 and (c, d) 800 μm below skin surface. Light skin pigmentation ($f_{\text{mel}} = 4\%$), $\lambda = 755 \text{ nm}$, and $D_{\text{th}} = 21 \text{ J/cm}^2$.




Fig. 7 (a) % Damage to blood vessels containing μ NETs or nNETs ($f_{\text{NETs}} = 10\%$) as a function of $\mu_{a,\text{BV}}$ and $\mu_{a,\text{NETs}}$ (markers) fitted with exponential or sigmoidal functions (solid curves). Damage profiles shown in red ($\Omega \geq 1$) for blood vessels located at depths of (b, c) 500 and (d, e) 800 μm below skin surface containing μ NETs, or nNETs at $f_{\text{NETs}} = 10\%$ with $\mu_{a,\text{BV}} = 3 \text{ mm}^{-1}$. Moderate skin pigmentation ($f_{\text{mel}} = 15\%$), $\lambda = 755 \text{ nm}$, and $D_{\text{th}} = 6 \text{ J/cm}^2$.

exceeds $\approx 3 \text{ mm}^{-1}$ with either μ NETs or nNETs, % damage to blood vessels ranging in depth between 500 and 800 μm is maximized to the same level of $\approx 95\%$.

The therapeutic advantage of NETs in conjunction with 755-nm laser irradiation over current treatment at 585 nm without NETs was more evident with moderate skin pigmentation. For example, blood vessels with $\mu_{a,\text{BV}} = 6 \text{ mm}^{-1}$, and containing μ NETs or nNETs had $>98\%$ damage to vessels at either depths [Fig. 7(a)], whereas blood vessels without NETs irradiated at 585 nm had 34 and 11% damage to vessels at depths of 500 and 800 μm , respectively (Fig. 5).

For heavily pigmented skin ($f_{\text{mel}} = 50\%$), there was a dramatic decrease in damage to blood vessels containing NETs with $\mu_{a,\text{BV}} = 1$ to 6 mm^{-1} ($f_{\text{NETs}} = 10\%$) in response to 755-nm laser irradiation at $D_{\text{th}} = 3 \text{ J/cm}^2$ (results not shown). When f_{NETs} was increased to 25%, blood vessels located at depths of 500 and 800 μm showed 60 and 36% damage, respectively, in presence of μ NETs ($\mu_{a,\text{BV}} = 18 \text{ mm}^{-1}$) [Figs. 8(a) and 8(c)]. At this value of f_{NETs} , blood vessels containing nNETs ($\mu_{a,\text{BV}} = 18 \text{ mm}^{-1}$) showed 36 and 27% damage to vessels at depths of 500 and 800 μm , respectively [Figs. 8(b) and 8(d)]. For $\mu_{a,\text{BV}} = 18 \text{ mm}^{-1}$, the corresponding ICG concentrations in the fabrication is $\approx 1258 \mu\text{M}$ [Eq. (12)].

These results demonstrate that blood vessels containing μ NETs had greater therapeutic efficacy as compared to nNETs. The increase in % damage to a blood vessel containing μ NETs is due to higher energy deposition values deeper into the




Fig. 8 Damage profiles shown in red ($\Omega \geq 1$) for blood vessels containing μ NETs, or nNETs at $f_{\text{NETs}} = 25\%$ with $\mu_{a,BV} = 18 \text{ mm}^{-1}$, and located at depths of (a, b) 500 and (c, d) 800 μm below skin surface. Heavy skin pigmentation ($f_{\text{mel}} = 50\%$), $\lambda = 755 \text{ nm}$, and $D_{\text{th}} = 3 \text{ J/cm}^2$.

blood vessels resulting from greater amount of light scattered in the forward direction by μ NET ($g = 0.99$, Table 2) as compared to nNETs ($g = 0.55$, Table 2). Given their larger diameter, the optical behavior of μ NETs is consistent with Mie scattering,²⁹ allowing for higher forward scattering.

3.2 PWS Geometry Based on Patient OCT Image

A 3-D rendering of the patient PWS geometry is shown in Fig. 9(a) (see also multimedia files). The PWS vessel diameters ranged from 60 to 650 μm with most around 200 μm and as deep as 1 mm below the skin surface. To test different skin pigmentations, epidermal layers were added to the geometry with light ($f_{\text{mel}} = 4\%$), moderate ($f_{\text{mel}} = 15\%$), and heavy ($f_{\text{mel}} = 50\%$) pigmentations. Due to the greater therapeutic efficacy of μ NETs for simulated PWS vessel geometries, we chose to use μ NETs for patient PWS vessels with $\mu_{a,BV} = 6 \text{ mm}^{-1}$ and $f_{\text{NETs}} = 10\%$ for lightly pigmented skin ($f_{\text{mel}} = 4\%$), and $\mu_{a,BV} = 18 \text{ mm}^{-1}$ and $f_{\text{NETs}} = 25\%$ for moderately ($f_{\text{mel}} = 15\%$) and heavily ($f_{\text{mel}} = 50\%$) pigmented skins. Damage profiles are shown for lightly, moderately, and heavily pigmented




Fig. 9 3-D rendering of (a) human PWS obtained by OCT (Video 1, MPEG, 2.3 MB [URL: <https://doi.org/10.1117/1.JBO.23.12.121616.1>]) and (b–f) damage profiles to PWS blood vessels. Red regions correspond to values of $\Omega \geq 1$. (b, c) Damage profiles in response to 585-nm laser irradiation without NETs for light ($f_{\text{mel}} = 4\%$) (Video 2) and moderate ($f_{\text{mel}} = 15\%$) (Video 3) pigmentation levels, respectively (Video 2, MPEG, 2.3 MB [URL: <https://doi.org/10.1117/1.JBO.23.12.121616.2>]); Video 3, MPEG, 2.3 MB [URL: <https://doi.org/10.1117/1.JBO.23.12.121616.3>]. (d–f) Damage profiles in response to 755-nm laser irradiation in presence of μ NETs for light ($f_{\text{mel}} = 4\%$) (Video 4), moderate ($f_{\text{mel}} = 15\%$) (Video 5), and heavy ($f_{\text{mel}} = 50\%$) (Video 6) pigmentation levels, respectively (Video 4, MPEG, 3.3 MB [URL: <https://doi.org/10.1117/1.JBO.23.12.121616.4>]); Video 5, MPEG, 3.3MB [URL: <https://doi.org/10.1117/1.JBO.23.12.121616.5>]; Video 6, MPEG, 3.3 MB [URL: <https://doi.org/10.1117/1.JBO.23.12.121616.6>]. Parameters for patient PWS blood vessels containing μ NETs were $\mu_{a,BV} = 6 \text{ mm}^{-1}$, $f_{\text{NETs}} = 10\%$ for (d), and $\mu_{a,BV} = 18 \text{ mm}^{-1}$, $f_{\text{NETs}} = 25\%$ for (e, f).

skin irradiated with 755 nm for PWS vessels with μ NETs, or 585 nm for PWS vessels without NETs [Figs. 9(b)–9(f)].

The therapeutic effectiveness of NETs in conjunction with 755-nm irradiation over 585-nm irradiation without NETs can be seen from these results. We calculated the % damage for vessels up to the depth of 1000 μ m below the surface on a pixel-by-pixel basis within the simulated 3-D PWS geometry obtained by OCT imaging. This was accomplished by comparing each pixel in the 3-D damage profile with that of the 3-D geometry. The number of pixels showing values of $\Omega \geq 1$ were summed and divided by the total number of pixels in the geometry to obtain an estimate of % damage. Lightly, moderately, and heavily pigmented skins showed % damage of 88%, 30%, and 4%, respectively, in presence of μ NETs. For PWS vessels without NETs, there was 32 and 2% damage for lightly and moderately pigmented skins, respectively. In response to the irradiation parameters investigated, current therapeutic approach, based on 585-nm laser irradiation, is most effective in treating patients with light pigmentation by damaging blood vessels to a depth of ≈ 1.2 mm but still with some blood vessels remaining intact in the lateral direction [Fig. 10(b)]. Depth of damage is reduced to ≈ 0.8 mm and accompanied by a reduction in lateral vascular damage in moderately pigmented PWS skin [Fig. 10(c)].

Under the parameters investigated, laser irradiation at 755 nm in conjunction with μ NETs can increase the damage depth to ≈ 1.4 mm and also induce complete damage to the vascular plexus in the lateral direction in patients with light pigmentation [Fig. 10(d)]. For patients with moderate and heavy pigmentation levels, blood vessels to depth of ≈ 1.2 mm can be damaged when using 755-nm laser irradiation in conjunction with NET, but the extent of lateral vascular damages decreases

[Figs. 10(e) and 10(f)]. These results suggest that similar to the optical response of single blood vessel, there may be shadowing effects to reduce the energy deposited to deeper blood vessels.⁴⁹ Such effects can occur if vessels are as close as a few microns.⁵⁰ Optimizing the ICG concentration in NETs may provide a means to allow for more uniform light energy distributions.

Use of free ICG in conjunction with pulsed 808- and 810-nm laser irradiation for treatment of patients with PWS has been reported.^{11,12} While promising clinical results have been reported in these studies, delivering NETs to PWS blood vessels is more advantageous than free ICG as it provides a method to increase the circulation time of ICG, and can consequently, elongate the therapeutic window of time over which laser irradiation can be performed. For example, nanoconstructs (≈ 80 nm diameter) composed of poly (lactic-co-glycolic acid) core coated with erythrocyte-derived membranes were retained in blood for 3 days with circulation half-life of ≈ 8 h in mice.⁵¹ Piao et al.⁵² reported that the circulation half-life of gold materials cloaked with erythrocyte membranes (≈ 90 nm diameter) was ≈ 9.5 h. Increased circulation time of erythrocyte-coated particles can be attributed to the presence of “self-marker” membrane proteins to allow evasion by immune cells.^{53,54} One important “self-marker” membrane protein on the surface of the erythrocytes is CD-47 glycoprotein, whose interaction with immunoreceptor signal regulatory protein alpha (SIRP α), expressed by macrophages, results in inhibition of phagocytosis.⁵⁵ We previously demonstrated that CD-47 remains on the surface of the NETs,⁵⁶ providing a mechanism for increased circulation time of NETs. Furthermore, NETs can serve as a biocompatible and nontoxic platform for delivery of ICG. In a recent study, Rao et al.⁵⁷ reported the absence of systemic toxicity at 15 days




Fig. 10 2-D x-z cross section of (a) human PWS obtained by OCT and (b–f) damage to PWS blood vessels. Red regions correspond to values of $\Omega \geq 1$. (b, c) Damage profiles in response to 585-nm laser irradiation without NETs for light ($f_{mel} = 4\%$), and moderate ($f_{mel} = 15\%$) pigmentation levels, respectively. (d–f) Damage profiles in response to 755-nm laser irradiation in presence of μ NETs for light ($f_{mel} = 4\%$), moderate ($f_{mel} = 15\%$), and heavy ($f_{mel} = 50\%$) pigmentation levels, respectively. Parameters for patient PWS blood vessels containing μ NETs were $\mu_{a,BV} = 6 \text{ mm}^{-1}$, $f_{NETs} = 10\%$ for (d), and $\mu_{a,BV} = 18 \text{ mm}^{-1}$, $f_{NETs} = 25\%$ for (e, f).

postintravenous injection of erythrocyte membrane-coated upconversion nanoparticles in mice.

Results of this theoretical study indicate that use of NETs in conjunction with pulsed 755-nm laser irradiation can provide a personalized approach for effective treatment of PWS blood vessels by customizing the optical properties of NETs to match the type of vascular plexus (e.g., depth, size, and distribution of blood vessels) and induce the necessary photothermal effects. The theoretical model presented in this study can be used in guiding the fabrication of NETs with an appropriate level of ICG and development of laser dosimetry parameters.

4 Conclusions

Using theoretical models, we have demonstrated that NETs can be beneficial for NIR laser treatment of PWS. NETs can be fabricated with patient-specific optical properties to allow for a personalized treatment based on the size and depth of the blood vessels as well as the pigmentation of the individual's skin. The use of NIR lasers in combination with NETs addresses key challenges in vascular phototherapy by reducing epidermal damage and increasing penetration to deeper blood vessels.

Disclosures

Authors have no relevant financial interests in this paper.

Acknowledgments

This study was supported in parts by grants from the National Institute of Arthritis and Musculoskeletal and Skin Diseases (1R01-AR068067-01A1), U.S. National Science Foundation (CBET-1509218), and the Slovenian Research Agency (ARRS; P1-0192). Institutional support from the Beckman Laser Institute Endowment, Arnold and Mabel Beckman Foundation, and the Packard Foundation is also acknowledged.

References

- K. N. Kanada et al., "A prospective study of cutaneous findings in newborns in the United States: correlation with race, ethnicity, and gestational status using updated classification and nomenclature," *J. Pediatr.* **161**(2), 240–245 (2012).
- G. Kautz et al., "Treatment of resistant port wine stains (PWS) with pulsed dye laser and non-contact vacuum: a pilot study," *Lasers Med. Sci.* **25**(4), 525–529 (2010).
- R. Rajaratnam, S. A. Laughlin, and D. Dudley, "Pulsed dye laser double-pass treatment of patients with resistant capillary malformations," *Lasers Med. Sci.* **26**(4), 487–492 (2011).
- Z. F. Jasim and J. M. Handley, "Treatment of pulsed dye laser-resistant port wine stain birthmarks," *J. Am. Acad. Dermatol.* **57**(4), 677–682 (2007).
- B. Greve and C. Raulin, "Prospective study of port wine stain treatment with dye laser: comparison of two wavelengths (585 nm vs. 595 nm) and two pulse durations (0.5 milliseconds vs. 20 milliseconds)," *Lasers Surg. Med.* **34**(2), 168–173 (2004).
- M. J. Koster et al., "Assessment of clinical outcome after flashlamp pumped pulsed dye laser treatment of port-wine stains: a comprehensive questionnaire," *Plast. Reconstr. Surg.* **102**(1), 42–48 (1998).
- K. M. Kelly et al., "Description and analysis of treatments for port-wine stain birthmarks," *Arch. Facial Plast. Surg.* **7**(5), 287–294 (2005).
- S. L. Jacques, "Optical properties of biological tissues: a review," *Phys. Med. Biol.* **58**(11), R37–R61 (2013).
- J. V. Frangioni, "In vivo near-infrared fluorescence imaging," *Curr. Opin. Chem. Biol.* **7**(5), 626–634 (2003).
- M. V. Marshall et al., "Near-infrared fluorescence imaging in humans with indocyanine green: a review and update," *Open Surg. Oncol. J.* **2**(2), 12–25 (2010).
- A. Klein et al., "A randomized controlled trial to optimize indocyanine green-augmented diode laser therapy of capillary malformations," *Lasers Surg. Med.* **45**(4), 216–224 (2013).
- A. Klein et al., "Indocyanine green-augmented diode laser treatment of port-wine stains: clinical and histological evidence for a new treatment option from a randomized controlled trial," *Br. J. Dermatol.* **167**(2), 333–342 (2012).
- P. E. Liggett et al., "Preliminary results of combined simultaneous trans-pupillary thermotherapy and ICG-based photodynamic therapy for choroidal melanoma," *Ophthalmic Surg. Lasers Imaging* **36**(6), 463–470 (2005).
- A. R. Montazerabadi et al., "The effects of combined treatment with ionizing radiation and indocyanine green-mediated photodynamic therapy on breast cancer cells," *J. Photochem. Photobiol. B* **109**, 42–49 (2012).
- T. Desmettre, J. M. Devoisselle, and S. Mordon, "Fluorescence properties and metabolic features of indocyanine green (ICG) as related to angiography," *Surv. Ophthalmol.* **45**(1), 15–27 (2000).
- B. Bahmani et al., "Effect of polyethylene glycol coatings on uptake of indocyanine green loaded nanocapsules by human spleen macrophages in vitro," *J. Biomed. Opt.* **16**(5), 051303 (2011).
- V. Saxena, M. Sadoqi, and J. Shao, "Polymeric nanoparticle delivery system for indocyanine green: biodistribution in healthy mice," *Int. J. Pharm.* **308**(1–2), 200–204 (2006).
- M. A. Yaseen et al., "Stability assessment of indocyanine green within dextran-coated mesocapsules by absorbance spectroscopy," *J. Biomed. Opt.* **12**(6), 064031 (2007).
- X. Zheng et al., "Enhanced tumor treatment using biofunctional indocyanine green-containing nanostructure by intratumoral or intravenous injection," *Mol. Pharm.* **9**(3), 514–522 (2012).
- A. K. Kirchherr, A. Briel, and K. Mäder, "Stabilization of indocyanine green by encapsulation within micellar systems," *Mol. Pharm.* **6**(2), 480–491 (2009).
- V. B. Rodriguez et al., "Encapsulation and stabilization of indocyanine green within poly(styrene-alt-maleic anhydride) block-poly(styrene) micelles for near-infrared imaging," *J. Biomed. Opt.* **13**(1), 014025 (2008).
- L. Wu et al., "Hybrid polypeptide micelles loading indocyanine green for tumor imaging and photothermal effect study," *Biomacromolecules* **14**(9), 3027–3033 (2013).
- J. C. Kraft and R. J. Y. Ho, "Interactions of indocyanine green and lipid in enhancing near-infrared fluorescence properties: the basis for near-infrared imaging in vivo," *Biochemistry* **53**(8), 1275–1283 (2014).
- S. Mordon et al., "Thermal damage assessment of blood vessels in a hamster skin flap model by fluorescence measurement of a liposome-dye system," *Proc. SPIE* **2975**, 20–31 (1997).
- T. Toyota et al., "Near-infrared-fluorescence imaging of lymph nodes by using liposomally formulated indocyanine green derivatives," *Bioorg. Med. Chem.* **22**(2), 721–727 (2014).
- P. Sharma et al., "Gadolinium-doped silica nanoparticles encapsulating indocyanine green for near infrared and magnetic resonance imaging," *Small* **8**(18), 2856–2868 (2012).
- G. Kim et al., "Indocyanine-green-embedded PEBBLES as a contrast agent for photoacoustic imaging," *J. Biomed. Opt.* **12**(4), 044020 (2007).
- B. Bahmani, D. Bacon, and B. Anvari, "Erythrocyte-derived phototheranostic agents: hybrid nano-vesicles containing indocyanine green for near infrared imaging and therapeutic applications," *Sci. Rep.* **3**, 2180 (2013).
- J. M. Burns et al., "Optical properties of biomimetic probes engineered from erythrocytes," *Nanotechnology* **28**(3), 035101 (2017).
- M. Milanič and B. Majaron, "Three-dimensional Monte Carlo model of pulsed-laser treatment of cutaneous vascular lesions," *J. Biomed. Opt.* **16**(12), 128002 (2011).
- B. Majaron, M. Milanič, and J. Premru, "Monte Carlo simulation of radiation transport in human skin with rigorous treatment of curved tissue boundaries," *J. Biomed. Opt.* **20**(1), 015002 (2015).
- G. Liu et al., "In vivo, high-resolution, three-dimensional imaging of port wine stain microvasculature in human skin," *Lasers Surg. Med.* **45**(10), 628–632 (2013).
- S. L. Jacques, "Skin optics," 1998, <http://omlc.org/news/jan98/skinoptics.html>

34. H. Lu et al., "Melanin content and distribution in the surface corneocyte with skin phototypes," *Br. J. Dermatol.* **135**(2), 263–267 (1996).
35. A. Kienle and R. Hibst, "A new optimal wavelength for treatment of port wine stains?" *Phys. Med. Biol.* **40**(10), 1559–1576 (1995).
36. R. Graaff et al., "Optical properties of human dermis in vitro and in vivo," *Appl. Opt.* **32**(4), 435–447 (1993).
37. N. Bosschaart et al., "A literature review and novel theoretical approach on the optical properties of whole blood," *Lasers Med. Sci.* **29**(2), 453–479 (2014).
38. F. A. Duck, *Physical Properties of Tissue: A Comprehensive Reference*, Academic Press, San Diego, London (1990).
39. J. S. Nelson et al., "Dynamic epidermal cooling during pulsed laser treatment of port-wine stain. A new methodology with preliminary clinical evaluation," *Arch. Dermatol.* **131**(6), 695–700 (1995).
40. B. M. Pikkula et al., "Effects of droplet velocity, diameter, and film height on heat removal during cryogen spray cooling," *Ann. Biomed. Eng.* **32**(8), 1133–1142 (2004).
41. T. Dai et al., "Comparative study of cryogen spray cooling with R-134a and R-404a: implications for laser treatment of dark human skin," *J. Biomed. Opt.* **11**(4), 041116 (2006).
42. T. Dai et al., "Comparison of human skin opto-thermal response to near-infrared and visible laser irradiations: a theoretical investigation," *Phys. Med. Biol.* **49**(21), 4861–4877 (2004).
43. J. W. Tunnell, L. V. Wang, and B. Anvari, "Optimum pulse duration and radiant exposure for vascular laser therapy of dark port-wine skin: a theoretical study," *Appl. Opt.* **42**(7), 1367–1378 (2003).
44. G. Aguilar et al., "Influence of nozzle-to-skin distance in cryogen spray cooling for dermatologic laser surgery," *Lasers Surg. Med.* **28**(2), 113–120 (2001).
45. B. Majaron et al., "Intermittent cryogen spray cooling for optimal heat extraction during dermatologic laser treatment," *Phys. Med. Biol.* **47**(18), 3275–3288 (2002).
46. J. A. Weaver and A. M. Stoll, "Mathematical model of skin exposed to thermal radiation," *Aerosp. Med.* **44**(3), 316 (1969).
47. J. R. Lepock et al., "Relationship of hyperthermia-induced hemolysis of human erythrocytes to the thermal denaturation of membrane proteins," *Biochim. Biophys. Acta* **980**(2), 191–201 (1989).
48. L. Izikson and R. R. Anderson, "Treatment endpoints for resistant port wine stains with a 755 nm laser," *J. Cosmet. Laser Ther.* **11**(1), 52–55 (2009).
49. T. Lister, P. A. Wright, and P. H. Chappell, "A new Monte Carlo program for simulating light transport through port wine stain skin," *Lasers Med. Sci.* **29**(3), 1017–1028 (2014).
50. T. J. Pfefer et al., "Modeling laser treatment of port wine stains with a computer-reconstructed biopsy," *Lasers Surg. Med.* **24**(2), 151–166 (1999).
51. C. M. J. Hu et al., "Erythrocyte membrane-camouflaged polymeric nanoparticles as a biomimetic delivery platform," *Proc. Natl. Acad. Sci. U. S. A.* **108**(27), 10980–10985 (2011).
52. J. G. Piao et al., "Erythrocyte membrane is an alternative coating to polyethylene glycol for prolonging the circulation lifetime of gold nanocages for photothermal therapy," *ACS Nano* **8**(10), 10414–10425 (2014).
53. P. A. Oldenburg, "CD47: a cell surface glycoprotein which regulates multiple functions of hematopoietic cells in health and disease," *ISRN Hematol.* **2013**, 1–19 (2013).
54. P. L. Rodriguez et al., "Minimal 'self' peptides that inhibit phagocytic clearance and enhance delivery of nanoparticles," *Science* **339**(6122), 971–975 (2013).
55. P. Burger et al., "CD47 functions as a molecular switch for erythrocyte phagocytosis," *Blood* **119**(23), 5512–5521 (2012).
56. J. T. Mac et al., "Erythrocyte-derived nano-probes functionalized with antibodies for targeted near infrared fluorescence imaging of cancer cells," *Biomed. Opt. Express* **7**(4), 1311–1322 (2016).
57. L. Rao et al., "Erythrocyte membrane-coated upconversion nanoparticles with minimal protein adsorption for enhanced tumor imaging," *ACS Appl. Mater. Interfaces* **9**(3), 2159–2168 (2017).

Joshua M. Burns is a PhD candidate in bioengineering at the University of California, Riverside, California, USA. He received his MS degree in bioengineering from the University of California, Riverside, California, USA, and his BS degrees in physics and biological sciences from the University of California, Irvine, California, USA. His research focuses on the development of biological optical constructs and instrumentation for biomedical imaging and phototherapy of cancer and vascular diseases.

Wangcun Jia, PhD, is an associate researcher at the Beckman Laser Institute and Medical Clinic at the University of California, Irvine, California, USA. He received his PhD in mechanical engineering from the Hong Kong University of Sciences and Technology. His research interests include thin-film thermal sensor, environment-friendly cryogen spray cooling, numerical, preclinical, and clinical investigation of laser treatment of port-wine stain birthmarks, and erythrocyte-based drug delivery and associated near-infrared laser therapy.

J. Stuart Nelson, MD, PhD, is a professor of surgery and biomedical engineering at the University of California, Irvine, California, USA, and a medical director of the Beckman Laser Institute and Medical Clinic. He served as a president of the American Society for Laser Medicine and Surgery (ASLMS) 2001 to 2002 and editor-in-chief of the ASLMS journal *Lasers in Surgery and Medicine* 2005 to 2015. He has published more than 340 scientific articles and 20 book chapters.

Boris Majaron is a senior staff member at Jožef Stefan Institute in Ljubljana, Slovenia, and an associate professor of physics at the University of Ljubljana. The focus of his research efforts has gradually shifted from optical spectroscopy of solid-state laser materials to biomedical optics, e.g., advancement of various laser treatments, characterization of biological tissues by photothermal radiometry and/or diffuse reflectance spectroscopy, and development of nanostructures for bioimaging and phototherapy.

Bahman Anvari, PhD, is a professor in the Department of Bioengineering at UC Riverside. His research interests are in photomedicine with current focus on the engineering of delivery platforms for light-based theranostic applications. He has authored more than 350 scientific articles. He is a fellow of the American Association for the Advancement of Science (AAAS), American Institute for Medical and Biological Engineering (AIMBE), Biomedical Engineering Society (BMES), and SPIE.