

# Lawrence Berkeley National Laboratory

## Recent Work

### Title

CHOICE OF COORDINATE SYSTEM IN ELLIPSOMETRY

### Permalink

<https://escholarship.org/uc/item/0hp157rm>

### Author

Muller, Rolf H.

### Publication Date

1967-09-01

c.7

# University of California

## Ernest O. Lawrence Radiation Laboratory

CHOICE OF COORDINATE SYSTEM IN ELLIPSOMETRY

Rolf H. Muller

September 1967

TWO-WEEK LOAN COPY

*This is a Library Circulating Copy  
which may be borrowed for two weeks.  
For a personal retention copy, call  
Tech. Info. Division, Ext. 5545*

Berkeley, California

UCRL-17841

c.2

## **DISCLAIMER**

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

UCRL-17841  
UC-32 Math. and Comp.  
TID-4500 (50th Ed.)

UNIVERSITY OF CALIFORNIA  
Lawrence Radiation Laboratory  
Berkeley, California  
AEC Contract No. W-7405-eng-48

CHOICE OF COORDINATE SYSTEM IN ELLIPSOMETRY

Rolf H. Müller

September 1967

CONFIDENTIAL - SECURITY INFORMATION  
EXCLUDED FROM AUTOMATIC DOWNGRADING AND  
DECLASSIFICATION

Printed in the United States of America  
Available from

Clearinghouse for Federal Scientific and Technical Information  
National Bureau of Standards, U.S. Department of Commerce  
Springfield, Virginia 22151

Price: Printed Copy \$3.00; Microfiche \$0.65

## CHOICE OF COORDINATE SYSTEM IN ELLIPSOMETRY

Rolf H. Muller

Inorganic Materials Research Division, Lawrence Radiation Laboratory  
University of California, Berkeley, California

The choice of coordinate system made in two previous reports<sup>1,2</sup> for the reflection of light affects the results derived in an important way which deserves more comment. Due to the choice made for the positive direction of the reflected electric field component  $\vec{E}_p$  " polarized parallel to the plane of incidence (see Fig. 1) the coordinate system  $(\vec{E}_p \vec{E}_s \vec{k}_o)$  for the incident wave is right handed while that for the reflected wave  $(\vec{E}_p \vec{E}_s \vec{k}_o)$  is left handed and all the data derived pertain to this system. Therefore, if the usual x-y coordinate system is employed for the incident wave (looking into the beam) and p and s components are substituted for the case of reflection on a vertical surface (Fig. 2a), then the positive p direction appears reversed after reflection (Fig. 2b).

Since it is preferable to use a right-handed coordinate system as a laboratory frame of reference to analyze the elliptic polarization, and since the previous analysis of ellipse shape and orientation<sup>3</sup> has been based on the usual right hand orientation of the x and y coordinates, either of the following two equivalent coordinate transformations for incident or reflected light, respectively, may be made:

- a. Change the coordinate system of the incident light to a left-handed system: For incident linearly polarized light this entails the simple replacement of the  $+45^\circ$  azimuth by  $-45^\circ$  in order to produce s and p components which are in phase and of equal amplitude. For the reflected wave, shape, orientation and sense of rotation

of the ellipse can then be considered in the usual right-handed coordinate system (Figs. 5 and 6, ref. 3) with the values for  $\psi$  and  $\Delta$  derived by the theory.

b. Change the coordinate system for the reflected light to a right-handed system: A reflection of the ellipse around the  $\vec{E}_s$  axis is necessary for this transformation. Since this operation changes orientation and sense of rotation of the ellipse in the laboratory frame of reference, the value of  $\theta$  is replaced by  $(180^\circ - \theta)$  and the sign of  $\gamma$  is reversed (Fig. 3). Thus for given values of  $\psi$  (e.g.  $\sim 60^\circ$ ) and  $\Delta$  (e.g.  $\sim -45^\circ$ ) the ellipse expected to be of the form of Fig. 3a from Fig. 5, ref. 3 is observed as in Fig. 3b and described by the right-hand system of Fig. 3c. Considered in the same coordinate system, the first and the last ellipse show the same amplitude ratio  $\tan\psi$  but differ by  $\pi$  in phase difference  $\Delta$ . Thus, this second coordinate transformation is equivalent to the addition (or subtraction) of  $\pi$  to the computed value of  $\Delta$  in order to predict the correct ellipse orientation to be observed. Thus, the dependence of ellipse orientation on phase difference  $\Delta$  shown in Fig. 4 is obtained from the previously derived Fig. 5 (Fig. 6, ref. 3).


**Acknowledgment:**

This work was performed under the auspices of the United States Atomic Energy Commission.

REFERENCES


1. J. R. Mowat and R. H. Muller, Reflection of Polarized Light from Absorbing Media, UCRL-11813, August 1966).
2. J. R. Mowat and R. H. Muller, Reflection of Polarized Light from Film-Covered Surfaces, UCRL-17128, February 1967.
3. R. H. Muller and J. R. Mowat, Analysis of Elliptically Polarized Light, UCRL-11802, October 1965.






MUB-12905

Fig. 1 Sign conventions for the electric field (positive direction of  $\vec{E}_s$  and  $\vec{E}_p$  in incident, reflected and refracted waves indicated by arrows). The complex index of refraction of the absorbing medium is designated  $n(1+i\kappa)$ . The subscript p stands for polarization parallel and the subscript s for polarization normal to the plane of incidence.



XBL679-5276


Fig. 2. Coordinate system for incident (a) and reflected (b) waves as seen looking toward the source for reflection on a vertical surface.



XBL679-5277


Fig. 3 Coordinate systems for reflected light. Derivation of ellipse from computed values of  $\psi$  and  $\Delta$  (assumed here approximately  $60^\circ$  and  $-45^\circ$  respectively).

- a. Ellipse expected in right-hand system on the basis of  $\psi$  and  $\Delta$  values computed for left hand system
- b. observed ellipse due to left hand system used in computation
- c. ascribed orientation of observed ellipse in right hand laboratory system



MUB-8168-A

Fig. 4 Ellipse orientation observed in right-hand laboratory coordinate system with phase difference  $\Delta$  computed for left-hand coordinate system.



XBL679-5278

Fig. 5 Ellipse orientation in right-hand coordinate system with phase difference  $\Delta$  for right-hand system.

This report was prepared as an account of Government sponsored work. Neither the United States, nor the Commission, nor any person acting on behalf of the Commission:

- A. Makes any warranty or representation, expressed or implied, with respect to the accuracy, completeness, or usefulness of the information contained in this report, or that the use of any information, apparatus, method, or process disclosed in this report may not infringe privately owned rights; or
- B. Assumes any liabilities with respect to the use of, or for damages resulting from the use of any information, apparatus, method, or process disclosed in this report.

As used in the above, "person acting on behalf of the Commission" includes any employee or contractor of the Commission, or employee of such contractor, to the extent that such employee or contractor of the Commission, or employee of such contractor prepares, disseminates, or provides access to, any information pursuant to his employment or contract with the Commission, or his employment with such contractor.

