Lawrence Berkeley National Laboratory

Recent Work

Title

An automated multiplexed turbidometric and data collection system for measuring growth kinetics of anaerobes dependent on gaseous substrates.

Permalink

<https://escholarship.org/uc/item/26g1r28r>

Authors

Hunt, Kristopher A
Forbes, Jonathan
Taub, Fred
et al.

Publication Date

2021-09-01

DOI

10.1016/j.mimet.2021.106294

Peer reviewed

**An automated multiplexed turbidometric and data collection system for measuring growth
kinetics of anaerobes dependent on gaseous substrates**

Kristopher A. Hunt ^a, Jonathan Forbes ^a, Fred Taub ^a, Nicholas Elliott ^a, Jessica Hardwicke ^a,
Robert Petersen ^a, Nejc Stopnisek ^b, David A. C. Beck ^c, and David A. Stahl ^a

^a Civil and Environmental Engineering, University of Washington Seattle, More Hall Box
352700, Seattle, WA 98195

^b Department of Microbial Ecology, Netherlands Institute of Ecology, 6708 PB Wageningen, the
Netherlands

^c Department of Chemical Engineering, University of Washington Seattle, Benjamin Hall 440,
Seattle, WA 98105

CORRESPONDING AUTHOR

Kristopher A. Hunt
Civil & Environmental Engineering
University of Washington
201 More Hall, Box 352700
Seattle, WA 98195-2700
Email: hunt0362@uw.edu

KEYWORDS: anaerobic, multiplex, automated, optical density, microbial

Abstract

Standard methods of monitoring the growth kinetics of anaerobic microorganisms are generally impractical when there is a protracted or indeterminate period of active growth, and when high numbers of samples or replications are required. As part of our studies of the adaptive evolution of a simple anaerobic syntrophic mutualism, requiring the characterization of many isolates and alternative syntrophic pairings, we developed a multiplexed growth monitoring system using a combination of commercially available electronics and custom designed circuitry and materials. This system automatically monitors up to 64 sealed, and as needed pressurized, culture tubes and reports the growth data in real-time through integration with a customized relational database. The utility of this system was demonstrated by resolving minor differences in growth kinetics associated with the adaptive evolution of a simple microbial community comprised of a sulfate reducing bacterium, *Desulfovibrio vulgaris*, grown in syntrophic association with *Methanococcus maripaludis*, a hydrogenotrophic methanogen.

Highlights

- The ODIn supports parallel quantification of up to 64 cultures under a defined gas headspace.
- High throughput culture monitoring resolved difficult to measure differences in growth rate.
- Real-time data visualization allows for immediate feedback on experiment progress.

Introduction

A variety of automated systems for monitoring microbial growth based on changes in turbidity or fluorescence are available, using small reactor systems (Takahashi et al., 2015; Toprak et al., 2013; Wong et al., 2018), microtiter plate readers (Duetz et al., 2000), or more specialized commercial systems (Vuono et al., 2019). Such systems however are generally not suitable for monitoring the growth of anaerobic microbial cultures, particularly those that require a gaseous substrate for growth or depend on a close hydrogen-based syntrophic coupling, upon which many anaerobic microbial food webs depend. In addition, the slow growth of many fastidious anaerobes requires that cultures be monitored continuously at regular intervals over multiple day periods. As part of ongoing studies on simple microbial communities composed of the hydrogenotrophic methanogen *Methanococcus maripaludis* S2 coupled with a facultatively syntrophic sulfate-reducing bacterium, *Desulfovibrio vulgaris* Hildenborough, we fabricated an automated multiplexed system to monitor growth. Essential advantages of the new system are compatibility with the Balch-type anaerobic culture tubes commonly used in the cultivation of fastidious anaerobes and a capacity for highly replicated data collection by simultaneous monitoring of up to 64 culture tubes at high temporal resolution. An important feature that distinguishes it from other available devices is the automated control of the sample holder platform, providing for both maximum gas exchange for culture growth in the down position and accurate optical density readings in the up position. Thus, the optical density readings and growth kinetics are directly comparable to data collected using the more cumbersome manual format.

The Optical Density Instrument (ODIn) was designed to consistently agitate sealed culture vessels, monitor growth of cultures via a noninvasive metric, and maintain relatively constant

conditions. This instrument will accommodate up to 64 pressurized tubes on a remotely controlled shaking platform, which automatically adjusts the angle of the tubes to accommodate periodic readings during continuous growth. To achieve good mixing and gas exchange of the culture medium during shaking, the tubes are positioned horizontally. Optical density is measured by temporarily arresting the shaking of the sample holder platform, raising the sample platform to a near vertical position, and then serially collecting optical density readings from all 64 channels using light emitter and sensor pairs on opposing sides of each culture tube. Sensor output data is reported automatically through a wired Ethernet connection and recorded on a remote database server where a custom data collection program displays the sensor data in real-time, converts sensor data to optical density, calculates changing growth rates, and offers direct comparisons of growth kinetics from individual cultures within or between different growth experiments. Thus, growth is quantified with very little disruption of the sample during monitoring periods that may be up to days for the slow growing cultures analyzed in this report. We here present a detailed description of the construction and operation of ODIn, demonstrating its application for quantifying and resolving small differences in the growth kinetics that arise through the adaptive evolution of a model microbial community composed of the hydrogenotrophic methanogen *Methanococcus maripaludis* growing in syntrophic association with *Desulfovibrio vulgaris*.

Materials and Methods

System overview. The ODIn system was constructed using commercially available electronic components, including sensing and control components operated by a microcontroller to

99 automate the collection of optical density measurements over a time series (Fig. 1). The system
100 is designed to house 64 closed Balch tubes (18x150 mm, Bellco Glass Inc., Vineland, NJ), each
101 tube sealed with a crimp-closed rubber stopper for gas retention. Tubes are housed in an array of
102 eight identical machined modular resin tube racks, each designed to accommodate eight
103 individual tubes in a linear array secured by a removable top cover and securing O-rings (Fig. 2).
104 The racks are fastened to a shaker table either using compression straps or by insertion into a
105 custom housing.

106 Using a microcontroller to direct operations, the optical density of each sample is quantified
107 using individually paired light emitting diodes (LEDs) and receiving phototransistors (Optek
108 Inc., Galena, OH) integrated into the base of each rack (Fig. 2B). The electrical leads for the
109 sensor sets are attached to a custom printed circuit board (PCB, ExpressPCB, Mulino, OR)
110 secured in a recess at the rack base by a removable bottom plate. The linear range of each sensor
111 set was manually established prior to installation by quantifying the milliamps (mA) output at
112 known optical density at 600 nm (Hach Co. Loveland, CO) using McFarland standards covering
113 a range of optical densities (Fig. 3).

114

Fig. 1. Schematic diagram of control and sensing elements for the ODIn system. The system combines commercially available components with custom circuitry through an Arduino microcontroller platform to coordinate actions of the shaker platform, turbidity measurement frequency, and data reporting.

Fig. 2. Tube rack (**A**) machined from acetyl resin to house eight culture tubes in individual sensor channels. Sensors consisting of a LED and phototransistor (PT) are positioned opposite one another near the bottom of each rack (**B**) and connected to a PCB embedded in the base of the rack.

Fig. 3. Relationship between sensor output (milliamps) and corresponding spectrophotometric readings of McFarland turbidity standards.

The tube racks are secured to a hinged acrylic lid attached to a raised platform (Fig. 4) secured on a New Brunswick Innova 2300 orbital shaker (Eppendorf, Hamburg, DE). The lid of the platform is manipulated by a linear actuator (Servocity, Winfield, KS) controlled by a separate circuit board (Canakit, North Vancouver, BC) through a tethered cable connector. During shaking, tubes are maintained in a horizontal position. At user defined intervals, the shaker is stopped, and the tubes are raised to approximately 50-degrees above the horizontal prior to taking OD readings. Sequential readings of the 64 sample tubes requires a period of two-three minutes before lowering the platform and reinitiating shaking. A minimum of 5-mL of medium is needed for an accurate reading of culture turbidity. Reading frequency is controlled by microcontroller software using programmed default settings of 5-, 20-, 60-, and 120-minute reading intervals selected with a rotary knob on the system control box (Fig. 5).

Fig. 4. Construction of platform (A) and linear actuator (B) for raising tube racks in preparation for periodic measurements of turbidity. Following shaker table deactivation, the linear actuator unfolds a scissor hinge, transitioning tubes from a horizontal to a near vertical position.

Fig. 5. Control box. Controls for different functions are located on the lid and include a system status feedback display LCD screen, a knob controlling a four-position rotary switch for selecting preset sampling time intervals, an eight-position rotary knob to select a single tube rack for display of sensor output values during run and debug modes, and a push button “Pause” switch for system pausing and activation.

The system control box is separate from sensing components of the system (Fig. 5). A custom acrylic box houses all system control components including the microcontroller, communication devices, power supply, motor controller, and control interfaces, all of which communicate through a custom PCB controller (ExpressPCB, Mulino, OR). Access to internal components is by a hinged lid on the control box. Communication with each tube rack is through a serial cable connection to eight DE-9 female receptacles (NorComp, Charlotte, NC) located on the side of the control box (Fig. 1). System power is provided by a 12VDC 100W panel mount power supply powered through a standard external 120 VAC female appliance coupler to a 120 VAC wall outlet. Operation control interfaces are located on the control box lid and include a system status feedback display LCD screen (Adafruit Industries, LLC New York, NY), a four-position rotary knob for selecting preset sampling time intervals, an eight-position rotary knob to select a single tube rack for display of sensor output values during a debugging mode, a two position toggle switch for selecting between device modes “Run” and “Debug”, and a push button “Pause” switch for system pausing and activation. Direct communication with the Arduino Mega microcontroller is through the USB Type-B embedded port (Adafruit Industries, LLC New York, NY) accessed on the side of the control box. Local data storage is performed using an SD card Data Logger Shield (Adafruit Industries, LLC New York, NY), which provides an SD card

receptacle for a standard SD card accessed on the side of the control box. An Ethernet connection for data output is through an Ethernet Shield (Adafruit Industries, LLC New York, NY) using a static IP address programmed into the Arduino and accessed on the side of the control box. The orbital shaker is plugged into the side of the control box and controlled by a AC interrupting relay via the Arduino.

During normal operation, the interval between OD readings is controlled by presets on a selectable rotary knob. At the selected time intervals, the microcontroller deactivates the orbital shaker. Following a ten second delay to allow the orbital shaker to come to a full stop, the hinged acrylic platform lid is raised, and sensor readings are taken sequentially for the 64 channels. Those data are stored on a local SD card as well as reported to a remote server through the Ethernet connection. The data can be visualized and archived in real-time for all 64 channels using custom software developed on FileMaker Advanced version 14 (Claris International, Inc., USA), the host application has been tested through FileMaker Server 15 and the client application has been tested through FileMaker Pro 17 (Fig. 6A). The FileMaker algorithm processes ODIn data collected through a server interface but operates independently from the ODIn hardware. Upon completion of the data acquisition and reporting procedure, the linear actuator retracts, and the shaker reactivates. A full description of the device, fabrication, and operation is provided in supplementary information.

Fig. 6. Continuous monitoring of growth using the FileMaker interface provides direct feedback of experimental progress (A) and is indistinguishable from manually collected measurements (B). Example FileMaker display of growth curves for eight channels of a single tube rack where data can be displayed as milliamps or converted to optical density values after completion of the experiment. Growth is tracked in real-time and updated as new data packets are received from the sensor channels.

System operation. Before initiating a growth study, the system is held in a paused state to allow LED light output to stabilize at the operating temperature. This occurs within four hours when the ambient temperature is 37°C. A two-position toggle switch on the lid of the control box is used to select between the operational mode used for data collection during an experimental run and a debugging program mode for continuous displaying of sensor output in real-time on the control box system status LCD screen. In the debugging routine, the tube racks are elevated using the hinged lifting platform and the output from each tube in a tube rack is continuously displayed on the control box LCD screen, with selection of readings from each of the eight racks controlled by a rotary knob on the control box (Fig. 5). This procedure allows for immediate sensor value data to be displayed from any sensing channel and is a useful diagnostic tool for identifying faults and determining optical working ranges.

210

211 An experimental run is initiated by depressing the pause button on the control box when the
212 toggle switch is in the “up” position. This action raises the hinged lifting platform acrylic lid and
213 initiates the calibration procedure on each of the 64 sensing channels. Once calibration is
214 complete, the program takes the first measurement of each sample, and then reads the status of
215 the four-position rotary switch on the control box that determines the sampling interval of
216 programmable preset values, either 5, 20, 40, or 60 minutes. These intervals are configurable in
217 the program software and may be readjusted during an experiment by switching to any of the
218 four values using the selectable dial. The linear actuator then retracts, lowering the lifting
219 platform acrylic lid to return to a horizontal position and a countdown commences until the next
220 data collection point. The control box system status LCD screen displays the selected interval
221 time in minutes and the active countdown in milliseconds until the next sampling. If the systems
222 pause button is depressed during an active experiment it will continue to count down to the set
223 data collection time but will not collect data until the system is resumed. This allows samples to
224 be removed during an active experiment for external sampling.

225

226 When the time interval between samplings is reached, the orbital shaker is depowered through
227 the interrupting relay and the tube racks are raised for the next measurement. To mitigate the
228 effects of random spikes in sensor readings, a total of 20 ADC values are taken for each sensing
229 circuit at each sampling point and averaged for the reported value. All 64 data points, along with
230 the elapsed time since testing began, are sent to the data collection server via UDP datagram
231 transmission over an Ethernet network connection. The measurements are also recorded on an

onboard SD card located in the SD Data Logger Shield in the control box as a backup in a timestamped .CSV file. An onboard real-time clock with coin cell battery backup keeps track of the current date and time. Data is both reported and recorded in the .CSV file format and arranged with the output of each channel's reading in separate columns. Output data are milliamp values as recorded by the ADC. Upon completion of the experiment, each culture is measured at 600-nm in a spectrophotometer and pre- and post-experiment OD₆₀₀ values used to convert milliamp values to an OD₆₀₀ (Fig. 6). A step-by-step protocol of general operation for an experiment using the FileMaker interface is described with images below.

General operational protocol

Step 1: Turn on the ODIn control box and leave “Paused” for equilibration

Step 2: Setup a run in the FileMaker interface

Step 2.1: Click the “+” button in the bottom left corner of the interface

Step 2.2: Enter a unique name for the experiment

Step 2.3: Return to the home screen and click “start” for the experiment

Step 3: Setup the ODIn

Step 3.1: Prepare sealed Balch tubes with at least 5-ml of media

Step 3.2: Record first manual OD₆₀₀ measurements for 2-point calibration

Step 3.3: Load Balch tubes into the ODIn tube racks

Step 3.4: Press the “Pause” button on the ODIn control box to initiate ODIn

Monitor mA values reported at intervals set on the control box for “live” feedback of run

Step 4: Terminate the experiment

Step 4.1: Click the “Stop” button in the FileMaker interface

Step 4.2: Turn off the ODIn control box and remove Balch tubes

Step 4.3: Take second manual OD₆₀₀ measurement for 2-point calibration

Step 4.4: Enter metadata under the appropriate experiment using the FileMaker interface

Step 5: Analyze data for desired parameters

243

244 **Cultivation and data analysis.** Culture media were prepared as previously described (Lim et

245 al., 2014) containing 7.5 mM lactate and 5 mM sulfite, 30 mM lactate and 15 mM sulfate, and 10

mM acetate or 30 mM lactate to facilitate growth of *D. vulgaris* and *M. maripaludis* monocultures and co-cultures, respectively. Balch tubes were filled with between 10-mL and 20-mL of anaerobically prepared liquid media with a headspace of 80% N₂/20% CO₂ for both *D. vulgaris* and syntrophic cocultures and 80% H₂/20% CO₂ pressurized to 30 psi for *M. maripaludis* cultures. Replicate culture tubes were inoculated with cultures recovered from freezer stocks as previously described (Hillesland and Stahl, 2010) and grown at 37 °C with continuous shaking at 300 rpm. Sensor readings (mA) collected at 20-minute intervals were converted to optical density (OD₆₀₀) by relating a spectrophotometric measurement taken for each culture tube at run completion with the final mA value. Growth kinetics were analyzed using the logistic fit option of the grofit (Kahm et al., 2010) packages developed for R-project (R version 3.2.3, <https://www.R-project.org/>).

Results and Discussion

The utility of the system was evaluated using comparative analysis of available cultures previously shown to exhibit minor to significant differences in growth rate and yield. Using this sensor design the ODIn system can characterize growth patterns with a precision and sensitivity difficult to achieve by manual reading (Fig. 6). The system was also used to examine changes in growth kinetics associated with the evolution of a simple microbial mutualism established between *D. vulgaris* Hildenborough and *M. maripaludis* S2 (Turkarslan et al., 2021). Prior studies have shown that this forced syntrophic mutualism, based on interspecies hydrogen transfer, improved rapidly through adaptive evolution, increasing several fold in growth rate and yield within a few hundred generations of initial pairings (Hillesland and Stahl, 2010). Ongoing

analysis of mutations that have accumulated in replicated cultures suggests that both common and divergent patterns of mutation accumulation in different populations are associated with growth improvement. For example, many *D. vulgaris* in replicated evolution lineages lost the ability for sulfate respiration due to nonsense mutations in genes coding for sulfate activation and reduction to sulfite (Turkarslan et al., 2021).

Resolving the contributions of those and other mutations to community improvement has been complicated by the large number of mutations present at low frequency in each evolution lineage reflecting by the emergence of multiple genotypes of evolved *D. vulgaris* and *M. maripaludis* populations (Hillesland et al., 2014). In order to identify mutations within a single evolved genotype contributing to community growth improvement, clones of evolved *D. vulgaris* and *M. maripaludis* at different generations of evolution were isolated from different replicated evolution lineages. Growth kinetics of individual clones in monoculture, as well as when syntrophically paired within and between different evolution lineages were measured using the ODIn system (Fig. 7). Unlike the highly reproducible kinetics of growth in monoculture (Fig. 6), growth was much more variable when individual evolved organisms were paired, presumably reflecting minor differences in initial conditions. Thus, analysis of many replicates, as was made feasible with the ODIn system, was essential for resolving minor differences in growth kinetics.

Fig. 7. Growth parameters for the slow growing *Desulfovibrio vulgaris* Hildenborough (DvH) and *Methanococcus maripaludis* S2 (Mmp) cocultures were determined for different ancestral and evolved strains indicating fitness benefits of coevolution challenging to resolve through manual measurements. Colors indicate different combinations of strains and error bars indicate one standard error of measurement with number of replicates (N) indicated.

The ODIn system automates the collection of optical density readings over the extended growth periods of days for these slow growing cultures, eliminating the constraints of experimenter fatigue and variability of performing manual measurements. An exemplary growth curve of *D. vulgaris* using ODIn data and manual measurements over 26 hours demonstrates the agreement of the new device and traditional methods (Fig. 6B). Thus, in addition to a much greater capacity for multiplexing than other available systems for monitoring the growth of anaerobes, the current format fully replicates that of established manual methods. The format also opens other developmental opportunities, for example implementing tunable wavelengths in the device

for monitoring of optical density at more specialized wavelengths, such as 550, 650, and 660 nm, designed to avoid confounding signals from substrate or products or optimized to differentiate biomass components. The highly replicated collection of growth data made feasible by the ODIn system is now providing an essential foundation to identify mutations and combinations of mutations contributing to improved mutualistic growth of the evolved co-cultures. More generally, the system offers broad utility for microbiological studies quantifying the growth kinetics of fastidious slow growers and those dependent on gaseous substrates for growth, such as methanotrophs and Knallgas bacteria, and those dependent interspecies exchange of gaseous metabolites as examined in this demonstration of the ODIn system operation.

Acknowledgements

This research by ENIGMA- Ecosystems and Networks Integrated with Genes and Molecular Assemblies (<http://enigma.lbl.gov>), a Science Focus Area Program at Lawrence Berkeley National Laboratory is based upon work supported by the U.S. Department of Energy, Office of Science, Office of Biological & Environmental Research under contract number DE-AC02-05CH11231.

References

Duetz, W.A., Rüedi, L., Hermann, R., O'Connor, K., Büchs, J., Witholt, B., 2000. Methods for Intense Aeration, Growth, Storage, and Replication of Bacterial Strains in Microtiter Plates. Appl. Environ. Microbiol. 66, 2641–2646. <https://doi.org/10.1128/AEM.66.6.2641->

- 324 2646.2000
- 325 Hillesland, K.L., Lim, S., Flowers, J.J., Turkarslan, S., Pinel, N., Zane, G.M., Elliott, N., Qin, Y.,
326 Wu, L., Baliga, N.S., Zhou, J., Wall, J.D., Stahl, D.A., 2014. Erosion of functional
327 independence early in the evolution of a microbial mutualism. *Proc. Natl. Acad. Sci.* 111,
328 14822–14827. <https://doi.org/10.1073/pnas.1407986111>
- 329 Hillesland, K.L., Stahl, D.A., 2010. Rapid evolution of stability and productivity at the origin of
330 a microbial mutualism. *Proc. Natl. Acad. Sci.* 107, 2124–2129.
331 <https://doi.org/10.1073/pnas.0908456107>
- 332 Kahm, M., Hasenbrink, G., Lichtenberg-Fraté, H., Ludwig, J., Kschischo, M., 2010. grofit :
333 Fitting Biological Growth Curves with R. *J. Stat. Softw.* 33.
334 <https://doi.org/10.18637/jss.v033.i07>
- 335 Lim, S., Stolyar, S., Hillesland, K., 2014. Culturing Anaerobes to Use as a Model System for
336 Studying the Evolution of Syntrophic Mutualism. pp. 103–115. [https://doi.org/10.1007/978-](https://doi.org/10.1007/978-1-4939-0554-6_8)
337 [1-4939-0554-6_8](https://doi.org/10.1007/978-1-4939-0554-6_8)
- 338 Takahashi, C.N., Miller, A.W., Ekness, F., Dunham, M.J., Klavins, E., 2015. A Low Cost,
339 Customizable Turbidostat for Use in Synthetic Circuit Characterization. *ACS Synth. Biol.*
340 4, 32–38. <https://doi.org/10.1021/sb500165g>
- 341 Toprak, E., Veres, A., Yildiz, S., Pedraza, J.M., Chait, R., Paulsson, J., Kishony, R., 2013.
342 Building a morbidostat: an automated continuous-culture device for studying bacterial drug
343 resistance under dynamically sustained drug inhibition. *Nat. Protoc.* 8, 555–567.
344 <https://doi.org/10.1038/nprot.nprot.2013.021>

- 345 Turkarslan, S., Stopnisek, N., Thompson, A.W., Arens, C.E., Valenzuela, J.J., Wilson, J., Hunt,
346 K.A., Hardwicke, J., de Lomana, A.L.G., Lim, S., Seah, Y.M., Fu, Y., Wu, L., Zhou, J.,
347 Hillesland, K.L., Stahl, D.A., Baliga, N.S., 2021. Synergistic epistasis enhances the co-
348 operativity of mutualistic interspecies interactions. *ISME J.* [https://doi.org/10.1038/s41396-](https://doi.org/10.1038/s41396-021-00919-9)
349 [021-00919-9](https://doi.org/10.1038/s41396-021-00919-9)
- 350 Vuono, D.C., Lipp, B., Staub, C., Loney, E., Harrold, Z.R., Grzymiski, J.J., 2019. A real-time
351 multiplexed microbial growth intervalometer for capturing high-resolution growth curves.
352 *Front. Microbiol.* <https://doi.org/10.3389/fmicb.2019.01135>
- 353 Wong, B.G., Mancuso, C.P., Kiriakov, S., Bashor, C.J., Khalil, A.S., 2018. Precise, automated
354 control of conditions for high-throughput growth of yeast and bacteria with eVOLVER.
355 *Nat. Biotechnol.* 36, 614–623. <https://doi.org/10.1038/nbt.4151>
- 356

Supplementary Information

Sensor Circuit Design. The sensor system for each culture tube contains a snap-in mount 880-nm LED, a paired phototransistor and a four-section amplifier bias circuit (Fig. S1), which together are used to measure signal attenuation resulting from increasing culture turbidity. One section of this circuit uses a 2.501 VDC power supply to deliver a constant 833.67 μ A current to the LED. This constant current supply to the emitter provides constant light output throughout the experiment. The phototransistor produces a direct electrical current, proportional to the amount of light being transmitted through the culture tube by the LED. A clear culture tube would cause the phototransistor to produce a large electrical current, whereas a very turbid culture tube would result in only a small electrical current being generated. Each LED/phototransistor pair in the device is characterized and selected so that no more than 2.48 mA of current is generated for a tube of uninoculated medium and at least 0.54 mA is generated for a culture of 1.350 OD₆₀₀ (the highest density for the target experiments). A transresistance amplifier converts each 1 mA of current input to a 1 VDC + 2.501 VDC output. The added 2.501 VDC is then immediately removed in the next stage of the analog circuit to provide electrical isolation from power supply fluctuations. The signal at this stage of the analog circuit is 1 VDC for every 1 mA of current generated in the phototransistor. The signal is then amplified, and finally subtracted from the calibration voltage set for that sensor at the start of each experiment. This adjustment is done both to maximize sensor resolution by generating a signal that accommodates the range of voltages read by the 16-bit 5 VDC analog to digital converter (ADC), as well as to invert the signal so that small and large ADC values correspond to low and high turbidity, respectively. Between the sensing circuit and the ADC, an eight-channel multiplexer is used to sequentially measure each of the eight cultures in each tube

rack. Suppliers and components used are described in the supplemental file
 “Notes_on_Materials”.

Fig. S1. Circuit Diagram for a single sensing channel. The emitter is powered by a constant current to maintain consistent illumination of the tube sample. The resulting current produced by the sensor is converted to a voltage, isolated from any noise polluting the signal, and then inverted and adjusted by the calibration voltage before being fed into the ADC for measurement.

Sensor calibration. At initiation of each growth study, all 64 LED/phototransistor sensors are powered and allowed to reach a stable operating temperature by maintaining the system in the “Paused” state for a minimum of four hours (overnight is recommended). The inoculated culture tubes are then secured in tube racks and each sensor calibrated independently against its sample tube, with the objective of adjusting sensing circuit outputs to be near 0.3 VDC, corresponding to an ADC value of approximately 1600. By calibrating all sensors to begin each growth study at an output of 0.3 VDC, a small buffer is provided to compensate for possible small decreases in OD₆₀₀ at the beginning of operation. The calibration voltage for a sensing circuit was set using a digital to analog convertor (DAC) with a programmable range of 0 VDC to 5 VDC. Calibration begins with the DAC set to provide half of its maximum voltage (2.5 VDC) to the sensing circuit

398 being calibrated, and the ADC value of the sensor circuit output is recorded. If the ADC value is
399 lower than the target of 1600, then the final calibration voltage cannot be lower than the voltage
400 that was just tested, and the DAC is then reprogrammed to provide a voltage that is halfway
401 between the previous voltage tested and the maximum possible voltage for this calibration. A
402 similar adjustment is made if the ADC value is higher than the target of 1600. Once the
403 adjustment to the calibration voltage being tested is made, the ADC value is once again
404 measured and recorded. Calibration continues in this manner until the minimum and maximum
405 possible calibration voltages converge on a voltage that results in a sensor output that
406 corresponds closest to the target ADC value of 1600. Following this calibration, the value used
407 to program the DAC to the identified calibration voltage for the sensor circuit is stored in
408 program memory so that the calibration voltage for each sensing channel can be set during each
409 measurement taken during the growth study (Fig. S2). The final ADC values following
410 calibration are sent to the data collection server as the first data points. Individual sensor sets
411 commonly calibrate to consistent beginning ADC values over the course of many experiments,
412 therefore deviations from usual calibration values can be used to identify faults occurring on a
413 specific sensing channel.

Fig. S2. Program logic for calibrating each sensor channel. Each sensor is calibrated in series until the ADC reports the value closest to 1600, within the range of voltages available for the DAC to input into each sensor channel.

Data collection. The ODIn database archives the data generated by the ODIn system hardware and facilitates its analysis. The database, developed with FileMaker Advanced version 14 and hosted by FileMaker Server v14, operates independently from the ODIn hardware. FileMaker Advanced is a cross-platform, relational database development application published by FileMaker, Inc (Santa Clara, CA). The FileMaker Server hosts databases created with FileMaker Advanced making them available to multiple simultaneous users via FileMaker applications for Macintosh, Windows, iOS, or browser-based clients. At its top level, the database is organized by individual experiments. Each experiment corresponds to a single run of the ODIn hardware

with sensing channel output data collected for each of the 64 Balch tubes. Metadata for each experiment includes experiment name, description, start and stop timestamps, and the number of measurements collected. This information is displayed in the database's experiment view. Each experiment in the ODIn database is organized by rack, and each of the eight racks can be given additional descriptive metadata as inputted by the user. In the database structure, each of the eight racks is divided into records for each of its eight Balch tubes. More detailed metadata can be entered for each tube, including description, media composition, electron donor, electron acceptor, and organism(s). In addition, a notes field is available for each tube, allowing essentially unlimited text entry. Associated with the notes are three image fields, which can be used to store photographs or other images related to a tube. A variety of auto-fill tools have been created to facilitate the user entry of metadata for each of the 64 culture tubes. These tools include functions for replicating a tube's metadata entry, with automatic sequential numbering and a "fill-down" button, which will copy the current tube's metadata to the remaining tubes in the rack. A clone of FileMaker interface and database containing presented data is available at [10.5281/zenodo.4646431](https://doi.org/10.5281/zenodo.4646431).

Ease-of-use was a paramount design concern, with the intention of achieving greater accuracy and metadata entry completion. The ODIn database's rack view ([Fig. S3](#)) is designed to quickly show the metadata for each tube in a rack. Many experiments include biological replicates, in which two or more tubes may be inoculated from the same sample. While it can be valuable to view results for each replicate individually, it is preferable to average the results of replicates and display them as one. The ODIn database simplifies the grouping of replicates together and a color-coding system makes the replicate-grouping apparent to the viewer. Up to 32 replicate

450 groups can be created for each experiment. When working in the ODIn database experiment
451 view, clicking the right-arrow button associated with an experiment triggers two events. If the
452 experiment is still running, the ODIn database will query the data collection server to determine
453 if new data is available and if so, that data will be imported. Subsequently, or if the experiment
454 has already been completed, ODIn will switch to a summary view, showing data for all 64 tubes,
455 organized into eight graphs, one for each rack. The ODIn database can graphically display the
456 summary of results, organized by rack, a single rack's results, a single tube's results, or the
457 results of any combination of tubes and replicate sets from a single or from multiple experiments.
458 Additionally, a user can choose between graphs and specify filters that restrict the graphs to
459 specific time intervals or that specify minimum and maximum values for OD₆₀₀, mA, or growth
460 rate. Any combination of filters may be applied simultaneously, including removal of spurious
461 readings and linear vs log axes. The growth rate graphs display the change in growth rate over
462 time and the values are averaged using a user-selectable number of readings before and after
463 each time point. Finally, the ODIn database includes an export tool that allows one-click
464 exportation of raw data, OD values, or mA values to save on a local machine as a .CSV file for
465 manual data analysis.

Fig. S3. Tube rack sample metadata display. Selectable tabs detail information for all tube rack channels for a complete eight channel rack. Entry fields are provided for sample description, media composition, organism(s) grown, electron donor, electron acceptor, initial OD₆₀₀, Final OD₆₀₀. Row headings are selectable for grouping by assigning colors to those within a group.

The FileMaker algorithms and database constructed for use with the ODIn apparatus is available at 10.5281/zenodo.4646431. The data server software is available at 10.5281/zenodo.4619754.

Commercially assembled system components. Atmel ATmega2560V microcontroller in an Arduino Mega 2560 and Arduino Ethernet Shield (Adafruit Industries, LLC New York, NY) are used for sensor operation, sampling protocol, data acquisition, data reporting and mechanical

operations. One IEC 60320 C13 120 VAC female panel mount connector (Qualtek Electronics Corp., Mentor, OH) is provided for AC power supply and one NEMA 5-15P 120 VAC outlet (Molex LLC, Lisle, IL) connected to a G2R PCB mount relay (Omron Corporation, Kyoto, Japan) for shaker table activation/deactivation. One L298 H-Bridge dual bidirectional motor controller (Canakit Corporation North Vancouver, BC, Canada) is used for operation of an HDA4-30 4-inch stroke linear actuator (Servocity, Winfield, Kansas). A Delta Electronics PMT-12V100W1A 100 W AC/DC converter (Delta Electronics, Inc., Taoyuan, Taiwan) is used to provide 12 VDC power for the linear actuator power source, and to be down converted for use by the system. Two Texas Instruments PTN78060WAH DC-DC converters (Texas Instruments Incorporated, Dallas, Texas) take 12 VDC and supply 5.2 VDC power for the sensor circuit and PCB power source, and 2.501 VDC to drive the emitters and provide a reference voltage for power isolation. ADS1115 16-Bit 4-channel Analog to Digital converter and MCP4725 12-bit Digital-to-Analog converter (Adafruit Industries LLC, New York, NY) are used to take measurements and provide calibration voltages, respectively. A full list of all materials used is available in the bill of materials file titled “Notes_on_Materials.xlsx”. Circuit boards production files are available for both the control box board and tube box board as “Control Board mark 1.pcb” and “Tube_Rack_Boards.zip” respectively. Electrical connections between all control box system components were made using solid core 22AWG hookup wire and detailed in the available file “ODIn Control Box Pinout.xlsx”.

Tube rack custom printed circuit board. Tube rack PCBs were manufactured by ExpressPCB.com using the digital file “OD_mark4_revision_2.pcb” (Fig. S4). The following components were used to complete the assembly; 10K ohms resistors, 3K ohms resistors, 1K

ohms resistors (Susumu International USA, Palisades Park, NJ), opamps (Microchip Technology Inc., Chandler, AZ), female DE-9 connector (TE Connectivity Ltd., Schaffhausen, Switzerland), and 8-channel multiplexer (Texas Instruments Incorporated, Dallas, Texas). Resistor placement is described in Table S1 and Fig. S5. Each opamp is placed on the PCB to position the circle indicator (see Fig. S6) to the closest edge of the circuit board relative to its solder pads. All eight opamps are positioned in the same orientation. The multiplexer is placed in a position to locate the gold bar indicator on one short edge of the chip (see Fig. S7) closest to the silkscreen “Output Test” on the board. Terminal blocks (Phoenix Contact USA, Middletown, PA), as delivered, are block sets of two separate spring-loaded wire connectors; red and white. Sets of four blocks are connected by removal of the red baffle plate on the red terminal block, revealing molded attachment pins, and connected to another terminal block of two by the newly exposed pins. This action is repeated until a set of a total of eight blocks is fabricated. This set of eight blocks now has 16 separate wire connectors and can be attached to the circuit board by pacing the wire leads on the bottom of the terminal blocks through holes proximal to the opposite board side of the PCB. The terminal block assembly wire input holes face the resistor array on the PCB. The female DE-9 serial connector is placed in the middle of the PCB with the serial cable connector side facing out towards the opamps. All connections were made using 60/40 solder. Tube rack circuit boards were also printed and assembled at Technical & Assembly Services Corporation (TASC, Seattle, WA) using the files (“OD_mark4_revision2”) and bill of material (“BOM_of_OD_mark4_revision2”) provided.

Fig. S4. Tube rack PCB. Front side layout of tube rack PCB with silkscreen text displayed. PCB produced from “OD_mark4_revision_2.pcb” file and assembled from the components listed on the “Tube Rack PCB” sheet of the “ODIn_BOM.xlsx” file.

Fig. S5. Resistor placement for each sensing channel circuit on Tube Rack PCB. Resistor values and part numbers are provided on the “Tube Rack PCB” sheet of the “ODIn_BOM.xlsx” file and the assembly placement detailed in Table S1.

Fig. S6. Opamp solder Tube Rack PCB placement. Sixteen leads from the opamp are soldered on to each of the eight solder pads on the Tube Rack PCB. Circle indicator on top side of opamp is oriented toward the closest edge of the PCB.

Fig. S7. Multiplexer placement on Tube Rack PCB. Sixteen leads from multiplexer are soldered to the single centrally located solder pad on each Tube Rack PCB. Multiplexer is oriented with the gold bar indicator oriented opposite the “7” x 1”” silkscreen text.

Table S1. Susumu (Palisades Park, NJ) resistor placement on Tube Rack PCB

Resistor	Ω Value	Part Number	Manufacturer
R1	10K	RR05P10.0KDDKR-ND	Susumu
R2	10K	RR05P10.0KDDKR-ND	Susumu
R3	1K	RR05P1.0KDDKR-ND	Susumu
R4	1K	RR05P1.0KDDKR-ND	Susumu
R5	3K	RR0510P-302-D	Susumu
R6	10K	RR05P10.0KDDKR-ND	Susumu
R7	10K	RR05P10.0KDDKR-ND	Susumu
R8	3K	RR0510P-302-D	Susumu

LEDs and phototransistors are connected to the terminal blocks by securing the leads to the terminal blocks in the order specified by the silkscreen on the PCB. Phototransistor (OPB-100Z) leads are green and white and connected to “(W) PTC” and “(G) PTE” labeled terminal blocks.

546 LED (OPB-100) leads are red and black and connected to the “(R) LED+” and “(B) LED-”
547 labeled terminal blocks.

548

549 **Control box custom printed circuit board.** A single control box PCB was manufactured by
550 ExpressPCB.com using the digital file “Control board mark 2.pcb”. Assembly of the circuit
551 board uses the components listed in the “Control Box” tab of “Notes_on_Materials.xlsx”.
552 Components are placed according to the location specified by the silkscreen text on the PCB
553 face. Unlike the Tube Box PCBs, the Control Box PCB identifies the component to be soldered
554 to the board at each solder pad and through hole. Silkscreen text details the specific component
555 to attach by its type and electronic value all of which are detailed in the spreadsheet. D-sub
556 serial connectors were attached to the PCB by socket cap screws and nuts from the screw set on
557 the “Tube Box Board” sheet of “Notes_on_Materials.xlsx”.

558

559 **Construction of tube racks and control box housing with interface components.** All eight
560 tube racks were machined using a CNC mill (Bolton Tools Corp. Cerritos, CA). The tube rack
561 was designed to accommodate up to eight Balch tubes having dimensions of 18x150mm. All
562 tube racks and tube rack cabinet were made using the submitted CAD file schematics;
563 “tube_rack_base_opp.dxf”, “tube_rack_base_cover.dxf”, “tube_rack_base.dwg” and
564 “tube_holder_assembly.dwg”.

565

566 Placement of interface components (switches, knobs, and screen) was done without the use of a
567 guide for specific placement but rather spaced evenly around the centrally positioned system
568 status LCD screen located on the lifting lid of the control box. Component placement in this
569 manner allowed unencumbered access to the underside of the interface components for attaching
570 and soldering wire leads. Cutouts for attaching components to the acrylic were made using a
571 CNC mill and cut to the specifications listed on the component datasheet provided with their
572 delivery. The acrylic was cut with a rotary cutting tool to provide access to the Arduino stack
573 interfaces (Arduino Mega 2560, SD Data Logging Shield, and Arduino Ethernet Shield (Adafruit
574 Industries, LLC New York, NY)). The lid for the control box was attached to the box using the
575 continuous hinge and secured using countersunk 3/16" sheet metal screws. System status LCD
576 screen was attached to the lid using the brass screws provided with its delivery. Placement of the
577 components in the interior of the control box was done using metal standoffs secured to the base
578 of the box interior (Fig. S8). Components secured in this manner were the 12 VDC power
579 supply, 120 VAC relay, and the Arduino stack. The Control Box PCB was secured inside the
580 control box by attachment through the acrylic side wall of the right-angle D-sub 9 pin connectors
581 using their provided mounting screws.

Fig. S8. Control box with lid open showing placement of all control box components.

Control box wiring. Solid core 22 AWG hookup wire was used for control box wiring. All components (motor controller, control box PCB, Arduino Mega, and SD Data Logging Shield) were connected to the PCBs using the header connectors (Sullins Connector Solutions, San Marcos, CA). The Canakit motor controller was modified by removal of the right-angle male pin header set and replacement with a straight through female header connector. All header connector positions were wired to specified components according to the connection guide in the file “ODIn Control Box Pinout.xlsx”. Connections from the control box to the eight tube racks used 10’ DE-9 M/F serial cables.

Construction of raised platform. The raised platform used to secure the tube rack to the New Brunswick orbital shaker was constructed according to the follow specifications. The dimensions of the box are 18" l x 18" w x 5" h. Side walls of the raised platform were made from ½" thick aluminum stock cut to lengths of 18" x 4" and 17" x 4" and assembled using countersunk sheet metal screws to achieve the 18" x 18" square frame. The base of the platform was produced by attachment of an 18" l x 18" w x 1/2" d sheet of stainless steel to the aluminum walled frame using countersunk stainless-steel screws. The lifting lid of the platform was constructed from an 18" l x 18" w x 1/2" d sheet of transparent cast acrylic and attached to the base of platforms aluminum wall frame using a continuous hinge secured to the acrylic and aluminum sidewall by self-tapping sheet metal screws.

Linear actuator and scissor hinge lifting mechanism. The linear actuator is positioned inside the raised platform opposite to the lifting edge of the acrylic lid with its stationary end attached by a linear actuator pivoting mounting bracket (Fig. 4 of the main text). To raise the lid via the linear actuator a scissor-hinge mechanism was manufactured from stainless steel stock. 3-mm stainless steel sheets were cut into rounded-end rectangular strips measuring 160-mm in length. Hinge mechanism was generated by drilling 4-mm holes in each end of the rounded-end rectangular steel strips and attaching two strips to each side of the linear actuator piston rod through hole using a 40-mm x 4mm threaded-end pivot pin and securing with a lock nut with wave washers and bushings positioned between each metal-metal contact. Aluminum cubes (27-mm) were machined and attached to the interior front edge of the lifting platform and lifting edge underside of the acrylic lid and used to attach the opposite end of each rounded end rectangular strip using 40-mm threaded end pivot pins, wave washers, bushings, and lock nuts. Anchoring

the steel strips to the platform interior base, the piston rod of the linear actuator, and the underside of the lifting platform acrylic lid (using pivot pins, bushings, and wave washers) produces the lifting action of linear actuator extension. The linear actuator is connected to the 5-pin Conec threaded connector (American CONEC Corporation, Garner, NC) using the datasheet delivered with the device and connected to the control box for operation.

Arduino software, code, and libraries. The “arduino-1.0.6-windows.exe” development environment was used to program operation of the Arduino Mega microcontroller. ODIn software project-specific libraries were developed for operation of the ADC, DAC, and LCD in place of libraries provided by the stock Arduino development environment, which if used will be incompatible with the ODIn software. The following actions are necessary to allow the programs to compile, and load on the Arduino: Locate where the Arduino program is installed on the computer communicating with the Arduino Mega (default 64-bit Windows location is “C:\Program Files (x86)\Arduino” Installation to the standard Program Files folder might be required on a 32-bit machine). Locate in that "Arduino" folder containing the Arduino software, re-name the existing "libraries" folder to "stock libraries" and from the provided .zip file, "NECESSARY LIBRARIES", unzip and copy the "libraries" folder found there to the "Arduino" folder where the Arduino software was installed. Finally, restart the Arduino sketch program and the software should compile with the newly installed libraries. All libraries and software for operation of the ODIn apparatus is available at [10.5281/zenodo.4663185](https://zenodo.org/record/4663185).