

UCLA

UCLA Previously Published Works

Title

Analytical Evaluation of the Abbott RealTime CT/NG Assay for Detection of Chlamydia trachomatis and Neisseria gonorrhoeae in Rectal and Pharyngeal Swabs

Permalink

<https://escholarship.org/uc/item/2hs6m7k5>

Journal

Journal of Molecular Diagnostics, 22(6)

ISSN

1525-1578

Authors

Adamson, Paul C
Pandori, Mark W
Doernberg, Sarah B
et al.

Publication Date

2020-06-01

DOI

10.1016/j.jmoldx.2020.03.004

Peer reviewed



Analytical Evaluation of the Abbott RealTime CT/NG Assay for Detection of *Chlamydia trachomatis* and *Neisseria gonorrhoeae* in Rectal and Pharyngeal Swabs

Paul C. Adamson,^{*} Mark W. Pandori,[†] Sarah B. Doernberg,[‡] Lauren Komarow,[§] Zoe Sund,[¶] Thuy Tien T. Tran,[§] David Jensen,[¶] Ephraim L. Tsalik,^{||**} Carolyn D. Deal,^{††} Henry F. Chambers,[‡] Vance G. Fowler, Jr.,[¶] Scott R. Evans,[§] Robin Patel,^{‡‡} and Jeffrey D. Klausner^{**§§} for the Antibacterial Resistance Leadership Group

From the Division of Infectious Diseases,^{*} David Geffen School of Medicine, University of California, Los Angeles, Los Angeles, California; the Alameda County Department of Public Health,[†] Oakland, California; the Division of Infectious Diseases,[‡] School of Medicine, University of California, San Francisco, San Francisco, California; The Biostatistics Center,[§] George Washington University, Rockville Maryland; the Office of Regulatory Affairs and Quality,[¶] School of Medicine, and the Division of Infectious Diseases,^{||} Department of Medicine, Duke University, Durham, North Carolina; the Emergency Medicine Service,^{**} Durham Veterans Affairs Medical Center, Durham, North Carolina; the Enteric and Sexually Transmitted Infections Branch,^{††} National Institute of Allergy and Infectious Diseases, Bethesda, Maryland; the Division of Clinical Microbiology,^{‡‡} Department of Laboratory Medicine and Pathology, Mayo Clinic, Rochester, Minnesota; and the Department of Epidemiology,^{§§} Fielding School of Public Health, University of California, Los Angeles, Los Angeles, California

Accepted for publication
March 10, 2020.

Address reprint requests to Paul C. Adamson, M.D., M.P.H., or Jeffrey D. Klausner, M.D., M.P.H., Division of Infectious Diseases, David Geffen School of Medicine, University of California, Los Angeles, 10833 Le Conte Ave., Los Angeles, CA 90095. E-mail: padamson@mednet.ucla.edu or jdklausner@mednet.ucla.edu.

Chlamydia trachomatis and *Neisseria gonorrhoeae* infections in the rectum and pharynx are important extragenital reservoirs of infection. Few assays approved by the US Food and Drug Administration are commercially available to diagnose pharyngeal or rectal infections. The current study reports on the analytical performance of the Abbott RealTime CT/NG assay, including the limit of detection, inclusivity, and analytical specificity for *C. trachomatis* and *N. gonorrhoeae* in rectal and pharyngeal specimens. The limit of detection was performed using known concentrations of organisms, elementary bodies per milliliter (EB/mL) for *C. trachomatis* and colony-forming units per milliliter (CFU/mL) for *N. gonorrhoeae*, in clinical rectal and pharyngeal swab matrices. Inclusivity was performed against 12 serovars of *C. trachomatis* and seven strains of *N. gonorrhoeae*. The analytical specificity was performed using 28 different bacteria and viruses. The limit of detection for *C. trachomatis* was 2.56 EB/mL in pharyngeal specimens and 12.8 EB/mL in rectal specimens. The limit of detection for *N. gonorrhoeae* was 0.0256 CFU/mL for both pharyngeal and rectal specimens. The inclusivity and analytical specificity were 100% for both rectal and pharyngeal specimens. These analytical performance data demonstrate that the Abbott CT/NG RealTime assay is an accurate, sensitive, and specific assay in rectal and pharyngeal specimens, supporting the potential of the assay for detection of rectal and pharyngeal *C. trachomatis* and *N. gonorrhoeae* infections. (*J Mol Diagn* 2020, 22: 811–816; <https://doi.org/10.1016/j.jmoldx.2020.03.004>)

Supported in part by National Institute of Allergy and Infectious Diseases (NIAID) award UM1A1104681, NIMH training grant T32MH080634 (P.C.A.), and NIAID mid-career mentoring award K24-AI093969 (V.G.F.).

Disclosures: C.D.D. of the NIAID was involved in study conceptualization and design, and members of the NIAID team participated in team meetings and communication. The content is solely the responsibility of the

authors and does not necessarily represent the official views of the Antibacterial Resistance Leadership Group or the NIH.

Current address of M.W.P., Pathology and Laboratory Medicine, University of Nevada, Reno School of Medicine and Nevada State Public Health Laboratory, Reno, NV; of T.T.T.T., Merck & Co. Inc., Upper Gwynedd, PA.

Chlamydia trachomatis and *Neisseria gonorrhoeae* are the two most commonly reported sexually transmitted infections in the United States, and the number of cases is increasing.^{1,2} The rectum and pharynx are important extra-genital reservoirs of infection because rectal infections are associated with increased risk of HIV acquisition,³ and oropharyngeal infections are hypothesized to be a source of gonococcal antimicrobial resistance due to decreased tissue penetration of antimicrobial agents.⁴ Rectal and pharyngeal infections are typically asymptomatic; thus, detection and treatment of infections at these sites rely on effective screening tests. The US Centers for Disease Control and Prevention (CDC) recommends screening of all sexually active men who have sex with men (MSM) for rectal *C. trachomatis* and *N. gonorrhoeae* infections and pharyngeal *N. gonorrhoeae* at least annually.¹

Because of their high sensitivity and specificity, nucleic acid amplification tests have become the recommended method for detecting *N. gonorrhoeae* and *C. trachomatis* infections of the urogenital tract.⁵ Until recently, no commercial nucleic acid amplification test was approved by the US Food and Drug Administration (FDA) for use on pharyngeal or rectal specimens. However, a recent clinical trial performed by the authors evaluated the performance of three commercially available assays for detection of *C. trachomatis* and *N. gonorrhoeae* in pharyngeal and rectal specimens (ClinicalTrials.gov Identifier: NCT02870101). The results of that study were used to support FDA clearance for the detection of *C. trachomatis* and *N. gonorrhoeae* in two of the three assays for rectal and pharyngeal swabs.⁶

The Abbott RealTime CT/NG assay was included in the aforementioned study and used to create the anatomic site infected standard.⁷ The Abbott RealTime CT/NG assay is an *in vitro* PCR assay for direct, qualitative detection of plasmid DNA of *C. trachomatis* and genomic DNA of *N. gonorrhoeae*. The assay is currently FDA approved for testing of urogenital specimens (Abbott Molecular Inc. package insert: Abbott RealTime Chlamydia trachomatis/Neisseria gonorrhoeae Assay. Des Plaines, IL). The current study reports on the analytical performance of the Abbott RealTime CT/NG assay for detection of *C. trachomatis* and *N. gonorrhoeae* in rectal and pharyngeal specimens.

Materials and Methods

All tests were performed on the Abbott m2000 RealTime system (Abbott Laboratories, Abbott Park, IL). Testing was performed by the Alameda County Department of Public Health Laboratory (Oakland, CA), except those tests performed as part of the verification process, described below. Prior interference studies for FDA approval in urogenital specimens were performed previously and not repeated in the current study (Abbott Molecular Inc. package insert: Abbott RealTime Chlamydia trachomatis/Neisseria gonorrhoeae Assay. Des Plaines, IL).

All tests were performed using rectal and pharyngeal matrices purchased from a commercial laboratory (Bio-Collections, Miami, FL); they were obtained from asymptomatic individuals and had documentation of negative test results for *C. trachomatis* and *N. gonorrhoeae* by commercially available assays that the FDA approved for urogenital specimens. The matrices were provided in M2000 multi-collection vials that were frozen at -45°C . The specimens were subsequently thawed and pooled into 50-mL volumes for testing.

Limit of Detection

The limit of detection for *C. trachomatis* was determined by testing dilutions of *C. trachomatis* (Serovar LGV-II, 434 Strain, Advanced Biotechnologies Inc., Eldersburg, MD) in clinical rectal swab and pharyngeal swab matrices. The quantified stock [320 elementary bodies (EB)/mL] was obtained from a commercial laboratory (Advanced Biotechnologies Inc., Eldersburg, MD), and the elementary bodies and the DNA copy number of the stock were quantified by electron microscopy and quantitative PCR. This strain was chosen as the reference standard because it had been extensively characterized by a number of methods: PCR (genomically) quantified, the EBs were quantified by electron microscopy, and the strain is validated and known to be inactivated. The stock was subsequently diluted in a series of 10 fivefold dilutions, resulting in a range of 64 EB/mL to 3.3×10^{-5} EB/mL. To estimate the limit of detection, each dilution was tested in duplicate. Subsequently, six of the fivefold dilutions were identified across the apparent limit of detection (three above and three below) and assessed in replicates of 20 by the Abbott RealTime PCR for CT/NG on the M2000 system. The highest dilution of EB detectable in $\geq 95\%$ (19 of 20) of the replicates was defined as the limit of detection.

The limit of detection for *N. gonorrhoeae* was determined using plate counts for quantification. *N. gonorrhoeae* ATCC 49226 was first streaked for purity on chocolate agar. A single colony was isolated and diluted in isotonic nutrient broth in a series of 10-fold dilutions. Each dilution was split into two aliquots, with one aliquot being heat inactivated at 95°C for 15 minutes and 100 μL of the other aliquot simultaneously plated on chocolate agar plates and cultured overnight. Plate counts were then assessed in colony-forming units per milliliter (CFU/mL) to quantify the concentrations of organisms in the heat inactivated aliquot. The heat inactivated aliquot of *N. gonorrhoeae* corresponding to a plate count of 16 CFU/mL was used to make a series of fivefold dilutions of heat inactivated *N. gonorrhoeae*, ranging from 16 to 0.000205 CFU/mL, in rectal or pharyngeal matrices. Those heat inactivated aliquots of *N. gonorrhoeae* ATCC 49226 were evaluated on the Abbott M2000 system in replicates of at least 20. The highest dilution detectable at $\geq 95\%$ frequency was defined as the limit of detection and reported in CFU/mL.

Inclusivity

Inclusivity was performed to determine whether the Abbott RealTime CT/NG Assay detected different bacterial strains. Inclusivity for *C. trachomatis* was performed by diluting 12 *C. trachomatis* serovars (A, B, C, D, E, F, G, H, I, J, K, and L2) in clinical rectal and pharyngeal swab matrices to a concentration of four times the limit of detection. Each evaluated serovar dilution was tested in 20 replicates. For *N. gonorrhoeae*, inclusivity was performed using seven reference isolates, four obtained from the CDC based on antimicrobial susceptibility⁸ (F-28, CDC 10328, CDC 10329, and SPL-4) and three selected from a commercially available collection (FA1090, F-18, and MHD 340; ATCC, Manassas, VA). Each strain was tested at five times the limit of detection and tested in replicates of 20.

Analytical Specificity

The analytical specificity (cross-reactivity) of the Abbott RealTime CT/NG assay was previously tested against an extensive list of other organisms, including *Neisseria flava*, *Neisseria perflava*, and *Neisseria meningitidis* types A through D, which were all found to be negative on the assay (Abbott Molecular Inc. package insert: Abbott RealTime Chlamydia trachomatis/Neisseria gonorrhoeae Assay, Des Plaines, IL). The current study evaluated an additional 28 microorganisms commonly found in the pharynx and rectum and not already evaluated in the prior cross-reactivity studies. Bacteria were tested at 2×10^6 CFU/mL. The tested concentrations for the viruses were 2.5×10^6 genomic copies per assay for Epstein-Barr virus, 2.5×10^6 DNA copies for human coxsackie virus A2 strain FI, 2.5×10^6 DNA copies for adenovirus (strain 41), and 3.0×10^6 RNA copies for norovirus (genogroup G2). All assays were performed in pharyngeal and rectal matrices known to be negative for *C. trachomatis* and *N. gonorrhoeae*. Each organism was tested once.

Assay Verification

To assess accuracy, four sets of 20 specimens were made by spiking *C. trachomatis* or *N. gonorrhoeae* to achieve concentrations of analyte 10 to 10,000 times the limit of detection in rectal and pharyngeal matrices negative for *C. trachomatis* and *N. gonorrhoeae*. The *C. trachomatis* specimens served as a negative control for *N. gonorrhoeae* and vice versa. To assess precision, large-volume specimens were created to contain an analyte concentration 10 times higher than the limit of detection and were run in quintuplicate. Replicate panels and large-volume samples were tested on the Abbott M2000 platform at BioCollections Worldwide Inc. (Miami, FL).

Results

Limit of Detection

The limit of detection of the Abbott RealTime CT/NG assay for *C. trachomatis* was 2.56 EB/mL in pharyngeal specimens, with all 20 specimens testing positive at this level (100%; 95% CI, 83.2%–100%), and 12.8 EB/mL in rectal specimens, with all 20 specimens testing positive at this level (100%; 95% CI, 83.2%–100%). The limit of detection for *N. gonorrhoeae* was 0.0256 CFU/mL for both pharyngeal and rectal specimens, with 19 pharyngeal specimens (95%; 95% CI, 75.1%–99.9%) and 20 rectal specimens (100%; 95% CI, 83.2%–100%) testing positive at this level (Table 1).

Inclusivity

The inclusivity for *C. trachomatis* was 100% because all 12 *C. trachomatis* serovars tested at four times the limit of detection were detected in 100% (20 of 20) rectal and pharyngeal matrices. The inclusivity for *N. gonorrhoeae* also achieved 100% (20 of 20) inclusivity for detection of each of the seven strains tested.

Analytical Specificity

In total, there were 28 microorganisms spiked into both pharyngeal and rectal specimens and tested by the Abbott RealTime CT/NG assay. There was no cross-reactivity observed, corresponding to an analytical specificity of 100% (Table 2).

Assay Verification

In total, there were 80 specimens tested by the second laboratory. All 40 rectal and pharyngeal specimens spiked with *C. trachomatis* and all 40 rectal and pharyngeal specimens spiked with *N. gonorrhoeae* tested positive. *C. trachomatis* was detected in 0 of the 20 *N. gonorrhoeae*–spiked samples from each anatomic site and *N. gonorrhoeae* was detected in 0 of the 20 *C. trachomatis*–spiked samples from each anatomic site. For the precision testing, all 10 specimens spiked with *C. trachomatis* (100%) tested positive; all 10 specimens spiked with *N. gonorrhoeae* also tested positive by the external laboratory.

Discussion

This study reports the findings of an analytical evaluation of the Abbott RealTime CT/NG assay for detection of *C. trachomatis* and *N. gonorrhoeae* in rectal and pharyngeal specimens. The findings indicate that the assay had 100% inclusivity and specificity among the serotypes and microorganisms tested and that it performed well in the verification study, with 100% accuracy and precision among

Table 1 Serial Dilutions of *Chlamydia trachomatis* and *Neisseria gonorrhoeae* in Rectal and Pharyngeal Specimens Used to Determine the Limit of Detection of the Abbott RealTime CT/NG Assay

Assay type	Source	Concentration*	No. of positive samples	No. of negative samples	Percentage of positive samples (95% CI)
<i>Chlamydia trachomatis</i> (EB/mL)	Pharyngeal	320	20	0	100.0 (83.2–100.0)
		64	20	0	100.0 (83.2–100.0)
		12.8	20	0	100.0 (83.2–100.0)
		2.56	20	0	100.0 (83.2–100.0)
		0.512	17	3	85.0 (62.1–96.8)
		0.1024	0	20	0.0 (0.0–16.8)
		0.02048	0	22	0.0 (0.0–15.4)
		0.004096	0	19	0.0 (0.0–17.7)
	Rectal	320	20	0	100.0 (83.2–100.0)
		64	21	0	100.0 (83.9–100.0)
		12.8	20	0	100.0 (83.2–100.0)
		2.56	18	2	90.0 (68.3–98.8)
		0.512	16	4	80.0 (56.3–94.3)
		0.1024	5	15	25.0 (8.7–49.1)
		0.02048	3	17	15.0 (3.2–37.9)
		0.004096	0	20	0.0 (0.0–16.8)
<i>Neisseria gonorrhoeae</i> [CFU/mL]	Pharyngeal	16	6	0	100.0 (54.1–100.0)
		3.2	22	0	100.0 (84.6–100.0)
		0.64	21	0	100.0 (83.9–100.0)
		0.128	20	0	100.0 (83.2–100.0)
		0.0256	19	1	95.0 (75.1–99.9)
		0.00512	4	16	20.0 (5.7–43.7)
		0.001024	0	20	0.0 (0.0–16.8)
		0.000205	0	20	0.0 (0.0–16.8)
	Rectal	16	6	0	100.0 (54.1–100.0)
		3.2	21	0	100.0 (83.9–100.0)
		0.64	21	0	100.0 (83.9–100.0)
		0.128	20	0	100.0 (83.2–100.0)
		0.0256	20	0	100.0 (83.2–100.0)
		0.00512	2	18	10.0 (1.2–31.7)
		0.001024	1	19	5.0 (0.1–24.9)
		0.000205	0	20	0.0 (0.0–16.8)

Bold values indicate the limit of detection, where $\geq 95\%$ of the replicates were detected.

*Concentrations are given in elementary bodies per milliliter (EB/mL) for *C. trachomatis* and colony-forming units per milliliter (CFU/mL) for *N. gonorrhoeae*.

specimens tested. In addition, the assay was able to detect low bacterial loads.

The limits of detection claim for the Abbott RealTime CT/NG assay in urogenital specimens is 320 copies of *C. trachomatis* target DNA and 320 copies of *N. gonorrhoeae* target DNA per assay, corresponding to approximately 30 to 40 organisms per assay (Abbott Molecular Inc. package insert: Abbott RealTime Chlamydia trachomatis/Neisseria gonorrhoeae Assay. Des Plaines, IL). Results of the current study showed the limit of detection for *C. trachomatis* to be 12.8 EB/mL in rectal specimens, 2.56 EB/mL in pharyngeal specimens, and 0.0256 CFU/mL in both rectal and pharyngeal specimens for *N. gonorrhoeae*. Although it is difficult to compare precisely copies of DNA with EB or CFU, these findings suggest that the limits of detection for *C. trachomatis* and *N. gonorrhoeae* in rectal and pharyngeal specimens are similar or slightly lower than the reported

claims for the Abbott RealTime CT/NG assay in urogenital specimens.

The limits of detection of the Abbott RealTime CT/NG assay for *C. trachomatis* in both rectal and pharyngeal specimens were somewhat lower than those reported by another PCR-based assay, the Xpert CT/NG assay (Cepheid, Sunnyvale, CA), an FDA-approved assay for rectal and pharyngeal specimens. The Xpert CT/NG assay limit of detection was 88 to 161 EB/mL in rectal and 161 to 225 EB/mL in pharyngeal specimens by probit analysis (Cepheid package insert: Xpert CT/NG. Sunnyvale, CA). For *N. gonorrhoeae*, the limit of detection on the Abbott RealTime CT/NG assay was estimated to be 0.0256 CFU/mL in both rectal and pharyngeal specimens, which also compared favorably to the Xpert CT/NG assay, with limits of detection reported to be 4.9 to 5.3 CFU/mL in rectal specimens and 6.4 to 7.1 CFU/mL in pharyngeal

Table 2 Analytical Specificity of Abbott RealTime CT/NG Assay against 28 Bacteria and Viruses in Rectal and Pharyngeal Matrices

Microorganism	Concentration	Pharyngeal specimens		Rectal specimens	
		<i>Chlamydia trachomatis</i>	<i>Neisseria gonorrhoeae</i>	<i>Chlamydia trachomatis</i>	<i>Neisseria gonorrhoeae</i>
<i>Acinetobacter baumannii</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Arcanobacterium hemolyticum</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
Adenovirus	2.5 × 10 ⁶ DNA copies/assay	Negative	Negative	Negative	Negative
<i>Bacteroides oralis</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Bordetella pertussis</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Burkholderia cepacia</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Campylobacter rectus</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Citrobacter koseri</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Clostridium difficile</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Corynebacterium diphtheriae</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Corynebacterium pseudodiphtheriticum</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
Enterovirus	2.5 × 10 ⁶ DNA copies/assay	Negative	Negative	Negative	Negative
Epstein-Barr virus	2.5 × 10 ⁶ genomic copies/assay	Negative	Negative	Negative	Negative
<i>Fusobacterium necrophorum</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Hemophilus parahaemolyticus</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Hemophilus parainfluenzae</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Helicobacter pylori</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Legionella micdadei</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Neisseria meningitidis</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
Norovirus	3.0 × 10 ⁶ RNA copies/assay	Negative	Negative	Negative	Negative
<i>Peptostreptococcus micros</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Providencia alcalifaciens</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Staphylococcus hemolyticus</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Stenotrophomonas maltophilia</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Streptococcus anginosus</i> group	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Streptococcus bovis</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Streptococcus mitis</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative
<i>Veillonella parvula</i>	1.0 × 10 ⁶ cells/assay	Negative	Negative	Negative	Negative

specimens by probit analysis (Cepheid package insert: Xpert CT/NG, Sunnyvale, CA). The limits of detection for both assays were performed on heat inactivated specimens corresponding to a CFU/mL broth dilution. The genetic targets of the assay exist in numerous copies within each cell and are released at heat inactivation, thus making it possible to detect >1 CFU/mL, which is a method for quantification of viability. In addition, limit of detection data from the Abbott RealTime CT/NG assay also compare favorably with other PCR platforms for the detection of *C. trachomatis* and *N. gonorrhoeae* in vaginal and urine specimens. For example, the FDA-approved BD Max CT/GC/TV assay has a limit of detection as low as 5 to 100 EB/mL in urine specimens for *C. trachomatis* and as low as 60 to 117 cells/mL in vaginal specimens for *N. gonorrhoeae* (Becton Dickinson and Company Life Sciences package insert: BD MAX CT/GC/TV Assay, Sparks, MD).

Previous studies have evaluated the bacterial loads associated with *N. gonorrhoeae* and *C. trachomatis* infections at rectal and pharyngeal sites. In 2013, Bissessor et al⁹ found that bacterial loads of *N. gonorrhoeae* obtained from MSM were higher in the rectum (median, 18,960 copies per swab) compared with the pharynx (median, 2100 copies per swab)

and highest among men with symptoms of proctitis (median, 278,000 copies per swab), whereas Chow et al¹⁰ reported a median 170,493 copies per swab among men testing positive by nucleic acid amplification tests. In a systematic review that assessed bacterial loads for *C. trachomatis*, Vodstrcil et al¹¹ found considerable variation of bacterial loads by anatomic site, sex, and quantification technique, which limits comparisons. However, more recently, Phillips et al¹² evaluated bacterial loads of *C. trachomatis* in MSM and found a median of 1204 copies per swab in the posterior oropharynx. Two studies that evaluated *C. trachomatis* bacterial loads in rectal infections found similar results, with a median log of 3.4 copies/mL (2512) *C. trachomatis* among women¹³ and mean log of 3.5 copies/mL (3162) *C. trachomatis* among MSM and women who reported anal sex.¹⁴ One recent large study found higher bacterial loads in rectal infections compared with pharyngeal infections and no differences between men and women tested but reported only cycle threshold levels.¹⁵ Although it is difficult to define the precise number of bacteria needed to cause infections at these sites, the limit of detection for the Abbott RealTime CT/NG assay reported here would be expected to detect infections for the bacterial loads in those reports.

The clinical performance of the Abbott RealTime CT/NG in rectal and pharyngeal specimens was recently evaluated in a multicenter study of 2598 symptomatic and asymptomatic patients, alongside two other assays that have now been FDA approved for detection of *C. trachomatis* and *N. gonorrhoeae* at rectal and pharyngeal sites.⁷ The negative predictive value of the Abbott RealTime CT/NG assay was high: 98.7% and 99.0% for *N. gonorrhoeae* at pharyngeal and rectal sites, respectively, and 99.7% and 98.3% for *C. trachomatis* at pharyngeal and rectal sites, respectively. Those estimates were similar to the other two assays tested and correspond to high sensitivity of the assays. The limit of detection findings also support the high sensitivity of the test.

The data presented herein demonstrate that the Abbott CT/NG RealTime assay is an accurate, sensitive, specific, and reliable assay in laboratory-derived rectal and pharyngeal specimens. These findings, alongside clinical performance data, support potential claims of use of the assay for detection of *C. trachomatis* and *N. gonorrhoeae* infections in rectal and pharyngeal specimens.

Acknowledgments

We thank the members of the Duke Clinical Research Institute and the Georgetown Data Coordinating Center for their assistance with data collection, cleaning, management, and analysis.

References

1. Workowski KA, Bolan GA: Sexually transmitted diseases treatment guidelines, 2015. *MMWR Recomm Rep* 2015, 64:1–137
2. Centers for Disease Control and Prevention: Sexually Transmitted Disease Surveillance 2017. Atlanta, US Department of Health and Human Services, 2018
3. Bernstein KT, Marcus JL, Nieri G, Philip SS, Klausner JD: Rectal gonorrhea and chlamydia reinfection is associated with increased risk of HIV seroconversion. *J Acquir Immune Defic Syndr* 2010, 53: 537–543
4. Unemo M, Shafer WM: Antimicrobial resistance in *Neisseria gonorrhoeae* in the 21st century: past, evolution, and future. *Clin Microbiol Rev* 2014, 27:587–613
5. Centers for Disease Control and Prevention: Recommendations for the laboratory-based detection of *Chlamydia trachomatis* and *Neisseria gonorrhoeae*—2014. *MMWR Recomm Rep* 2014, 63:1–19
6. Food and Drug Administration: FDA Clears First Diagnostic Tests for Extragenital Testing for Chlamydia and Gonorrhea. White Oak, MD, Food and Drug Administration, March 23, 2019
7. Doernberg SB, Komarow L, Tran TTT, Sund Z, Pandori MW, Jensen D, Tsalik EL, Deal CD, Chambers HF, Fowler VG Jr, Evans SR, Patel R, Klausner JD: Antibacterial resistance leadership group: simultaneous evaluation of diagnostic assays for pharyngeal and rectal *Neisseria gonorrhoeae* and *Chlamydia trachomatis* using a master protocol. *Clin Infect Dis* 2019, [Epub ahead of print] doi: 10.1093/cid/ciz1105
8. Centers for Disease Control and Prevention: *Neisseria gonorrhoeae* Reference Strains for Antimicrobial Susceptibility Testing. Atlanta, Centers for Disease Control and Prevention, 2005
9. Bissessor M, Tabrizi SN, Fairley CK, Danielewski J, Whitton B, Bird S, Garland S, Chen MY: Differing *Neisseria gonorrhoeae* bacterial loads in the pharynx and rectum in men who have sex with men: implications for gonococcal detection, transmission, and control. *J Clin Microbiol* 2011, 49:4304–4306
10. Chow EPF, Tabrizi SN, Phillips S, Lee D, Bradshaw CS, Chen MY, Fairley CK: *Neisseria gonorrhoeae* bacterial DNA load in the pharynx and saliva of men who have sex with men. *J Clin Microbiol* 2016, 54:2485–2490
11. Vodstrcil LA, McIver R, Huston WM, Tabrizi SN, Timms P, Hocking JS: The epidemiology of *Chlamydia trachomatis* organism load during genital infection: a systematic review. *J Infect Dis* 2015, 211:1628–1645
12. Phillips TR, Fairley CK, Maddaford K, Danielewski J, Hocking JS, Lee D, Williamson DA, Murray G, Kong F, De Petra V, Bradshaw CS, Chen MY, Wigan R, Snow A, Howden BP, Garland SM, Chow EPF: Bacterial load of *Chlamydia trachomatis* in the posterior oropharynx, tonsillar fossae, and saliva among men who have sex with men with untreated oropharyngeal *Chlamydia*. *J Clin Microbiol* 2019, [Epub ahead of print] doi:10.1128/JCM.01375-19
13. Dirks JAMC, Van Liere GAFS, Hoebe CJPA, Wolffs P, Dukers-Muijers NHTM: Genital and anal *Chlamydia trachomatis* bacterial load in concurrently infected women: a cross-sectional study. *Sex Transm Infect* 2019, 95:317–321
14. Van Liere GAFS, Dirks JAMC, Hoebe CJPA, Wolffs PF, Dukers-Muijers NHTM: Anorectal *Chlamydia trachomatis* load is similar in men who have sex with men and women reporting anal sex. *PLoS One* 2015, 10:e0134991
15. Wijers JNAP, Dukers-Muijers NHTM, Van Liere GAFS, Dirks JAMC, Wolffs PFG, Hoebe CJPA: Men and women have an equal oropharyngeal and anorectal *Chlamydia trachomatis* bacterial load: a comparison of 3 anatomic sites. *J Infect Dis* 2019, [Epub ahead of print] doi:10.1093/infdis/jiz668