UC Irvine

UC Irvine Previously Published Works

Title

Single-fiber optical tweezers for cellular micro-manipulation

Permalink

<https://escholarship.org/uc/item/3tv744rr>

Journal

Optics and Photonics News, 19(12)

ISSN

1047-6938

Authors

Mohanty, SK
Mohanty, KS
Berns, MW

Publication Date

2008

DOI

10.1364/OPN.19.12.000042

Copyright Information

This work is made available under the terms of a Creative Commons Attribution License, available at <https://creativecommons.org/licenses/by/4.0/>

Peer reviewed

Single-Fiber Optical Tweezers for Cellular Micro-Manipulation

Samarendra K. Mohanty, Khyati S. Mohanty
and Michael W. Berns

The short working distance of microscope objectives has severely restricted the application of optical tweezers and scissors at large depths. Therefore, researchers are paying more and more attention to the use of optical fiber for this purpose. Recently, in-depth single fiber optic trapping of low- and high-index particles has been demonstrated using micro-axicon-tip fibers.^{1,2}

The shape of the cone angle at the axicon's tip enabled fiber-optic trapping in the near-field.² Further, we have demonstrated controlled guidance of neuronal growth cones as well as the trapping and stretching of neurons using fiber-optic tweezers.³ The cells could be stretched³ by the combined action of two forces—an attractive gradient force due to fiber-optic tweezers at high beam powers pulling the membrane and a scattering force on the membrane as reported in dual-fiber trapping.

We also observed alignment of intracellular dark (high refractive index) material along the direction of laser beam propagation.³ By mode-locking, the beam of the fiber-optic tweezers was converted to fiber-optic scissors, enabling the dissection of neuronal processes.³

This microscopic-controlled nano-dissection of neurons followed by a process of resealing and repair could serve as a useful tool for basic and applied studies on neuronal damage, repair and regeneration. When the femtosecond fiber-optic microbeam was at reduced average power, we could microinject impermeable exogenous materials into the trapped cells. At high average powers, we accomplished lysis of a three dimensionally trapped cell.³

In the figure, we show optical trapping as well as lysis of biological cells using a single axicon tip fiber. The cell, distant from the fiber tip (marked by arrow in a) is attracted toward the fiber tip

Trapping, transport and lysis of a biological cell using axicon-tip single-fiber tweezers and scissors. All images are in the same magnification. Scale bar: 10 μm .

(b) at a power of 95 mW and was stably trapped very close to the axicon tip (c). The trapped cell could be transported to a new location (d, e) by maneuvering the fiber tip. Switching the laser beam on and off alternatively allowed the cell to move close (g) or away (f) from the fiber tip, ruling out the possibility of non-optical attraction between the cell and the fiber.

By mode-locking the near infrared laser beam, we could deliver femtosecond pulses (about 200 fs, 76 MHz), and the same fiber probe could be used for lysis of the trapped cells (h, i) in a timescale of 600 ± 200 ms. This feature is required in many assays to terminate biochemical reactions immediately, thus preventing measurement artifacts.

The non-invasive micro-axicon-tipped optical fiber can also be used in multi-functional mode for in-depth trapping, stretching, rotation, sorting, microinjec-

tion and ablation as well as for exciting fluorophores. The depth attainable by optical micromanipulation is enhanced by a single microfabricated fiber device. Moreover, this technology could lead to sophisticated sensing and imaging capabilities that can be applied to live cells.⁴ ▲

Samarendra K. Mohanty (smohanty@uci.edu), Khyati S. Mohanty and Michael W. Berns are with the Beckman Laser Institute, University of California-Irvine, Irvine, Calif., U.S.A. K.S. Mohanty and M.W. Berns are also affiliated with the department of biomedical engineering at the University of California-Irvine in Irvine, Calif., U.S.A.

References

1. K.S. Mohanty et al. "In depth fiber optic trapping of low-index microscopic objects," *Appl. Phys. Lett.* **92**, 151113 (2008).
2. S.K. Mohanty et al. "Organization of microscale objects using a microfabricated optical fiber tip," *Opt. Lett.* **33**, 2155-7 (2008).
3. S.K. Mohanty et al. *J. Biomed. Opt.* **13**, AIP ID code: 046805JBO (2008).
4. Y. Verma et al. "Tapered single mode fiber tip high lateral resolution optical coherence tomography," *Las. Phys. Lett.* **4**, 686-9 (2007).