

Lawrence Berkeley National Laboratory

Recent Work

Title

RELATIVE INTENSITY CALIBRATION OF A Ge(Li) GAMMA-RAY SPECTROMETER

Permalink

<https://escholarship.org/uc/item/5bb6c0qj>

Author

Jardine, Leslie J.

Publication Date

1971-05-01

Submitted to Nuclear Instruments
and Methods

RECEIVED
LAWRENCE
RADIATION LABORATORY

UCRL-20435
Preprint

c.2

DOCUMENTS SECTION

RELATIVE INTENSITY CALIBRATION OF A
Ge(Li) GAMMA-RAY SPECTROMETER

Leslie J. Jardine

May 1971

AEC Contract No. W-7405-eng-48

TWO-WEEK LOAN COPY

*This is a Library Circulating Copy
which may be borrowed for two weeks.
For a personal retention copy, call
Tech. Info. Division, Ext. 5545*

LAWRENCE RADIATION LABORATORY
UNIVERSITY of CALIFORNIA BERKELEY

UCRL-20435

c.2

DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

RELATIVE INTENSITY CALIBRATION OF A Ge(Li) GAMMA-RAY SPECTROMETER*

Leslie J. Jardine

Lawrence Radiation Laboratory
University of California
Berkeley, California 94720

May 1971

Abstract

Portions of the gamma-ray spectra of ^{182}Ta and $^{180\text{m}}\text{Hf}$ have been remeasured with a Ge(Li) spectrometer system to determine accurate relative intensities. For ^{182}Ta the eleven strongest transitions in the energy range of 100-1300 keV were measured. The $I_{\gamma}(443)/I_{\gamma}(501)$ relative gamma-ray branch in $^{180\text{m}}\text{Hf}$ was remeasured to be 5.70 ± 0.15 . The intensity values derived from these measurements are recommended for use with a IAEA standard intensity set in the calibration of relative photopeak efficiencies for Ge(Li) detectors. An overall accuracy of $\pm 4\%$ for the relative intensity calibration over the energy range of 100-1300 keV can be expected and $\pm 3\%$ for the 500-2800 keV energy range.

*Work performed under the auspices of the U. S. Atomic Energy Commission.

1. Introduction

The development of energy calibration standards for Ge(Li) detectors has advanced to a state where many standards in the energy range of 60-2800 keV are determined with errors of less than 0.1 keV. However, the absolute or relative gamma-ray intensities of these standards are often uncertain by as much as 5 %. In a few exceptional cases relative intensity determinations have been performed with errors of less than 2 % by careful consideration of gamma-ray cascades corrected for internal conversion¹⁰). The criteria of either simple cascades with no cross-over transitions or highly accurate decay schemes place a severe restriction on both the availability and on the use of absolute transition intensities for intensity calibrations. However, relative transition intensities for complex decay schemes can be determined accurately, and these data are generally more easily applied.

It is the purpose of this paper to suggest ^{182}Ta and $^{180\text{m}}\text{Hf}$ as isotopes spanning the energy range of 100-1300 keV to supplement the IAEA¹) or similar standard intensity source sets for relative photopeak efficiency determinations of Ge(Li) detectors. Relative intensity measurements have been performed on ^{182}Ta in the energy range of 100-1300 keV and are compared with others^{3,4,23}). The absolute gamma-ray intensities for $^{180\text{m}}\text{Hf}$ are presented and compared with values reported in the literature^{10,30,33}) based upon the remeasured $I_{\gamma}(443)/I_{\gamma}(501)$ gamma-ray branching ratio. Relative intensity values for isotopes are included in the tables in order to collect the best sets of data necessary to calibrate Ge(Li) detectors over the energy range of 100-2800 keV to ± 4 %.

2. Experimental

The 115.1d ^{182}Ta was produced in the Berkeley Research Reactor by irradiation of 1.0 and 0.5 mil foils of natural Ta ($99.9877\% \text{ }^{181}\text{Ta}$)²⁾ which has a large thermal neutron cross-section of 21 barns. The small 0.07 barn cross-section to produce the 16.5 m ^{182m}Ta did not interfere after a short decay period following irradiation. The intensities of the eleven strongest gamma-rays were measured on a calibrated 10-cm³ Ge(Li) detector. This detector had been calibrated with two sets of absolute intensity sources obtained from the IAEA. ^{180m}Hf was also used to define the efficiency curve in the critical 100-335 keV region. The accuracy of the photopeak efficiency determination was checked by measuring several radioisotopes whose relative intensities are known^{10,11)} to 2%. These measured intensities agreed to 3% with those in Kane and Mariscotti¹⁰⁾ and Donnelly et al.¹¹⁾. The efficiency curve for the 10-cm³ detector obtained in this way is shown in fig. 1. The detector resolution ranged from 1.4 keV at 122 keV to 2.3 keV at 1332.5 keV. The areas of photopeaks were determined using two separate computer codes, SAMPO⁶⁾ and SPECT⁷⁾, which use gaussian functions with exponential tails to approximate the experimental photopeak shapes. Photopeak areas from the codes agreed to within 1% in the selected peaks. Measurements of photopeak intensities were made for each of the two ^{182}Ta foils which were corrected for attenuation⁴⁵⁾ due to source thickness, and these were verified with an isotopically separated ("mass free") source. The results are given in Table 1 and are compared to other results^{4,5,24)}.

The measured ^{182}Ta relative intensities should be accurate to 3% over the energy range of 100-1300 keV. Previously, Edwards et al.⁴⁾ quoted errors

of 4-5% over the energy range of 100-264 keV. White et al.⁵⁾ measured the relative intensities of ^{182}Ta gamma-rays in the energy range of 152-1300 keV with errors of 4-5%. Over the smaller energy range of 1000-1300 keV, errors quoted were in the range, 1.5-2 %.

$^{180\text{m}}\text{Hf}$ was produced by reactor irradiation of the separated isotope ^{179}Hf . $^{180\text{m}}\text{Hf}$ is very convenient^{10,11)} in the energy region between 93-501 keV where calibration points for efficiency curves are sparse. The decay scheme^{31,33)} of $^{180\text{m}}\text{Hf}$ (fig. 2) allows absolute gamma-ray intensities to be derived by correcting for internal conversion if the $I_{\gamma}(443)/I_{\gamma}(501)$ gamma-ray intensity branching is accurately known. With reference to fig. 2 and the $^{180\text{m}}\text{Hf}$ decay scheme the following equations apply.

$$I_t(93) = I_t(215) = I_t(332) = I_t(443) + I_t(501) \quad (1)$$

$$I_t(443) = I_t(501) = (1 + \alpha(443))I_{\gamma}(443) + (1 + \alpha(501))I_{\gamma}(501) \quad (2)$$

I_t and I_{γ} are the transition and gamma intensities, respectively; α is the total internal conversion coefficient. The $I_{\gamma}(443)/I_{\gamma}(501)$ relative gamma-ray branch was measured to be 5.70 ± 0.15 on the above detector in agreement with ref. 31. The mixing ratio of 3.5% M2 and 96.5% E3 reported by Bodenstein et al.³⁴⁾ for the 500.7 keV transition was used in calculating absolute gamma intensities from the transition intensities of eq. (1). The other transitions are of E2 multipolarity. Theoretical conversion coefficients used were those of Hager and Seltzer³⁵⁾. The calculated absolute gamma-intensity results obtained from eq. (1) are shown in Table 2 with other results^{10,30,33)}. Included are the 93.3 and 500.7 keV transitions not given by ref. 10. The calculated $14.5 \pm 0.4\%$

for the 500.7 keV transition is in agreement with the values $\sim 15\%$ given by Gvozdev et al.³⁷⁾ and $14.8 \pm 0.8\%$ of Goldhaber and McKeown³¹⁾ and Paul et al.³²⁾.

Table 3 shows the intensity values used for the IAEA calibrated set and ²⁴Na. The intensity values are those recommended by the IAEA. They are given to collect the necessary intensity values, along with the ¹⁸²Ta and ^{180m}Hf, into one paper to be used for the efficiency calibration of Ge(Li) spectrometer systems. ²⁴Na is also listed since it is useful for extending the curve to higher energies.

3. Conclusions

The ^{182}Ta intensities determined in this work were used to derive the relative photopeak efficiency curve of a 40-cm^3 coaxial detector together with $^{180\text{m}}\text{Hf}$, ^{24}Na , and the IAEA calibrated set for which the results are shown in fig. 3. A third detector has been recently calibrated⁹⁾ at this laboratory using this method and showed similar results. These isotopes and intensity values, when combined with the IAEA intensity sources and ^{24}Na , should determine the relative photopeak efficiency of Ge(Li) spectrometers to $\pm 4\%$ over the energy range of 100-500 keV and $\pm 3\%$ over the energy range of 500-2800 keV.

Acknowledgements

The author owes thanks to Drs. S. G. Prussin and C. M. Lederer for helpful discussions on the problems associated with this study. He also owes thanks to Dr. J. M. Hollander for support. He is indebted to Jeff Gallup for loan of his isotopically separated ^{182}Ta source.

References

- 1) International Atomic Energy Agency, Vienna (Set of calibrated to $\leq 1\%$ intensity sources listed in Table 3).
- 2) F. A. White, T. L. Collins, and F. M. Burke, Phys. Rev. 97 (1955) 566.
- 3) D. H. White and R. E. Birkett, Nucl. Phys. A136 (1969) 657.
- 4) W. F. Edwards, F. Boehm, J. Rogers, and E. J. Seppi, Nucl. Phys. A63 (1965) 97.
- 5) D. H. White, R. E. Birkett, and T. Thomson, Nucl. Instr. Methods 77 (1970) 261.
- 6) J. T. Routti and S. G. Prussin, Nucl. Instr. Methods 72 (1969) 125.
- 7) C. M. Lederer, University of California, Lawrence Radiation Laboratory Report No. UCRL-18948, unpublished (1969).
- 8) L. J. Jardine, unpublished results (1970).
- 9) L. J. Jardine and E. Temple, unpublished results (1970).
- 10) W. R. Kane and M. A. Mariscotti, Nucl. Instr. Methods 56 (1967) 189.
- 11) D. P. Donnelly, H. W. Baer, J. J. Reidy, and M. L. Wiedenbeck, Nucl. Instr. Methods 57 (1967) 219.
- 12) T. Yamazaki and J. M. Lollander, Nucl. Phys. 84 (1966) 505.
- 13) L. B. Magnusson, Phys. Rev. 107 (1957) 161.
- 14) W. Bambynek, European Atomic Energy Community Report No. EUR 2632d (1965).
- 15) A. Peghaire, Nucl. Instr. Methods 75 (1969) 66.
- 16) G. D. Spouse and S. S. Hanna, Nucl. Phys. 74 (1965) 177.
- 17) W. L. Croft, B. G. Pettersson, and J. H. Hamilton, Nucl. Phys. 48 (1963) 267.
- 18) R. W. Fink, R. C. Jopson, H. Mark, and C. D. Swift, Rev. Mod. Phys. 38 (1966) 513.

- 19) C. M. Lederer, J. M. Hollander, and I. Perlman, Table of Isotopes, Wiley, New York (1967).
- 20) E. Vatai, D. Varga, and J. Vehin, Nucl. Phys. A116 (1968) 637.
- 21) H. H. Hansen, G. Lowenthal, A. Spornol, W. van der Eijk, and R. Vaninbroukx, Z. Physik 218 (1969) 25.
- 22) J. H. Hamilton, S. R. Amtey, B. van Nooijen, A. V. Ramayya, and J. J. Pinajian, Phys. Letters 19 (1966) 682.
- 23) S. Raman, Z. Physik 228 (1969) 387.
- 24) J. I. Rhode, O. E. Johnson, and W. G. Smith, Phys. Rev. 129 (1963) 815.
- 25) N. H. Lazar, E. Eichler, and G. D. O'Kelley, Phys. Rev. 101 (1956) 727.
- 26) M. Sakai, T. Yamazaki, and J. M. Hollander, Nucl. Phys. 84 (1966) 302.
- 27) D. C. Hall and R. G. Albright, Nucl. Phys. A91 (1967) 495.
- 28) W. Rubinson and K. P. Gopinathan, Phys. Rev. 170 (1968) 969.
- 29) J. J. Sapyta, E. G. Funk, and J. W. Mihelich, Nucl. Phys. 139 (1969) 161.
- 30) S. C. Gujrathi and J. M. D'Auria, Nucl. Phys. A161 (1971) 410.
- 31) G. Scharff-Goldhaber and M. McKeown, Phys. Rev. 158 (1967) 1105.
- 32) H. Paul, M. McKeown, and G. Scharff-Goldhaber, Phys. Rev. 158 (1967) 1112.
- 33) W. F. Edwards and F. Boehm, Phys. Rev. 121 (1961) 1499.
- 34) E. Bodenstedt, H. J. Körner, E. Gerdau, J. Radeloff, C. Günther, and G. Strube, Z. Physik 165 (1961) 57.
- 35) R. S. Seltzer and E. C. Seltzer, Nucl. Data A4 (1968) 1.
- 36) L. J. Jardine, University of California, Lawrence Radiation Laboratory Report No. UCRL-20476, unpublished (1971).
- 37) V. S. Gvozdev, L. I. Rusinov, Yu. I. Filimonov, and Yu. L. Knazov, Nucl. Phys. 6 (1958) 561.

- 38) F. L. Oetting and S. R. Gunn, *J. Inorg. Nucl. Chem.* 29 (1967) 2659.
- 39) R. E. Stone and E. K. Hulet, *J. Inorg. Nucl. Chem.* 30 (1968) 2003.
- 40) S. C. Anspach, L. M. Cavallo, S. B. Garfinkel, J. M. R. Hutchinson, and C. N. Smith, *Natl. Bureau Stds. Misc. Publ. Report No. 260-9* (1965).
- 41) F. Lagourine, Y. LeGallic, and J. Legrand, *Int. J. Appl. Rad. Isotopes* 19 (1968) 475.
- 42) K. F. Walz and H. W. Weiss, *Z. f. Naturforschung* 25a (1970) 921.
- 43) E. I. Wyatt, S. A. Reynolds, T. H. Handley, W. S. Lyon, and H. A. Parker, *Nucl. Sci. Engr.* 11 (1961) 74.
- 44) J. S. Merritt and J. G. V. Taylor, *Atomic Energy of Canada, Ltd. Report No. AECL-3333* (1969) 32.
- 45) Kai Siegbahn, Alpha-, Beta-, and Gamma-Ray Spectroscopy, Volume 1, North-Holland Publishing Co., Amsterdam (1965) p. 827.

Table 1. Energies and Relative Intensities of ^{182}Ta in Energy Range 100-1300 keV

E_{γ}^a	Edwards et al. ⁴⁾	White et al. ⁵⁾	Sapyta et al. ²³⁾	Present Work
	I_{γ}^b	I_{γ}^c	I_{γ}^c	I_{γ}^c
100.104±0.002	40.2		40.7±4.1	40.2±1.0
152.435±0.003	20.5±0.8	21.3±1.0	19.5±2.0	20.5±0.5
156.387±0.003	8.04±0.4	8.07±0.4	7.5±0.75	7.6±0.2
179.393±0.004	9.2±0.4	9.57±0.5	8.7±0.9	8.8±0.3
222.109±0.005	22.5±0.9	22.6±1.2	21.2±2.1	21.3±0.55
229.322±0.005	11.1±0.5	10.9±0.5	10.5±1.1	10.3±0.3
264.072±0.009	10.8±0.5	10.6±0.4	10.3±1.0	10.1±0.3
E_{γ}^d				
1121.298±.013		100.	100.	100.
1189.046±.013		47.4±0.7	46.3±3.2	46.5±0.7
1221.399±.013		79.3±1.2	77.3±5.4	77.3±1.2
1231.010±.013		33.4±0.5	32.7±2.3	32.8±0.5

^aEnergies listed in keV are those reported by Edwards et al.⁴⁾

^bIntensities⁴⁾ are renormalized to the 100 keV transition of present work

^cIntensities are normalized to 1121 keV transition

^dThe energies listed in keV are those reported by White and Birkett³⁾.

Table 2. Gamma Energies and Absolute Gamma Intensities of ^{180m}Hf in the Range of 57-501 keV

E_{γ}^a	I_{γ}^a	I_{γ}^b	I_{γ}^c	I_{γ}^d
57.442	47.5±3.4		48.6±0.86	
93.263	18.40±0.16		16.7±0.33	17.47±0.17
215.241	81.1±2.4	81.4	83.4±2.3	81.4±0.8
332.272	94.4±4.0	94.4±0.8	94.4±4.0	94.4±0.9
443.168	84.8	83.0±1.2	81.9±4.3	82.8±1.5
500.702	15.7±1.3		17.05±5.2 ^e	14.5±0.4

^aGujrathi and D'Auria³⁰) results with stated energy errors of ±0.015 keV. The measured gamma intensities were renormalized to 332 keV.

^bKane and Mariscotti¹⁰) renormalized results.

^cEdwards and Boehm³³) results using a bent crystal gamma spectrometer.

^dPresent work where ~ 1% error was arbitrarily assigned for the 93, 215, and 332 keV transitions due to uncertainty in the theoretical E2 conversion coefficients.

^eThis was deduced from feeding and not measured directly (ref. 33).

Table 3. Energies and Intensities of the IAEA Standards and ^{24}Na

Isotope	Half-Life ^b	E_{γ}^a	I_{γ}^b
^{241}Am	$432.9 \pm 0.8\text{y}$ ^{38,39)}	59.538 ± 0.008	35.9 ± 0.6 ^{12,13,14,15)}
^{57}Co	$271.6 \pm 0.5\text{d}$ ⁴⁰⁾	122.055 ± 0.013	85.0 ± 1.7 ^{8,16,27,28)}
		136.47 ± 0.008	10.65 ± 0.4 ⁸⁾
		136.47 ± 0.008	11.4 ± 1.3 ^{16,27,28)}
^{203}Hg	$46.8 \pm 0.2\text{d}$ ^{40,41)}	72.873 ± 0.001	9.7 ± 0.5 ^{17,18)}
		82.5 ± 0.2	2.8 ± 0.2
		279.186 ± 0.009	81.55 ± 0.15
^{22}Na	$2.602 \pm 0.005\text{y}$ ⁴⁰⁾	511.0041 ± 0.0016	181.1 ± 0.2 ^{19,20)}
		1274.55 ± 0.04	99.95 ± 0.02
^{137}Co	$29.90 \pm 0.05\text{y}$ ⁴²⁾	32.1 ± 0.1	5.7 ± 0.2 ^{12,21)}
		36.5 ± 0.1	1.3 ± 0.1
		661.618 ± 0.028	85.1 ± 0.4
^{54}Mn	$312.6 \pm 0.3\text{d}$ ^{15,19,43)}	834.81 ± 0.03	100.0 ²²⁾
^{60}Co	$5.275 \pm 0.005\text{y}$ ^{41,42,44)}	1173.231 ± 0.024	99.87 ± 0.05 ²³⁾
		1332.501 ± 0.021	99.999 ± 0.001

(continued)

Table 3. (continued)


Isotope	Half-Life	E_{γ}^a	I_{γ}^b
^{88}Y	$107.4 \pm 0.8\text{d}$ ^{40,43)}	898.021 ± 0.023	91.4 ± 0.7 ^{24,25,26)}
		1836.129 ± 0.031	99.4 ± 0.1
^{24}Na	$15.00 \pm 0.02\text{h}$	1368.526 ± 0.044	100.0
		2754.098 ± 0.183	100.0

^aEnergy values are adopted values from Jardine³⁶⁾.

^bIntensity and half-life values recommended by Nuclear Data and the IAEA¹⁾.

Figure Captions

- Fig. 1. The relative photopeak efficiency curve for the 10-cm³ Ge(Li) detector as a function of gamma-ray energy.
- Fig. 2. The ^{180m}Hf decay scheme used in the photopeak efficiency determination. The energies (keV) are based on the measurements of Gujrathi and D'Auria³⁰). The absolute gamma-ray intensities expressed in per cent from this study are shown in parenthesis.
- Fig. 3. The relative photopeak efficiency curve for a 40-cm³ true coaxial Ge(Li) detector as a function of energy obtained using the isotopes and intensities reported in the study.



XBL715-3458

Fig. 1





Fig. 2



XBL715-3459

Fig. 3

LEGAL NOTICE

This report was prepared as an account of work sponsored by the United States Government. Neither the United States nor the United States Atomic Energy Commission, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness or usefulness of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights.

TECHNICAL INFORMATION DIVISION
LAWRENCE RADIATION LABORATORY
UNIVERSITY OF CALIFORNIA
BERKELEY, CALIFORNIA 94720