Lawrence Berkeley National Laboratory

Recent Work

Title

TREATMENT PLANNING FOR CHARGED PARTICLE RADIOTHERAPY

Permalink

<https://escholarship.org/uc/item/5j17h648>

Author

Chen, G.T.Y.

Publication Date

1979-05-01

C.2

TREATMENT PLANNING FOR CHARGED PARTICLE RADIOTHERAPY

George T. Y. Chen, Rajinder P. Singh, Joseph R. Castro,
John T. Lyman, and Jeanne M. Quivey

May 1979

Prepared for the U. S. Department of Energy
under Contract W-7405-ENG-48

TWO-WEEK LOAN COPY

*This is a Library Circulating Copy
which may be borrowed for two weeks.
For a personal retention copy, call
Tech. Info. Division, Ext. 6782*

RECEIVED
LAWRENCE
BERKELEY LABORATORY

JUN 28 1979

LIBRARY AND
DOCUMENTS SECTION

LBL-7493 C.2

DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

TREATMENT PLANNING FOR CHARGED PARTICLE RADIOTHERAPY

George T.Y. Chen, Ph.D.
Rajinder P. Singh, Ph.D.
Joseph R. Castro, M.D.
John T. Lyman, Ph.D.
Jeanne M. Quivey, M.D.

Radiotherapy Section, Bldg.55, Rm. 121
Lawrence Berkeley Laboratory
Berkeley, CA. 94720

Present at the 19th Annual Meeting, American Society of Therapeutic Radiologists,
Denver, CO. (1977)

Supported by National Institutes of Health-National Cancer Institute Grants No.,
5P01CA19138 and 2P01CA19138 and by Energy Research and Development Agency Contract
W-74-5 Eng 48

Accepted for publication to the International Journal of Radiation Oncology,
Biology and Physics.

Reprint requests to G.T.Y. Chen, Ph.D.

ABSTRACT

A computerized treatment planning system based on quantitative information in computerized tomographic scans has been developed for the heavy ion radiotherapy trials at the Lawrence Berkeley Laboratory. The CT number of each pixel is converted to a water equivalent length and used in the calculation of isoeffect dose distributions from multiport heavy charged particle irradiation. The water equivalent length of pixels in the beam path has also been histogrammed and used in quantitatively studying the range shortening effects by inhomogeneities. Treatment plans for representative lesions of the pelvis, abdomen, and thorax are presented.

KEY WORDS

Heavy ion radiotherapy, computerized tomography
Treatment planning

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

INTRODUCTION

The use of heavy charged particles in radiotherapy may provide a therapeutic gain as a result of an improved dose distribution and the potential of a biologic advantage(3). In order to test the value of this new modality, the technical problems of dose delivery must be adequately understood and solved. One such aspect is the development of computerized treatment planning, which will reflect realistically the isoeffect distributions from multiport heavy ion therapy in the presence of inhomogeneities. In this article, we describe the development of a computerized tomography based treatment planning program and present typical treatment treatment plans for several sites of interest to the heavy ion radiotherapy trials under way at the Lawrence Berkeley Laboratory (4).

METHODS AND MATERIALS

Beam Properties

Finite Range

The finite range of charged particles is the principal source of an improved physical dose distribution over photon therapy. Beam depth of penetration is determined by the initial energy of the particles and the properties of the medium traversed. Accelerated beams of helium, carbon, neon, and argon ions are available with sufficient energy to pass through beam scattering foils and ridge filters and still have a penetration range in the patient of at least 25 cm of water. From our experience with helium ions, a maximum residual range of about 26 cm is adequate for most deep seated lesions in the abdomen and pelvis, including the range shortening effects of bone. A frequency analysis of residual ranges used in 33 patients showed an average range of 20.5 ± 3.9 cm of water. The physical dose rates for these beams as measured at the proximal peak, range from 250 rad/min/liter for carbon to 35 rad/min/liter for argon. Folding in the high relative biological effectiveness (RBE) for these heavy ions increases the effectiveness of these beams relative to conventional radiations by a factor of from 2 to 3 (2). This results in treatment times of typically one to two minutes.

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

By appropriate use of a variable water column and three dimensional compensation, the depth of penetration may be adjusted precisely to cover the tumor volume adequately and minimize the dose to critical tissues distal to the target. However, in order to exploit this potential advantage properly, the appropriate water equivalent depth of penetration must be determined accurately. This parameter is the single most important quantity to be defined in order to position the beam precisely in depth. Because of the sharp dose gradient at the distal end of the beam, an incorrect calculation of the depth of penetration could lead to a significant underdose of the target or an undesirable overdose of sensitive tissues distal to the tumor. The quantitative information necessary to characterize the inhomogeneities in the beam path and to determine the appropriate treatment depth is obtained from CT data. Details of the method used are presented later.

Ridge Filtered Beams

The narrow Bragg peaks of heavy ions are broadened to therapeutically useful dimensions by spiral ridge filters. (11) Since the RBE increases toward the end of the range, the physical dose is decreased near the distal end of the ridge filtered beam in order to produce a region of uniform biological effect. The biophysical model used to calculate the physical dose distribution required for isosurvival over a desired depth is described elsewhere (4). The physical dose distribution of a family of ridge filtered carbon beams is shown in Figure 1a. This series of ridge filtered beams is designed to produce regions of uniform biological effect from 2 to 14 cm in depth in 2 cm increments. The curves are arbitrarily normalized at the distal edge. The corresponding Cobalt-Rad-Equivalent (CoRE) dose distributions is shown in Figure 1b. A CoRE distribution is the low LET depth dose profile

equivalent to the biological effect produce by the heavy charged particle beam. These curves are also called isoeffect depth dose distribution because they produce uniform biological effect over their flat top regions. The physical dose distributions of the carbon beams in Figure 1a are designed to produce the isoeffect distributions in Figure 1b. Isoeffect distributions for helium and neon are shown in Figures 1c and 1d. These isoeffect depth dose distributions are the principal source of beam data used in treatment planning calculations.

The carbon and neon beams exhibit tails in the dose distributions which result from the production of nuclear fragments as the primary beam traverses the medium. These lighter fragments have a slightly longer range than the parent nuclei and result in dose past the stopping point of the primary beam. For carbon, the tail dose is initially 30% and falls to 15% within 10 additional centimeters of water. The specific amount of fragmentation is dependent on the initial energy of the beam. A lower energy beam will result in a lower fragmentation dose, but would have a shorter residual range.

The entrance dose of heavy ion beams is dependent upon the ridge filter specification and ion type. The helium skin dose from a single port ranges from 60% for a 6 cm ridge filter, to 80% for a 14 cm ridge filter, while for neon, the corresponding values are 75% and 92% respectively. The use of opposing or multiple ports with particle beams reduces the skin dose to well-tolerated levels.

Sharp Penumbra

Heavy ion beams exhibit sharp lateral penumbras. In the plateau region of the depth dose curve, the lateral dose has been measured to fall from

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

90% to 10% in 3.5 mm for the helium beam. As the beam energy decreases, multiple Coulomb scattering increases, and the penumbra in the isosurvival region of the ridge filtered helium beam increases to 10 mm. The carbon and neon penumbras are a factor of about 3 smaller than the helium values because of the increased particle mass. In treatment planning calculations, the helium penumbra is modelled by a Gaussian fall off with parameters obtained from experimental measurements. For carbon and neon, the penumbras are extremely sharp, and are smaller than the grid spacing of the calculation matrix. In these cases, the penumbra dose is calculated to fall from the 100% level to the 0% level within one lateral pixel.

USE OF CT DATA

CT data are used to define the position of the treatment volume relative to critical structures and to quantify the inhomogeneities which perturb the particle beams. The CT value of each pixel must be converted to a water equivalent path length in order to provide the expected change in residual range of the beam. CT numbers are a measure of the linear attenuation coefficient of each pixel at diagnostic photon energies. At these energies, the linear attenuation coefficient is dependent upon the effective atomic number, which contributes to the photoelectric effect, and the electron density. The equation relating these quantities is shown below (14):

$$\mu = \rho N_0 \frac{Z}{A} [\sigma^{KN}(E) + \sigma^{PE}(E, Z)] \quad (1)$$

where ρ = density

N_o = Avogadro's number

Z = atomic number

A = atomic weight

σ_{KN} = Klein Nishina cross section per electron

σ_{PE} = photoelectric cross section per electron

The quantity $\rho N_o Z/A$ is the electron density of the material in electrons per cubic cm. The photoelectric cross section contributes to the observed CT number with a Z^3 dependence. At 70 KeV (equivalent to 140 KVP), the photoelectric effect constitutes 5% of the interactions in muscle and 26% of the interactions in compact bone (9).

In contrast, the equation governing energy loss of heavy charged particles in a medium (1) is:

$$\frac{dE}{dX} = \frac{k_1 \rho N_o Z}{\beta^2 A} \left[\ln \frac{k_2 \beta^2}{I(1-\beta^2)} - \beta^2 \right] \quad (2)$$

where k_1 = constant including particle charge

β = relativistic velocity v/c

k_2 = constant

I = adjusted ionization potential of the medium

This equation defines a direct relation between energy loss and the electron density of the medium, with a logarithmic dependence on the adjusted ionization potential. The ionization potential of different human tissues produces a small variation in the logarithmic term. The dominant factor in determining the range shortening of charged particles is the electron density.

Since human tissues are characterized by a number of different atomic

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

numbers and densities, the problem is to extract the water equivalent length per pixel from CT data. The required electron density may be extracted from a dual energy scan as suggested by others (12). However, a precise knowledge of the effective energy and pixel registration between the dual scans complicates the use of this technique.

If approximations are made, it is possible to correlate the water equivalent length of each pixel with the CT number from a single energy scan with sufficient accuracy for clinical use. Specifically, the method used involves defining a calibration curve relating the CT numbers of tissue analogues with the observed water equivalent length. The model involves conceptualizing tissues of various densities as mixtures of two appropriate materials (6,10,16). In order to simplify the extraction of water equivalent length per pixel, the CT number interval is divided into two sub-regions.

For CT numbers less than 0, the tissue producing the observed CT number is assumed to be a mixture of air and water (10). Since both of these materials have very nearly the same Z and Z/A , the CT number is therefore proportional to the water equivalent path length. Fat, which has a Z of 6.2 (13) is the potential problem in this interval. However, using equation (1), calculations show that the lower Z is offset somewhat by a higher Z/A , resulting in a net error in the calculated water equivalent length for fat of about 1.5%.

For CT numbers greater than zero, the mixture is assumed to have a well defined correlation between density and atomic number. This assumption reduces the number of independent variables to one, thus allowing for the water equivalent range of bone to be extracted from a single energy CT scan. The rationale for this model rests on the assumption that intermediate density bone may be thought of as a mixture of water and compact bone. Furthermore, a realistic bone analogue material is needed for the calibration. Dibasic potassium phosphate is such a material, and is a water soluble bone analogue material (16) with physical density and attenuation properties very similar

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

to the principal mineral in bone, namely calcium hydroxapatite.

Calibration curves converting CT number to water equivalent path length per pixel appear in Figure 2, ^{for the} General Electric 7800 Scanner and the EMI 5005 Scanner. These data were obtained by CT scanning the phantom materials and then placing the samples in a helium ion beam to measure the relative water equivalent range. Tissue phantom materials used in the calibration included 1) solutions of water and potassium phosphate to simulate a variety of bone densities, 2) alcohol and water mixtures to simulate adipose tissues, and 3) phantom lung and cork to simulate low density tissues. The GE scanner was operated at 120 KVP and the EMI scanner was run at 140 KVP during both calibration and patient scans. Both curves exhibit two general slopes, one associated with CT values less than zero, and a lower slope correlated with the high CT range. The decreased slope for bone may be explained by the Z^3 dependence of the photoelectric effect contributing strongly to CT values, while Z dependent effects contribute to the water equivalent length only weakly through the log term in equation 2.

Note that both scanners have a limited dynamic range for high density measurements. A CT number of 500 corresponds to a water equivalent range of 1.5. Furthermore, bone with a higher water equivalent length is interpreted by the scanner as having the same CT number, since this is the maximum value recorded. Fortunately the number of pixels with a CT number equal to 500 is negligibly small. An analysis of scan numbers of a coronal head scan through the base of the skull showed that less than .5% of the pixels had CT numbers greater than 490. Although this analysis is subject to errors from finite slice thickness and partial averaging, the result suggests that the amount of truly dense bone (with a density greater than 1.5) is quite small and distributed in thin shells rather than concentrated in a single volume.

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

The quantitative CT data are sensitive to additional factors, such as sampling position, systematic calibration drift, and artifacts. The calibration data were taken in a 25 cm diameter cylindrical water phantom, with sample vials symmetrically placed within the circumference. In order to test the influence of body shape on CT values, selected samples were also placed in an anthropomorphically shaped phantom and scanned. When identical bone solutions were placed in different locations of the body phantom, three percent differences in CT values were measured. No corrections for these variations are made in conversion of CT data to water equivalent range. Calibration shifts were also detected over a period of about a year. These drifts have been measured to be as large as 3%. To circumvent this source of error, calibration scans are taken before patient scans.

Cupping artifacts are found in CT data near bone/tissue interfaces and other high/low density regions. This effect is manifested by an apparent decrease of CT numbers extending several pixels into the bone with a corresponding increase of CT numbers in the region from the interface into the tissue. Range calculations of beams perpendicular to the interface will not be influenced by the artifact since the CT effects on either side of the interface tend to balance each other. Calculational errors produced on tangential rays by this effect have yet to be studied.

The accuracy of the single energy technique in range determination was tested by CT scanning anthropomorphic Alderson Rando phantom, calculating the anterior-posterior water equivalent range of the body section using the calibration curve, and comparing these results with experimentally measured ranges using the Bragg peak shift seen in the helium ion beam. The results of this study appear in Figure 3. In general, agreement between calculation and experiment is within 3%, with a suggestion that a systematic error of

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

3 mm is present.

This systematic error of 1.5% may be due to the difference in the relative stopping power between water and the phantom material. The phantom is made from isocyanate rubber, which is oxygen deficient. Over 50% of the oxygen content in water or muscle is replaced with carbon in this compound (15). The largest difference in Figure 3 between calculated and experimental range appear to be about 5%.

The present level of accuracy is sufficient for most abdominal and pelvis irradiations. A 3% error with a typical residual range of 20 cm yields a 6 mm uncertainty in the stopping point of the beam. This amount is ordinarily included within the margins of the target volume as defined by the radiotherapist. Improvements in the calibration curve and elimination of the systematic difference could improve the stopping point uncertainty by a factor of 2.

TREATMENT PLANNING CODE

A computer program was written to calculate isoeffect or physical dose distributions for heavy ion irradiations. The quantities specified for each port include particle type, direction, field size, residual range, ridge filter, weight, field position and data files describing external compensation. Any number of static ports may be specified. The density matrix of water equivalent path lengths is generated by use of the calibration curve. This array is an 80x80 matrix obtained from compressing the original 320x320 scan data. The pixel dimensions are a function of the specific scanner, but generally are about .5 cm after the matrix compression. The principal beam transport algorithm is based on range shortening by tissues and inhomogeneities. Given the residual range, the water equivalent length of each successive pixel is subtracted to obtain the new residual range. For oblique fields, the partial path through each pixel is calculated. The

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

dose at each pixel is calculated by a look up table relating residual range with central axis dose. An additional quantity calculated in each plan is the percent integral dose. This is defined as the sum over all pixels of the percent dose times the water equivalent factor of each pixel. This differs from the quantity of integral dose in that it is calculated only within the particular slice. However, the percent integral dose per slice is useful in evaluating the advantages of various multiple port configurations in the selection of an optimal field arrangement. The treatment plan giving 100% to the target volume with the lowest percent integral dose is generally preferred over other plans, assuming critical structures are avoided.

The dose calculation algorithm does not include effects from multiple Coulomb scattering. Heavy charged particles undergo small angle scattering as do all charged particles. However, for heavy ions, the scatter angles and lateral displacements are small relative to lighter particles such as pions and electrons. Under most conditions, multiple Coulomb scattering does not affect the resulting isodose distributions. Regions where this effect is noticeable include regions near the boundary of tissues with greatly differing densities, such as the mediastinum-lung interface. In this region, scattered particles may traverse drift spaces and affect the dose laterally adjacent to beam rays. Dosimetry studies are needed to quantify these effects.

The dose enhancement and depletion near sharp edges has been studied experimentally with the helium ion beam to determine the magnitude of this effect and the need for its inclusion in dose calculation schemes. Dose perturbations near a 1 inch slab of lucite in air measure about 6% in magnitude. The geometry used for these measurements are described elsewhere (8). This value is not nearly as large as the theoretically possible $\pm 50\%$ dose perturbations (5) because of the non-parallel nature of the beam. The small amount of beam divergence which degrades the dose perturbation results from

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

small angle scatter as the beam passes through beam elements such as the ridge filter and water column. Since the edge effect is small for our beams, algorithms for calculating these perturbing effects are not included in the dose calculations.

A mini computer with 128 K words of core memory, floating point processor, a 33 megabyte disk drive and graphics display are the components used for treatment planning. The program is written in Fortran IV, and a typical treatment plan requires 6 minutes of computation to produce the dose matrix. An additional 5 minutes are needed to obtain a hard copy of the treatment plan on an electrostatic printer-plotter. In this form, isodose and isoeffect contour lines are superimposed on a grey level image of the CT slice. Calculations are set up in batch mode to facilitate the computation of several planes in the treatment volume without operator intervention.

RESULTS

Representative Treatment Plans

Treatment plans are included which show representative portals for the irradiation of the outlined target volumes. These plans however, do not show the only possible portal arrangements utilized in clinical irradiation since patient condition and previous treatments may dictate the use of particular fields. Only one slice per site is displayed here although in practice, multiple slices are calculated in order to appreciate the dose distribution in three dimensions. In these plans, the CT data are assumed to depict the correct relative location of tumor to normal tissues. This is not strictly ^{the} ~~true~~ in our case since patients presently are scanned in a recumbent position, while many of them are treated upright because of the limitation of a fixed horizontal beam.

With these limitation in mind, treatment plans are now presented for the irradiation of the pancreas, para-aortic nodes, pelvis and esophagus. Each

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

plan contains isoeffect contours ranging from 10 to 100% in 10% increments of CoRE dose. The irradiation parameters for each port are also given, including ion type, beam direction, weight, and field width, residual range in centimeters of water, and ridge filter in cm. The percent integral dose of the plan for each slice is also calculated.

Pancreas

Figure 4 shows treatment plan for the irradiation of a pancreatic lesion with carbon ions. Equally weighted anterior and lateral ports are used for this particular tumor location. In addition to a surface flattening bolus, a simple wedge shaped compensator is used in both ports to shape the high dose region around the distal contour of the target volume. The artificial step like dose gradient posterior and distal to the target results from the finite pixel size used to describe the compensation. In practice, the compensator would be constructed as a smooth wedge, not as a step wedge.

In the abdomen, the major types of tissues and inhomogeneities encountered are soft tissue, fat, rib and bowel gas. From an analysis of pixels in the beam path, the areas traversed by the charged particle beams have an average density of 1.0 g/cc. Varying amounts of fat along rays will produce range variations of approximately 0.5 cm. Although ribs, are present in the lateral portal, they perturb the beam minimally, on the order of 3 mm. This is consistent with an expected 1 cm. path length through a given rib, which typically would have a density of 1.3 g/cc. in its densest areas.

The target volume in this plan is irradiated uniformly to 100% while tissues proximal to the target are dose to 40%. The spinal cord, one kidney and portions of the gut receive up to a 10% exit dose from the beam fragmentation tail. These critical structures received significantly less dose with charged particles compared with photon irradiation using standard treat-

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

ment plans. A three field variation of this plan was also calculated, but appeared to produce no advantage over the two field for the target shown. Although the proximal areas were given about 10% less dose in the three field plan, a large region of intestinal tract was necessarily irradiated to about 30%. In addition, the percent integral dose for the three field plan was 8% higher. If the lesion were more centered with respect to midline, a three field plan would be acceptable.

Esophagus

The primary inhomogeneities involved in the irradiation of an esophageal lesion are the lungs, trachea, bronchi, sternum and ribs. As Figure 5 shows, the thorax presents a number of regions with vastly different densities. This histogram of pixels within the beam paths reveals a soft tissue peak at 1.02 g/cc. with a large peak resulting from lung at .27 g/cc. The pixels with intermediate densities between .5 and .77 g/cc. are located at the edges of the lung and at the anterior body surface. These densities result from both partial sampling and respiratory motion during the CT scan.

A neon ion treatment plan for the esophagus is shown in Figure 6. Three equally weighted and compensated fields are used in this comedown irradiation. The anterior and right portals have wedge shaped compensators which pull the dose gradient away from the spinal cord. Again, the target volume is uniformly treated to 100% with an average mediastinal dose of 25% and average lung dose of 35%. The spinal cord receives 10% from the fragmentation tail of the neon beam.

Aside from the wedge compensators, only a flattening bolus is used to produce isodose shaping. Although the path length through lung may vary significantly the resulting isodose distribution from 3 ports is quite

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

suitable for this target volume. Note the irregular isoeffect contour in the lungs resulting from inhomogeneities.

This particular site is amenable to in vivo dosimetry. Typically, a multidiode probe housed in a naso-gastric tube is passed down the esophagus to the region of interest, and the charged particle beam is swept through the body (by sweeping the water column) until the distal edge of the beam is detected. Such measurements give information on the required residual range, and furnish experimental verification of calculated residual ranges. These techniques were reported by us elsewhere (5).

Para-aortic Nodes

Para-aortic nodes are technically difficult to irradiate with photons because in most techniques either the spinal cord or kidneys reach tolerance dose before an adequate dose is given to the nodes. These limiting factors may be reduced by the use of heavy ion beams. A proposed scheme for the irradiation of the para-aortic nodes with helium ions is presented in Figure 7. A pair of parallel opposed fields are used, weighted 2:1 in favor of the posterior field. The target volume is uniformly irradiated to 100% with this field configuration while the anterior intestinal tract receives 30% and the spinal cord received about 60%. In this plan, compensation was used to shape the high dose region around the irregularly shaped target. The use of a cord compensator in the anterior field pulls the steep dose gradient away from the spinal cord and cauda equina while the tumor is still covered adequately.

The use of a direct posterior field necessarily includes irradiation through the vertebral body. This bony inhomogeneity may pull back the beam by as much as 1.5 cm. The calculated range shortening by a vertebral body is shown in Figure 8. The image of this bony inhomogeneity is represented

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

in hexadecimal form, where each character represents a particular water equivalent range. The pixel size in this image is 1 mm x 1 mm. The level of the display has been elevated to show only the vertebral body. With the posterior entrance surface bolussed flat, the residual range was set to give sufficient penetration to totally traverse the inhomogeneity. The asterisks at the top of the figure represent the calculated stopping points of the beam. A comparison of those rays which passed through the vertebral body with those which passed only through soft tissue shows a maximum range shortening here of 1.0 cm. The calculated range shortening for a number of vertebral bodies has been compared with experimental measurements using film stack techniques (5), and both experiment and calculation agree within expected uncertainties. The observed range shortening spans the interval 1.0 to 1.6 cm.

Pelvis

A treatment plan for the irradiation of a large pelvic lesion with carbon. Figure 9 shows. In this site, the beams must penetrate through the iliac crest before reaching the target. A pixel by pixel analysis of the CT scan shows that within the beam paths, approximately 25% of the pixels are fat, and about 17% of the pixels are bone with an average density of 1.17 g/cc. Data in the form of a histogram appear in Figure 10.

A three field portal arrangement has been chosen for this irradiation. The carbon beam is perturbed most in the posterior port, which must traverse the high density bone in the posterior region of the left iliac crest. From this direction, the range shortening may be as great as 2 cm. as compared with soft tissue paths. Compensation has been added to shape the high dose region to a circularly shaped tumor. The use of heavy ions here has reduced the dose to the bladder and part of the interestinal tract.

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen, et al

DISCUSSION

This effort represents the first version of a charged particle treatment planning project. Several areas of the program will be improved in the near future. The treatment planning process will be modified to allow for interactive use through target volume delineation with a cursor and CT graphics display hardware. This will enable the therapist to enter the desired target volume on several sequential CT slices. Such information will be used to shape the appropriate portal collimators and the contour information will be used to design three dimensional compensation. Increase in calculational speed will permit the therapist and physicist to rapidly calculate a selection of portal arrangements and display intermediate isodose distributions on the graphics unit before final selection of the optimal plan.

Other areas for program improvement include the incorporation of respiration. Computer simulation of beam scanning systems is also planned. The present technology of dose delivery requires the selection of a ridge filter whose width is determined by the maximum depth of the tumor along the beam axis. This constrains us to use a ridge filter which has an inappropriate width in other planes of interest. One solution to this constraint is to eliminate the use of fixed ridge filter and use a raster scanning pencil beam with variable ridge filter. At each raster point, the depth of penetration and dose, as well as the ridge filter width would be altered dynamically to produce the resulting desired dose distribution. With computer simulation of this system, it will be possible to explore the advantages of beam scanning over fixed ridge filter systems, and to define the performance specifications suitable for charged particle therapy.

ACKNOWLEDGEMENTS

The authors wish to thank Dr. Douglas Boyd of the University of California, San Francisco for the use of his CT phantom and for many valuable comments, and Dr. Samuel Pitluck for his assistance in computer programming.

References

1. Bischel, H. Passage of charged particles through matter. American Institute of Physics Handbook, sec. 8, p.142, McGraw-Hill, 1972.
2. Blakely, E.A., Tobias, C.A., Yang, T.C. and Smith, K.A. Cell survival parameters for extended bragg peaks. Radiat. Res. 74, 588 (1978) (abstract).
3. Castro, J.R. and Quivey, J.M. Clinical experience and expectations with helium and heavy ion irradiation. Int. J. Radiat. Oncol. Biol. Phys. 3, 127-131 (1977).
4. Castro, J.R., Quivey, J.M., Lyman, J.T., Chen, G.T.Y., Kanstein, L. and Walton, R. Heavy ion radiotherapy, Lawrence Berkeley Laboratory, Report 5610, 198-218 (1977).
5. Chen, G.T.Y., Lyman, J.T., Quivey, J.M., and Castro, J.R. Analysis of dose perturbations by vertebral bodies in charged particle radiotherapy. Med. Phys. 4 (1977) (abstract)
6. Genant, H.K. and Boyd, D. Quantitative bone mineral analysis using dual energy computerized tomographic scanning. Invest. Radiol. 12, 545 (1977) (abstract)
7. Goitein, M. A technique for calculating the influence of thin inhomogeneities on charged particle beams. Med. Phys. 5, 258-264 (1978).
8. Goitein, M., Chen, G.T.Y., Ting, J.Y., Schneider, R.J. and Sisterson, J.M. Measurements and calculations of the influence of thin inhomogeneities on charged particle beams. Med. Phys. 5, 265-273 (1978).
9. Hubbell, J.H. Photon cross sections, attenuation coefficients and energy absorption coefficients from 10 KeV to 100 KeV. NSRDA-NBS 29, p. 70 (1969). U.S. Government Printing Office, Washington, D.C.
10. Kijewski, P.K. and Bjarngard, B.E. The use of computed tomography data for radiotherapy dose calculations. Int. J. Radiat. Oncol. Biol. and Phys. 4, 429-435 (1978).
11. Raju, M.R., Gnanapurani, M., Martins, B., Howard, J. and Lyman, J.T. Measurement of OER and RBE of the 910 MeV helium ion beam using cultured cells (TI). Radiology 102, 425-478 (1972).
12. Rutherford, R.A., Pullan, B.R. and Isherwood, I. Measurement of effective atomic number and electron density using an EMI scanner. Neurorad. 11, 15-21 (1976).
13. Spiers, J. Effective atomic number of tissues and phantom materials. Brit. J. Radiol. 19, 52-63 (1945).

HEAVY ION RADIOTHERAPY TREATMENT PLANNING...Chen et al.

14. Weber, J. and van den Berge, D.J. The effective atomic number and the calculation of the composition of phantom materials. Brit. J. Radi .
42, 378-383 (1969).
15. White, D. R. Tissue substitutes in experimental radiation physics. Med.
Phys. 5, 467-479 (1978)
16. Witt, R.M. Bone standards for the intercomparison of photon absorption-
metric bone mineral measuring systems. DHEW Pub. No.75-883, Proc. Int.
Conference on Bone Mineral Measurements, Chicago, 1973, p.114-122. Pub-
lished by U.S. Dept. HEW, Washington, D.C.

LEGEND

Figure 1. (a) Physical depth dose distribution of a family of carbon ridge filtered beams. (b) Corresponding isoeffect distributions for carbon beams. (c) Isoeffect distributions for helium beams (d) Isoeffect distributions for neon beams.

Figure 2. Calibration curve relating CT number from EMI and GE scanners to water equivalent range per pixel.

Figure 3. A comparison of the CT predicted water equivalent range (line) and measured range of a Rando phantom slice of the abdomen (points). The beam was passed through the section from an anterior port.

Figure 4. A treatment plan for the irradiation of the pancreas with a carbon beam.

Figure 5. Histogram of the water equivalent ranges of pixels within the irradiated volume of an esophageal lesion.

Figure 6. A three field treatment plan for an esophageal lesion using a neon beam.

Figure 7. A treatment plan for the para-aortic nodes. The posterior field is weighted 2:1 relative to the anterior field.

Figure 8. The calculated range shortening by a vertebral body.

Figure 9. Treatment plan for a pelvic mass with carbon beams.

Figure 10. Histogram of water equivalent ranges of each pixel within the irradiated volume in the pelvis case.

XBL787-3372

FIGURE 1

XBL793-3293

FIGURE 2

XBL793-3294
FIGURE 3

BEAM: CARBON
PORT: ANTERIOR
FIELD SIZE (RTX) 12.77 CM
RESIDUAL RANGE: 12.70 CM
RIDGE FILTER: 0.00 CM
HEIGHT: 0.5

BEAM: CARBON
PORT: RIGHT LATERAL
FIELD SIZE (RTX) 9.00 CM
RESIDUAL RANGE: 10.00 CM
RIDGE FILTER: 12.00 CM
HEIGHT: 0.5

INTEGRAL DOSE: 107.00

FIGURE 4

XBB 787-8406

XBL793-3292

FIGURE 5

DEPH: MEON
PORT: ANTERIOR
FIELD SIZE (MTH) 6.18 CM
RESIDUAL RANGE: 15.50 CM
RIDGE FILTER: 6.00 CM
WEIGHT: 0.3

DEPH: MEON
PORT: RIGHT LATERAL
FIELD SIZE (MTH) 6.18 CM
RESIDUAL RANGE: 16.20 CM
RIDGE FILTER: 6.00 CM
WEIGHT: 0.3

DEPH: MEON
PORT: LEFT LATERAL
FIELD SIZE (MTH) 6.18 CM
RESIDUAL RANGE: 13.00 CM
RIDGE FILTER: 6.00 CM
WEIGHT: 0.3

INTEGRAL DOSE: 100.23

0.00 2.00 4.00 6.00
1 CM = 2.00 CM

FIGURE 6

XBB 787-8407

ANGLE: 90.0°
 BEAM: HELIUM
 FIELD SIZE (MTH) 8.50 CM
 RESIDUAL RANGE: 18.50 CM
 RIDGE FILTER: 8.00 CM
 WEIGHT: 0.3
 INTEGRAL DOSE: 50.95

ANGLE: 270.0°
 BEAM: HELIUM
 FIELD SIZE (MTH) 8.50 CM
 RESIDUAL RANGE: 17.50 CM
 RIDGE FILTER: 8.00 CM
 WEIGHT: 0.7
 INTEGRAL DOSE: 96.69

TOTAL INT DOSE: 147.34

FIGURE 7

XBB 787-8403

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30

KEY

DENSITY OF STACK = 1.000

0	..	1.0500	TO	1.0000
1	..	1.0000	TO	1.1100
2	..	1.1100	TO	1.1400
3	..	1.1400	TO	1.1700
4	..	1.1700	TO	1.2000
5	..	1.2000	TO	1.2300
6	..	1.2300	TO	1.2600
7	..	1.2600	TO	1.2900
8	..	1.2900	TO	1.3200
9	..	1.3200	TO	1.3500
A	..	1.3500	TO	1.3800
B	..	1.3800	TO	1.4100
C	..	1.4100	TO	1.4400
D	..	1.4400	TO	1.4700
E	..	1.4700	TO	1.5000

POSTERIOR SURFACE BOLUSSED
FLAT

XBL 787-9623

FIGURE 8

BEAM: CARBON
PORT: ANTERIOR
FIELD SIZE (MTH) 10.30 CM
RESIDUAL RANGE: 19.00 CM
RIDGE FILTER: 12.00 CM
HEIGHT: 0.3

BEAM: CARBON
PORT: LEFT LATERAL
FIELD SIZE (MTH) 8.85 CM
RESIDUAL RANGE: 17.00 CM
RIDGE FILTER: 10.00 CM
HEIGHT: 0.3

BEAM: CARBON
PORT: POSTERIOR
FIELD SIZE (MTH) 10.30 CM
RESIDUAL RANGE: 21.00 CM
RIDGE FILTER: 12.00 CM
HEIGHT: 0.3

INTEGRAL DOSE: 182.55

0.00 2.00 4.17 6.25
1 CM = 2.00 CM

FIGURE 9

XBB 787-8405

XBL793-3295

FIGURE 10

This report was done with support from the Department of Energy. Any conclusions or opinions expressed in this report represent solely those of the author(s) and not necessarily those of The Regents of the University of California, the Lawrence Berkeley Laboratory or the Department of Energy.

Reference to a company or product name does not imply approval or recommendation of the product by the University of California or the U.S. Department of Energy to the exclusion of others that may be suitable.

TECHNICAL INFORMATION DEPARTMENT
LAWRENCE BERKELEY LABORATORY
UNIVERSITY OF CALIFORNIA
BERKELEY, CALIFORNIA 94720