

UCSF

UC San Francisco Previously Published Works

Title

Developing current procedural terminology codes that describe the work performed by machines

Permalink

<https://escholarship.org/uc/item/6q46h3cv>

Journal

npj Digital Medicine, 5(1)

ISSN

2398-6352

Authors

Frank, Richard A

Jarrin, Robert

Pritzker, Jordan

et al.

Publication Date

2022

DOI

10.1038/s41746-022-00723-5

Copyright Information

This work is made available under the terms of a Creative Commons Attribution License, available at <https://creativecommons.org/licenses/by/4.0/>

Peer reviewed

PERSPECTIVE OPEN



Developing current procedural terminology codes that describe the work performed by machines

Richard A. Frank^{1,2}✉, Robert Jarrin^{1,3}, Jordan Pritzker^{1,4}, Michael D. Abramoff^{1,5,6}, Michael X. Repka^{1,7}, Pat D. Baird^{1,8}, S. Marlene Grenon^{1,9}, Megan Ruth Mahoney^{1,10}, John E. Mattison^{1,11} and Ezequiel Silva III^{1,12,13}

The “Taxonomy of Artificial Intelligence for Medical Services and Procedures” became part of the Current Procedural Terminology (CPT®) code set effective January 1, 2022. It provides a framework for discrete and differentiable CPT codes which; are consistent with the features of the devices’ output, characterize interaction between the device and the physician or other qualified health care professional, and foster appropriate payment. Descriptors include “Assistive”, “Augmentative”, and “Autonomous”. As software increasingly augments the provision of medical services the taxonomy will foster consistent language in coding enabling patient, provider, and payer access to the benefits of innovation.

npj Digital Medicine (2022)5:177; <https://doi.org/10.1038/s41746-022-00723-5>

INTRODUCTION

Artificial Intelligence (AI) in health care had been discussed frequently within the American Medical Association’s House of Delegates when in 2019, this group accepted a report “Augmented Intelligence in Health Care”^{1,2}. One of the ways that the American Medical Association (AMA) responded to this report is that the AMA’s Digital Medicine Payment Advisory Group (DMPAG) formed a workgroup with specific expertise in AI. The workgroup developed the Taxonomy of Artificial Intelligence for Medical Services and Procedures and made application to the Current Procedural Terminology (CPT®) Editorial Panel for this taxonomy to be included as an appendix within the CPT® code set³. The Editorial Panel accepted this application during their September 2021 meeting, and Appendix S was made effective January 1, 2022.

CPT® is the most widely accepted medical nomenclature used across the country, offering physicians and health care professionals a uniform language for coding professional services and procedures. Public and private insurers rely on CPT codes to track and pay for medical services⁴. CPT codes are designed to describe the work performed by the physician. When the service is provided in a nonfacility setting (e.g., physician office), CPT codes also account for the practice expense⁵, both fixed and specific to the individual patient for an individual service. The factors considered in determining the value of physician work are the time it takes to perform the service, the necessary technical skill and physical effort, the required mental effort and clinical judgment, and stress due to the potential risk to the patient⁶. Appendix S provides language to describe the “work done by machines” in relation to the work of the physician. Although this concept is specific to CPT coding, the descriptors in Appendix S can help standardize terminology across healthcare; differentiation of the components of software-based medical service will facilitate labeling, valuation, coverage, and payment⁷ (Fig. 1). This article provides context and background on the taxonomy so that innovators, payors, regulators, medical professionals, and patient

organizations can better understand it as they develop CPT code change applications (CCAs) for novel AI products and services.

The workgroup engaged in an heuristics exercise to identify AI products and services currently available or in development and compared them to a limited number of existing AI-related CPT codes such as multianalyte assays with algorithmic analyses and computer-aided detection imaging. The workgroup noted the large number of FDA-cleared devices not yet represented by a CPT code indicated a clear and present need for applicable descriptors to be used in CCAs⁸. The FDA has to date reviewed and authorized 343 “Software as a Medical Device (SaMD)”⁹ products that companies legally market via 510(k)clearance, de novo process, or pre-market approval. Developers and innovators have clinically beneficial products to bring to market and finding their place within CPT is a logical next step to ensure patients have access to the benefits of those software-based services.

The workgroup also considered a product for autonomous point of care diabetic retinal exam which had been cleared by FDA through their de novo pathway. For lack of any precedent in the code set, CPT code 92229 had been created using the word “automated” instead of “autonomous”. At their February 2022 meeting, shortly after Appendix S was made effective, the CPT Editorial Panel considered and adopted a CCA for revision of code 92229 by removing the term “automated” and replacing it with “autonomous”.

UNIQUE FEATURES OF AI

The workgroup concluded that clinical integration and payment would require terminology and categorization for AI CPT code descriptor. With all this in mind the workgroup determined that, to describe the “work done by machines”, CPT codes for AI should be consistent with the technological features of the device or its output, characterize the interaction between the device and the physician or other qualified health care professional (QHP), be discrete and differentiable from other codes and descriptors,

¹American Medical Association’s Digital Medicine Payment Advisory Group (DMPAG) AI Workgroup, American Medical Association, Chicago, IL, USA. ²Frank Healthcare Advisors, LLC, Gainesville, FL, USA. ³The Omega Concern, LLC, Washington, DC, USA. ⁴CVS/Aetna, Inc., Hartford, CT, USA. ⁵University of Iowa, Iowa City, IA, USA. ⁶Digital Diagnostics, Inc, Coralville, IA, USA. ⁷Johns Hopkins University School of Medicine, Baltimore, MD, USA. ⁸Philips, Cambridge, MA, USA. ⁹University of California San Francisco, Innovation Ventures, San Francisco, CA, USA. ¹⁰University of California San Francisco, San Francisco, CA, USA. ¹¹Arsenal Capital Partners, New York, NY, USA. ¹²South Texas Radiology Group, San Antonio, TX, USA. ¹³UT Health San Antonio, Long School of Medicine, Department of Radiology, San Antonio, TX, USA. ✉email: RichardFrankCPT@gmail.com

foster correct valuation, coverage, and appropriate payment, and support a reasonable range of business models and sites of service.

The taxonomy is applicable to both Category I and Category III CPT codes. Category I CPT codes provide a uniform language for coding medical services and procedures and Category III CPT codes are a set of temporary codes that allow data collection for emerging technologies, services, procedures, and service paradigms. As the “work done by machines” increasingly contributes to the provision and intensity of medical services, the taxonomy is meant to facilitate proper and accurate CPT coding, so the integration of AI into tracking, coverage, and payment systems evolves in a consistent manner. This will help facilitate patient access to the technology and accurate payment to physicians.

In some cases, rather than reducing physician input, the complexity of the output from AI may counterintuitively increase physician work, which currently may not be transparent in coding, coverage, and payment. The taxonomy is a framework intended to engender descriptors which are discrete and differentiable. Ideally, the resultant coding will do more than simply characterize the “work done by the machine” along the spectrum of “assistive,” “augmentative,” and “autonomous,” but will also delineate the machine’s part in the medical service and impact on the subsequent human work in such a way that the coding, coverage, and payment for the services will fairly reflect the contribution of each as shown in Table 1.




Fig. 1 The continuum of AI CPT code descriptors. Appendix S is intended to enable discrete and differentiable code descriptors by accurately characterizing the “work done by machines” relative to the work of the physician or other qualified health care professional. Therefore, these same descriptors may be useful also in regulatory labeling of novel products like “Software as a Medical Device” (SaMD) and in valuation, coverage and payment policy for the associated novel services or procedures, like “Software as a Service” (SaaS).

CURRENT CONTEXT

The AI Taxonomy, published on the AMA website and made effective from January 2022, provides guidance for choices of language for descriptors for AI-related CCAs³. It is intended that all AI-related CCA’s would be drafted and assessed with knowledge of the AI taxonomy henceforth, and descriptors used consistently across like codes.

When applicants submit a CCA for an AI-related code, the code change requestor should include technical details necessary to support the CPT Editorial Panel’s conclusion as to where the technology fits in the CPT-defined hierarchy of “assistive”, “augmentative”, and “autonomous”. Developers should consult the AI Taxonomy when developing descriptors for other highly sophisticated software. For example, Magnetic Resonance Spectroscopy is a highly sophisticated software, not the result of machine learning as is typical of products like the IDx-DR diabetic retinopathy device discussed *supra*, but instead is an “expert system” which renders data understandable in terms of intermediary metabolism associated with pathophysiology¹⁰. Such analysis of data is quite different, and adds more value, than simple image reconstruction. These types of “expert systems” may meet the criteria within the hierarchy of the taxonomy. As another example, deriving an estimate of Fractional Flow Reserve from coronary CT is considered augmentative because the quantitative output is clinically meaningful, ie understood by physicians to be directly applicable in optimizing the patient’s care pathway.

WHY NO DEFINITION OF THE TERM “AI”?

In health care, there is no single product, procedure, nor service for which the term “AI” is sufficient or necessary to describe its intended clinical use or utility; and therefore, the specific term “AI” is not defined in the AI Taxonomy nor elsewhere in the CPT code set. The term “AI” is not intended to encompass nor constrain the full scope of innovations performing “work done by machines”, nor are these anthropomorphic terms intended to convey that the physician is being “replaced”¹¹. Classification of AI medical services and procedures as assistive, augmentative, and autonomous is based on the clinical procedure or service provided to the patient, and the work performed by the machine on behalf of the physician or other QHP.

GRANULARITY FOR EACH OF ASSISTIVE, AUGMENTATIVE, AND AUTONOMOUS

This AI taxonomy is guidance for classifying various applications of AI (expert systems, machine learning, software-based services, etc.) medical services and procedures into one of three categories:

Table 1. The relationship between components of the service and categorization.

Service components	AI Category: Assistive	AI Category: Augmentative	AI Category: Autonomous
Primary Objective	Detects clinically relevant data	Analyzes and/or quantifies data in a clinically meaningful way	Interprets data and independently generates clinically meaningful conclusions
Provides independent diagnosis and/or management decision	No	No	Yes
Analyzes data	No	Yes	Yes
Requires physician or other QHP interpretation and report	Yes	Yes	No
Examples in CPT code set	Computer-aided detection (CAD) imaging (77048, 77049, 77065–77067, 0042 T, 0174 T, 0175 T)	Non-invasive estimate of coronary fractional flow reserve derived from coronary CT (7X005, effective Jan 2024, https://www.ama-assn.org/system/files/september-2022-cpt-summary-panel-actions.pdf)	Retinal imaging (92229)

assistive, augmentative, and autonomous. The workgroup selected these terms to characterize the work done by the machine relative to the work performed by the physician or other QHP. These words should become intuitive to stakeholders describing healthcare AI services in the CPT context.

Assistive services or procedures detect clinically relevant data; analysis and interpretation remain to be done and the physician work is not reduced. The data may be from other acquisition devices and transmitted via a third device. The service or procedure performed by the device “assists” the physician by bringing to their attention something which may or may not impact the diagnosis or management decision but is clinically relevant and might otherwise be missed. The physician does the remainder of the work—analyzes, determines the clinical importance of, and reports their interpretation of the data.

Augmentative services or procedures provide clinically meaningful analyses and/or quantification of data, without interpretation or reporting. The data may be acquired, transmitted, and detected by other devices. The physician may perform additional analyses but interprets the output from the analysis and reports conclusions based on their interpretation of the data. “Clinically meaningful” is in the judgement of a prudent physician that the output of the device will be directly applicable in optimizing their patient’s care pathway.

Autonomous services or procedures include interpretation and clinical conclusions which are independently reported and may lead directly to a recommendation and action, depending on the level of autonomy. A device performing procedures or services autonomously may or may not have acquired, transmitted, or detected the data.

The work performed by the machine for the physician or other QHP is autonomous when the machine automatically interprets data and independently generates clinically meaningful conclusions without concurrent physician or other QHP involvement. An autonomous medical service includes interrogating and analyzing data. The clinically meaningful conclusion may be a characterization of data (e.g., likelihood of pathophysiology) used to establish a diagnosis or to implement a therapeutic intervention. As illustrated in Table 2, there are three levels of autonomous AI medical services and procedures with varying physician or other QHP involvement:

1. The device interprets data, draws conclusions, and offers diagnosis and/or management options. These options are contestable and require action on the part of a physician or other QHP to implement.
2. The device interprets data, draws conclusions, and initiates diagnosis and/or management options with alert/opportunity for override, may require physician or other qualified QHP action to implement.
3. The device interprets data, draws conclusions, and initiates diagnosis and/or management options without requiring action by physician or other QHP to implement. The physician or QHP would have to take initiative to intervene or contest the action taken by the device.

WHY IS “AUTOMATED” NOT IN THE TAXONOMY?

As discussed previously, the CPT Panel members used the term “automated” in 92229 because there was no precedent for autonomous as a descriptor; however, the term “automated” does

Level of autonomy	I	II	III
Interprets	+	+	+
Recommends		+	+
Acts			+

not appropriately capture the “work done by machines” in AI-related services like the retinal exam. The term automated connotes a workflow efficiency obtained by utilizing a machine. Moreover, if a task or analysis is conducted automatically, then it may not be separable, there may have been no clinical judgement applied to the procedure or service for that patient, and, therefore, the clinical necessity may be unclear. For example, 3D volumetrics may be performed automatically on all chest CT images loaded to a PACS system, whereas the clinical utility of 3D volumetrics is only established for a subset of the patients undergoing that image acquisition.

FUTURE APPLICATION

As technology develops and evolves, the application of guidance provided in the AI taxonomy should help to maintain consistency across CPT codes that describe the work performed by machines. Further, it may be necessary to modify this taxonomy framework. Consistency in the depiction of the interaction between physician and machine will be important in valuation even as innovation in this relationship between human and machine is likely to evolve.

Implementation of this taxonomy within the CPT code set will require intentionality by the CPT Editorial Panel, so the precedents are used consistently among similar healthcare AI. The AMA will continue to gather technical expertise to keep pace with technologies expected to improve patient and population outcomes, expand access to care, enhance patient experience, and promote physician satisfaction.

Reporting summary

Further information on research design is available in the Nature Research Reporting Summary linked to this article.

DATA AVAILABILITY

This perspective is based on the authors’ experience in developing CPT Appendix S: Artificial Intelligence Taxonomy for Medical Services and Procedures. Typically, recollections are not appropriate for a data repository. Questions may be directed to the corresponding author.

Received: 17 June 2022; Accepted: 15 November 2022;

Published online: 03 December 2022

REFERENCES

1. Bodenheimer, T. & Sinsky, C. From triple to quadruple aim: care of the patient requires care of the provider. *Ann. Fam. Med.* **12**, 573–576 (2014).
2. Board of Trustees. American Medical Association. Augmented intelligence (AI) in healthcare. Proceedings of the 2019 Annual Meeting of the American Medical Association. 138–158. <https://www.ama-assn.org/system/files/2019-08/a19-bot-reports.pdf>. Accessed April 14, 2022.
3. CPT Editorial Panel. CPT appendix S: AI taxonomy for medical services and procedures. American Medical Association. January 1, 2022. <https://www.ama-assn.org/practice-management/cpt/cpt-appendix-s-ai-taxonomy-medical-services-procedures>. Accessed April 13, 2022.
4. CPT overview and code approval. American Medical Association. <https://www.ama-assn.org/practice-management/cpt/cpt-overview-and-code-approval>. Accessed August 31, 2022.
5. Medicare physician payment schedules. American Medical Association. <https://www.ama-assn.org/practice-management/medicare-medicare-physician-payment-schedules>. Accessed August 31, 2022.
6. RBRVS overview. American Medical Association. <https://www.ama-assn.org/about/rvs-update-committee-ruc/rbrvs-overview>. Accessed August 31, 2022.
7. Abramoff, M. D. et al. A reimbursement framework for artificial intelligence in healthcare. *npj Digit. Med.* **5**, 72 (2022).
8. U.S. Food and Drug Administration. Artificial intelligence and machine learning (AI/ML) enabled medical devices. <https://www.fda.gov/medical-devices/software-medical-device-samd/artificial-intelligence-and-machine-learning-ai/ml-enabled-medical-devices>. Accessed April 4, 2022.

9. U.S. Food and Drug Administration. Software as a medical device (SaMD). <https://www.fda.gov/medical-devices/digital-health-center-excellence/software-medical-device-samd>. Accessed April 14, 2022.
10. Abràmoff, M. D. et al. Collaborative Community on Ophthalmic Imaging Executive Committee and Foundational Principles of Ophthalmic Imaging and Algorithmic Interpretation Working Group. Foundational considerations for artificial intelligence using ophthalmic images. *Ophthalmology* **129**, e14–e32 (2022).
11. Grimm C.M. The danger of anthropomorphic language in robotic AI systems. Tech Stream Published June 18, 2021. <https://www.brookings.edu/techstream/the-danger-of-anthropomorphic-language-in-robotic-ai-systems/> Accessed April 11, 2022.

ACKNOWLEDGEMENTS

The AI Workgroup is grateful for the support of the American Medical Association's Digital Medicine Payment Advisory Group, AMA staff and for the significant work of the CPT Editorial Panel in maintenance of the CPT code set. The thoughts and opinions of the authors are their own and do not necessarily represent AMA policy or the position of the CPT Editorial Panel. Information provided in this paper is intended to complement Appendix S of the Current Procedural Terminology and does not supersede specific coding guidance published in the CPT codebook, which is subject to change and further modification by the CPT Editorial Panel. The AMA along with each author, disclaims responsibility for any errors or omissions and for any consequences attributable to or related to any use, reliance, or interpretation of the content and information in this publication. CPT © 2022 American Medical Association. All rights reserved. CPT is a registered trademark of the AMA.

AUTHOR CONTRIBUTIONS

R.A.F., R.J., J.P., M.D.A., M.X.R., E.S. made substantial contributions to the conception or design of the work, drafted the work or substantively revised it and approved the submitted version. P.B.D., S.M.G., M.R.M., J.E.M. made substantial contributions to the conception or design of the work and approved the submitted version.

COMPETING INTERESTS

R.A.F. reports the following competing interests: Principal, Frank Healthcare Advisors, LLC; member, CPT Editorial Panel. Sits on the Board of and is a minor shareholder in a Canadian innovator, Optina Diagnostics, which has no marketed products. Serves as Entrepreneur in Residence of DigitalDx Ventures I, L.P. Provides expert testimony to FDA, CMS, US Congress, WHO, IAEA, et alia. M.D.A. reports the following competing

interests: Investor, Director, Consultant, Founder, Digital Diagnostics Inc; patents and patent applications assigned to the University of Iowa and Digital Diagnostics that are relevant to the subject matter of this manuscript; Chair Healthcare AI Coalition, Washington DC; member, American Academy of Ophthalmology (AAO) AI Committee; member. J.M. reports the following competing interests: Operating Partner and Chief Medical Information Officer of Arsenal Capital Partners which invests in companies that use AI to enhance clinical products and services. E.S. reports the following competing interests: Chair, RVS Update Committee (RUC). Other authors do not report financial and non-financial competing interests as defined by Nature Research, that might be perceived to influence the interpretation of the paper.

ADDITIONAL INFORMATION

Supplementary information The online version contains supplementary material available at <https://doi.org/10.1038/s41746-022-00723-5>.

Correspondence and requests for materials should be addressed to Richard A. Frank.

Reprints and permission information is available at <http://www.nature.com/reprints>

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2022