

# UCLA

## UCLA Previously Published Works

### Title

Carotid Doppler Measurement Variability in Functional Hemodynamic Monitoring: An Analysis of 17,822 Cardiac Cycles.

### Permalink

<https://escholarship.org/uc/item/734179cw>

### Journal

Critical Care Explorations, 3(6)

### Authors

Kenny, Jon-Émile  
Barjaktarevic, Igor  
Mackenzie, David  
[et al.](#)

### Publication Date

2021-06-01

### DOI

10.1097/CCE.0000000000000439

Peer reviewed

OPEN

# Carotid Doppler Measurement Variability in Functional Hemodynamic Monitoring: An Analysis of 17,822 Cardiac Cycles

**OBJECTIVES:** Carotid Doppler ultrasound is used as a measure of fluid responsiveness, however, assessing change with statistical confidence requires an adequate beat sample size. The coefficient of variation helps quantify the number of cardiac cycles needed to adequately detect change during functional hemodynamic monitoring.

**DESIGN:** Prospective, observational, human model of hemorrhage and resuscitation.

**SETTING:** Human physiology laboratory at Mayo Clinic.

**SUBJECTS:** Healthy volunteers.

**INTERVENTIONS:** Lower body negative pressure.

**MEASUREMENTS AND MAIN RESULTS:** We measured the coefficient of variation of the carotid artery velocity time integral and corrected flow time during significant cardiac preload changes. Seventeen-thousand eight-hundred twenty-two cardiac cycles were analyzed. The median coefficient of variation of the carotid velocity time integral was 8.7% at baseline and 11.9% during lowest-tolerated lower body negative pressure stage. These values were 3.6% and 4.6%, respectively, for the corrected flow time.

**CONCLUSIONS:** The median coefficient of variation values measured in this large dataset indicates that at least 6 cardiac cycles should be averaged before and after an intervention when using the carotid artery as a functional hemodynamic measure.

**KEY WORDS:** carotid Doppler; corrected flow time; fluid responsiveness; measurement variability; velocity time integral

Jon-Émile S. Kenny, MD<sup>1</sup>

Igor Barjaktarevic, MD, PhD<sup>2</sup>

David C. Mackenzie, MDCM<sup>3,4</sup>

Mai Elfarnawany, PhD<sup>5</sup>

Zhen Yang, BMath

Andrew M. Eibl, BComm<sup>1</sup>

Joseph K. Eibl, PhD<sup>1,6</sup>

Chul Ho Kim, PhD<sup>7</sup>

Bruce D. Johnson, PhD<sup>7</sup>

Functional hemodynamic monitoring (FHM) in the ICU has gained traction over the last 2 decades (1). A requisite for FHM is stroke volume (SV) assessment—or some surrogate—following a change in cardiac preload (2). FHM is superior to commonly used static indices such as filling pressure, ejection fraction, or absolute cardiac output when predicting SV response to preload (3).

Quantitative Doppler ultrasonography of the descending aorta, left ventricular outflow tract (LVOT), and peripheral arteries such as the common carotid have been used, successfully, to measure the impact of altered cardiac preload on ventricular output (3). Importantly, however, detecting change in the Doppler waveform with statistical confidence depends upon the change threshold as well as variation introduced by physiology (e.g., respiratory-induced changes in SV) and measurement (e.g., human factors). Accordingly, the coefficient of variation (CV) partly determines beat sample size. For example, if the Doppler

Copyright © 2021 The Authors. Published by Wolters Kluwer Health, Inc. on behalf of the Society of Critical Care Medicine. This is an open-access article distributed under the terms of the Creative Commons Attribution-Non Commercial-No Derivatives License 4.0 (CCBY-NC-ND), where it is permissible to download and share the work provided it is properly cited. The work cannot be changed in any way or used commercially without permission from the journal.

DOI: 10.1097/CCE.0000000000000439

velocity time integral (VTI) CV is 5% and a change threshold of 10% is desired, then measuring 3–4 cardiac cycles before and after preload modification affords statistical confidence; this is the assumption for the LVOT VTI (4).

Given that the flow profile in a peripheral artery like the carotid is generally “blunted parabolic” (i.e., mixed velocities) and not “plug” (i.e., near uniform velocity, as in the ascending aorta) and that changing preload mediates SV variation, assuming a constant CV of 5% may be inaccurate. Using a novel, wireless, wearable Doppler ultrasound patch, we quantified the CV of commonly employed Doppler measures at resting baseline and during altered preload induced by lower body negative pressure (LBNP). Establishing the CV of carotid Doppler metrics establishes the number beats sampled to detect change with statistical confidence.

## METHODS

### Clinical Setting

We recruited 11 healthy, adult volunteers with no known cardiovascular history on no regular cardiovascular medications. Written informed consent was obtained for all subjects, and the study was approved by the Research Ethics Board of the Mayo Clinic, Institutional Review Board number 19-010136.

### Adherent Doppler System

The U.S. Food and Drug Administration (FDA) cleared Doppler ultrasound patch (Flosionics Medical, Sudbury, ON, Canada) is a wireless, wearable, continuous wave 4 MHz ultrasound. It was placed by palpation below the angle of the jaw; when an audible Doppler shift and spectrogram consistent with the common carotid artery were obtained, the patch was adhered to the neck. The duration of systole, in seconds, was used to calculate the corrected flow time (FTc), as described previously (5, 6).

### Lower Body Negative Pressure and Stroke Volume

All subjects underwent a seven-stage protocol in duplicate. Each stage was 5 minutes beginning with resting baseline. LBNP was reduced by 15 mm Hg

per stage down to and including –60 mm Hg and then by 10 mm Hg down to and including –80 mm Hg, as tolerated. The final stage was a 5-minute recovery following release of LBNP to atmospheric pressure. For this investigation, we restricted analysis to data at resting baseline (T1), the lowest stage tolerated for each subject (T2), and return to atmospheric pressure (T3). Accordingly, T1 to T2 diminished cardiac preload while T2 to T3 increased cardiac preload. The Nexfin (Bmeye, Amsterdam, The Netherlands) was applied to subjects in the supine position; Nexfin is a U.S. FDA cleared non-invasive SV monitor that uses “volume clamp” to calculate SV.

### Coefficient of Variation

Doppler cardiac cycles with artifact or during LBNP stage transition were excluded. For each subject, the CV was calculated as the SD divided by the mean of all cardiac cycles within a given stage.

## RESULTS

The average age of the subjects was 29.5 years and 39% were female. The average body mass index was 24.0 kg/m<sup>2</sup>. **Table 1** lists vital signs, SV, carotid Doppler measures, and their change during decreased (T1 to T2) and increased (T2 to T3) cardiac preload. A representative example of Doppler variation and the CV values for all subjects for each stage are shown in **Figure 1**.

## DISCUSSION

In total, 17,822 cardiac cycles were analyzed to measure the CV of commonly employed carotid Doppler metrics. As the average carotid VTI<sub>max</sub> CV was greater than 5% and changed significantly with preload, simply averaging 3–4 cardiac cycles before and after an intervention may not be statistically adequate. For example, given a VTI<sub>max</sub> CV of 9%, a change threshold of 15% (5), a type I error of 0.05, and power 0.80 (7), roughly six cardiac cycles should be obtained before and after preload modification to confidently capture change. Importantly, the number of beats sampled varies as a function of CV and the threshold to detect. For example, detecting a 3% change in FTc (6) with a CV of 4.5% demands roughly 36 beats sampled before and after an intervention.

**TABLE 1.**  
Average Hemodynamic Changes Across Three Stages for All Protocols

Measure	Baseline (T1)	↓Preload (T2)		↑Preload (T3)	
	Mean (sd)	Mean (sd)	% Δ From T1	Mean (sd)	% Δ From T2
Heart rate (beats/min)	63.2 (6.9)	109.6 (14.2)	+73.4% <sup>a</sup>	60.4 (7.9)	-44.9% <sup>a</sup>
Systolic blood pressure (mm Hg)	127.3 (15.6)	120.9 (16.6)	-5.0% <sup>a</sup>	136.9 (15.9)	+13.2% <sup>a</sup>
Diastolic blood pressure (mm Hg)	79.0 (8.5)	86.9 (12.3)	+10.0% <sup>a</sup>	85.3 (8.8)	-1.8% <sup>a</sup>
Mean arterial pressure (mm Hg)	97.2 (10.8)	98.8 (13.8)	+1.6% <sup>a</sup>	105.8 (12.3)	+7.1% <sup>a</sup>
Respiratory rate (breaths/min)	16.2 (5.1)	18.8 (6.8)	+16.0% <sup>a</sup>	16.4 (5.6)	-12.3% <sup>a</sup>
Stroke volume (mL)	96.6 (12.0)	59.1 (8.2)	-38.8% <sup>a</sup>	98.0 (12.4)	+65.7% <sup>a</sup>
Velocity time integral (cm)	35.5 (5.9)	15.8 (4.4)	-55.5% <sup>a</sup>	36.6 (7.3)	+132.0% <sup>a</sup>
Corrected flow time (ms)	319.5 (22.8)	270.0 (19.3)	-15.5% <sup>a</sup>	320.6 (23.7)	+18.7% <sup>a</sup>
Cardiac cycles	5,610	7,702		4,510	

<sup>a</sup> $p \leq 0.0001$ .


Importantly, the aforementioned pertains to assessing carotid Doppler before and after a relatively prolonged hemodynamic stimulus (e.g., passive leg raising [8], preload administration) rather than when the respiratory cycle itself is the stimulus used to infer functional state (e.g., peak velocity variation) (9). In general, respiratory cycle-induced FHM is limited to patients completely passive with the mechanical ventilator, receiving relatively large tidal volumes, free from left ventricular dysfunction, and in sinus rhythm (3). Nevertheless, even though our subjects were spontaneously breathing, when cardiac preload was lowest (i.e., T2), variation was greatest for both VTI and FTc.

In a previous investigation, we found that carotid VTI slightly outperformed FTc at detecting a 10% SV change in healthy volunteers (5). In that study, data were averaged over 10-second windows, capturing roughly 10–15 cardiac cycles before and after cardiac preload change. Based on the CV values reported herein, this sample size could partly explain the smaller false negative rate of VTI relative to FTc (i.e., 5% vs 10%, respectively) for detecting increased SV.

Notably, our current findings approximate inherent, physiologic variation devoid of human factors because the Doppler ultrasound patch is adherent. That is, the

presented CVs are likely “best-case” given that manual manipulation of an ultrasound probe introduces additional error. Angle error and sample volume selection can increase the CV by four- to five-fold even in expert technicians (10). Errors introduced by manual manipulation of the ultrasound probe are important in peripheral arteries like the carotid for a few reasons. First, at 60° insonation, a 5° error engenders 15% velocity error (10). Second, the carotid typically has blunted parabolic flow; that is, a greater velocity gradient across the vessel lumen that amplifies sampling error (11). Finally, compressible neck tissue might accentuate geometric spectral broadening by changing aperture-to-depth ratio (12). These limitations are in contradistinction to the LVOT where the angle of insonation is low, the velocity profile is uniform (i.e., “plug”), and the depth of insonation is more constant; thus, 5% CV for the LVOT VTI is understandable.

The relatively small number of healthy volunteers without cardiovascular disease are limitations of this research correspondence; it was performed as a convenience sample in a physiology laboratory. However, the large number of cardiac cycles captured, absent human measurement error, are clear advantages; the reported intrinsic variation of the carotid Doppler pulse informs beat sample size calculations for recently initiated studies in patients.



**Figure 1.** Inherent variation of carotid Doppler. **A**, Representative spectrogram across a respiratory cycle. **B**, Corresponding velocity time integral (VTI) across the respiratory cycle, each *bar* represents a single beat VTI, *horizontal dashed line* shows average VTI across the nine pictured beats. **C**, Coefficient of variation (CV) of the maximal VTI across two preload changes during lower body negative pressure. **D**, CV for the corrected flow time (FTc) for the same changes in preload. In **(C)** and **(D)**, each *point* represents the CV from a single subject during the indicated stage; each subject performed the protocol in duplicate. Std = sp of the CV about its mean.

In conclusion, our results are anticipated based on the underlying physiologic principles of FHM (1). Preload dependent subjects are expected to increase SV variation with changes in venous return. While

this has physiologic implications, the *statistical* consequences might be ignored; increased Doppler variation in a peripheral artery demands a larger sample size to detect change with confidence.

## ACKNOWLEDGMENTS

We wish to acknowledge Matt Myers, Briana Ziegler, Alex Carlson, and Brad Cierzan for their technical assistance, recruitment, scheduling, and data management

- 1 Health Sciences North Research Institute, Sudbury, ON, Canada.
- 2 Division of Pulmonary and Critical Care, Department of Medicine, David Geffen School of Medicine at UCLA, Los Angeles, CA.
- 3 Department of Emergency Medicine, Maine Medical Center, Portland, ME.
- 4 Tufts University School of Medicine, Boston, MA.
- 5 Flosionics Medical, Toronto, ON, Canada.
- 6 Northern Ontario School of Medicine, Sudbury, ON, Canada.
- 7 Human Integrative and Environmental Physiology Laboratory, Department of Cardiovascular Diseases, Mayo Clinic, Rochester, MN.

This work was supported by the Office of the Assistant Secretary of Defense for Health Affairs through the Congressionally Directed Medical Research Program under Award Number W81XWH1910591.

Dr. Kenny, Dr. Elfarnawany, Ms. Yang, Mr. Eibl, and Dr. Eibl work for Flosionics Medical (Sudbury, ON, Canada), a start-up building the wearable Doppler ultrasound. Dr. Barjaktarevic reports consulting fees for GE Healthcare. The remaining authors have disclosed that they do not have any potential conflicts of interest.

Opinions, interpretations, conclusions, and recommendations are those of the authors and are not necessarily endorsed by the Department of Defense.

For information regarding this article, E-mail: [jon-emile@heart-lung.org](mailto:jon-emile@heart-lung.org)

## REFERENCES

1. Michard F, Teboul JL: Using heart-lung interactions to assess fluid responsiveness during mechanical ventilation. *Crit Care* 2000; 4:282–289
2. Monnet X, Marik P, Teboul JL: Passive leg raising for predicting fluid responsiveness: A systematic review and meta-analysis. *Intensive Care Med* 2016; 42:1935–1947
3. Monnet X, Marik PE, Teboul JL: Prediction of fluid responsiveness: An update. *Ann Intensive Care* 2016; 6:111
4. Jozwiak M, Mercado P, Teboul JL, et al: What is the lowest change in cardiac output that transthoracic echocardiography can detect? *Crit Care* 2019; 23:116
5. Kenny J-ÉS, Barjaktarevic I, Eibl AM, et al: A carotid Doppler patch accurately tracks stroke volume changes during a preload-modifying maneuver in healthy volunteers. *Crit Care Explor* 2020; 2:e0072
6. Kenny JS, Barjaktarevic I, Mackenzie DC, et al: Diagnostic characteristics of 11 formulae for calculating corrected flow time as measured by a wearable Doppler patch. *Intensive Care Med Exp* 2020; 8:54
7. Van Belle G: *Statistical Rules of Thumb*. Hoboken, NJ, John Wiley & Sons, 2011
8. Marik PE, Levitov A, Young A, et al: The use of bioactance and carotid Doppler to determine volume responsiveness and blood flow redistribution following passive leg raising in hemodynamically unstable patients. *Chest* 2013; 143:364–370
9. Ibarra-Estrada MÁ, López-Pulgarín JA, Mijangos-Méndez JC, et al: Respiratory variation in carotid peak systolic velocity predicts volume responsiveness in mechanically ventilated patients with septic shock: A prospective cohort study. *Crit Ultrasound J* 2015; 7:12
10. Lui EY, Steinman AH, Cobbold RS, et al: Human factors as a source of error in peak Doppler velocity measurement. *J Vasc Surg* 2005; 42:972–979
11. Gill RW: Measurement of blood flow by ultrasound: Accuracy and sources of error. *Ultrasound Med Biol* 1985; 11:625–641
12. Hoskins PR: Estimation of blood velocity, volumetric flow and wall shear rate using Doppler ultrasound. *Ultrasound* 2011; 19:120–129