UCLA

UCLA Previously Published Works

Title

Digital microfluidics for automated hanging drop cell spheroid culture.

Permalink

<https://escholarship.org/uc/item/74v4329w>

Journal

Journal of laboratory automation, 20(3)

ISSN

2211-0682

Authors

Aijian, Andrew P
Garrell, Robin L

Publication Date

2015-06-01

DOI

10.1177/2211068214562002

Peer reviewed

1Digital Microfluidics for Automated Hanging Drop Cell Spheroid Culture

2A.P. Aijian,^a and R. L. Garrell^{a,b,c,*}

3^a Department of Bioengineering, University of California, Los Angeles, CA 90095-1600,

4USA

5^b Department of Chemistry and Biochemistry, University of California, Los Angeles, CA

690095-1569, USA

7^c California NanoSystems Institute(CNSI), University of California, Los Angeles, CA 90095-

87227, USA.

9* Corresponding Author: E-mail: garrell@chem.ucla.edu; Tel: +1 310-825-4383

10

11Keywords: Digital Microfluidics, Spheroids, Electrowetting, 3D Cell Culture, Cell-based

12Assays

13

14

15

16

17

18

19Abstract

20Cell spheroids are multicellular aggregates, grown in-vitro, that mimic the three-
21dimensional morphology of physiological tissues. While there are numerous benefits
22to using spheroids in cell-based assays, the adoption of spheroids in routine
23biomedical research has been limited, in part, by the tedious workflow associated
24with spheroid formation and analysis. Here we describe a digital microfluidic
25platform that has been developed to automate liquid handling protocols for the
26formation, maintenance, and analysis of multicellular spheroids in hanging drop
27culture. We show that droplets of liquid can be added to and extracted from through-
28holes, or ‘wells,’ fabricated in the bottom plate of a digital microfluidic device,
29enabling the formation and assaying of hanging drops. Using this digital microfluidic
30platform, spheroids of mouse mesenchymal stem cells were formed and maintained
31in-situ for 72 h, exhibiting good viability (>90%) and size uniformity (%CV <10%,
32intra-experiment, <20% inter-experiment). A proof-of-principle drug-screen was
33performed on human colorectal adenocarcinoma spheroids to demonstrate the ability
34to recapitulate physiologically relevant phenomena such as insulin-induced drug
35resistance. With automatable and flexible liquid handling, and a wide range of in-situ
36sample preparation and analysis capabilities, the digital microfluidic platform
37provides a viable tool for automating cell spheroid culture and analysis.

40Introduction

41Cell spheroids are multicellular compact aggregates, grown in-vitro, that have a three-
42dimensional (3D), spherical morphology. Unlike cells grown in two-dimensional
43adherent monolayers, cells grown in three dimensions possess a high degree of
44intercellular interactions and exhibit relatively complex nutrient and metabolic mass
45transport gradients. These lead to cellular heterogeneity within the 3D aggregate and to
46gene and protein expression patterns that more closely mimic in-vivo tissues.¹⁻⁶ The
47differential expression profiles result in significant differences in cellular behaviors
48such as drug sensitivity, differentiation capacity, malignancy, function, and viability.
49For example, hepatocellular carcinoma cells grown as spheroids exhibit more
50physiologically relevant levels of cytochrome P450 activity and albumin secretion
51compared to cells grown in monolayers.⁷ Another example is mammary epithelial
52cells, which exhibit basement membrane-induced apoptosis resistance when grown in
53three dimensions, but are susceptible to apoptosis in monolayer culture.⁸ Because of
54their enhanced physiological relevance compared to monolayer cell cultures, cell
55spheroids can provide more accurate models for cell-based assays and screens.
56Improved tissue and disease models not only enhance basic research, but can be
57extremely valuable in commercial research, particularly in the pharmaceutical
58industry, where the failure rates for drug candidates entering clinical trials are typically
59>80%.⁹⁻¹³

60

61Despite the known advantages of 3D cell cultures, their use in cell-based assays and
62screens has been limited. It is estimated that <30% of cancer and molecular biologists

utilize 3D cell culture and that <20% of drug leads generated by the pharmaceutical industry are developed through cell-based phenotypic assays.^{14, 15} A major reason for the relatively low adoption of 3D cell models is the limited number of user-friendly, flexible, and automated methods for performing spheroid culture and analysis.^{16, 17} While a variety of technologies and methods are available for culturing 3D micro-tissues, each approach has limitations that make it unsuitable for routine assays and screens.¹⁸ Non-automated spheroid culture methods, such as the manual hanging-drop technique or the use of micro-molds, are inexpensive and relatively simple, but interrogating individual spheroids requires manual transfer to a separate vessel. Additionally, the non-automated methods often require a significant amount of manual sample handling, which can be tedious, time-consuming, and prone to variability and error.

75

A number of microfluidic techniques for spheroid culture have been developed that can enable high-throughput and massively parallel spheroid formation, but which do not support the interrogation of spheroids individually on the same device.¹⁹⁻²² Rotary vessels and spinner flasks can also be used to generate a large number of spheroids, but provide limited control over spheroid size and also do not allow for in-situ assaying of individual spheroids. Alternatively, specially engineered well plates, such as those capable of supporting hanging drop culture²³ or those with non-adhesive surfaces designed to induce cell aggregation,²⁴ are compatible with robotic liquid handling equipment, enabling automation and high-throughput processing. However, the ability to automate spheroid culture and analysis using these methods requires

86access to robotic liquid handling equipment, which can be prohibitively expensive to
87acquire, operate, and maintain for many research labs, particularly those in academic
88settings where the emphasis may not be on high-throughput experiments.
89Additionally, functionalities necessary for spheroid culture and analysis, such as in-
90situ microscopy, mixing, and temperature control, require additional, often expensive,
91hardware to be added to the liquid handling instrument. Thus, there is a need for a
92spheroid culture and analysis technology that provides some of the advantages of
93automation, in a platform that is more accessible than current automated methods. We
94propose that digital microfluidics (D μ F), a flexible and precise microfluidic liquid
95handling technology, can be used to automate cell spheroid culture and analysis as
96well as provide some unique benefits over existing automated techniques.

97

98Digital microfluidics is a type of microfluidic platform that enables the manipulation
99of discrete droplets of liquid in either an air or liquid ambient medium through the
100spatially and temporally-controlled application of electric fields.²⁵⁻²⁷ The application
101of an electric potential across the solid-liquid contact line generates a combination of
102electrostatic and/or dielectrophoretic forces, depending on the frequency of the applied
103field and the relative permittivities of the liquid and ambient phases. Sequentially
104applying an electric potential to an array of planar electrodes can enable the translation
105of droplets across the array,²⁸ and can also be used to split, merge and mix droplets.

106

107Here we present a digital microfluidic device that enables the formation of hanging
108drops to allow in-situ cell spheroid culture. With the ability to automate liquid

109handling, and with a wide range of in-situ bioanalytical techniques developed for the
110D μ F platform, D μ F can ultimately provide a powerful tool for automation of spheroid-
111based assays and screens.

112

113**Materials and Methods**

114Bone marrow-derived mouse mesenchymal stem cells (ATCC® CRL-12424™) were
115generously donated by Prof. Tatiana Segura (UCLA). HT-29, human colorectal
116adenocarcinoma cells (ATCC® HTB-38™) and BJ, human foreskin fibroblasts (ATCC®
117CRL-2522™) were purchased from ATCC. Leibovitz L-15 cell culture medium,
118Dulbecco's Modified Eagle Medium (DMEM), penicillin-streptomycin (P/S) solution
119(10,000 U Pen., 10 mg Strep./mL), L-glutamine, fetal bovine serum (FBS), and the
120LIVE/DEAD® Viability/Cytotoxicity Kit for mammalian cells were obtained from
121Life Technologies (Carlsbad, CA). Pluronic® F-68 was purchased from Sigma (St.
122Louis, MO). Cytop® (CTL-809M) and CYSOLV-180 were purchased from Bellex
123International Corporation (Wilmington, DE). Human recombinant insulin was
124purchased from R&D Systems, Inc. (Minneapolis, MN) and Irinotecan HCL was
125purchased from BIOTANG, Inc. (Lexington, MA).

126

127Device fabrication was conducted in the CNSI Integrated System Nanofabrication
128Cleanroom at UCLA. Briefly, water white glass substrates (LabScientific, Inc. CAT#
1297787) were coated with 1100 Å indium tin oxide (ITO) via sputtering and were
130patterned with electrodes via photolithography and reactive ion etching. For this
131work, the substrate with the patterned electrode array was used as the top-plate and an

un-patterned ITO-coated slide was used as the bottom-plate. Prior to coating with the dielectric, through-holes were manually drilled into specific locations on the bottom-plate using a benchtop drill press and diamond-coated drill bits. Through-holes were also drilled into the footprint of the reservoir electrodes in the top-plate to provide a world-to-chip interface. The top-plates were then coated with 3–4 μm of dielectric polymer parylene-C (Specialty Coating Systems) by vapor deposition. The top and bottom-plates were rendered hydrophobic by spin coating $\sim 300\text{--}400$ nm of Cytop on each. Prior to use, the walls of the wells in the bottom-plate were gently scraped with a diamond-coated drill bit to remove the Cytop coating and expose the hydrophilic glass surface. A schematic of a D μ F device assembly is shown in Figure 1.

142

All microfluidic liquid handling was performed using a custom LabView application to control electrode actuation. Liquid handling was performed at 100–115 V_{pp} AC and at a frequency of 18.5 kHz. Analysis of hanging droplet liquid exchange was performed by measuring the absorption of a standardized solution of brilliant blue dye prepared in water before and after liquid exchange cycles using a Thermo Scientific NanoDrop 2000c UV-Vis spectrophotometer. The liquid exchange process is described in more detail below.

150

For the preparation of cell solutions for use on the D μ F device, cells were thawed and seeded in polystyrene dishes in growth medium (DMEM, 4 mM L-glutamine, 10% FBS, 100 U/mL P/S solution). Cells were grown to $\sim 80\%$ confluency, trypsinized, and re-suspended in spheroid growth medium (Leibovitz L-15 medium, 4 mM L-

155glutamine, 7.5% FBS, 100 U/mL P/S, 0.04% Pluronic® F-68) at cell densities ranging
156from $\sim 7.5 \times 10^5$ – 1×10^6 cells/mL for culture on the device.

157

158Detailed schematics of the experimental setup, as used in this work, are shown in
159Figure S1 in the Electronic Supplemental Information. Prior to use, the devices were
160sterilized by dipping them in a 70% aqueous ethanol solution and gently drying with
161compressed air. For device operation, the bottom-plate was placed on an aluminum
162holding plate that contained a milled recess below the location of the wells to allow
163hanging drops to form beneath the device. The bottom-plate of the device was sealed
164to the aluminium plate using silicone grease (Dow Corning High Vacuum Grease).
165The bottom of the recess was enclosed with a glass slide to prevent exposure of the
166hanging drops to the laboratory environment during drop formation. To minimize
167evaporation, 1.5 μ L of 10 cst silicone oil was pre-loaded into each well prior to the
168formation of hanging drops. Additionally, a small amount of water was placed in the
169enclosed recess to create a humidified environment. The top-plate was secured to
170another aluminum plate and was interfaced with the bottom-plate such that particular
171electrodes in the top-plate aligned with the location of the wells in the bottom-plate.
172The two plates were separated by a custom designed adhesive silicone spacer (Grace
173Biolabs, Bend, OR) to create a gap height of 300 μ m and were secured using binder
174clips. Drops of cell-suspension were added to the reservoir electrodes via through-
175holes drilled into the top-plate.

176

177Hanging drop and spheroid formation were achieved by dispensing droplets of cell
178suspension from the reservoir and moving the droplets to the location of a well. Upon
179contact with the hydrophilic wall of the well, droplets were pulled into the well via
180capillary forces. Addition of multiple droplets to a well resulted in the formation of a
181hanging drop. Exchange of the medium within the hanging drop was achieved by
182performing the following sequence of steps one or more times per well: (1) delivering
183a drop of fresh medium to a well, (2) using electrowetting actuation to repeatedly pull
184out and release a liquid finger from the well to facilitate mixing of the liquid in the
185well, (3) extracting a drop from the well of twice the volume of the amount initially
186delivered, and (4) adding another drop of fresh medium to the well. Devices were
187kept in an incubator at 37 °C and 95% relative humidity at all times except during
188liquid handling.

189

190For confocal imaging, spheroids were stained with fluorescent markers by incubation
191in imaging medium for 2 h at 4 °C followed by 30 min at 37 °C to ensure enhanced
192staining of the interior of the spheroid.²⁹ The imaging medium consisted of 2 µM
193calcein-am and 4 µM ethidium homodimer-1 (Life Technologies, LIVE/DEAD®
194Viability/Cytotoxicity Kit, for mammalian cells) in Hank's Balanced Salt Solution
195(HBSS, Life Technologies) supplemented with 1 mg/mL ascorbic acid, 25 mM HEPES
196buffer solution – pH 7, 100 U/mL P/S, 100 µM non-essential amino acids, and 4 mM
197L-glutamine; 1N NaOH was used to adjust the pH to 7.2. Following staining with
198imaging medium, the spheroids were washed with HBSS containing 1 mg/mL ascorbic
199acid. A custom PDMS imaging chamber was secured beneath the bottom plate such

that flooding the wells with the HBSS solution caused the hanging drops to detach from the device into individual wells within the PDMS chamber, enabling the spheroids to be imaged directly from the device. Figure S1 in the Electronic Supplemental Information shows a schematic of the device interfaced with the PDMS imaging chamber.

Confocal imaging was performed using a Leica TCS SP2 confocal microscope. Spheroid images were constructed by creating a maximum projection of multiple z-plane sections spaced 3–7 μm apart. The proportion of living cells within a spheroid was estimated by counting the number of live (green) and dead (red) cells in five different, equally spaced z-planes throughout the spheroid. ImageJ was used for all image analysis, which included hanging drop volume, as well as spheroid viability, diameter, and aspect ratio measurements.

Results and Discussion

Device design and operation

To enable hanging drop formation, through-holes, or ‘wells,’ were fabricated into strategic locations in the bottom-plate of the device. The schematic in Figure 1 shows the basic principle of μF hanging drop formation along with typical device dimensions. Hanging drops are formed when droplets of liquid are delivered to the location of a well and, upon making contact with the hydrophilic walls of the well, get pulled into the well spontaneously via capillary forces (Figure 2a). Adding multiple

222drops to a well results in the formation of a curved liquid-air interface that protrudes
223beneath the bottom-plate, similar to a hanging drop (Figure 2b). To ensure the
224formation of a hanging drop, the wells were designed such that the Bond number (Bo ,
225a dimensionless parameter describing the ratio of gravitational to surface tension
226forces) of the system is greater than ~ 0.3 , which is within the range where
227gravitational forces begin to influence the shape of the meniscus.³⁰⁻³² A $Bo \geq 0.3$
228requires a well diameter of ≥ 2.4 mm.

229

230To simplify device fabrication protocols, the top plate contained the actuating
231electrodes and the bottom-plate contained the ground electrode. While either the
232actuating or ground plate can be modified with through-holes and used as the bottom-
233plate to support hanging drop formation, we found that incorporating the wells into
234the plate containing the actuating electrodes was more difficult because the holes
235needed to be drilled precisely within the footprint of an electrode, which occasionally
236resulted in damaging the electrode. Additionally, decoupling the wells and actuating
237electrodes allows for the actuating top-plate to be removed and replaced in the case of
238dielectric breakdown, without disrupting the hanging drops in the wells in the bottom-
239plate.

240

241Dielectric breakdown occurs when the electric field across the dielectric layer exceeds
242the dielectric strength of the material, resulting in localized, physical destruction of
243the dielectric layer. Dielectric breakdown typically occurs on the actuating plate,
244where charges within the drop and electric field lines outside the drop concentrate

near the droplet edge closest to the actuated electrode, creating a region of locally elevated electric field strength.³³⁻³⁵ If the degree of dielectric breakdown is minor, droplets can still be transported normally across the location of the breakdown. Significant dielectric breakdown can cause electrolysis of aqueous solutions as a result of to current flow into the drop, and can also damage critical electronic connections on the device, thereby impeding droplet movement (“pinning”). For the top-plates used in this work, the entire spheroid culture process, which required ~800-1200 total electrode actuation steps for the culture of 6-8 spheroids, could typically be achieved without the occurrence of dielectric breakdown. Approximately one out of every four devices showed evidence of dielectric breakdown at some point during the culture protocol, typically during the 48 h medium exchange process, i.e., after the hanging drops had been formed and all the cell handling had been completed. Because the spheroids were maintained in hanging drops beneath the bottom plate and were relatively far from either of the interior surfaces of the top- and bottom-plates, dielectric breakdown did not disrupt or affect the spheroids within the hanging drops.

260

To allow visualization of droplet handling, the actuating electrodes in the top plate were made from a transparent conductive material, indium tin oxide (ITO). Videos of liquid handling and hanging drop formation are provided in the Electronic Supplemental Information.

265

The wells in the bottom plate contain a tapered opening on the top side to aid in the insertion of drops into the well. Droplets that reach the edge of a well can experience

268cathotaxis, or pinning at the intersection of two interfaces, due to both the change in
269geometry at the well edge and the difference in surface properties between the
270hydrophobic surface of the bottom plate and the hydrophilic interior of the well walls.
271By tapering the inside walls of the well to form an acute angle with the surface of the
272bottom plate, as opposed to a right angle formed by a cylindrical through-hole, the
273pinning effect on a drop of liquid at the edge of the well is reduced, facilitating droplet
274insertion into the well.

275

276The volume of a hanging drop is determined by the number of drops dispensed from a
277reservoir and added to a well. Thus, the volume and reproducibility of droplet
278dispensing from the reservoirs are critical to the volume and reproducibility of the
279hanging drops. To determine the variation in dispensed drop volumes, we used image
280analysis to measure the volumes of 144 drops of an aqueous surfactant solution
281(0.04% Pluronic® F-68) dispensed from different reservoirs across three different
282devices using a programmed dispensing sequence. An aqueous surfactant solution was
283used so that the surface tension of the liquid and, consequently, the volume of the
284dispensed drops, would be similar to that of the growth medium solution subsequently
285used in the cell culture experiments, which also contained 0.04% Pluronic® F-68. The
286dispensed droplet volume was determined by measuring the area of the drop in contact
287with the top-plate using device features of known dimensions as a scale, and
288multiplying by the known distance of the inter-plate gap. While in actuality the
289sidewalls of the droplet are curved, to simplify the volume measurements we used the
290straight-wall, cylindrical approximation to calculate droplet volumes, which is a

reasonable approximation considering the relatively small droplet aspect ratio on our devices ($h/w \sim 0.15$) and the contact angle of approximately 100° .³⁶ The distribution of droplet volume measurements is shown in Figure 3. The average volume of a single dispensed drop was $1.75 \pm 0.13 \mu\text{L}$ (7.7% CV; %CV = coefficient of variation). This degree of droplet volume variation is consistent with reproducibility values from other electrowetting devices that do not utilize capacitance metering to control dispensing volumes. Droplet dispensing reproducibility can be improved to low single-digit %CV by employing capacitance metering methods or by optimizing reservoir and dispensing electrode design.³⁷⁻³⁹ The variation in the volumes of hanging drops was also determined (Figure 4). Hanging drops formed from 4, 5, and 6 dispensed drops had average volumes of $7.4 \pm 0.5 \mu\text{L}$, $8.8 \pm 0.8 \mu\text{L}$, and $10.2 \pm 0.5 \mu\text{L}$, respectively, corresponding to %CV range of 5–9%. This volume range was chosen because, for the devices used in this work, at least 4 drops are required to fill a well and form the curved surface necessary for cell aggregation.

It should be noted that the volume and reproducibility data shown here represent results from a particular dispensing sequence and device arrangement (i.e., gap height = $300 \mu\text{m}$). Various droplet volumes can be dispensed on a D μ F device by simply altering the gap height and/or changing the dispensing sequence. The devices used in this work support the formation of hanging drops up to $\sim 55 \mu\text{L}$ before the drops detach from the well due to their weight. We observed that hanging drops of larger volumes can be supported by varying the thickness of the bottom plate, the well geometry, or

the surface tension of the liquid comprising the drop (determined experimentally, data not shown).

315

Cell spheroids require ~50% medium exchange every 48 h for optimal growth.^{40, 41}

Thus, to enable long term hanging drop spheroid culture, protocols for in situ medium exchange using digital microfluidic liquid handling were developed. Medium exchange requires extracting the spent medium from a hanging drop and replacing the spent medium with fresh medium. Liquid can be extracted from a hanging drop by using the electrodes adjacent to a well to pull out a drop of liquid. Repeating the process of extracting and adding drops of medium to a well, as described in the Materials and Methods section, results in the exchange of the medium within the well. Assuming the hanging drop initially contains the volume of four dispensed drops, the medium exchange protocol theoretically allows for exchange of 40% and 64% of the initial drop volume after one and two exchange cycles, respectively (according to the dilution rate for this particular exchange protocol: $C = 0.6^n$, where C = the concentration of spent medium in the drop, and n = the number of exchange cycles). The video “Hanging Drop Liquid Exchange” in the Electronic Supplemental Information shows two cycles of the liquid exchange protocol. Using a hanging drop of a standardized brilliant blue dye solution to mimic spent medium and DI water to represent fresh medium, we assessed the degree of exchange by measuring the change in dye concentration of the hanging drop after successive exchange cycles by visible spectrophotometry. Figure 5 shows that the dye concentration calculated from UV-Vis absorption are consistent with the theoretical predictions, indicating that >50%

336medium exchange can be achieved with one or more exchange cycles. These data also
337indicate that D μ F can provide precise control over the composition of the hanging
338drop, which is critical for performing cell-based assays and screens.

339

340*Cell Spheroid Culture*

341After establishing the ability to form a hanging drop and conduct medium exchange, a
342complete cell spheroid culture protocol was performed to demonstrate proof-of-
343principle for fully automated D μ F cell spheroid culture. Droplets of mouse MSC
344suspension in growth medium were delivered to wells to form hanging drops of ~7-10
345 μ L (~5250-7500 cells/drop). Pluronic® F-68 was included in the growth medium to
346minimize the adsorption of proteins to the hydrophobic surface of the device, which
347can impede the movement of proteinaceous solutions.⁴² At 0.04%, Pluronic® F-68 is
348known to be non-cytotoxic.⁴³ Leibovitz L-15 medium was used for spheroid culture
349because it is buffered by phosphates and free-base amino acids instead of sodium
350bicarbonate. This medium allows cell growth in the absence of a controlled CO₂
351atmosphere; our current digital microfluidic setup is operated outside of an incubator
352at ambient atmospheric conditions. During liquid handling, the microfluidic apparatus
353was kept at ~37 °C by placing a thin-film polyimide heater in contact with the
354aluminium device holder. After liquid handling, devices were transferred to an
355incubator at 37 °C and relative humidity of 95%. To prevent fluctuations in
356atmospheric conditions between the liquid-handling and incubation periods, the
357incubator was also maintained at ambient atmosphere (i.e., without CO₂ control).

358

359Medium exchange was performed once daily. During culture, the spheroid sits at the
360bottom of the hanging drop, which is ~ 1.8 mm below the top opening of the well
361(assuming a $7.4 \mu\text{L}$ drop in a 2.5-mm diameter well). Because liquid from the drop is
362extracted from the top opening of the well and medium exchange never requires
363extraction of more than 25% of the initial hanging drop volume, the spheroid remains
364settled within the hanging drop throughout the medium exchange protocol and does
365not get extracted from the well.

366

367Figure 6a shows confocal micrographs of typical spheroids of mouse mesenchymal
368stem cells cultured on the D μ F device over the course of 72 h using automated sample
369handling protocols. The spheroids were stained with calcein-AM and ethidium
370homodimer-1 to indicate living (green) and dead (red) cells, respectively. Counting
371the number of living and dead cells at various z-planes within the spheroid indicated
372that the spheroids exhibited $>90\%$ cell viability. The spheroid diameter was measured
373at 24, 48 and 72 h (Figure 6b) following hanging drop formation using a USB-
374microscope (Dino-Light AD4013TL). A seeding density of 7.5×10^5 cells/mL produced
375spheroids of up to $\sim 400 \mu\text{m}$ after 72 h in culture. The size and viability of the
376spheroids generated on the D μ F platform are consistent with those obtained through
377other hanging drop techniques over the same timeframe using similar cell number
378conditions.²³ Intra-device spheroid diameter variation was $\sim 8\%$; this is comparable to
379other hanging drop techniques, which exhibit a %CV range of $\sim 3\%$ (for robotic liquid
380handlers) to 15% (for manual methods), and is superior to spheroid generation on non-
381adhesive flat-bottom well plates, which show spheroid diameter variation of up to 40–

38260%.^{44, 45} Because the cell density is the same for each hanging drop, the intra-
383experiment spheroid diameter variation is attributable to the variation in the volumes
384of the hanging drops. The inter-device variation in spheroid diameter (i.e., for
385spheroids grown on different devices) was 14%, 18%, and 18% for spheroids at 24,
38648, and 72 h, respectively (Figure 6b). The relatively larger inter-experiment variation
387compared to the intra-experiment results is likely due to variations in cell densities
388between the different experiments. For this work, cell suspensions of $\sim 7.5 \times 10^5$ cells/mL
389were prepared based on hemocytometer measurements, which can exhibit variability
390of 10-40% depending on cell concentration.^{46, 47} More precise cell-density
391measurement techniques, which can achieve a %CV of $< 3\%$,⁴⁸ would reduce the inter-
392experiment spheroid diameter variability. Figure 6c shows the distribution in spheroid
393aspect ratio (ratio of spheroid major axis to minor axis) for 77 spheroids of various cell
394numbers and types. The average aspect ratio for the spheroids cultured on the D μ F
395platform was 1.15 ± 0.09 , corresponding to a CV of $\sim 8\%$. The spheroid aspect ratio
396was measured after at least 48 h in culture to allow spheroid compaction to occur.
397Table 1 summarizes performance characteristics for spheroid culture conducted on a
398D μ F device.

399

400With the ability to initiate and maintain viable spheroids in culture as well as freely
401add, mix, and extract liquid from a hanging drop, the D μ F platform enables
402automation of spheroid-based assays and screens. To demonstrate this capability, we
403performed a proof-of-principle spheroid-based drug screen, using D μ F to examine the
404impact of insulin exposure on the chemosensitivity of colon cancer cells to treatment

405with the chemotherapeutic agent irinotecan. Insulin has been shown to cause
406resistance to chemotherapy in certain colon cancer cell lines via activation of the
407PI3K/Akt pathway.^{49, 50} For the drug screening assay, hanging drops of HT-29 human
408colon adenocarcinoma cells were initiated and maintained on a D μ F device for 48 h to
409allow for the formation of compact spheroids. After 48 h of culture, the medium for
410some spheroids was exchanged for medium containing 500 nM insulin, while the
411remaining spheroids received normal growth medium. The insulin-induced drug
412resistance effect has been observed in HT-29 cells in-vitro at insulin doses of 100-1000
413nM.^{50, 51} Spheroids were allowed to incubate in their respective medium for 24 h after
414which the medium was exchanged for medium containing 100 μ M irinotecan, or, for
415controls, normal growth medium. Previous studies had shown that HT-29 spheroids
416exhibit ~20-50% cell death upon exposure to 100 μ M irinotecan.^{40, 52, 53} The D μ F drug
417screening assay workflow is depicted in Figure 7a. To evaluate drug toxicity, the
418diameter of each spheroid was measured at 48, 72, and 96 h. Figure 7b shows the
419average normalized diameter (the ratio of spheroid diameter at 96 h to the diameter at
42048 h) of the spheroids for the different assay conditions. Spheroids that received just
421the drug treatment exhibited a ~20% decrease in diameter, while those that were
422exposed to insulin prior to drug treatment did not exhibit any decrease in size. These
423results are consistent with the insulin-induced drug resistance effect observed in HT-29
424cells in-vitro and, to our knowledge, represent the first time this effect has been
425demonstrated using a three-dimensional HT-29 colon cancer model.

427 Another interesting result of the colon cancer spheroid-based drug screen was the
428 formation of ‘colonospheres’: spherical structures composed of several colonic
429 mucosal epithelial cells that appear as rounded-off epithelial cysts.⁵⁴ The colonosphere
430 morphology signifies a reorganization from a spherical aggregate into one that more
431 closely mimics the morphology of the colon epithelium, which contains numerous
432 glandular and crypt structures. Figure 8 shows examples of colon cancer spheroids
433 exhibiting the colonosphere morphology. This phenotype is of particular interest in
434 cancer research, as literature suggests that colonospheres exhibit a relatively high
435 proportion of cells with a cancer-stem-cell phenotype, which is critical to tumor
436 formation and growth.⁵⁵

437

438 The work presented here advances on previous D μ F cell culture studies that
439 established the ability to seed and maintain cells in adherent monolayer culture on a
440 D μ F device over an extended period of time.⁵⁶⁻⁵⁹ Those studies confirmed that the
441 electric fields used to drive droplet movement have negligible detrimental impact on
442 cell viability, and developed protocols for the manipulation of complex biological
443 solutions. Other work has demonstrated the ability to encapsulate a suspension of cells
444 within hydrogel posts between the plates of a D μ F device.⁶⁰⁻⁶² The encapsulation of
445 cells within hydrogel posts provides a useful tool for modelling cell-matrix
446 interactions, which are key to understanding the cellular microenvironment and
447 important physiological processes such as the epithelial-mesenchymal transition.⁶³
448 However, while the use of hydrogel posts enables cell growth in three dimensions,
449 there are certain limitations associated with these techniques. When using inter-plate

450gel posts, the thickness of the cell aggregates within the gel is limited to the thickness
451of the gap between the plates of the digital microfluidic device (typically $\leq \sim 300\text{ }\mu\text{m}$).
452Additionally, when cells suspensions are encapsulated in gel-posts, the cells are
453randomly distributed throughout the gel, providing little control over the size and
454morphology of the aggregates that form. Lastly, these methods require the use of a
455scaffold or matrix to support 3D cell culture; in some cases, this can be
456disadvantageous, because the scaffold materials may require extra sample preparation
457steps, can be expensive, are susceptible to lot-to-lot variability, may consist of non-
458physiological materials, can complicate sample recovery/analysis, can restrict the
459movement of cells or nutrient transportation, can interfere with screening compounds,
460and may not allow recapitulation of processes that rely on a high degree of cell-cell
461interactions such as embryogenesis, morphogenesis, or tumorigenesis.⁶⁴⁻⁶⁸ The
462platform described here allows scaffold-free three-dimensional cell culture. That said,
463because solutions can be freely added to or extracted from a hanging drop, cell
464suspensions or compact spheroids within a hanging drop could be encapsulated in a
465scaffold material if desired. This platform also enables the growth of spheroids that
466exceed the thickness of the inter-plate gap of a D μ F device, allowing for the formation
467of spheroids that exhibit physiologically relevant morphologies specific to large
468aggregates, such as the development of a necrotic, hypoxic core that can occur within
469spheroids $>400\text{--}500\text{ }\mu\text{m}$ in diameter.^{18, 69-71} Because not all cell types form spheroids,
470and because the behavior of individual cells or small cell clusters encapsulated within
471an extracellular matrix can provide interesting physiological insights, the hanging-drop

472and gel-post techniques for three-dimensional cell culture on a D μ F device are
473complementary.

474

475The platform described here also provides a number of unique advantages compared to
476existing spheroid culture techniques. The primary advantages of the D μ F system are
477automation and the flexibility of the liquid handling protocols. By automating liquid
478handling, digital microfluidics can enable increased throughput and minimize hands-
479on time compared to manual spheroid culture methods, potentially reducing variability
480and human-error in spheroid culture and assay protocols. Digital microfluidics also
481allows droplets to be manipulated either sequentially or simultaneously and droplet
482handling can be pre-programmed for complete automation, or can be controlled in
483real-time allowing for assay flexibility and reconfigurability. Because D μ F provides
484temporal and spatial control over the handling of discrete drops of liquid, any type of
485solution can be added to or extracted from any particular well at will. Thus, spheroids
486can be exposed to a wide variety of stimuli such as drug candidates, different cell
487types, differentiation factors, genetic modulators, and cell secretions in a highly
488controlled fashion. Additionally, because liquid movement on a D μ F platform is not
489confined to channels, liquid can be freely exchanged from one hanging drop to
490another, allowing controlled communication between different spheroids on a device.
491The ability to extract solution from a well allows for in-situ or ex-situ analysis of
492secretions or extracellular conditions from distinct spheroids at any point throughout
493the spheroid culture. This precise control over the composition and analysis of the
494spheroid microenvironment is difficult or impossible to achieve using other spheroid

495culture techniques. For example, while flow-based microfluidic techniques are
496advantageous for massively parallel and/or high throughput spheroid culture protocols,
497such methods are non-ideal for assays that require flexible or reconfigurable liquid
498handling or precise and selective control over the microenvironment of individual
499spheroids. Likewise, microarray or micro-well techniques, in which cells passively
500aggregate in defined locations on a patterned substrate, allow for high-throughput and
501uniform spheroid formation, but do not allow for compartmentalization or
502interrogation of individual spheroids.

503

504While robotic liquid handling systems do allow for automation of spheroid culture and
505analysis, digital microfluidics enables unique liquid handling capabilities that are
506difficult or impossible to achieve using robotic liquid handling. For example, digital
507microfluidics allows for the interrogation of hanging drops either individually or in
508parallel, enables handling of very small volumes of liquid (pL– μ L),^{72, 73} allows for
509magnetic or dielectrophoretic sorting of cells or beads,⁷⁴⁻⁷⁷ enables programmable and
510spatially controlled heating of individual or multiple locations,⁷⁸ supports rapidly
511sequential delivery of reagents to single or multiple locations,⁷⁹ allows for in-situ
512electrochemical detections,⁸⁰⁻⁸² and allows for the formation of hydrogels with
513controllable geometry and orientation.^{61, 83, 84} Additionally, a wide range of
514bioanalytical capabilities including mass spectrometry sample preparation,⁸⁵ PCR,⁸⁶
515qPCR,⁸⁷ immunoassays,⁸⁸ surface plasmon resonance imaging,⁸⁹ and fluorescence
516imaging,⁹⁰ have been developed for the D μ F platform, providing in-situ analytical and

517multiplexing functionalities that could be challenging to incorporate into a robotic
518liquid handling spheroid-culture workflow.

519

520The D μ F platform described here does have certain limitations compared to other
521automated spheroid culture techniques. The primary limitation of the system is the
522relatively low throughput compared to robotic liquid handling systems. Because all of
523the liquid handling in digital microfluidics is performed in the same two-dimensional
524plane, the device must accommodate both the wells and the transportation electrodes,
525which limits the number of wells that can be placed on a device. By contrast, the
526liquid handling path for robotic liquid handling systems usually occurs on a different
527plane than the well-plate, allowing the wells to be packed closer together. Thus digital
528microfluidics cannot achieve the same well density that is possible using hanging-drop
529well-plates, and is best suited for research environments in which medium-throughput
530processing is sufficient. While the prototype devices used in this work are limited to
53146 actuating electrodes, which enables the formation of up to eight hanging drops, a
532more advanced D μ F setup, such as the DropBot,⁹¹ an open source D μ F hardware and
533software system which allows for hundreds of individually addressable electrodes,
534would enable >50 spheroids to be maintained and addressed on a single device.
535Although this well density is considerably lower than commercially available 96- or
536384 hanging drop well plates (3DBiomatrix Inc., InSphero), operating multiple D μ F
537devices simultaneously would increase the throughput. Digital microfluidics also
538operates at lower working volumes than other automated spheroid culture methods (7–
53912 μ L hanging drops on this D μ F platform compared to 20–30 μ L for a 384 well

540hanging drop plate²³), which, while advantageous in some respects, can also present
541challenges. Specifically, smaller drops are more susceptible to evaporation, which can
542alter the composition of the hanging drop. Smaller hanging drops also require a higher
543cell density than larger hanging drops to achieve a spheroid of the same size. Working
544with higher cell densities requires more precise liquid handling as spheroid size is
545related directly to hanging drop volume and cell density. These challenges, however,
546can be mitigated by employing humidity controls and droplet dispensing monitoring
547and control systems. Lastly, D μ F devices are susceptible to dielectric breakdown
548during prolonged operation, which can interfere with assay procedures. However,
549dielectric and hydrophobic material selection and deposition techniques are active
550areas of research in digital microfluidics and many design parameters, such as the
551materials, thicknesses, and organization of the dielectric and hydrophobic layers, as
552well as the ambient medium (i.e. air vs. oil) and operating voltage and frequency can
553be optimized to minimize the chance of dielectric breakdown.⁹²⁻⁹⁵ The performance of
554optimized devices can support at least 25,000 droplet actuation steps without dielectric
555breakdown, which is sufficiently reliable for commercial applications.⁹⁶

556

557The work presented here demonstrates that digital microfluidics, with highly flexible
558and automated liquid handling capabilities, and compatibility with a variety of in-situ
559analytical techniques, has the potential to serve as a powerful tool for automated cell
560spheroid culture. Ultimately, a digital microfluidic platform that facilitates cell
561spheroid culture and analysis may help increase adoption of three-dimensional cell-
562based assays and screens in routine biomedical research.

563

564Acknowledgements

565The authors thank Professor Heather Maynard (UCLA, Dept. of Chemistry and
566Biochemistry) for providing access to cell culture facilities and equipment. All
567confocal laser scanning microscopy was performed at the CNSI Advanced Light
568Microscopy/Spectroscopy (ALMS) Shared Resource Facility at UCLA, supported
569with funding from NIH-NCRR shared resources grant (CJX1-443835-WS-29646) and
570NSF Major Research Instrumentation grant (CHE-0722519). The authors also thank
571Dr. Matthew J. Schibler (ALMS Shared Resource Facility) for assistance with the
572confocal microscopy performed for this work. This work was supported by NSF-DGE
573UCLA IGERT Materials Creation Training Program (MCTP) [0654431], institutional
574funds, and a Dissertation Year Fellowship (to APA).

575

576† Electronic Supplementary Information (ESI) available online.

577

578References

5791. Ridky, T. W.; Chow, J. M.; Wong, D. J., et al., Invasive three-dimensional organotypic neoplasia from multiple normal human epithelia.
580*Nat Med* **2010**, *16* (12), 1450-5.
5812. Souza, G. R.; Molina, J. R.; Raphael, R. M., et al., Three-dimensional tissue culture based on magnetic cell levitation. *Nature*
582*Nanotechnology* **2010**, *5* (4), 291-296.
5833. Birgersdotter, A.; Baumforth, K. R.; Porwit, A., et al., Three-dimensional culturing of the Hodgkin lymphoma cell-line L1236 induces a
584HL tissue-like gene expression pattern. *Leuk Lymphoma* **2007**, *48* (10), 2042-53.
5854. De Witt Hamer, P.; Van Tilborg, A.; Eijk, P., et al., The genomic profile of human malignant glioblastoma is altered early in primary cell
586culture and preserved in spheroids. *Oncogene* **2008**, *27*, 2091 - 2096.

5875. Ernst, A.; Hofmann, S.; Ahmadi, R., et al., Genomic and expression profiling of glioblastoma stem cell-like spheroid cultures identifies
588novel tumor-relevant genes associated with survival. *Clin Cancer Res* **2009**, *15*, 6541 - 6550.

5896. Cheng, N. C.; Chen, S. Y.; Li, J. R., et al., Short-term spheroid formation enhances the regenerative capacity of adipose-derived stem
590cells by promoting stemness, angiogenesis, and chemotaxis. *Stem Cells Transl Med* **2013**, *2* (8), 584-94.

5917. Chang, T. T.; Hughes-Fulford, M., Monolayer and spheroid culture of human liver hepatocellular carcinoma cell line cells demonstrate
592distinct global gene expression patterns and functional phenotypes. *Tissue Eng Part A* **2009**, *15* (3), 559-67.

5938. Boudreau, N.; Werb, Z.; Bissell, M. J., Suppression of apoptosis by basement membrane requires three-dimensional tissue organization
594and withdrawal from the cell cycle. *Proc Natl Acad Sci U S A* **1996**, *93* (8), 3509-13.

5959. DiMasi, J. A.; Feldman, L.; Seckler, A., et al., Trends in Risks Associated With New Drug Development: Success Rates for
596Investigational Drugs. *Clin Pharmacol Ther* **2010**, *87* (3), 272-277.

59710. DiMasi, J. A.; Reichert, J. M.; Feldman, L., et al., Clinical approval success rates for investigational cancer drugs. *Clin Pharmacol Ther*
598**2013**, *94* (3), 329-35.

59911. Ledford, H., Translational research: 4 ways to fix the clinical trial. *Nature* **2011**, *477* (7366), 526-8.

60012. Jayasundara, K. S.; Keystone, E. C.; Parker, J. L., Risk of failure of a clinical drug trial in patients with moderate to severe rheumatoid
601arthritis. *J Rheumatol* **2012**, *39* (11), 2066-70.

60213. Hay, M.; Thomas, D. W.; Craighead, J. L., et al., Clinical development success rates for investigational drugs. *Nat Biotechnol* **2014**, *32*
603(1), 40-51.

60414. Huttmacher, D. W., Biomaterials offer cancer research the third dimension. *Nat Mater* **2010**, *9* (2), 90-3.

60515. Lee, J. A.; Berg, E. L., Neoclassic drug discovery: the case for lead generation using phenotypic and functional approaches. *J Biomol*
606**Screen** **2013**, *18* (10), 1143-55.

60716. Ho, W. Y.; Yeap, S. K.; Ho, C. L., et al., Development of multicellular tumor spheroid (MCTS) culture from breast cancer cell and a
608high throughput screening method using the MTT assay. *Plos One* **2012**, *7* (9), e44640.

60917. Kunz-Schughart, L.; Freyer, J.; Hofstaedter, F., et al., The use of 3-D cultures for high-throughput screening: the multicellular spheroid
610model. *J Biomol Screen* **2004**, *9*, 273 - 285.

61118. Lin, R.-Z.; Chang, H.-Y., Recent advances in three-dimensional multicellular spheroid culture for biomedical research. *Biotechnol J*
612**2008**, *3* (9-10), 1172-1184.

61319. Patra, B.; Chen, Y.-H.; Peng, C.-C., et al., A microfluidic device for uniform-sized cell spheroids formation, culture, harvesting and flow cytometry analysis. *Biomicrofluidics* **2013**, 7 (5), -.
61520. Wu, L.; Di Carlo, D.; Lee, L., Microfluidic self-assembly of tumor spheroids for anticancer drug discovery. *Biomed Microdevices* **2008**, 10 (2), 197-202.
61721. Yu, L.; Chen, M. C. W.; Cheung, K. C., Droplet-based microfluidic system for multicellular tumor spheroid formation and anticancer drug testing. *Lab Chip* **2010**, 10 (18), 2424-2432.
61922. Lee, K.; Kim, C.; Young Yang, J., et al., Gravity-oriented microfluidic device for uniform and massive cell spheroid formation. *Biomicrofluidics* **2012**, 6 (1), -.
62123. Tung, Y. C.; Hsiao, A. Y.; Allen, S. G., et al., High-throughput 3D spheroid culture and drug testing using a 384 hanging drop array. *Analyst* **2011**, 136 (3), 473-478.
62324. Yoshii, Y.; Waki, A.; Yoshida, K., et al., The use of nanoimprinted scaffolds as 3D culture models to facilitate spontaneous tumor cell migration and well-regulated spheroid formation. *Biomaterials* **2011**, 32 (26), 6052-6058.
62525. Frieder, M.; Jean-Christophe, B., Electrowetting: from basics to applications. *Journal of Physics: Condensed Matter* **2005**, 17 (28), R705.
62726. Nelson, W. C.; Kim, C.-J. C., Droplet Actuation by Electrowetting-on-Dielectric (EWOD): A Review. *Journal of Adhesion Science and Technology* **2012**, 26 (12-17), 1747-1771.
62927. Fair, R. B., Digital microfluidics: is a true lab-on-a-chip possible? *Microfluid Nanofluid* **2007**, 3 (3), 245-281.
63028. Jones, T. B., Microfluidic schemes using electrical and capillary forces. *Journal of Physics: Conference Series* **2008**, 142 (1), 012054.
63129. Wartenberg, M.; Acker, H., Quantitative recording of vitality patterns in living multicellular spheroids by confocal microscopy. *Micron* **1995**, 26 (5), 395-404.
63330. Concus, P., Static Menisci in a Vertical Right Circular Cylinder. *J Fluid Mech* **1968**, 34, 481-495.
63431. Chen, L.; Tian, Y. S.; Karayiannis, T. G., The effect of tube diameter on vertical two-phase flow regimes in small tubes. *Int J Heat Mass Tran* **2006**, 49 (21-22), 4220-4230.
63632. Pozrikidis, C., Stability of sessile and pendant liquid drops. *J Eng Math* **2012**, 72 (1), 1-20.
63733. Lee, H.; Yun, S.; Ko, S. H., et al., An electrohydrodynamic flow in ac electrowetting. *Biomicrofluidics* **2009**, 3 (4), -.

63834. Papathanasiou, A. G.; Boudouvis, A. G., Manifestation of the connection between dielectric breakdown strength and contact angle saturation in electrowetting. *Appl Phys Lett* **2005**, *86* (16), -.

64035. Drygiannakis, A. I.; Papathanasiou, A. G.; Boudouvis, A. G., On the Connection between Dielectric Breakdown Strength, Trapping of Charge, and Contact Angle Saturation in Electrowetting. *Langmuir* **2008**, *25* (1), 147-152.

64236. Liu, Y.; Banerjee, A.; Papautsky, I., Precise droplet volume measurement and electrode-based volume metering in digital microfluidics. *Microfluid Nanofluid* **2014**, *17* (2), 295-303.

64437. Nikapitiya, N. Y. J. B.; You, S. M.; Moon, H. In *Droplet dispensing and splitting by electrowetting on dielectric digital microfluidics*, Micro Electro Mechanical Systems (MEMS), 2014 IEEE 27th International Conference on, 26-30 Jan. 2014; 2014; pp 955-958.

64638. Ren, H.; Fair, R. B.; Pollack, M. G., Automated on-chip droplet dispensing with volume control by electro-wetting actuation and capacitance metering. *Sensors and Actuators B: Chemical* **2004**, *98* (2-3), 319-327.

64839. Hong, R.; Fair, R. B. In *Micro/nano liter droplet formation and dispensing by capacitance metering and electrowetting actuation*, Nanotechnology, 2002. IEEE-NANO 2002. Proceedings of the 2002 2nd IEEE Conference on, 2002; 2002; pp 369-372.

65040. Friedrich, J.; Seidel, C.; Ebner, R., et al., Spheroid-based drug screen: considerations and practical approach. *Nat Protoc* **2009**, *4* (3), 309-324.

65241. Cell Culture Media Exchange. In *HDP1384 Perfecta3D® 384-Well Hanging Drop Plates Protocol*, 3D Biomatrix Inc.: 2012.

65342. Luk, V. N.; Mo, G. C. H.; Wheeler, A. R., Pluronic additives: A solution to sticky problems in digital microfluidics. *Langmuir* **2008**, *24* (12), 6382-6389.

65543. Au, S. H.; Kumar, P.; Wheeler, A. R., A New Angle on Pluronic Additives: Advancing Droplets and Understanding in Digital Microfluidics. *Langmuir* **2011**.

65744. Drewitz, M.; Helbling, M.; Fried, N., et al., Towards automated production and drug sensitivity testing using scaffold-free spherical tumor microtissues. *Biotechnol J* **2011**, *6* (12), 1488-1496.

65945. Kelm, J. M.; Timmins, N. E.; Brown, C. J., et al., Method for generation of homogeneous multicellular tumor spheroids applicable to a wide variety of cell types. *Biotechnol Bioeng* **2003**, *83* (2), 173-180.

66146. Ongena, K.; Das, C.; Smith, J. L., et al., Determining Cell Number During Cell Culture using the Scepter Cell Counter. **2010**, (45), e2204.

66347. Nielsen, L. K.; Smyth, G. K.; Greenfield, P. F., Hemacytometer Cell Count Distributions: Implications of Non-Poisson Behavior. *Biotechnology Progress* **1991**, *7* (6), 560-563.

66548. Hsuing, F.; McCollum, T.; Hefner, E., et al., Comparison of Count Reproducibility, Accuracy, and Time to Results between a
666Hemocytometer and the TC20 Automated Cell Counter. In *Bulletin 6003 Rev B*, Bio-Rad Laboratories, Inc.: Hercules, CA, 2013; pp 1-4.

66749. Chen, J.; Katsifis, A.; Hu, C., et al., Insulin decreases therapeutic efficacy in colon cancer cell line HT29 via the activation of the
668PI3K/Akt pathway. *Curr Drug Discov Technol.* **2011**, *8* (2), 119-25.

66950. Chen, J.; Huang, X.-F.; Qiao, L., et al., Insulin caused drug resistance to oxaliplatin in colon cancer cell line HT29. *Journal of*
670*Gastrointestinal Oncology* **2011**, *2* (1), 27-33.

67151. Baricevic, I.; Roberts, D. L.; Renehan, A. G., Chronic Insulin Exposure does not Cause Insulin Resistance but is Associated with
672Chemo-resistance in Colon Cancer Cells. *Horm Metab Res* **2014**, *46* (02), 85-93.

67352. Hardelauf, H.; Frimat, J.; Stewart, J., et al., Microarrays for the scalable production of metabolically relevant tumour spheroids: a tool
674for modulating chemosensitivity traits. *Lab Chip* **2011**, *11*, 419 - 428.

67553. Friedrich, J.; Eder, W.; Castaneda, J., et al., A Reliable Tool to Determine Cell Viability in Complex 3-D Culture: The Acid Phosphatase
676Assay. *J Biomol Screen* **2007**, *12* (7), 925-937.

67754. Stelzner, M.; Helmrath, M.; Dunn, J. C. Y., et al., *A nomenclature for intestinal in vitro cultures*. 2012; Vol. 302, p G1359-G1363.

67855. Kanwar, S.; Yu, Y.; Nautiyal, J., et al., The Wnt/beta-catenin pathway regulates growth and maintenance of colonospheres. *Mol Cancer*
679**2010**, *9* (1), 212.

68056. Barbulovic-Nad, I.; Au, S. H.; Wheeler, A. R., A microfluidic platform for complete mammalian cell culture. *Lab Chip* **2010**, *10* (12),
6811536-1542.

68257. Shih, S. C.; Barbulovic-Nad, I.; Yang, X., et al., Digital microfluidics with impedance sensing for integrated cell culture and analysis.
683*Biosens Bioelectron* **2013**, *42*, 314-20.

68458. Eydelnant, I. A.; Uddayasankar, U.; Li, B. Y., et al., Virtual microwells for digital microfluidic reagent dispensing and cell culture. *Lab*
685*Chip* **2012**, *12* (4), 750-757.

68659. Au, S. H.; Fobel, R.; Desai, S. P., et al., Cellular bias on the microscale: probing the effects of digital microfluidic actuation on
687mammalian cell health, fitness and phenotype. *Integr Biol* **2013**, *5* (8), 1014-1025.

68860. George, S. M.; Moon, H., THREE DIMENSIONAL TISSUE BASED DIGITAL MICROFLUIDIC SCREENING PLATFORM. In *15th*
689*International Conference on Miniaturized Systems for Chemistry and Life Sciences*, Seattle, Washington, USA, 2011; pp 1545-1547.

69061. Eydelnant, I. A.; Betty Li, B.; Wheeler, A. R., Microgels on-demand. *Nat Commun* **2014**, *5*.

69162. Au, S. H.; Chamberlain, M. D.; Mahesh, S., et al., Hepatic organoids for microfluidic drug screening. *Lab Chip* **2014**, *14* (17), 3290-3299.

69363. Lee, J. M.; Dedhar, S.; Kalluri, R., et al., The epithelial–mesenchymal transition: new insights in signaling, development, and disease. *The Journal of Cell Biology* **2006**, *172* (7), 973-981.

69564. Demirbag, B.; Huri, P. Y.; Kose, G. T., et al., Advanced cell therapies with and without scaffolds. *Biotechnol J* **2011**, *6* (12), 1437-1453.

69665. Rimann, M.; Graf-Hausner, U., Synthetic 3D multicellular systems for drug development. *Current Opinion in Biotechnology* **2012**, *23* (5), 803-809.

69866. Lee, M.; Wu, B. M.; Dunn, J. C. Y., Effect of scaffold architecture and pore size on smooth muscle cell growth. *J Biomed Mater Res A* **2008**, *87A* (4), 1010-1016.

70067. Kim, J. B., Three-dimensional tissue culture models in cancer biology. *Semin Cancer Biol* **2005**, *15* (5), 365-377.

70168. Youssef, J.; Morgan, J. R., Self-Assembled Scaffold-Free Three-Dimensional Micro-Tissues for Tissue Repair. In *Advances in Wound Care*, October 2007 ed.; Sen, C. K., Ed. Mary Ann Liebert, Inc.: U.S.A., 2010; Vol. 1, pp 451-456.

70369. Kunz-Schughart, L. A.; Freyer, J. P.; Hofstaedter, F., et al., The use of 3-D cultures for high-throughput screening: The multicellular spheroid model. *J Biomol Screen* **2004**, *9* (4), 273-285.

70570. Mehta, G.; Hsiao, A. Y.; Ingram, M., et al., Opportunities and challenges for use of tumor spheroids as models to test drug delivery and efficacy. *J Control Release* **2012**, *164* (2), 192-204.

70771. Groebe, K.; Mueller-Klieser, W., On the relation between size of necrosis and diameter of tumor spheroids. *International Journal of Radiation Oncology*Biophysics* **1996**, *34* (2), 395-401.

70972. Nelson, W. C.; Kim, J. Y., Monolithic Fabrication of EWOD Chips for Picoliter Droplets. *Microelectromechanical Systems, Journal of* **2011**, *20* (6), 1419-1427.

71173. Lin, Y.-Y.; Welch, E. R. F.; Fair, R. B., Low voltage picoliter droplet manipulation utilizing electrowetting-on-dielectric platforms. *Sensors and Actuators B: Chemical* **2012**, *173* (0), 338-345.

71374. Yuejun, Z.; Ui-Chong, Y.; Sung Kwon, C., Microparticle Concentration and Separation by Traveling-Wave Dielectrophoresis (twDEP) for Digital Microfluidics. *Microelectromechanical Systems, Journal of* **2007**, *16* (6), 1472-1481.

71575. Fan, S. K.; Huang, P. W.; Wang, T. T., et al., Cross-scale electric manipulations of cells and droplets by frequency-modulated dielectrophoresis and electrowetting. *Lab Chip* **2008**, *8* (8), 1325-1331.

71776. Ng, A. H. C.; Choi, K.; Luoma, R. P., et al., Digital Microfluidic Magnetic Separation for Particle-Based Immunoassays. *Anal Chem* **7182012**, *84* (20), 8805-8812.

71977. Chang-An, C.; Chiun-Hsun, C.; Ghaemmaghami, A. M., et al. In *Separation of dendritic and T cells using electrowetting and dielectrophoresis*, Nano/Micro Engineered and Molecular Systems (NEMS), 2012 7th IEEE International Conference on, 5-8 March 2012; **7212012**; pp 183-186.

72278. Nelson, W. C.; Peng, I.; Lee, G. A., et al., Incubated protein reduction and digestion on an electrowetting-on-dielectric digital microfluidic chip for MALDI-MS. *Anal Chem* **2010**, *82* (23), 9932-7.

72479. Nichols, K. P.; Gardeniers, J. G. E., A Digital Microfluidic System for the Investigation of Pre-Steady-State Enzyme Kinetics Using Rapid Quenching with MALDI-TOF Mass Spectrometry. *Anal Chem* **2007**, *79* (22), 8699-8704.

72680. Malic, L.; Veres, T.; Tabrizian, M., Biochip functionalization using electrowetting-on-dielectric digital microfluidics for surface plasmon resonance imaging detection of DNA hybridization. *Biosens. Bioelectron.* **2009**, *24* (7), 2218-2224.

72881. Lederer, T.; Clara, S.; Jakoby, B., et al., Integration of impedance spectroscopy sensors in a digital microfluidic platform. *Microsyst Technol* **2012**, *18* (7-8), 1163-1180.

73082. Karuwan, C.; Sukthang, K.; Wisitsoraat, A., et al., Electrochemical detection on electrowetting-on-dielectric digital microfluidic chip. *Talanta* **2011**, *84* (5), 1384-1389.

73283. Min-Yu, C.; Shih-Kang, F. In *Electric manipulations of hydrogel on a digital microfluidic platform*, Nano/Micro Engineered and Molecular Systems (NEMS), 2012 7th IEEE International Conference on, 5-8 March 2012; **7332012**; pp 407-410.

73484. Chiang, M.-Y.; Shih-Kang, F., Generation and Assembly of Cell-Laden Hydrogels on a Digital Microfluidic Platform. In *International Conference on Miniaturized Systems for Chemistry and Life Sciences*, Okinawa, Japan, 2012; pp 575-577.

73685. Aijian, A. P.; Chatterjee, D.; Garrell, R. L., Fluorinated liquid-enabled protein handling and surfactant-aided crystallization for fully in situ digital microfluidic MALDI-MS analysis. *Lab Chip* **2012**, *12* (14), 2552-2559.

73886. Chang, Y. H.; Lee, G. B.; Huang, F. C., et al., Integrated polymerase chain reaction chips utilizing digital microfluidics. *Biomed Microdevices* **2006**, *8* (3), 215-25.

74087. Sista, R.; Hua, Z.; Thwar, P., et al., Development of a digital microfluidic platform for point of care testing. *Lab Chip* **2008**, *8* (12), **7412091-2104**.

74288. Sista, R. S.; Eckhardt, A. E.; Srinivasan, V., et al., Heterogeneous immunoassays using magnetic beads on a digital microfluidic platform. *Lab Chip* **2008**, *8* (12), 2188-2196.

74489. Malic, L.; Veres, T.; Tabrizian, M., Biochip functionalization using electrowetting-on-dielectric digital microfluidics for surface plasmon
745resonance imaging detection of DNA hybridization. *Biosensors and Bioelectronics* **2009**, *24* (7), 2218-2224.

74690. Barbulovic-Nad, I.; Yang, H.; Park, P. S., et al., Digital microfluidics for cell-based assays. *Lab Chip* **2008**, *8* (4), 519-526.

74791. Fobel, R.; Fobel, C.; Wheeler, A. R., DropBot: An open-source digital microfluidic control system with precise control of electrostatic
748driving force and instantaneous drop velocity measurement. *Appl Phys Lett* **2013**, *102* (193513), 1-5.

74992. Dhindsa, M.; Heikenfeld, J.; Weekamp, W., et al., Electrowetting without electrolysis on self-healing dielectrics. *Langmuir* **2011**, *27* (9),
7505665-70.

75193. Schultz, A.; Chevalliot, S.; Kuiper, S., et al., Detailed analysis of defect reduction in electrowetting dielectrics through a two-layer
752'barrier' approach. *Thin Solid Films* **2013**, *534* (0), 348-355.

75394. Liu, H.; Dharmatilleke, S.; Maurya, D. K., et al., Dielectric materials for electrowetting-on-dielectric actuation. *Microsyst. Technol.*
754**2009**, *16* (3), 449-460.

75595. Lin, Y.-Y.; Evans, R. D.; Welch, E., et al., Low voltage electrowetting-on-dielectric platform using multi-layer insulators. *Sensors and*
756*Actuators B: Chemical* **2010**, *150* (1), 465-470.

75796. Srinivasan, V.; Pamula, V. K.; Fair, R. B., An integrated digital microfluidic lab-on-a-chip for clinical diagnostics on human
758physiological fluids. *Lab Chip* **2004**, *4* (4), 310-315.

759

760

761

762

763

764Tables

Table 1. Performance characteristics of devices used in this work

Performance Characteristic	Value
Hanging drop volume reproducibility ^a	~8%
Maximum hanging drop volume ^b	~55 uL
Intra-experiment spheroid diameter variation ^c	~8%
Inter-experiment spheroid diameter variation ^d	~16%
Spheroid size (diameter) range ^e	50-700+ µm
Average spheroid aspect ratio ^f	1.15 ± 0.09

- a. Average %CV of droplet volume for hanging drops.
- b. Volume range specific to devices used for this work. Max volume = volume above which droplets detach from the device; determined experimentally.
- c. Average %CV of spheroids formed on the same chip. N = 30 total spheroids across 8 different experiments.
- d. Average %CV of spheroids formed on different chips. N = 30 total spheroids across 8 different experiments.
- e. Values represent the smallest and largest spheroid diameters obtained in this work over a period of 24-72 h on the DµF devices. These values do not represent minimum or maximum possible sizes.
- f. Aspect ratio = ratio of a spheroid's major axis to minor axis. N = 77 spheroids of different cell types

765

766

767

768

769

770

771

772

773

774

775

776

777

778

779

780

781

Figure 1. Device schematic and dimensions. Through-holes in the top plate allow for the addition of solutions to on-chip reservoirs, while through-holes, or ‘wells,’ in the bottom plate allow for the formation of hanging drops. Drops that are delivered to a well are drawn into the well spontaneously upon contact with the hydrophilic well wall. Addition of multiple drops to a well allows for the formation of a hanging drop with a curved air-liquid interface. Cells suspended in the drop can aggregate at this interface, forming a single spheroid within the drop.

Figure 2. Hanging drop formation on a digital microfluidic device. (a) A series of images showing a top-down view of the insertion of drops of cell media (dyed blue for enhanced visualization, $\sim 1.2 \mu\text{L}$) into a well on the device. (b) A series of images showing a side-view of a well after the addition of multiple drops to the well. The drops insert spontaneously into the well and, after a sufficient volume has been added, form a hanging drop with the curved interface necessary to induce cell aggregation.

Figure 3. Distribution of dispensed drop volumes. 144 drops were dispensed from different reservoirs across 3 different devices using a pre-programmed droplet dispensing sequence. The average drop volume was $1.75 \mu\text{L}$ and the %CV of the volume of all drops dispensed was $\sim 8\%$.

Figure 4. Size and variation of hanging drop volumes. Hanging drops comprised of 4, 5, and 6 dispensed drops had volumes of $7.4 \pm 0.5 \mu\text{L}$, $8.8 \pm 0.8 \mu\text{L}$, and $10.2 \pm 0.5 \mu\text{L}$, respectively ($N = 8$ hanging drops formed for each condition). For each condition, the variation in hanging drop volume was $<10\%$. Hanging drops of volumes up to $\sim 55 \mu\text{L}$ can

be formed on the devices used in this work, however, only 7-10 μL is needed to form cell spheroids via the hanging drop technique.

Figure 5. Extent of liquid exchange, predicted and experimental results. The extent of liquid exchange after one and two exchange cycles was monitored by measuring the change in absorbance of the dyed hanging drop solution and calculating the concentration from a standard curve. The dilution of a hanging drop after each cycle can be seen in the images above the plot. The agreement between the measured concentrations and the predicted values indicates that thorough mixing of the hanging drop is achieved during each exchange cycle and that D μ F provides good control over the composition of the hanging drop. Error bars indicate the standard deviation of measurements from three different experiments.

Figure 6. Cell spheroids formed by D μ F. (a) Representative images of spheroids of mouse mesenchymal stem cells (mMSC) grown on a digital microfluidic device after 24, 48, and 72 h of in-situ incubation. Each image is of a different spheroid. Spheroids exhibit >90% viability during this time-frame as determined by staining with calcein-AM/ethidium homodimer-1 to visualize living (green) and dead (red) cells. (b) The spheroids of mMSC formed from cell suspensions of $\sim 7.5 \times 10^5$ cells/mL had diameters of 249 ± 34 μm after 24 h (N=12 spheroids), 327 ± 58 μm after 48 h (N=10 spheroids), and 425 ± 75 (N=8 spheroids) after 72 h in culture. The data here represent average spheroid sizes at each time point from 6 separate experiments. (c) The distribution and average aspect ratio of spheroids grown on the D μ F platform (N=77 spheroids). The data represent measurements from HT-29 colorectal carcinoma spheroids, BJ fibroblast spheroids, and mouse MSC spheroids with diameters ranging from 100 to 1000 μm (N=12). Solid line indicates

827the average aspect ratio value (1.15), the dashed lines indicate +/- one standard deviation
828(0.09).

829**Figure 7. Spheroid-based drug screening protocols and results.** (a) Diagram illustrating
830the workflow for a spheroid-based drug screen performed on human colorectal
831adenocarcinoma cells (ATCC HT-29). After 48 h of compaction, spheroids received either
832normal or insulin-containing medium. After another 24 h, the spheroids received either
833normal or drug-containing (irinotecan) medium, and were incubated for another 24 h. (b)
834Comparison of the normalized spheroid diameter (D_{96h}/D_{48h}) for spheroids exposed to the
835different drug screen assay conditions. The error bars indicate +/- one standard deviation
836from the average for each condition (medium + medium: N = 4, insulin + medium: N = 4,
837medium + drug: N = 6, insulin + drug: N = 8).

838**Figure 8. Colonosphere morphologies.** HT-29 colon adenocarcinoma spheroids exhibited
839colonosphere morphologies after 96 h of in-vitro hanging drop culture on a Dlture on a D6 h
840of in-vitro hanging drop culture on a D the average for each condition (medium + medium:pt-
841like folds, similar to the morphology of the colon epithelium. Image (d) is a confocal cross-
842section image taken at a z-plane ~75 μm into the spheroid interior. The image clearly shows
843the large lumen/crypt structure that is indicative of colonosphere morphology. The spheroids
844in these images are stained with calcein-AM/ethidium homodimer-1 to visualize living
845(green) and dead (red) cells. Scale bars correspond to 200 μm .