Lawrence Berkeley National Laboratory

Recent Work

Title

LIFETIME OF THE 186-KEV LEVEL OF THORIUM-231

Permalink

<https://escholarship.org/uc/item/7g9281f8>

Author

Strominger, D.

Publication Date

1958-10-01

UNIVERSITY OF
CALIFORNIA

*Radiation
Laboratory*

TWO-WEEK LOAN COPY

*This is a Library Circulating Copy
which may be borrowed for two weeks.
For a personal retention copy, call
Tech. Info. Division, Ext. 5545*

BERKELEY, CALIFORNIA

Non Std. Dist.
For Publ. in Phys. Rev.

DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

UCRL-8507

UNIVERSITY OF CALIFORNIA

Radiation Laboratory
Berkeley, California

Contract No. W-7405-eng-48

LIFETIME OF THE 186-KEV LEVEL OF THORIUM-231

D. Strominger

October 1958

Printed for the U. S. Atomic Energy Commission

LIFETIME OF THE 186-KEV LEVEL OF THORIUM-231

D. Strominger

Radiation Laboratory and Department of Chemistry
University of California, Berkeley, California

October 1958

ABSTRACT

The lifetime of the 186-kev level of Th^{231} has been studied following U^{235} alpha decay. The half-life value of this level is 0.77 ± 0.12 msec.

LIFETIME OF THE 186-KEV LEVEL OF THORIUM-231^{*}

D. Strominger

Radiation Laboratory and Department of Chemistry
University of California, Berkeley, California

October 1958

INTRODUCTION

Strominger and Rasmussen¹ summarized the available lifetime data for E1 transitions in the actinide region. This summary showed that the E1 transition lifetimes varied erratically from one isotope to another. Nilsson wave functions were considered for the intrinsic states in odd-proton nuclei, and it was demonstrated that the E1 transitions violated selection rules^{1,2} inherent in nuclei with very large spheroidal deformations. Similarly, it can be shown that all the expected low-lying E1 transitions between different intrinsic states in the actinide region violate the selection rules mentioned above. Hence all the low-lying E1 transitions are expected to have lifetimes longer by some orders of magnitude than the predictions of the single-proton lifetime formula.³

The alpha decay of U^{235} to the levels of Th^{231} is complex,⁴ and not all the details of the decay have been worked out. Figure 1 shows a partial decay scheme of U^{235} . Stephens⁵ has measured the K-conversion coefficients of the 186-kev and 143-kev transitions which depopulate the 186-kev level, and deduced that both are E1 transitions.

An upper limit of 1.5 μ sec for the half-life value of the 186-kev level of Th^{231} has been reported.¹ Since then, improved electronic techniques have been developed which make it possible to set lower lifetime limits. This paper describes the use of these new techniques to measure the half-life of the 186-kev level of Th^{231} .

* This work was done under the auspices of the U. S. Atomic Energy Commission.

EXPERIMENTAL TECHNIQUES AND PRELIMINARY EXPERIMENTS

Figure 2 is a block diagram of the apparatus used for these experiments. The radiations were detected with scintillation crystals coupled to RCA 6655 photomultiplier tubes. A potassium iodide (thallium iodide activated) crystal detected alpha particles, a sodium iodide (thallium iodide activated) crystal detected the gamma radiation, and a Lucite disc impregnated with terphenyl detected electrons.

Except for the time-to-pulse-height converter the rest of the electronic equipment has been described adequately in the literature.⁶

The first time-to-pulse-height converter used was modified from the circuit described by Weber *et al.*⁷ This instrument utilizes the double control grid feature of the 6BN6 vacuum tube. The 6BN6 tube is biased to cutoff by a negative voltage applied to one of the control grids. When the first input pulse (here converted into a positive pulse) arrives at the biased control grid the 6BN6 tube starts to conduct. The plate voltage drops linearly with time for the duration of the pulse (300 μ sec) or, if there is a coincidence, until the second input arrives at the second control grid and cuts the tube off. We usually delay the second input pulse about 100 μ sec to get a large signal from the instrument.

The output of the time-to-pulse-height converter is displayed on the 50-channel pulse-height analyzer whenever the selected events are detected by the two single-channel pulse-height analyzers. The time calibration is done by inserting cables of known length between the wide band amplifiers and the time-to-pulse-height converter.

This instrument was used to measure the half-lives of the Dy^{161} levels populated by Tb^{161} beta decay. Our results⁸ are 29 ± 3 μ sec for the half-life value of the 26-kev level and 3.4 ± 0.6 μ sec for the half-life value of the 75-kev level. These results are in agreement with the previously reported values.^{9,10} However, this instrument is not stable enough to measure half-lives shorter than 1 μ sec.

A second time-to-pulse-height converter was built with essentially the same features as the circuit described by Jones and Warren.¹¹ In this instrument the two input pulses cut off two 6E8 vacuum tubes which have one common

plate resistance. A shorting stub is placed in the plate circuit to shape the single pulses to a duration of about 30 μsec . A bias voltage on a 6L11A diode is adjusted so that when both tubes are cut off the diode conducts, but the diode remains cut off when only one of the 6L80F tubes is cut off. Hence the width of the pulse feeding through the 6L11A diode is proportional to the overlap of the two input pulses. This pulse is then integrated, giving a pulse whose voltage is proportional to the overlap of the two input pulses. We find it convenient to arbitrarily delay one pulse about 10 μsec so that the true coincidence pulses have a duration of about 20 μsec .

This second time-to-pulse-height converter was used to measure the half-life of the 204-keV level of Mo^{95} . This level is preceded by a 584-keV transition¹² after the electron-capture decay of 60-day $\text{Tc}^{95\text{m}}$. A Na^{22} sample was used for comparison purposes in order to do a centroid shift analysis.^{13,14}

Sodium iodide crystals were used to detect both gamma rays. One single-channel pulse-height analyzer was set on the 584-keV peak and the other was set on the 204-keV peak. A delay curve was then obtained for $\text{Tc}^{95\text{m}}$. Similarly a delay curve was obtained for the Na^{22} standard without changing the single-channel analyzer settings. Then the settings on the two single-channel analyzers were interchanged, effectively interchanging the two sides of the delay curves. Once again delay curves were obtained for both the $\text{Tc}^{95\text{m}}$ and Na^{22} samples. These experiments were repeated several times.

Unfortunately, the two series of experiments did not agree with each other within the reproducibility of the individual runs. The Na^{22} delay curves were displaced toward the higher channels. The cause of this electronic shift has not been ascertained. It is somewhat, but not wholly, dependent upon counting rates.

The average of the two results yields a half-life value of 0.85 ± 0.16 μsec ,¹⁵ which is in agreement with Quidat's value of 0.77 ± 0.03 μsec obtained from $\text{Tc}^{95\text{m}}$ decay¹⁶ and the value of 0.76 ± 0.07 μsec calculated from Coulomb-excitation yield considerations.¹⁷ Therefore it is reasonable to assume that the electronic shift in the delay curves is canceled out when averages are taken in the above fashion.

EXPERIMENTAL RESULTS FOR U^{235}

The first series of runs were alpha-gamma delay curves. Radium-226 was the standard used for the comparison delay curves. Radium-226 decays primarily to the ground state of Rn^{222} , but 5.7% of the alpha decays populate a level 187 kev above the ground state.¹⁸ The Ra^{226} alpha particles also have an energy very similar to the corresponding alpha particles of U^{235} .

The slope of the U^{235} delay curves showed that the half-life of the 186-kev level of Th^{231} was less than or equal to 1.0 μ sec. However, the Ra^{226} comparison delay curve displayed a limiting slope equivalent to a half-life of 0.6 μ sec. A crude analysis of the slopes of the two delay curves indicated that the half-life value of the 186-kev level of Th^{231} is about 0.8 μ sec. Unfortunately the centroid shift analysis between Ra^{226} and U^{235} is not useful for two reasons. First, the half-life value of the 187-kev level in Rn^{222} is not known, and, secondly, the alpha decay of U^{235} (see Fig. 1) is too complicated for a simple centroid analysis from alpha-gamma delay curves.

Figure 1 shows that a transition of about 205 kev precedes the 186-kev level in about 6% of the U^{235} alpha decays. Hence, one can obtain a gamma-gamma delay curve by setting one single-channel analyzer on the 205-kev peak and the other analyzer on the 186-kev peak. Actually, the 186-kev peak is so much more intense than the 205-kev peak that only the 186-kev peak is seen in the singles spectrum. Accordingly, one single-channel analyzer is set on the high-energy side of the 186-kev peak while the other is set on the low-energy side of the peak. A 100-mg sample of metallic uranium enriched in U^{235} was used to obtain the U^{235} delay curves. A Na^{22} sample was run for the comparison delay curves. The settings on the two single-channel analyzers were interchanged and some additional centroid shifts were calculated.

Just as with Tc^{95m} , the Na^{22} delay curves were displaced toward the higher channels. However, the average of the two sets of runs yields the half-life value of $0.77 \pm 0.12 \mu$ sec for the 186-kev level of Th^{231} . This value agrees with the crude estimate obtained from the alpha-gamma delay curves.

The 186-kev E1 transition in Th^{231} is slower by a factor of 6×10^4 than the single-proton lifetime prediction.³ This hindrance can be understood in terms of a violation of selection rules described in more detail elsewhere.^{1,2}

ACKNOWLEDGMENTS

I would like to express my appreciation to Michiyuki Nakamura for designing the second time-to-pulse-height converter used in these experiments. Kenneth T. Faler and John P. Unik graciously permitted the use of their unpublished results.

REFERENCES

1. D. Strominger and J. O. Rasmussen, Nuclear Phys. 3, 197 (1957).
2. G. Alaga, Nuclear Phys. 4, 625 (1957).
3. S. A. Moszkowski, Beta- and Gamma-Ray Spectroscopy, K. Siegbahn, Ed. (Interscience, New York, 1955), Chap. XIII.
4. R. C. Pilger, Jr., (Thesis), UCRL-3877, July 1957 (unpublished); Pilger, Stephens, Asaro, and Perlman, Bull. Am. Phys. Soc. Ser. II, 2, 394 (1957).
5. F. S. Stephens, Jr., (Thesis), UCRL-2970, June 1955 (unpublished).
6. For example: P. R. Bell, in Beta- and Gamma-Ray Spectroscopy, K. Siegbahn, Ed. (Interscience, New York, 1955), Chap. V.
7. W. Weber, C. W. Johnstone, and L. Cranberg, Rev. Sci. Instr. 27, 166 (1956).
8. K. T. Faler and D. Strominger, unpublished results (1957).
9. M. Vergnes, J. phys radium 18, 579 (1957).
10. Gregers, Hansen, Nathan, Nielsen, and Sheline, Nuclear Phys. 6, 630 (1958).
11. G. Jones and J. B. Warren, J. Sci. Instr. 33, 429 (1956).
12. J. P. Unik, unpublished results quoted in Revs. Modern Phys. 50, 585 (1958).
13. T. D. Newton, Phys. Rev. 78, 490 (1950).
14. Z. Bay, Phys. Rev. 77, 419 (1950).
15. J. P. Unik and D. Strominger, unpublished results (1958).
16. J. Quidat, Compt. rend. 246, 2119 (1958).
17. F. K. McGowan and P. H. Stelson, Phys. Rev. 109, 901 (1958).
18. F. Asaro and I. Perlman, Phys. Rev. 88, 129 (1952).

MU-16189

Fig. 1. Partial decay scheme of U^{235} .

MJ-16190

Fig. 2. Block diagram of millimicrosecond time sorter.

This report was prepared as an account of Government sponsored work. Neither the United States, nor the Commission, nor any person acting on behalf of the Commission:

- A. Makes any warranty or representation, express or implied, with respect to the accuracy, completeness, or usefulness of the information contained in this report, or that the use of any information, apparatus, method, or process disclosed in this report may not infringe privately owned rights; or
- B. Assumes any liabilities with respect to the use of, or for damages resulting from the use of any information, apparatus, method, or process disclosed in this report.

As used in the above, "person acting on behalf of the Commission" includes any employee or contractor of the Commission to the extent that such employee or contractor prepares, handles or distributes, or provides access to, any information pursuant to his employment or contract with the Commission.