UC Irvine

UC Irvine Previously Published Works

Title

Photocoagulation of dermal blood vessels with multiple laser pulses in an in vivo microvascular model.

Permalink

<https://escholarship.org/uc/item/7mq4d09m>

Journal

Lasers in surgery and medicine, 44(2)

ISSN

0196-8092

Authors

Jia, Wangcun
Tran, Nadia
Sun, Victor
et al.

Publication Date

2012-02-01

DOI

10.1002/lsm.22000

Copyright Information

This work is made available under the terms of a Creative Commons Attribution License, available at <https://creativecommons.org/licenses/by/4.0/>

Peer reviewed

Published in final edited form as:

Lasers Surg Med. 2012 February ; 44(2): 144–151. doi:10.1002/lsm.22000.

Photocoagulation of Dermal Blood Vessels With Multiple Laser Pulses in an In Vivo Microvascular Model

Wangcun Jia, PhD^{1,*}, Nadia Tran, BS¹, Victor Sun, BS¹, Marko Marinček, PhD², Boris Majaron, PhD³, Bernard Choi, PhD^{1,4,5}, and J. Stuart Nelson, MD, PhD^{1,4}

¹Beckman Laser Institute and Medical Clinic, Department of Surgery, University of California, Irvine, California 92617

²Fotona d.d., Ljubljana, Slovenia

³Jožef Stefan Institute, Ljubljana, Slovenia

⁴Department of Biomedical Engineering, University of California, Irvine, California 92697

⁵Edwards Lifesciences Center for Advanced Cardiovascular Technology, University of California, Irvine California 92697

Abstract

Background/Objectives—Current laser therapy of port wine stain (PWS) birthmarks with a single laser pulse (SLP) does not produce complete lesion removal in the majority of patients. To improve PWS therapeutic efficacy, we evaluated the performance of an approach based on multiple laser pulses (MLP) to enhance blood vessel photocoagulation.

Study Design—The hamster dorsal window chamber model was used. Radiant exposure (RE), pulse repetition rate (f_r), total number of pulses (n_p), and length of vessel irradiated were varied. Blood vessels in the window were irradiated with either SLP with RE of 4–7 J/cm² or MLP with RE per pulse of 1.4–5.0 J/cm², f_r of 0.5–26.0 Hz, and n_p of 2–5. The laser wavelength was 532 nm and pulse duration was 1 ms. Either a 2 mm vessel segment or entire vessel branch was irradiated. Digital photographs and laser speckle images of the window were recorded before and at specific time points after laser irradiation to monitor laser-induced blood vessel structural and functional changes, respectively.

Results—We found that: (1) for a SLP approach, the RE required to induce blood vessel photocoagulation was 7 J/cm² as compared to only 2 J/cm² per pulse for the MLP approach; (2) for MLP, two pulses at a repetition rate of 5 Hz and a RE of 3 J/cm² can induce photocoagulation of more than 80% of irradiated blood vessel; and (3) irradiation of a longer segment of blood vessel resulted in lower reperfusion rate.

Conclusions—The MLP approach can induce blood vessel photocoagulation at much lower RE per pulse as compared to SLP. The 5 Hz f_r and the need for two pulses are achievable with modern laser technology, which makes the MLP approach practical in the clinical management of PWS birthmarks.

Keywords

laser speckle imaging; dorsal window chamber; laser dermatologic surgery; vascular malformation; port wine stain; angiogenesis

INTRODUCTION

Port wine stain birthmarks (PWS) are congenital, progressive vascular malformations of human skin with an incidence rate of 3 per 1,000 live births [1,2]. PWS are characterized by an increase in blood vessel size and a decrease in perivascular nerve density [3–5]. The current treatment of choice for PWS is the pulsed dye laser (PDL) [6,7] with dynamic skin cooling [8,9]. However, complete PWS blanching is rarely achieved for many patients even after numerous PDL treatments [10–12]. Many factors contribute to incomplete PWS blanching, including the presence of large blood vessels which can only be partially photocoagulated by PDL due to high superficial light absorption by hemoglobin [13–15], epidermal melanin that reduces light delivery to targeted PWS vessels, and regeneration of photocoagulated blood vessels due to angiogenesis [16–19].

The multiple laser pulses (MLP) approach has been successfully used to treat facial and leg telangiectasia [20–22]. Histological evaluation of laser-irradiated normal and PWS skin revealed that MLP induced coagulation of deeper vasculature as compared to the single laser pulse (SLP) approach [23,24]. The introduction of multiple cryogen spurts applied intermittently with MLP makes this approach even more appealing because the epidermis can be effectively cooled between the consecutive laser pulses, thus enabling safe delivery of higher total light dosages [15,25,26]. From a thermophysical point of view, the rationale of the MLP approach is that the delivery of subsequent pulses before the blood vessels cool to the baseline temperature can increase the temperature of blood vessels to a greater degree as compared to a SLP [15,25]. From a biological point of view, researchers proposed that MLP induce summation of irreversible, thermal injury from a series of lower-peak temperature heating cycles [27]. The different views on the MLP approach explain why the inter-pulse interval varied from 0.25 seconds to 30 minutes and the total number of pulses (n_p) varied from 2 to 81 in previous studies of the MLP approach [15,20–26]. We believe that the large variation in the treatment parameters hinders the practical application of the MLP approach. To address this issue, we used an *in vivo* microvascular model to elucidate the mechanism of MLP-induced photocoagulation. The objective of the present study is to investigate the effects of radiant exposure (RE), pulse repetition rate (f_r), n_p , and irradiated vessel length on the short-term photocoagulation and long-term removal of blood vessels using the MLP approach as compared to SLP.

STUDY DESIGN/MATERIALS AND METHODS

In Vivo Animal Model

All experiments were conducted under a protocol approved by the Institutional Animal Care and Use Committee, University of California, Irvine. Adult male Golden Syrian hamsters (90–120 g) were used in this study. A dorsal window chamber (DWC) was installed on each animal. This model, first described by Algire [28], consists of a lengthwise fold of dorsal skin with an implanted clear glass window that permits *in vivo* visualization and irradiation of the subdermal blood vessels. The window chamber, when properly prepared, provides excellent viewing of subdermal blood vessels up to 4 weeks [17,29]. Details of the chamber structure and surgical procedure can be found elsewhere [30–32]. Briefly, after the animal was anesthetized with a cocktail of ketamine/xylazine, the dorsal skin was shaved, epilated, and lifted to form a skinfold. A pair of titanium window frames was attached to the front and back sides of the dorsal skinfold with screws and sutures. One layer of skin and subcutis with the panniculus carnosus was completely removed within the circular area of the frame's observation window to expose the subdermal blood vessels in the underlying intact skin. A thin glass window (12 mm diameter, 0.2 mm thickness) was then inserted into the window frame to protect the subdermis from dehydration and contamination. The window frames

were strategically placed on the backs of the animals to enable visualization of a tree-like vascular network for the experiments.

Laser Irradiation

Laser irradiation was performed on the window glass (subdermal) side of the preparation. Two millimeter blood vessel segments or the entire vessel branches were irradiated with a frequency-doubled Nd:YAG laser (Dualis VP+, Fotona, Ljubljana, Slovenia) which emits single or multiple pulses at a wavelength of 532 nm. The duration of an individual pulse is 1 ms, and the RE could be varied from 1.4 to 7 J/cm² with the 2 mm spot used in this study. For MLP, the n_p could be varied from 2 to 5 and the f_r from 0.5 to 26 Hz. Laser pulse energies were verified using an energy meter (FL250A-SH with Nova display, Ophir, Logan, UT).

Color and Laser Speckle Imaging

Digital color photos and laser speckle images (LSI) of the windows were acquired prior to, shortly after laser irradiation, and daily thereafter for 2 weeks. After the day of window implantation, hamsters were anesthetized with a mixture of oxygen and isoflurane (3%) through a nosecone. Color images of the subdermal sides of windows were taken while white light illumination was delivered to the subdermal side or through the epidermal side of the window. Although color images can document blood vessel structural changes after laser irradiation, they cannot be reliably used to judge whether blood flow has completely stopped. Therefore, LSI was used to determine blood flow dynamics in the window [17,33–35]. During LSI, the window was trans-illuminated with a CW HeNe laser to produce a speckle pattern. When blood flow is present, the speckle pattern varies with time; otherwise, the pattern is static. The speckle patterns were integrated over 10 ms with a CCD camera and processed with a sliding-window-based algorithm [36–38] to visualize blood flow in the window.

Data Analysis

We only analyzed the response of venules to laser irradiation, though arterioles and venules in the DWC usually run in pairs, because PWS are malformations of capillaries and/or post-capillary venules. We first identified blood vessels that were completely photocoagulated by comparing color images and LSI flow maps recorded before and after laser irradiation. The criterion was that flow was completely stopped in the photocoagulated vessels. The coagulated blood vessels were then tracked in the color images and flow maps in the following 2 weeks to determine whether reperfusion occurred.

RESULTS

Effect of Radiant Exposure on Blood Vessel Photocoagulation

Experiments were performed first to determine the threshold RE per pulse to coagulate 100–250 μ m diameter blood vessels in the DWC model for both SLP and MLP. Results are shown in Figure 1. For SLP, a RE of 7 J/cm² was required to induce photocoagulation in most (i.e., 8 out of 9) irradiated blood vessels. In contrast, MLP with five pulses at $f_r = 26$ Hz could induce photocoagulation in practically all irradiated blood vessels at RE = 2 J/cm² and higher. Fitting the Boltzmann's dose-response function to the observed coagulation percentage as a function of RE yields the 50% probability RE value of 1.78 J/cm² for the MLP approach, obviously very different from the value of 5.45 J/cm² obtained for the case of SLP.

Effect of Total Number of Pulses on Blood Vessel Photocoagulation

As shown in Figure 1, the threshold n_p was not more than 5 when RE was 2 J/cm². Further experiments were conducted using the DWC model to determine the threshold n_p when RE was 3–5 J/cm² and f_r was 10 Hz. A total of 9 animals, 3 for each RE, were used in this study group. One example, when RE was 3 J/cm², is shown in Figure 2. Figure 2a shows an image of the DWC before laser irradiation and Figure 2c is the LSI flow map. Vessel diameters, which ranged from 117 to 218 μ m, are also shown in Figure 2a. The yellow circles mark the sites of laser irradiation, which have a diameter of 2 mm. Figure 2b and d shows a color image and a LSI flow map of the DWC, respectively, after laser irradiation. When n_p was 2 or higher, a thermal coagulum (dark red color) [39–41] could be seen in the irradiated sites in Figure 2b, and no flow was detected in Figure 2d which indicates complete blood vessel coagulation. A summary of the results is shown in Table 1. It can be seen that more than 80% of the irradiated blood vessels were coagulated when n_p was 2 or higher.

Effect of Pulse Repetition Rate on Blood Vessel Photocoagulation

Results in Figure 1 and Table 1 confirm that effective blood vessel photocoagulation can be achieved when f_r was 26 and 10 Hz. More experiments were conducted to determine if blood vessels could be coagulated with a lower f_r at a RE per pulse of 4 J/cm² and a n_p of 2. A total of 4 animals were used in this study group. Figure 3 shows an example which demonstrates the effect of f_r on photocoagulation of blood vessels in the DWC model. Color image and LSI flow map of the DWC before laser irradiation are shown in Figure 3a and c, and the same DWC after laser irradiation is shown in Figure 3b and d. Three f_r values were used, 0.5, 1, and 5 Hz. Figure 3d shows that blood flow was clearly detected in the blood vessels irradiated at a f_r of 0.5 or 1 Hz. Alternatively, blood flow was completely stopped in the blood vessels irradiated at a f_r of 5 Hz. It can be noted that a blood coagulum was present downstream of the irradiation site for a blood vessel irradiated at a f_r of 0.5 Hz (arrows in Fig. 3b and d). This could imply that the coagulum was not large enough to block the entire vessel and was washed away from the irradiation site. A summary of the results is shown in Table 2. When f_r was 1 Hz or lower, only 1 out of 8 irradiated blood vessels were coagulated. In contrast, 5 out of 6 irradiated blood vessels were coagulated when f_r was 5 Hz.

Effect of Irradiated Vessel Length on Long-Term Blood Vessel Removal

A 2 mm segment of blood vessels was irradiated in the above animal experiments, although irradiation of a longer segment would be possible by using a larger laser spot size. However, unlike our DWC model, where blood vessels run perpendicularly to the laser beam axis, malformed post-capillary venules in PWS dermis may run nearly parallel with the laser beam, from papillary loops to the superficial horizontal plexus. As a result, coagulation of a long segment of a PWS blood vessel with a SLP is unlikely due to the limited PDL penetration depth in human skin. The MLP approach is capable of damaging vasculature located deeper in the skin [15,26], and thus, a longer vessel segment might be coagulated as compared to SLP. It is therefore of interest to study the effect of irradiated vessel length on long-term blood vessel removal.

In contrast with the first part of this study, where only isolated 2 mm vessel segments were irradiated (see Figs. 2a and 3a), we irradiated entire vessel branches (Fig. 4a) with MLP with a RE of 4 J/cm², a n_p of 5 and a f_r of 26 Hz. In an earlier study, it was found that 95% of the coagulated blood vessels reperfused within 2 weeks when isolated 2 mm vessel segments were irradiated at such conditions [42]. When the vessel branches were irradiated as shown in Figure 4, there was no evidence of reperfusion or regeneration of the coagulated vessel branches after 2 weeks in the presented example. From a total of 11 irradiated blood vessel

branches in 6 animals treated in the described manner, only 45% of the coagulated blood vessel branches reperfused within 2 weeks (Table 3).

DISCUSSION

Our animal study indicated that the threshold RE per pulse of 2 J/cm² for MLP to induce blood vessel photocoagulation is substantially lower as compared to that of 7 J/cm² for SLP (Fig. 1). Therefore, MLP approach may achieve a better therapeutic outcome of PWS in cases where the maximum permissible RE per pulse, limited by the epidermal damage threshold, is insufficient to induce persistent vascular shutdown. The data also indicate that the total quantity of light energy (e.g., cumulative RE in Fig. 1) that can be safely delivered with the MLP approach is considerably higher than with the SLP approach.

The n_p is a unique parameter for a laser based on the MLP approach. The selection of n_p is a balance between efficacy and safety for MLP. If n_p is too low, the largest PWS blood vessels may not be coagulated; if n_p is too high, heat diffusing from the vessels to the perivascular tissue may cause unwanted thermal injury. In theory, vessel diameter will certainly influence the selection of n_p because heat diffusion dynamics depend on the diameter of the vessel which serves as a heat source. In addition to vessel diameter, RE also affects n_p . Laser pulses with higher RE can heat the blood vessel to a sufficiently high temperature with a lower n_p . As shown in Table 1, more than 80% of the irradiated blood vessels were coagulated when two or more pulses were used in conjunction with a RE per pulse of 3 J/cm² or higher. Severe collateral skin damage such as ulcer was not observed in this study.

From a thermophysical point of view, the main criterion to determine f_r is that a significant portion of heat generated by successive laser pulses accumulates in the targeted PWS blood vessels, and thus, the core intravascular blood vessel temperature increases substantially with each subsequent laser pulse. Another factor influencing the selection of f_r is that the time interval between successive laser pulses must be long enough to allow sufficient cooling of the basal layer of the epidermis, which will depend primarily on the concentration and depth distribution of epidermal melanin. As shown in Table 2, blood vessels could be coagulated when f_r was 5 Hz or higher. The corresponding inter-pulse interval of 200 ms for $f_r = 5$ Hz is much longer than the 80 ms required for typical cryogen spray delivery in clinical practice, which allows for sufficient epidermal cooling and protection.

Our data showed that blood vessels could not be coagulated when f_r was 1 Hz or less. This outcome most likely is due to the excessive vessel temperature decrease that can occur during the 1-second interval between pulses. In the present study, a RE of 4 J/cm² and a n_p of 2 were employed. It is possible that a blood vessel can still be coagulated at such a low f_r when n_p is much higher. However, the possibility of collateral skin damage also increases with n_p [26] and thus we did not pursue that line of investigation.

Although the above threshold n_p and f_r might not be optimal for a given PWS treatment, they may represent a reasonable starting point. While most PDL have a f_r of 1–2 Hz, they could be modified to operate at 5 Hz or higher because lower RE per pulse are required to induce photocoagulation. Another choice is the frequency-doubled Nd:YAG laser used in this study which can produce laser pulses with much higher f_r . However, the epidermal melanin absorption at this laser's wavelength of 532 nm is higher than that at the customary 585 and 595 nm wavelengths emitted by clinical PDLs.

Due to the limited availability of blood vessels that can be identified clearly in the DWC model, we were only able to determine threshold RE, n_p , and f_r for a range of blood vessel diameters. It is certainly desirable to develop a more detailed correlation between the

optimal RE, n_p , f_r , and blood vessel diameter. However, the number of animals that would be required for this purpose could easily become prohibitive. A more practical approach to this problem is to validate current numerical models of laser–tissue interaction using the animal study results and predict the responses of various PWS blood vessels to laser irradiation using such validated numerical models.

Our data on the effect of irradiated vessel length demonstrate that the spatial extent of photocoagulation has considerable influence on the long-term removal of coagulated blood vessels. When a short segment of a blood vessel was photocoagulated, reperfusion of the blood vessel was consistently observed; when the length of the coagulated segment was increased, the probability of long-term removal increased substantially. Our data imply that, besides photocoagulation of the malformed capillary loops in PWS skin, damage of the blood vessels to and from the superficial horizontal plexus using a laser generating a deeper penetrating wavelength will increase long-term PWS therapeutic outcome.

CONCLUSIONS

A MLP approach can induce blood vessel photocoagulation at much lower RE per pulse as compared to SLP. The required f_r of 5 Hz and n_p of 2 are moderately low in terms of modern laser technology which may make the MLP approach practical in the clinical management of PWS. Our results also imply that spatial extent of photocoagulation might have considerable influence on the long-term PWS therapeutic outcome.

Acknowledgments

The authors wish to thank Thang Nguyen, Robert Gallegos, and David Lee for their assistance with the animal surgery. This work was supported by the National Institutes of Health (AR59244 and AR47551 to J.S.N., HD065536 to B.C., and P41-RR01192, Laser Microbeam and Medical Program), American Society for Laser Medicine and Surgery (Research Grant and Dr. Horace Furumoto Innovations Award to W.J.), and the Sturge–Weber Foundation (Pilot Research Grant to W.J.). Institutional support from the Beckman Laser Institute Endowment, Arnold and Mabel Beckman Foundation, and the Packard Foundation is also acknowledged. The prototype laser used in the animal experiments is on loan from Fotona, Ljubljana, Slovenia. This work was presented in part at the annual meeting of the American Society for Laser Medicine and Surgery, April 2009, in National Harbor, MD.

REFERENCES

1. Jacobs AH, Walton RG. The incidence of birthmarks in the neonate. *Pediatrics*. 1976; 58:218–222. [PubMed: 951136]
2. Alper JC, Holmes LB. The incidence and significance of birthmarks in a cohort of 4641 newborns. *Pediatr Dermatol*. 1986; 1:58–68. [PubMed: 6679890]
3. Barsky SH, Rosen S, Geer DE, Noe JM. The nature and evolution of port wine stains: A computer-assisted study. *J Invest Dermatol*. 1980; 74(3):154–157. [PubMed: 7359006]
4. Smoller BR, Rosen S. Port-wine stains—A disease of altered neural modulation of blood vessels. *Arch Dermatol*. 1986; 122(2):177–179. [PubMed: 3511859]
5. Selim MM, Kelly KM, Nelson JS, Wendelschafer-Crabb G, Kennedy WR, Zelickson BD. Confocal microscopy study of nerves and blood vessels in untreated and treated port wine stains: Preliminary observations. *Dermatol Surg*. 2004; 30(6):892–897. [PubMed: 15171768]
6. Morelli JG, Tan OT, Garden J, Margolis R, Seki Y, Boll J, Carney JM, Anderson RR, Furumoto H, Parrish JA. Tunable dye laser (577nm) treatment of port wine stains. *Lasers Surg Med*. 1986; 6(1): 94–99. [PubMed: 3959722]
7. Tan OT, Sherwood K, Gilchrist BA. Treatment of children with port-wine stains using the flashlamp-pulsed tunable dye laser. *N Engl J Med*. 1989; 320(7):416–421. [PubMed: 2913507]

8. Nelson JS, Milner TE, Anvari B, Tanenbaum BS, Kimel S, Svaasand LO, Jacques SL. Dynamic epidermal cooling during pulsed laser treatment of port-wine stain: A new methodology with preliminary clinical evaluation. *Arch Dermatol*. 1995; 131(6):695–700. [PubMed: 7778922]
9. Nelson JS, Milner TE, Anvari B, Tanenbaum BS, Svaasand LO, Kimel S. Dynamic epidermal cooling in conjunction with laser-induced photothermolysis of port wine stain blood vessels. *Lasers Surg Med*. 1996; 19(2):224–229. [PubMed: 8887927]
10. van der Horst CMAM, Koster PHL, de Borgie CAJM, Bossuyt PMM, van Gemert MJC. Effect of the timing of treatment of port-wine stains with the flash-lamp-pumped pulsed dye-laser. *N Engl J Med*. 1998; 338(15):1028–1033. [PubMed: 9535667]
11. Yohn JJ, Huff JC, Aeling JL, Walsh P, Morelli JG. Lesion size is a factor for determining the rate of port-wine stain clearing following pulsed dye laser treatment in adults. *Cutis*. 1997; 59(5):267–270. [PubMed: 9169268]
12. Katugampola GA, Lanigan SW. Five years' experience of treating port wine stains with the flashlamp-pumped pulsed dye laser. *Br J Dermatol*. 1997; 137(5):750–754. [PubMed: 9415235]
13. Lanigan SW. Port-wine stains unresponsive to pulsed dye laser: Explanations and solutions. *Br J Dermatol*. 1998; 139(2):173–177. [PubMed: 9767228]
14. Lucassen GW, Svaasand LO, Verkruysse W, van Gemert MJC. Laser energy threshold for thermal vascular injury in a port-wine stain skin model. *Lasers Med Sci*. 1995; 10(4):231–234.
15. Jia W, Choi B, Franco W, Lotfi J, Majaron B, Aguilar G, Nelson JS. Treatment of cutaneous vascular lesions using multiple-intermittent cryogen spurts and two-wavelength laser pulses: Numerical and animal studies. *Lasers Surg Med*. 2007; 39(6):494–503. [PubMed: 17659588]
16. Heger M, Beek JF, Moldovan NI, van der Horst C, van Gemert MJC. Towards optimization of selective photothermolysis: prothrombotic pharmaceutical agents as potential adjuvants in laser treatment of port wine stains—A theoretical study. *Thromb Haemost*. 2005; 93(2):242–256. [PubMed: 15711739]
17. Choi B, Jia W, Channul J, Kelly KM, Lotfi J. The importance of long-term monitoring to evaluate the microvascular response to light-based therapies. *J Invest Dermatol*. 2008; 128(2):485–488. [PubMed: 17657245]
18. Phung TL, Oble DA, Jia W, Benjamin LE, Mihm MC Jr, Nelson JS. Can the wound healing response of human skin be modulated after laser treatment and the effects of exposure extended? Implications on the combined use of the pulsed dye laser and a topical angiogenesis inhibitor for treatment of port wine stain birthmarks. *Lasers Surg Med*. 2008; 40(1):1–5. [PubMed: 18220264]
19. Nelson JS, Jia W, Phung TL, Mihm MC. Observations on enhanced port wine stain blanching induced by combined pulsed dye laser and rapamycin administration. *Lasers Surg Med*. 2011; 43(10):939–942. [PubMed: 22127673]
20. Rohrer TE, Chatrath V, Iyengar V. Does pulse stacking improve the results of treatment with variable-pulse pulsed-dye lasers? *Dermatol Surg*. 2004; 30(2):163–167. [PubMed: 14756644]
21. Mordon S, Brisot D, Fournier N. Using a “non uniform pulse sequence” can improve selective coagulation with a Nd:YAG Laser (1.06 mm) thanks to met-hemoglobin absorption: A clinical study on blue leg veins. *Lasers Surg Med*. 2003; 32(2):160–170. [PubMed: 12561051]
22. Fournier N, Brisot D, Mordon S. Treatment of leg telangiectases with a 532 nm KTP laser in multipulse mode. *Dermatol Surg*. 2002; 28(7):564–571. [PubMed: 12135506]
23. Tanghetti E, Sherr EA, Sierra R, Mirkov M. The effects of pulse dye laser double-pass treatment intervals on depth of vessel coagulation. *Lasers Surg Med*. 2006; 38(1):16–21. [PubMed: 16444693]
24. Aldanondo I, Boixeda P, Fernandez-Lorente M, Marquet A, Calvo M, Jaen P, Martin-Saez E. Selectivity of photothermolysis in the treatment of port wine stains using multiple pulses with a pulsed dye laser. *Actas Dermosifiliogr*. 2008; 99(7):546–554. [PubMed: 18682168]
25. Aguilar G, Diaz SH, Lavernia EJ, Nelson JS. Cryogen spray cooling efficiency: Improvement of port wine stain laser therapy through multiple-intermittent cryogen spurts and laser pulses. *Lasers Surg Med*. 2002; 31(1):27–35. [PubMed: 12124712]
26. Milanic M, Jia WC, Nelson JS, Majaron B. Numerical optimization of sequential cryogen spray cooling and laser irradiation for improved therapy of port wine stain. *Lasers Surg Med*. 2011; 43(2):164–175. [PubMed: 21384397]

27. Dierickx CC, Farinelli WA, Anderson RR. Multiple pulse photocoagulation of port-wine stain blood vessels (PWS) with a 585 nm pulsed dye laser. *Lasers Surg Med.* 1995; 7:56.
28. Algire GH. An adaptation of the transparent chamber technique to the mouse. *J Natl Cancer Inst.* 1943; 4:1–11.
29. Menger MD, Laschke MW, Vollmar B. Viewing the microcirculation through the window: Some twenty years experience with the hamster dorsal skinfold chamber. *Eur Surg Res.* 2002; 34(1–2): 83–91. [PubMed: 11867907]
30. Papenfuss HD, Gross JF, Intaglietta M, Treese FA. Transparent access chamber for the rat dorsal skin fold. *Microvasc Res.* 1979; 18(3):311–318. [PubMed: 537508]
31. Gourgouliaos ZF, Welch AJ, Diller KR, Aggarwal SJ. Laser-irradiation-induced relaxation of blood vessels in vivo. *Lasers Surg Med.* 1990; 10(6):524–532. [PubMed: 2263151]
32. Moy AJ, White SM, Indrawan ES, Lotfi J, Nudelman MJ, Costantini SJ, Agarwal N, Jia W, Kelly KM, Sorg BS, Choi B. Wide-field functional imaging of blood flow and hemoglobin oxygen saturation in the rodent dorsal window chamber. *Microvasc Res.* 2011; 82(3):199–209. [PubMed: 21787792]
33. Choi B, Kang NM, Nelson JS. Laser speckle imaging for monitoring blood flow dynamics in the in vivo rodent dorsal skin fold model. *Microvasc Res.* 2004; 68(2):143–146. [PubMed: 15313124]
34. Smith TK, Choi B, Ramirez-San-Juan JC, Nelson JS, Osann K, Kelly KM. Microvascular blood flow dynamics associated with photodynamic therapy, pulsed dye laser irradiation and combined regimens. *Lasers Surg Med.* 2006; 38(5):532–539. [PubMed: 16615132]
35. Channual J, Choi B, Osann K, Pattanachinda D, Lotfi J, Kelly KM. Vascular effects of photodynamic and pulsed dye laser therapy protocols. *Lasers Surg Med.* 2008; 40(9):644–650. [PubMed: 18951421]
36. Dunn AK, Bolay T, Moskowitz MA, Boas DA. Dynamic imaging of cerebral blood flow using laser speckle. *J Cereb Blood Flow Metab.* 2001; 21(3):195–201. [PubMed: 11295873]
37. Briers JD. Laser Doppler speckle and related techniques for blood perfusion mapping and imaging. *Physiol Meas.* 2001; 22(4):R35–R66. [PubMed: 11761081]
38. Ramirez-San-Juan JC, Ramos-Garcia R, Guizar-Iturbide I, Martinez-Niconoff G, Choi B. Impact of velocity distribution assumption on simplified laser speckle imaging equation. *Optics Express.* 2008; 16(5):3197–3203. [PubMed: 18542407]
39. Bezemer R, Heger M, van den Wijngaard JPH, Mordon SR, van Gemert MJC, Beek JF. Laser-induced (endo) vascular photothermal effects studied by combined brightfield and fluorescence microscopy in hamster dorsal skin fold venules. *Optics Express.* 2007; 15(14):8493–8506. [PubMed: 19547183]
40. Suthamjariya K, Farinelli WA, Koh W, Anderson RR. Mechanisms of microvascular response to laser pulses. *J Invest Dermatol.* 2004; 122(2):518–525. [PubMed: 15009739]
41. Heger M, Salles II, Bezemer R, Cloos MA, Mordon SR, Begu S, Deckmyn H, Beek JF. Laser-induced primary and secondary hemostasis dynamics and mechanisms in relation to selective photothermolysis of port wine stains. *J Dermatol Sci.* 63(3):139–147. [PubMed: 21664109]
42. Jia W, Sun V, Tran N, Choi B, Liu SW, Mihm MC, Phung TL, Nelson JS. Long-term blood vessel removal with combined laser and topical rapamycin antiangiogenic therapy: Implications for effective port wine stain treatment. *Lasers Surg Med.* 2010; 42(2):105–112. [PubMed: 20166161]

Fig. 1.

Effect of radiant exposure (RE) on blood vessel photocoagulation. The red curves are the Boltzmann's dose-response function fitted to the observed photocoagulation percentage as a function of RE. The numbers next to the columns are number of coagulated vessels versus number of irradiated vessels. SLP, single laser pulse; MLP, multiple laser pulses. For MLP, the total number of pulses was 5 and pulse repetition rate (f_r) was 26 Hz.

Fig. 2.
Effect of total number of pulses (numbers in the circles) on blood vessel photocoagulation.
a,b: Color images of the subdermal side of the window before and after laser irradiation.
c,d: LSI flow maps. Blood vessels were completely coagulated when the total number of pulses was 2 or higher. Radiant exposure (RE) was 3 J/cm^2 with a pulse repetition rate (f_r) of 10 Hz. The diameters of venules, which are larger and darker, are shown in (a).

Fig. 3. Effect of pulse repetition rate (f_r) on blood vessel photocoagulation. **a,b:** Color images of the subdermal side of the window before and after laser irradiation. **c,d:** LSI flow maps. Blood vessels were completely coagulated when f_r was 5 Hz. Radiant exposure (RE) was 4 J/cm^2 and total number of pulses (n_p) was 2. The diameters of venules, which are larger and darker, are shown in (a). Arrowheads: blood coagulum.

Fig. 4. Long-term response of blood vessels when entire vessel branches were photocoagulated. **a–d:** Color images of the subdermal side of the window before, after laser irradiation, 7 and 14 days later; arrows: photocoagulated blood vessel branches after irradiation; dashed lines: original location of the photocoagulated vessel branches. The photocoagulated branches disappeared completely at Day 14 (d) and thus LSI flow map was not included.

TABLE 1 Summary of the Effect of Total Number of Pulses (n_p)

n_p	Number of venules		
	Irradiated	Coagulated	Coagulation %
1	8	2	25
2	10	8	80
3	7	7	100
4	8	8	100
5	7	6	86

RE = 3–5 J/cm²; f_F = 10 Hz.

TABLE 2 Summary of the Effect of Pulse Repetition Rate (f_r)

f_r	Number of venules		
	Irradiated	Coagulated	Coagulation %
0.5	4	0	0
1	4	1	25
5	6	5	83

RE = 4 J/cm²; $n_p = 2$.

TABLE 3

Summary of the Effect of Irradiated Vessel Length

Animal #	Number of branches		
	Irradiated	Reperfused	Reperfusion %
1	1	1	100
2	2	0	0
3	2	0	0
4	2	2	100
5	2	1	50
6	2	1	50
Total	11	5	45

RE = 4 J/cm²; n_p = 5; f_r = 26 Hz.