Lawrence Berkeley National Laboratory

Recent Work

Title

THE EMISSION SPECTRUM OF CURIUM

Permalink

<https://escholarship.org/uc/item/85z108z4>

Authors

Conway, J.G.

Moore, M.F.

Crane, W.W.T.

Publication Date

1950-07-22

UNIVERSITY OF CALIFORNIA - BERKELEY

UCRL-815
DECLASSIFIED

TWO-WEEK LOAN COPY

*This is a Library Circulating Copy
which may be borrowed for two weeks.
For a personal retention copy, call
Tech. Info. Division, Ext. 5545*

RADIATION LABORATORY

DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

DECLASSIFIED

DECLASSIFIED

UNIVERSITY OF CALIFORNIA
Radiation Laboratory

Contract No. W-7405-eng-48

The Emission Spectrum of Curium

John G. Conway, Milton F. Moore, and W. W. T. Crane

July 22, 1950

CAUTION

This document contains information affecting the National Defense of the United States. Its transmission or the disclosure of its contents in any manner to an unauthorized person is prohibited and may result in severe criminal penalties under applicable Federal Laws.

Berkeley, California

DECLASSIFIED

DECLASSIFIED

Standard Distribution

Copy Nos.

Argonne National Laboratory	1-10
Atomic Energy Commission, Washington	11-12
Brookhaven National Laboratory	13-16
Carbide and Carbon Chemicals Division (K-25 Plant)	17-18
Carbide and Carbon Chemicals Division (Y-12 Plant)	19
General Electric Company, Richland	20-25
Hanford Operations Office	26
Iowa State College	27
Kellex Corporation	28
Knolls Atomic Power Laboratory	29-32
Los Alamos	33-35
Mound Laboratory	36-38
Naval Radiological Defense Laboratory	39
NEPA Project	40
New York Operations Office	41-42
Oak Ridge National Laboratory	43-48
Patent Branch, Washington	49
Technical Information Division, ORE	50-64
UCLA Medical Research Laboratory (Warren)	65
University of California Radiation Laboratory	66-68
University of Rochester	69-70

Total 70

Information Division
Radiation Laboratory
University of California
Berkeley, California

DECLASSIFIED

DECLASSIFIED

UCRL-815
Page 3

THE EMISSION SPECTRUM OF CURIUM

John G. Conway, Milton F. Moore, and W. W.T. Crane
Department of Chemistry and Radiation Laboratory
University of California, Berkeley, California

July 22, 1950

ABSTRACT

The wave lengths of the most prominent lines in the arc and spark spectra of curium have been measured. The region covered extended from 2516-5000 A. The wave lengths of the lines are given, together with a discussion of methods of spectrographic analysis of curium samples.

THE EMISSION SPECTRUM OF CURIUM

John G. Conway, Milton F. Moore, and W. W. Crane
Department of Chemistry and Radiation Laboratory
University of California, Berkeley, California

INTRODUCTION

In August of 1947 L. B. Werner¹ isolated a sufficient amount of curium that

(1) L. B. Werner, U. S. Atomic Energy Commission Declassified Document AECD-2729

a spectrochemical analysis could be performed. At that time the wave lengths of 54 lines were measured and these assigned to curium.² More recently W. W. Crane

(2) J. G. Conway and M. F. Moore, University of California Radiation Laboratory Report UCRL-78 (April, 1948).

isolated addition quantities of curium and a portion was made available for the further wave length measurements.

The handling of active elements such as curium requires many special protective techniques and equipment. To meet these requirements, a controlled atmosphere chamber and filter system was developed and constructed (see Figs. 1, 2, 3, and 4). Before sparking, the controlled atmosphere chamber is made air tight, so that during sparking, no activity can escape into the Berkeley glove box. To remove the activity created by the discharge during the sparking period, air is continuously flushed through the chamber and filter then exhausted onto the roof. The filter system is a six-foot train of loosely packed glass wool fibers connected to the chamber's air outlet. The Berkeley glove boxes are maintained with a separate filter system under negative pressure at all times.

Fig. 1. General view of Berkeley glove boxes and spectrograph.

FIG. 1

Fig. 2. Controlled atmosphere chamber in open position
for changing electrodes.

FIG. 2

Fig. 3. Controlled atmosphere chamber in sparking position mounted in Berkeley glove box.

- 10 -

FIG. 3

OZ 942

Fig. 4. A. Cu electrodes held by jacob's chucks.

B. Pyrex glass pipe 4" diameter.

C. Motors to adjust electrode separation (27 VDC).

D. Quartz window 5 1/8" diameter, 1/4" thick.

E. Air inlet.

F. Air outlet.

G. "O" ring gaskets.

FIG. 4

The curium was supplied in 0.1N HCl solution. One microgram aliquots were evaporated on 1/4" copper electrodes^{3,4} and sparked for 15 seconds using an

(3) M. Fred, N. H. Nachtrieb, and F. S. Tomkins, J. Optical Soc. Am., 37, 279 (1947).

(4) M. C. Bachelder, J. G. Conway, N. H. Nachtrieb, and B. S. Wildi, U. S. Atomic Energy Commission Declassified Document MDDC-511 (Dec., 1946).

Applied Research Laboratory unit with 0.021 mfd and 0.36 mh at 100 volts.

(See Fig. 5 - sample preparation box.) For arcing one microgram, aliquots were evaporated on 1/8" special spectroscopic graphite and burned at 7 amperes D.C. for 15 seconds. The Cynangen Bands were greatly reduced by the circulation of oxygen through the chamber during the arcing period.

The spectra were photographed in the first order on a Baird modified eagle mount spectrograph using a 25 micron slit. Eastman 103-0 and 103-F plates were used. The lines were measured with a Baush and Lomb spectrum measuring magnifier. Copper lines were used as wave length standards in the spark and iron lines in the arc.

RESULTS

See Table I. The wavelengths are good to 0.2 A. Background and band structure limit the number of lines that can be observed in the arc spectra.

Fig. 5. Sample preparation box.

FIG. 5

Table I

Wave Lengths of Curium

Wave Length	Intensity Spark	Wave Length	Intensity Spark	Wave Length	Intensity Spark
2516.0	2	2968.9	4	3473.5*	7
2562.3	1	2984.5	3	3492.9	2
2585.8	2	2996.2	3	3498.9*	5
2616.8	2	2998.3	3	3503.7	3
2617.2	3	2998.8	2	3510.3	3
2625.4	2	3000.8	2	3518.5	2
2628.2	2	3004.1	2	3546.1*	5
2636.2	3	3037.4	2	3550.3	3
2651.1	4	3044.7	6	3567.4*	5
2653.7	4	3046.6	5	3570.4*	2
2677.0	1	3047.5	2	3573.0*	5
2693.8	1	3050.8	2	3581.3*	3
2704.8	1	3059.0	2	3585.6*	3
2707.1	1	3081.6	4	3591.9*	2
2710.1	1	3092.8	4	3619.9	2
2725.6	2	3096.4	1	3639.9	1
2727.7	3	3101.5	1	3643.5	1
2736.8	3	3104.9	1	3670.4*	2
2748.0	1	3118.3	3	3673.2	3
2773.5	2	3143.9	1	3691.7	3
2792.0	1	3161.8	3	3692.5*	3
2796.2	1	3186.4	1	3696.4	3
2796.4	1	3190.3	1	3697.3	2
2811.5	5	3195.6	1	3708.8	1
2817.4	4	3197.0	3	3709.6	1
2821.5	4	3197.5	2	3723.3	3
2832.2	6	3198.0	3	3732.3	3
2833.7	4	3210.0	4	3738.2*	3
2862.2	2	3220.7	6	3739.3*	4
2862.8	2	3230.3	4	3750.5	3
2872.0	3	3238.6	2	3763.6	4
2882.3	1	3252.6*	2	3767.3	3
2888.5	2	3279.4*	2	3778.2*	4
2888.1	1	3296.7*	6	3788.0	1
2909.1	2	3304.9*	2	3799.5	3
2913.1	4	3367.9*	3	3842.1	3
2920.7	5	3374.3*	3	3846.9*	2
2929.0	2	3382.7	2	3903.9	7
2933.2	3	3392.0	3	3908.2	6
2934.9	3	3417.1*	6	3936.6*	5
2935.4	3	3424.2	1	3941.5	1
2944.5	2	3425.9*	6	3942.1	1
2953.0	3	3436.5	1	3947.9	3
2955.6	2	3439.9	3	3948.8	2
2957.5	4	3446.6	2	3953.3	1
2961.8	2	3458.3	1	3957.5	1
2963.0	2	3461.7	7	3960.8*	2

Wave Length	Intensity Spark	Wave Length	Intensity Spark	Wave Length	Intensity Arc
3962.2	2	4239.8	3	2832.2	9
3962.8	2	4283.0	1	3019.6	9
3965.0	2	4289.2	1	3943.7	9
3970.2	2	4293.0	1		
3974.9*	6	4299.1	1		
3984.6	2	4301.5	5		
3990.9	1	4309.7	1		
4003.3*	8	4313.1	4		
4016.2	1	4318.7	3		
4017.0	1	4323.4	5		
4017.9	2	4324.8	4		
4018.4	2	4357.2	2		
4024.5*	3	4433.7	3		
4028.0*	2	4434.2	2		
4031.6*	1	4435.3	2		
4036.4	2	4442.0	3		
4052.4	1	4446.5	1		
4061.6	1	4455.2	2		
4078.9*	4	4546.9	1		
4089.8	4	4570.1	2		
4106.9	3	4608.3	2		
4118.1	2	4632.1	1		
4128.5	2	4665.4	2		
4130.4	2	4670.4	2		
4141.3*	3	4681.8	2		
4148.2	1	4722.8	4		
4156.5	1	4761.6	1		
4156.0	4	4873.8	3		
4164.3	1	4880.5	3		
4188.4	2	4905.6	4		
4194.3	2	4906.3	1		
4202.4	1	4925.5	3		
4208.0	6	4989.9	1		
4218.6	3				

*Lines also found in 1947 experiment.

This work was performed under the auspices of the U. S. Atomic Energy Commission.