

UC Davis

UC Davis Previously Published Works

Title

Impact of the new ultra-high sensitivity mode in a long axial field-of-view PET/CT.

Permalink

<https://escholarship.org/uc/item/93b3c6jr>

Journal

Annals of Nuclear Medicine, 37(5)

Authors

Weidner, Sabine

Sari, Hasan

Buesser, Dorothee

et al.

Publication Date

2023-05-01

DOI

10.1007/s12149-023-01827-y

Peer reviewed



Impact of the new ultra-high sensitivity mode in a long axial field-of-view PET/CT

Clemens Mingels¹ · Sabine Weidner¹ · Hasan Sari^{1,2} · Dorothee Buesser¹ · Konstantinos Zeimpekis¹ · Kuangyu Shi¹ · Ian Alberts¹ · Axel Rominger¹

Received: 28 December 2022 / Accepted: 14 February 2023 / Published online: 13 March 2023
© The Author(s) 2023

Abstract

Objective Long axial field-of-view (LAFOV) PET/CT showed improved performance resulting from higher sensitivity. The aim was to quantify the impact of using the full acceptance angle (UHS) in image reconstructions with the Biograph Vision Quadra LAFOV PET/CT (Siemens Healthineers) compared to the limited acceptance angle (high sensitivity mode, HS).

Methods 38 oncological patients examined on a LAFOV Biograph Vision Quadra PET/CT were analysed. 15 patients underwent [¹⁸F]FDG-PET/CT, 15 patients underwent [¹⁸F]PSMA-1007 PET/CT, and 8 patients underwent [⁶⁸Ga]Ga-DOTA-TOC PET/CT. Signal-to-noise ratio (SNR) and standardised uptake values ($SUV_{\text{mean/max/peak}}$) were used to compare UHS and HS with different acquisition times.

Results The SNR was significantly higher for UHS compared to HS over all acquisition times (SNR UHS/HS [¹⁸F]FDG: 1.35 ± 0.02 , $p < 0.001$; [¹⁸F]PSMA-1007: 1.25 ± 0.02 , $p < 0.001$; [⁶⁸Ga]Ga-DOTA-TOC: 1.29 ± 0.02 , $p < 0.001$).

Conclusion UHS showed significantly higher SNR opening the possibility of halving short acquisition times. This is of advantage in further reduction of whole-body PET/CT acquisition.

Keywords Whole-body PET/CT · LAFOV PET/CT · Ultra-high sensitivity · High sensitivity · Acquisition time

Introduction

Nuclear medicine techniques especially in hybrid imaging have undergone rapid development since the first introduction of a clinical positron emission tomography/computed tomography (PET/CT) system [1]. The recently introduced LAFOV scanners in Philadelphia, Pennsylvania, USA (PenPET Explorer), in Davis, California, USA (uExplorer, United Imaging Healthcare America) and in Bern, Switzerland (Biograph Vision Quadra, Siemens Healthineers) demonstrated that the longer coverage of coincident photons results in higher count density and higher sensitivity of the LAFOV PET/CT systems. Accordingly, signal-to-noise ratio (SNR) of the scanner increases with longer FOV (2).

Recent work has shown that scan time can be reduced with a LAFOV PET system [3, 4]. Yet, at the time of the first publication of the latest LAFOV Biograph Vision Quadra PET/CT system, image reconstructions using the full acceptance angle (maximum ring difference (MRD) 322) were not available and reconstruction was restricted to a limited acceptance angle (MRD 85) [5, 6]. Detecting coincidences with full acceptance angle is known to be challenging in LAFOV system. Zhang et al. observed in the uEXPLORER total-body PET-scanner worsening axial resolution with increasing MRD since oblique lines-of-response (LOR) are more likely to be scattered in the patient with longer MRD [7]. Therefore, fast time-of-flight (TOF) resolution and accurate scatter correction are crucial to avoid misinterpretation of the coincident events [8, 9].

Here, we present the first clinical data with the new ultra-high sensitivity (UHS) reconstruction mode (MRD 322) in the Biograph Vision Quadra and aim to evaluate it with regard to noise rate, lesion quantification and acquisition time.

✉ Clemens Mingels
clemens.mingels@insel.ch

¹ Department of Nuclear Medicine, Inselspital, Bern University Hospital, University of Bern, Street: Freiburgstr. 18, 3010 Bern, Switzerland

² Advanced Clinical Imaging Technology, Siemens Healthcare AG, Lausanne, Switzerland

Materials and methods

Patient population

38 previously published patients undergoing clinically routine oncological PET/CT between October 2020 and December 2020 were analysed retrospectively [5]. Data from 15 patients receiving a [¹⁸F]FDG-PET/CT, 15 patients receiving [¹⁸F]PSMA-1007 PET/CT and 8 patients receiving [⁶⁸Ga]Ga-DOTA-TOC PET/CT were analysed (Fig. 1) [10]. 10 min list-mode (LM) data acquisition was utilised for all scans.

Imaging protocol

The PET LM data were calculated using 600, 360, 240, 120, 60, 30, 20, 10, 5 and 2 s durations to simulate shorter acquisition times. Long acquisition times were considered as > 120 s, short acquisitions as 120–30 s and very short acquisitions as < 30 s. Biograph Vision Quadra acquires the PET emission data using the maximum full ring difference (MRD 322), and offers two reconstruction modes: ultra-high sensitivity mode (UHS) with the full ring difference and full acceptance angle (54 degrees) and high sensitivity (HS) mode with limited ring difference (MRD 85) and acceptance angle (19 degrees) [11, 12]. In this work, we reconstructed the PET data with UHS and HS modes. Image reconstruction was performed using a dedicated image reconstruction software (e7-tools, Siemens Healthineers), which was available at our clinic at the time of analysis and will be provided by Siemens Healthineers to general users in the future. In all reconstructions, point-spread-function (PSF)-TOF method was used with four iterations and five subsets and were reconstructed with a 440 × 440 × 644 image matrix with a voxel size of 1.65 × 1.65 × 1.65 mm³. A Gaussian post-reconstruction filter with 2 mm full width at half maximum (FWHM) was applied to the images. Attenuation correction was performed using the low-dose non-enhanced CT data. 3D scatter correction was performed using a 3D residual-based method which was shown to be more accurate than 2D single scatter simulation method with large acceptance angles and oblique lines-of-response (i.e. MRD 322) [13]. All images in both MRD 85 and 322 were reconstructed using the 3D scatter correction.

Image evaluation

Image analysis and identification of target lesions (malignant tissue) were performed using a separate software (pmod; PMOD Technologies LLC, Zürich, Switzerland). Lesion uptake and metabolic tumour volumes were calculated by placing a volume-of-interest (VOI) with a 40%-iso-contour approach around the lesion as previously described [5, 14].

Peak and maximum standardised uptake values ($SUV_{\text{peak/max}}$) were used to evaluate target lesion knowing that SUV_{peak} has been shown to be less susceptible for variation at different acquisition times than SUV_{max} [15].

The background was measured by the placement of a 10 cm³ VOI in healthy liver tissue in the right liver lobe as previously described [16]. Using the PMOD software, VOIs were automatically applied to all different images obtained from different frame durations. The SNR was defined as the reciprocal coefficient of variation (COV) for the liver background, where σ was the standard deviation of the background VOI and μ was the SUV_{mean} of the background VOI [5].

Statistical analysis

Statistical analyses were performed using Graphpad Prism Version 9 (San Diego, California). The data are presented either as mean ± standard deviation (SD) or as median ± standard error (SE). Comparison between the different SNR and SUV measurements were characterised using paired Student's *T* test and applying Bonferroni correction. *p* values less than 0.0011 were considered statistically significant (marked with an asterisk “*”) according to Bonferroni correction.

Results

Patient examination

In total, 153 target lesions were identified in 38 patients in the reference standard of 10 min acquisition time. Patients' characteristics including the tumour type, mean administered radiopharmaceutical activity ± standard deviation, and age are outlined in Table 1.

Fig. 1 Study flowchart showing patient recruitment and the patients who were included





Table 1 Patients' characteristics of the included patients ($n=38$). Given are radiopharmaceuticals, tumour types, mean and SD of the activity (MBq) and age (mean)

Radiopharmaceutical	Tumour type	Activity (MBq)	Age (years)
[^{18}F]FDG	Lung: 6, lymphoma: 4, ORL: 3, breast: 2	265.6 ± 65.8	67.9
[^{18}F]PSMA-1007	Biochemical recurrence: 13, primary: 2	243.9 ± 14.0	75.5
[^{68}Ga]Ga-DOTA-TOC	SSTR-expressing neuroendocrine tumours: 8	154.1 ± 12.0	65.3

**Fig. 2** Example images of liver signal (SUV_{max}) in both high sensitivity and ultra-high sensitivity reconstructions (HS: A–D; UHS: E–H) in different acquisition times (30 s, 60 s, 120 s, 600 s)

Signal-to-noise ratio (SNR) in three different radiopharmaceuticals from 2 to 600 s

We report significantly higher SNR in all three examined radiopharmaceuticals for UHS compared to HS reconstructions (SNR UHS/HS ratio: [^{18}F]FDG: 1.35 ± 0.02 , $p < 0.001$; [^{18}F]PSMA-1007: 1.25 ± 0.02 , $p < 0.001$; [^{68}Ga]Ga-DOTA-TOC: 1.29 ± 0.02 , $p < 0.001$) over all acquisition times. Figure 2 shows examples of the reconstructed images in both HS and UHS with various acquisition times. As supported by the visual impression (example images Fig. 2), liver signal increased with increasing acquisition time in all three radiopharmaceuticals (Fig. 3 A–C). Furthermore, the SNR

in HS short acquisition time was comparable with half of its time counterpart in UHS ([^{18}F]FDG: HS 60 s: 4.58 ± 1.12 vs. UHS 30 s: 4.60 ± 1.24 ; HS 120 s: 6.32 ± 1.68 vs. UHS 60 s: 6.27 ± 1.72 and HS 240 s: 8.59 ± 2.345 vs. UHS 120 s: 8.34 ± 2.28). Detailed information of all three radiopharmaceuticals can be found in Fig. 3 D–F.

Target lesion values represented by SUV_{peak} and SUV_{max}


SUV_{peak} in HS was 5.34 ± 1.04 for [^{18}F]FDG and 5.48 ± 1.02 in UHS over all acquisition times. SUV_{peak} was constant over all acquisition times and there was no evidence of statistical significant difference between UHS and HS. Median \pm standard error SUV_{max} was 13.63 ± 1.84 in HS and 12.14 ± 1.62 in UHS. No evidence of statistical significant differences between HS and UHS SUV_{max} could be found. For [^{18}F]PSMA-1007, median \pm standard error SUV_{peak} in HS was 5.39 ± 1.23 and in UHS was 5.25 ± 1.22 over all acquisition times. SUV_{max} was 16.72 ± 3.20 in HS and 15.33 ± 2.87 in UHS. No statistical significant differences between HS and UHS $\text{SUV}_{\text{peak/max}}$ could be seen. Median \pm standard error SUV_{peak} for the patients with [^{68}Ga]Ga-DOTA-TOC in HS was 4.73 ± 1.2 and 4.74 ± 1.19 in UHS over all acquisition times. SUV_{max} was 14.05 ± 3.39 in HS and 12.68 ± 2.96 in UHS. No statistical significant differences between HS and UHS $\text{SUV}_{\text{peak/max}}$ could be seen. Details of all $\text{SUV}_{\text{max/peak}}$ at all acquisition times are outlined in Fig. 4.

Discussion

In this study, we show the first in-human evaluation of the ultra-high sensitivity mode (MRD 322) of the Biograph Vision Quadra PET/CT system. Previously published studies with the Biograph Vision Quadra were limited to a MRD of 85 crystal rings which limited the acceptance angle [5, 11, 17]. Using the full acceptance angle to detect coincidences (MRD 322), the NEMA sensitivity of the Biograph Vision Quadra PET/CT increased in phantom studies by the factor of two compared to MRD 85 (sensitivity MRD 85: 83.4 cps/kBq vs. MRD 322: 176 cps/kBq) [11].

In our clinical analysis, we report significant increase of the SNR as an indicator for the higher sensitivity of the

Fig. 3 Signal-to-noise ratios (SNR) of all three radiopharmaceuticals in high sensitivity (HS) and ultra-high sensitivity reconstruction (UHS) as reciprocal coefficient of the variance (COV). Given are the SNR in all three radiopharmaceuticals for 2 s, 5 s, 10 s, 20 s, 30 s, 60 s, 120 s, 240 s, 360 s and 600 s acquisition time (A–C). All SNR of UHS were significantly higher to HS mode. Shown are mean \pm SD. Comparable SNR could be found between 30 s UHS and 60 s HS, between 60 s UHS and 120 s HS and between 120 s UHS and 240 s HS in all three radiopharmaceuticals as visualised in the violin plots (D–F)



scanner in UHS mode (MRD 322) compared to HS mode (MRD 85). We examined the SNR in the liver of patients receiving either $[^{18}\text{F}]\text{FDG}$, $[^{18}\text{F}]\text{PSMA-1007}$ or $[^{68}\text{Ga}]\text{Ga-DOTA-TOC}$. In all three radiopharmaceuticals, SNR was significantly higher in UHS compared to HS mode. By increasing acquisition time, SNR also increased (Fig. 3 A–C). SNR of 30 s/60 s acquisition in UHS was comparable to 60 s/120 s in HS mode (Fig. 3 D–F). Visual impression (Fig. 2) supported those findings.


Individual standardised uptake values ($\text{SUV}_{\text{peak}/\text{max}}$) for all 153 tumour lesions were similar between the three radiopharmaceuticals over different acquisition times. Comparable $\text{SUV}_{\text{peak}/\text{mean}}$ even in short acquisitions (30–120 s) and high SUV_{max} (up to 20% higher) in short acquisitions are already described due to higher noise rates [18–20].

Accordingly, in our study, SUV_{peak} was stable from 2 s acquisition up to 600 s acquisition in both reconstruction modes and SUV_{max} was high in the very short acquisitions (ca. 100% higher than after 60 s at 2 s of acquisition). Using

UHS in short acquisition times (30–120 s) lead to reduced bias of SUV_{max} . After 30 s of scanning, SUV_{max} was comparable to the longer acquisitions in UHS whereas, in HS, comparable SUV_{max} could be found after 60 s of scanning ($[^{18}\text{F}]\text{FDG}$: UHS 30 s: 12.27 ± 1.55 , HS 120 s: 12.15 ± 1.52), only. The risk of misinterpreting the semi-quantitative SUV_{max} measurements in short acquisition times is a problem, which can be partially addressed using the UHS mode.

We note some limitations. UHS's sensitivity profile is non-uniform, unlike in HS. Sensitivity in MRD 85 is uniform up to the very last centimetres at the scanner whereas MRD 322's point of highest sensitivity is in the middle of the scanner and sensitivity decreases to the edges [11]. SNR might, therefore, differ in UHS when examining lesions at the edge of the scanner (e.g. brain or skin lesions at the lower extremities). Further experiments, especially phantom studies might be able to characterise the absolute Bq/ml in UHS and HS. This is not accounted for in our study where the analysis was limited to the liver, which is placed in the

Fig. 4 Median \pm standard error of standardised uptake values (SUV_{max} and SUV_{peak}) for tumour lesions in ^{18}F FDG (A, D), ^{18}F PSMA-1007 (B, E) and ^{68}Ga Ga-DOTA-TOC (C, F) in high sensitivity (HS) and ultra-high sensitivity (UHS) mode



middle of the LAVOV PET system and will be improved with the availability of continuous bed motion.

Conclusion

Using the UHS mode (MRD 322) on the LAFOV PET/CT Biograph Vision Quadra might be of clinical advantage due to significantly higher SNR relative to the HS mode (MRD 85) in all examined radiopharmaceuticals. This opens the possibility of halving the acquisition time using UHS in short PET acquisitions, leading to an improvement in clinical management of oncological patients.

Acknowledgements This work forms part of the doctoral thesis of Sabine Weidner.

Funding Open access funding provided by University of Bern.

Data availability All data are available and can be found at the corresponding author's address.

Declarations

Conflict of interest HS is a full-time employee of Siemens Healthcare AG, Switzerland. AR has received research support and speaker honoraria from Siemens. All other authors have no conflicts of interest to report.

Ethical approval The cantonal ethics committee approved this patient acquisition (KEK-Nr. 2020–01413). All patients provided written informed consent for inclusion. The study was performed in accordance with the Declaration of Helsinki.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes

were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

References

- Beyer T, Townsend DW, Brun T, Kinahan PE, Charron M, Roddy R, et al. A combined PET/CT scanner for clinical oncology. *J Nuclear Med Off Pub Soc Nucl Med.* 2000;41(8):1369–79.
- Cherry SR, Jones T, Karp JS, Qi J, Moses WW, Badawi RD. Total-body PET: maximizing sensitivity to create new opportunities for clinical research and patient care. *J Nucl Med Off Pub Soc Nucl Med.* 2018;59(1):3–12.
- Nardo L, Abdelhafez YG, Spencer BA, Badawi RD. Clinical implementation of total-body PET/CT at University of California. *Davis PET Clin.* 2021;16(1):1–7.
- Surti S, Pantel AR, Karp JS. Total Body PET: Why, How, What for? *IEEE Trans Radiat Plasma Med Sci.* 2020;4(3):283–92.
- Alberts I, Hünermund J-N, Prenosil G, Mingels C, Bohn KP, Viscione M, et al. Clinical performance of long axial field of view PET/CT: a head-to-head intra-individual comparison of the Biograph Vision Quadra with the Biograph Vision PET/CT. *Eur J Nucl Med Mol Imaging.* 2021;48(8):2395–404.
- Sachpekidis C, Pan L, Kopp-Schneider A, Weru V, Hassel JC, Dimitrakopoulou-Strauss A. Application of the long axial field-of-view PET/CT with low-dose [¹⁸F]FDG in melanoma. *Eur J Nucl Med Mol Imaging.* 2022;50(4):1158–67.
- Zhang X, Zhou J, Cherry SR, Badawi RD, Qi J. Quantitative image reconstruction for total-body PET imaging using the 2-meter long EXPLORER scanner. *Phys Med Biol.* 2017;62(6):2465–85.
- Daube-Witherspoon ME, Cherry SR. Scanner design considerations for long axial field-of-view PET systems. *PET Clin.* 2021;16(1):25–39.
- Ghabrial A, Franklin DR, Zaidi H. A Monte Carlo simulation study of scatter fraction and the impact of patient BMI on scatter in long axial field-of-view PET scanners. *Z Med Phys.* 2021;31(3):305–15.
- Mingels C, Bohn KP, Rominger A, Afshar-Oromieh A, Alberts I. Diagnostic accuracy of [(18)F]PSMA-1007 PET/CT in biochemical recurrence of prostate cancer. *Eur J Nucl Med Mol Imaging.* 2022;49(7):2436–44.
- Prenosil GA, Sari H, Fürstner M, Afshar-Oromieh A, Shi K, Rominger A, et al. Performance characteristics of the biograph Vision Quadra PET/CT system with a long axial field of view using the NEMA NU 2–2018 standard. *J Nucl Med Off Publ Soc Nucl Med.* 2022;63(3):476–84.
- Prenosil GA, Hentschel M, Weitzel T, Sari H, Shi K, Afshar-Oromieh A, et al. EARL compliance measurements on the biograph vision Quadra PET/CT system with a long axial field of view. *EJNMMI Phys.* 2022;9(1):26.
- Bal H, Panin VY, Schaefferkoetter J, Cabello J, Conti M (2021) Fully 3D scatter estimation in axially long FOV PETCT scanners: residual estimation approach. In: 2021 IEEE nuclear science symposium and medical imaging conference (NSS/MIC). Piscataway, NJ, USA, pp 1–4. <https://doi.org/10.1109/NSS/MIC44867.2021.9875665>
- Lee H, Paeng JC, Hong SH, Yoo HJ, Cheon GJ, Lee DS, et al. Appropriate margin thresholds for isocontour metabolic volumetry of fluorine-18 fluorodeoxyglucose PET in sarcoma: a hybrid PET/MRI study. *Nucl Med Commun.* 2016;37(10):1088–94.
- Sher A, Lacoeyille F, Fosse P, Vervueren L, Cahouet-Vannier A, Dabli D, et al. For avid glucose tumors, the SUV peak is the most reliable parameter for [(18)F]FDG-PET/CT quantification, regardless of acquisition time. *EJNMMI Res.* 2016;6(1):21.
- Boellaard R, Delgado-Bolton R, Oyen WJ, Giammarile F, Tatsch K, Eschner W, et al. FDG PET/CT EANM procedure guidelines for tumour imaging: version 20. *Eur J Nucl Med Mol Imaging.* 2015;42(2):328–54.
- Alberts I, Prenosil G, Mingels C, Bohn KP, Viscione M, Sari H, et al. Feasibility of late acquisition [68Ga]Ga-PSMA-11 PET/CT using a long axial field-of-view PET/CT scanner for the diagnosis of recurrent prostate cancer—first clinical experiences. *Eur J Nucl Med Mol Imaging.* 2021;48(13):4456–62.
- Morimoto M, Kudomi N, Maeda Y, Kobata T, Oishi A, Matsumoto K, et al. Effect of quantitative values on shortened acquisition duration in brain tumor 11C-methionine PET/CT. *EJNMMI Phys.* 2021;8(1):34.
- Prenosil GA, Weitzel T, Fürstner M, Hentschel M, Krause T, Cumming P, et al. Towards guidelines to harmonize textural features in PET: Haralick textural features vary with image noise, but exposure-invariant domains enable comparable PET radiomics. *PLoS ONE.* 2020;15(3): e0229560.
- Trägårdh E, Minarik D, Almquist H, Bitzén U, Garpered S, Hvitfelt E, et al. Impact of acquisition time and penalizing factor in a block-sequential regularized expectation maximization reconstruction algorithm on a Si-photomultiplier-based PET-CT system for (18)F-FDG. *EJNMMI Res.* 2019;9(1):64.

Publisher's Note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.