

UC Riverside

UC Riverside Previously Published Works

Title

Procedure for the development of multi-depth circular cross-sectional endothelialized microchannels-on-a-chip.

Permalink

<https://escholarship.org/uc/item/9hg1n4t6>

Authors

Li, Xiang
Mearns, Samantha Marie
Martins-Green, Manuela
et al.

Publication Date

2013

DOI

10.3791/50771

Peer reviewed

Video Article

Procedure for the Development of Multi-depth Circular Cross-sectional Endothelialized Microchannels-on-a-chip

Xiang Li¹, Samantha Marie Mearns¹, Manuela Martins-Green², Yuxin Liu¹¹Lane Department of Computer Science and Electrical Engineering, West Virginia University²Department of Cell Biology and Neuroscience, University of California at RiversideCorrespondence to: Yuxin Liu at Yuxin.Liu@mail.wvu.eduURL: <http://www.jove.com/video/50771>DOI: [doi:10.3791/50771](https://doi.org/10.3791/50771)

Keywords: Bioengineering, Issue 80, Bioengineering, Tissue Engineering, Miniaturization, Microtechnology, Microfluidics, Reflow photoresist, PDMS, Perfusion flow, Primary endothelial cells

Date Published: 10/21/2013

Citation: Li, X., Mearns, S.M., Martins-Green, M., Liu, Y. Procedure for the Development of Multi-depth Circular Cross-sectional Endothelialized Microchannels-on-a-chip. *J. Vis. Exp.* (80), e50771, doi:10.3791/50771 (2013).

Abstract

Efforts have been focused on developing *in vitro* assays for the study of microvessels because *in vivo* animal studies are more time-consuming, expensive, and observation and quantification are very challenging. However, conventional *in vitro* microvessel assays have limitations when representing *in vivo* microvessels with respect to three-dimensional (3D) geometry and providing continuous fluid flow. Using a combination of photolithographic reflowable photoresist technique, soft lithography, and microfluidics, we have developed a multi-depth circular cross-sectional endothelialized microchannels-on-a-chip, which mimics the 3D geometry of *in vivo* microvessels and runs under controlled continuous perfusion flow. A positive reflowable photoresist was used to fabricate a master mold with a semicircular cross-sectional microchannel network. By the alignment and bonding of the two polydimethylsiloxane (PDMS) microchannels replicated from the master mold, a cylindrical microchannel network was created. The diameters of the microchannels can be well controlled. In addition, primary human umbilical vein endothelial cells (HUVECs) seeded inside the chip showed that the cells lined the inner surface of the microchannels under controlled perfusion lasting for a time period between 4 days to 2 weeks.

Video Link

The video component of this article can be found at <http://www.jove.com/video/50771/>

Introduction

Microvessels, as a part of the circulation system, mediate the interactions between blood and tissues, support metabolic activities, define tissue microenvironment, and play a critical role in many health and pathological conditions. Recapitulation of functional microvessels *in vitro* could provide a platform for the study of complex vascular phenomena. However, conventional *in vitro* microvessel assays, such as endothelial cell migration assays, endothelial tube formation assays, and rat and mouse aortic ring assays, are unable to recreate the *in vivo* microvessels with respect to three-dimensional (3D) geometry and continuous flow control¹⁻⁸. Studies of microvessels using animal models and *in vivo* assays, such as corneal angiogenesis assay, chick chorioallantoic membrane angiogenesis assay, and Matrigel plug assay, are more time-consuming, high in cost, challenging with respect to observation and quantifications, and raise ethical issues^{1, 9-13}.

Advances in micromanufacturing and microfluidic chip technologies have enabled a variety of insights into biomedical sciences while curtailing the high experimental costs and complexities associated with animals and *in vivo* studies¹⁴, such as easily and tightly controlled biological conditions and dynamic fluidic environments, which would not have been possible with conventional macroscale techniques.

Here, we present an approach to construct an endothelialized microchannels-on-a-chip which mimics the 3D geometry of *in vivo* microvessels and runs under controlled continuous perfusion flow by using the combination of photolithographic reflowable photoresist technique, soft lithography, and microfluidics.

Protocol

1. Photolithography Fabrication of Photoresist Master Mold

The following protocol shows the process to fabricate the microchannels with diameters between 30-60 μm . To get a microchannel with a smaller diameter (less than 30 μm), a single spin-coating of photoresist is needed.

1. Transfer the reflow photoresist from the refrigerator at 4 °C to the cleanroom 24 hr prior to use and allow it to warm to room temperature.
2. Clean a silicon wafer and bake it for one hr at 150 °C to allow it to dehydrate. The dehydration will assist the photoresist adhesion to the silicon substrate.

- Spin-coat the first layer of photoresist using the following recipe:

Step	Time (Seconds)	Speed (rpm)	Acceleration
1	2	300	highest
2	10	0	
3	3	300	highest
4	60	600	highest
5	10	500	highest
6	10	600	highest

- Then, bake the wafer on a hotplate with a temperature of 110 °C for 90 sec. After the soft bake the thickness of the photoresist will be 20-30 μm .
- Spin coat a second layer of photoresist following the same recipe used for the first layer.
- Soft bake the wafer again by placing it on a hotplate with a temperature of 110 °C for 90 sec. After this soft bake the thickness of the photoresist will be 40-60 μm .
- Generate positive master patterns by exposing the photoresist to UV light with an exposure dose of 14,500 mJ/cm^2 through a film mask.
- Dilute the developer with deionized (DI) water (1:2 (v/v)). Rinse the wafer repeatedly in the solution until the pattern is fully developed. Then wash the wafer with DI water and dry it using nitrogen gas.
- Reflow: Place the wafer on a hotplate with a temperature of 120 °C for 4 min and cover with a glass Petri dish to prevent solvent evaporation. Remove the wafer from the hotplate and allow it to cool to room temperature. The photoresist thickness after reflow will be 50-60 μm .

2. Soft Lithography Fabrication of PDMS Microchannel Network

- Prepare polydimethylsiloxane (PDMS) solution at the weight ratio of 10:1 (base:curing agent) and mix it thoroughly using a planetary centrifugal mixer.
- Cast the PDMS solution onto the reflowed photoresist master mold. Place the casted PDMS in a desiccator for 15 min to degas. Use nitrogen gas to remove any remaining bubbles if necessary.
- Bake the PDMS in an oven at a temperature of 60 °C for 3 hr to allow it to cure. Then remove the cured PDMS layer from the master mold.
- Use a sharpened puncher to create inlet/outlet holes by punching holes in the channel network. Clean the surface of the PDMS using nitrogen gas.
- Treat two PDMS layers with oxygen plasma for 30 sec inside a plasma cleaner at an operating pressure of 4.5×10^{-2} Torr and an oxygen flow rate of 3.5 ft^3/min . Then, align the surfaces of the PDMS manually under an optical microscope. Use a drop of water if necessary for a better control of the alignment.
- Bake the device in an oven at 60 °C for 30 min to achieve permanent bonding.

3. HUVEC Culture and Seeding in the Chip

- Culture the HUVECs with culture medium with L-glutamine supplemented with 10% fetal bovine serum (FBS). For the experiment passages between 2 and 5 were used.
- Once the HUVECs are confluent, count the cells and harvest them by first rinsing the cells with HEPES buffered saline solution (HEPES-BSS), and then treat the cells with trypsin/EDTA and incubate for 2-6 min at 37 °C. After the trypsinization is complete, neutralize the trypsin/EDTA with trypsin neutralizing solution. Centrifuge and suspend the cells in culture medium with 8% dextran to collect them. Dextran was used to increase the medium viscosity to aid in better cell seeding and attachment.
- Treat the device with oxygen plasma for 5 min with an operating pressure of 4.5×10^{-2} Torr and an oxygen flow rate of 3.5 ft^3/min . Then load the device with DI water and treat with UV light for 8 hr in a laminar biosafety hood for sterilization.
- One day before the cells are ready, wash the device with 1x phosphate buffered saline (1x PBS) then coat with fibronectin (100 $\mu\text{g}/\text{ml}$, diluted with 1x PBS) and incubate in the refrigerator at 4 °C overnight.
- After the fibronectin coating, wash the device with 1x PBS then load with culture medium. Incubate the device at a temperature of 37 °C for 15 min.
- Load the HUVEC cells in 8% dextran culture medium with a cell concentration of $3-4 \times 10^6$ cells/ml. Place a 20 μl droplet of cells at one inlet of the device and tilt it to introduce the cells into the microfluidic channel. After 15-20 min the cells will begin to attach to the side walls of the channels. Rotate the device every 15 min to create a more uniform distribution of cells. If necessary, additional loading can be performed.

4. Long-term Perfusion Setup

After 5-6 hr of static culture, the attached HUVECs will start to fully spread out. Set up perfusion using a remote controlled syringe pump system with a steady flow of 10 $\mu\text{l}/\text{hr}$. Perfusion can be adjusted for a higher flow rate, and can last for a time period between 4 days to 2 weeks.

5. Cell Staining and Microscope Characterization

- When the cells reach confluence inside the device, firstly, wash the device with 1x PBS to thoroughly remove the medium. Then load the device with red dye diluted with diluent (4 μl -1 ml). Load the dye similar to the cell loading procedure. Incubate the device in the dark for 5 min at room temperature, and then wash the device with culture medium to stop the staining. Long incubation of the dye can cause cellular toxicity and disadhesion.

2. Load the device with blue dye diluted with 1x PBS (2 droplets per ml). Incubate in the dark for 5 min at room temperature then thoroughly wash the device with 1x PBS.
3. Examine the cell staining under an inverted optical microscope. If the staining was good, load the microchannels with fixing medium (3.5% paraformaldehyde diluted with 1x PBS), then submerge the device in fixing medium and completely cover it with aluminum foil. Store the device in the refrigerator at a temperature of 4 °C to prevent the device from drying out and photobleaching. The fixed device is now ready for confocal imaging, which can be done by a laser scanning confocal microscope.

Representative Results

Our approach to fabricate the multi-depth microchannel network mimics the complex 3D geometries of *in vivo* microvessels, in which the microchannels have rounded cross sections¹⁵. Additionally, the diameters of parent branching channels and the daughter channels approximately obey Murray's law for maintaining the fluid flow at a required level so that the overall channel resistance is low and flow velocities are more uniform throughout the network¹⁶⁻¹⁸. The processes and results for the fabrication of a semicircular photoresist master mold and a circular cross-sectional PDMS microchannel network were demonstrated in **Movie 1**, **Figure 2**, and **Movie 2**, respectively. The geometric features of the PDMS microchannel network were characterized and shown in **Figure 2**. Our results show that the photoresist reflow technique can create multi-depth branching channel networks in a more convenient approach by photoresist reflow techniques, and allow designing the microvascular biomimetic systems which approximately obey Murray's Law.

In many *in vitro* models, vascular cells are normally cultured on planar plates, filters, or these substrates coated with hydrogels. Under these conditions the microvessels are randomly generated by cellular self-assembly. In addition, conventional assays have inherent difficulties in achieving a constant flow over endothelial cells. The lack of a long-term and continuous media flow prevents the ability to maintain the stability of endothelial cell monolayer with appropriate barrier functions. In our model, the benefit of applying microfluidics is the convenience for fluid access and control (varying flow rates, duration and patterns) as well as getting rid of waste. We demonstrate the processes for primary human umbilical vein endothelial cells (HUVECs) seeding in the microchannels and setting a long-term fluid perfusion system for cell culture (**Figure 3**). Furthermore, because of the complex geometries of *in vivo* microvasculature, real-time monitoring of those small vessels is difficult. The developed PDMS based chip offers good optical properties and allows for high-quality and real-time imaging of the endothelialized microchannels. (**Figures 4** and **Movie 3**)

Movie 1. Schematic fabrication procedures for photoresist master molds. Initially, a pre-cleaned silicon substrate was prepared. The positive reflow photoresist layer was spin-coated onto the silicon substrate and was baked and dried. The photoresist was exposed to UV light through a patterned mask, and then, the patterned microchannels were developed. A semicircular cross-sectional microchannel network was created after the reflow at 120 °C for 4 min. [Click here to view movie.](#)




Figure 1. SEM shows the developed photoresist. a) before reflow; b) after reflow; c) molded PDMS shows the cross section of the resist before reflowing; the picture shows a rectangular cross section; d) molded PDMS showing a semicircular cross-section of the resist after reflowing; The cross section after reflow was controlled by the initial dimension design of the pattern and reflow temperature. (Reprinted with permission from Reference 15). [Click here to view larger figure.](#)

Movie 2. The schematic fabrication procedures for cylindrical microchannel network in PDMS. A PDMS solution was cast onto the photoresist mold and cured in the oven at a temperature of 60 °C for 3 hr. Two identical cured PDMS layers, each of which has a semicircular cross-sectional microchannel network, were aligned and bonded together to form microchannels with circular cross sections. [Click here to view movie.](#)




Figure 2. a) An aligned and bonded cylindrical microchannel network in PDMS. b-d) Circular cross-sections of PDMS molds show channel dimensions at each branching level (1-1', 2-2', and 3-3'). In addition, these figures show the creation of a multibranching and multidepth circular cross-sectional microfluidic channel network.




Figure 3. Schematic diagram showing the long-term perfusion through the chip using a remote controlled syringe pump.




Figure 4. Microscopy images using fluorescent cell membrane dye (red) and cell nuclear dye (blue) show that the HUVECs line the inner surface of a cylindrical microchannel network at different branching regions. a) Top, b) Middle, and c) Bottom. d) Confocal microscopy image shows the circular cross-sectional view of HUVECs lining the channel network. Scale bars: 100 μm .

Movie 3. Confocal movie showing the cell lining along the circular channel network. [Click here to view movie.](#)

Discussion

1. Master mold fabrication

One of the designing and guiding principles for vascular morphometry is known as Murray's law¹⁶, which states that the distribution of vessel diameters throughout the network is governed by minimum energy consideration. It also states that the cube of the diameters of a parent vessel at a bifurcation equals the sum of the cubes of the diameters of the daughter vessels ($d_0^3 = d_1^3 + d_2^3$)¹⁹. Additionally, Poiseuille's law has been used to estimate the magnitude of the shear stress in most of the vasculature as $\tau_s = \frac{4\mu Q}{\pi R^3}$, in which τ_s is the hemodynamic shear stress, μ is the blood viscosity, Q is the flow rate, and R is the radius of the vessel^{20,21}. For symmetric bifurcations, an important consequence based on Murray's law and Poiseuille's law is that the wall shear stress remains constant throughout the vascular network. Thus, for a fabricated symmetric microfluidic channel network with circular cross-sections, if the channel dimensions at bifurcations obey Murray's law, the shear stress experienced by the endothelial cells should be constant at different branching channels.

Many microfabrication techniques and processes have been applied to the fabrication of microchannels used for vascular cells²²⁻²⁵, however, the resulting microchannels resulted in rectangular, square or trapezoidal cross-sections of channels. The rectangular cross-sections of channels are constructed with sharp corners and abrupt transitions at bifurcations, which can impose widely varying fluid shear stresses and nonphysiological geometries on cells in different channel positions, thus resulting in variations in cell physiology^{26,27}.

A photoresist reflow process, resulting in a rounded channel profile, involves two procedures, 1) the melting of the patterned photoresist and the liquid resist surfaces are pulled into a shape which minimizes the energy of the system^{28,29} and 2) a cooling and solidification phase follows the melting process. The shapes of the reflowed channels are well approximated by a semicylindrical surface. Followed by reflowed master mold fabrication, a standard soft lithography approach was used to fabricate the PDMS microfluidic channel network with a circular cross-section. Our results showed that the photoresist reflow technique can create multi-depth branching microchannel networks in a more straight forward approach, and allows us to design microvascular biomimetic systems that approximately obey Murray's Law and mimic the geometry of *in vivo*

microvasculature, so that the overall channel resistance is low, and the shear stress variations at different branching levels can be minimized in the fabricated microfluidic channel network.

The physiologic blood microvessels adopt a roughly circular cross-section with radii between 30-300 μm ³⁰. The dimensions for the demonstrated microchannel network in this paper were varied around 60 μm to 100 μm at different branching levels. To fabricate microchannels with different range of diameters for mimicking *in vivo* microvessels, we recommend to use single spun layer reflow resist (limited to 30 μm) or other lower viscosity resists. For a microchannel with a larger diameter (30-60 μm), a double-coating procedure can be applied to get a thicker photoresist film¹⁵. Additional spin-coatings above two layers should be avoided to prevent nonuniform film thickness, which can result in an inaccurate exposure dose. For a film thickness above 60 μm , other higher viscous reflow photoresists are recommended.

In an ideal condition, to fabricate a resist mold with a semicircular cross section, the film thickness can be initially predetermined by giving a required width and focal length of the microchannel as $H = \frac{1}{2E} \left[\frac{R^2}{r} \sin^{-1} \left(\frac{r}{R} \right) - R + h \right]$, where H is the required spun-film thickness, r is half of the channel width, R is the constant radius of curvature, h is the central height of the curvature, and E is the ratio of resist volumes of the microchannel before and after reflowing³¹. For a cylindrical surface of a given width of microchannel, one particular volume of the resist is required. However, several parameters have been reported to affect the reflowed photoresist and make the resulted channel profiles more complex, such as critical angle and boundary movement between the photoresist and the solid substrate, material evaporation during reflow baking, temperature and timing for the reflow process, outgassing, substrate uniformity, and resist properties³²⁻³⁴. For example, the reflow temperature and timing can result in a volume change causing the boundary movement and hence varying the critical angle and final reflowed resist profile³³. As reported by our previous work¹⁵, the reflow process decreased the channel widths by 2% on average because of photoresist volume reductions and boundary movement. In addition, the volume for getting a cylindrical surface needs to make allowance for the effect of material evaporation during the reflow process^{35,36}. If cylindrical surfaces with different radii are required through a microchannel network it is necessary to vary the width of the microchannels, resulting in variations in the volumes at different regions in the network¹⁵. To successfully fabricate a cylindrical channel network with different diameters, the resists widths/volumes, reflow temperatures and timing, critical angles, boundary movement, resist viscosities and spin coating protocols, and substrate properties must be considered and tested.

2. Long-term cell culture

The importance of flow and the associated wall shear stress has been well recognized in regulating endothelial biology. This has been seen in areas such as inducing changes in cell shape and orientation, secretion and organization, proliferation and differentiation, intracellular signaling, cytoskeleton protein production and gene expression, vessel maturation and structure, and barrier functions³⁷⁻⁴⁷. The current *in vitro* methods for forming endothelial tubes generally rely on endothelial cells' (ECs) self-organization with or without mesenchymal cells in extracellular matrix (ECM) (type I collagen, fibrin, or Matrigel)⁴⁸⁻⁵⁴. Although these cultures have successfully modeled several microvascular behaviors, the artificial vessels formed through a random process of morphogenesis lack the desired spatial reproducibility and orientation. Additionally, the impact of flow on microvascular stability remains largely unknown because these self-assembled vessels do not easily combine luminal flow with a 3D tubular organization⁵⁵, posing a challenge to engineering blood vessels to have barrier functions and long-term vascular stability.

Microfluidic systems have proven to be practical and useful for introducing flows to a variety of biochemical and biological analysis^{56,57}. Our approach provides a convenient way to introduce fluid flow over the cultured cells. We used an advanced remote controlled syringe pump to achieve a convenient yet steady and accurate flow control through the microchannels for long-term cell culture (between 4 days and 2 weeks).

3. Cell culture in the chip

Plasma-assisted oxidation of the PDMS microchannels introduces silanol groups (SiOH) on the surfaces which renders the surface hydrophilic and aids in further protein coatings. Different ECM proteins (fibronectin, gelatin, and collagen) were tested for cell attachment, and the best result was found to be fibronectin coating. The HUVECs were prepared at a concentration of 3×10^6 cells/ml in culture media supplemented with 8% dextran (70 kDa) for seeding. Dextran increases the viscosity of the media to enable fine control over seeding density of ECs.

A confluent monolayer was developed between 2-4 days of the HUVECs being exposed to a constant medium flow. To visualize the cells after the cell culture, we labeled cells with membrane staining and nuclei staining dyes, these dyes exhibited lower cytotoxicity⁵⁸. We stained the cells as an adherent monolayer cultured in the chip. A complete 1x PBS wash is necessary to prevent extra dyes trapped inside the chip.

In summary, by the combination of reflow photoresist technique and PDMS replication, the developed multi-depth microchannel network was approximated by a circular surface and the channel diameters at each bifurcation approximately obey Murray's law. The shear stress variations at different branching levels can be minimized in the fabricated microfluidic channel network. In addition, the results from cell culture indicated the healthy condition of the endothelial cells. Thus, the developed endothelialized microchannels-on-a-chip provides a rapid and reproducible approach to create circular cross-sectional multi-branching and multidimensional microchannel networks, which mimics the geometry of *in vivo* microvessels. The procedure illustrates the use of unique capabilities in advanced micromanufacturing and microfluidic technologies to create a microvasculature model with a long-term, continuous perfusion control as well as high-quality and real-time imaging capability. With the increasing utility of microfluidic channels for cell biology, tissue engineering, and bioengineering applications, the endothelialized microchannels-on-a-chip is a potential assay for microvascular research.

Disclosures

The authors declare that they have no competing financial interests.

Acknowledgements

This research was partially supported by National Science Foundation (NSF 1227359), WVU EPSCoR program funded by the National Science Foundation (EPS-1003907), WVU ADVANCE office sponsored by the National Science Foundation (1007978), and WVU PSCoR, respectively.

The microfabrication work was done in WVU Shared Research Facilities (Cleanroom facilities) and Microfluidic Integrative Cellular Research on Chip Laboratory (MICRoChip Lab) at West Virginia University. The confocal imaging was done at WVU Microscope Imaging Facility.

References

1. Adair, T.H. *Angiogenesis: Integrated systems physiology: from molecule to function to disease*. Biota Publishing, 1st edition, (2011).
2. Goodwin, A.M. *In vitro* assays of angiogenesis for assessment of angiogenic and anti-angiogenic agents. *Microvasc. Res.* **74**, 172-183 (2007).
3. Smith, E.J. & Staton, C.A. Tubule formation assays. *Angiogenesis assays: A critical appraisal of current techniques*. John Wiley & Sons, Ltd, West Sussex, UK, 65-87 (2006).
4. Nakatsu, M.N., Davis J. J., & Hughes, C.C.W. Optimized fibrin gel bead assay for the study of angiogenesis. *J. Vis. Exp.* **3**, e186, doi:10.3791/186 (2007).
5. Nicosia, R.F. & Ottinetti, A. Growth of microvessels in serum-free matrix culture of rat aorta. A quantitative assay of angiogenesis *in vitro*. *Lab. Invest.* **63**, 115-122 (1990).
6. Aplin, A.C., Fogel, E., Zorzi, P., & Nicosia, R.F. The aortic ring model of angiogenesis. *Methods Enzymol.* **443**, 119-136 (2008).
7. Nicosia, R.F. The aortic ring model of angiogenesis: A quarter century of search and discovery. *J. Cell. Mol. Med.* **13**, 4113-4136 (2009).
8. Griffith L.G. & Swart, M.A. Capturing complex 3D tissue physiology *in vitro*. *Nat. Rev. Mol. Cell Biol.* **7**, 211-224 (2006).
9. Folkman, J. History of angiogenesis. In: *Angiogenesis: An integrative approach from science to medicine*. Springer, (2008).
10. Auerbach, R., Lewis, R., Shinnars, B., Kubai, L., & Akhtar, N. Angiogenesis assays: A critical overview. *Clin. Chem.* **49**, 32-40 (2003).
11. Auerbach, R., Akhtar, N., Lewis, R.L., & Shinnars, B.L. Angiogenesis assays: Problems and pitfalls. *Cancer Metastasis Rev.* **19**, 167-172 (2000).
12. Staton, C.A., Reed, M.W., & Brown, N.J. A critical analysis of current *in vitro* and *in vivo* angiogenesis assays. *Int. J. Exp. Pathol.* **90**, 195-221 (2009).
13. Staton, C.A., Stribbling, S.M., *et al.* Current methods for assaying angiogenesis *in vitro* and *in vivo*. *Int. J. Exp. Pathol.* **85**, 233-248 (2004).
14. Moraes, C., Mehta, G., Leshner-Perez, S.C., & Takayama, S. Organs-on-a-Chip: A focus on compartmentalized microdevices. *Ann. Biomed. Eng.* **40** (6), 1211-1227 (2012).
15. Huang, Z., Li, X., Martins-Green, M., & Liu, Y. Microfabrication cylindrical microfluidic channel networks for microvascular research. *Biomedical Microdevices.* **14** (5), 873-883 (2012).
16. Murray, C.D. The physiological principle of minimum work applied to the angle of branching of arteries. *J. Gen. Physiol.* **9** (6), 835-841 (1926).
17. Zamir, M. & Medeiros, J.A. Arterial branching in man and monkey. *J. Gen. Physiol.* **79**, 353-360 (1982).
18. Gafiychuk, V.V. & Lubashevsky, I.A. On the principles of the vascular network branching. *J. Theor. Biol.* **212**, 1-9 (2001).
19. Sherman, T.F. On connecting large vessels to small. The meaning of Murray's law. *J. Gen. Physiol.* **78** (4), 431-453 (1981).
20. Kamiya A., Bukhari R., & Togawa T. Adaptive regulation of wall shear stress optimizing vascular tree function. *Bull Math Biol.* **46**(1), 127-137 (1984).
21. LaBarbera M. Principles of design of fluid transport systems in zoology. *Science.* **249**, 992-1000 (1990).
22. Emerson, D.R., Cieslicki, K., Gu, X., & Barber, R.W. Biomimetic design of microfluidic manifolds based on a generalized Murray's law. *Lab Chip.* **6**, 447-454 (2006).
23. Lu, H., Koo, L.Y. *et al.* Microfluidic shear devices for quantitative analysis of cell adhesion. *Anal. Chem.* **76**, 5257-5264 (2004).
24. Shevkopyas, S.S., Gifford, S.C., Yoshida, T., & Bitensky, M.W. Prototype of an *in vitro* model of the microcirculation. *Microvasc. Res.* **65**, 132-136 (2003).
25. Kaihara, S., Borenstein, J., *et al.* Silicon micromachining to tissue engineer branched vascular channels for liver fabrication. *Tissue Eng.* **6**, 105-117 (2000).
26. Fisher, A.B., Chien, S., Barakat, A.I., & Nerem, R.M. Endothelial cellular response to altered shear stress. *Am. J. Physiol. Lung Cell. Mol. Physiol.* **281** (3), L529-533 (2001).
27. Nerem, R.M., Alexander, R.W. *et al.* The study of the influence of flow on vascular endothelial biology. *Am. J. Med. Sci.* **316** (3), 169-175 (1998).
28. Daly, D., Stevens, R.F., Hutley, M.C., & Davies, N. The manufacture of microlenses by melting photoresist. *Meas. Sci. Technol.* **1** (8), 759-766 (1990).
29. Schilling, A., Merz, R., Ossmann, C., & Herzig, H.P. Surface profiles of reflow microlenses under the influence of surface tension and gravity. *Opt. Eng.* **39** (8), 2171-2176 (2000).
30. Young, B. & Heath, J.W. *Wheater's functional histology: A text and colour atlas*. 4th edition. Churchill Livingstone, (2000).
31. O'Neill, F.T. & Sheridan, J.T. Photoresist reflow method of microlens production. Part 1: Background and experiments. *Optik Int. J. Light Electron Opt.* **113**, 391-404 (2002).
32. de Gennes P.G. Wetting: statics and dynamics. *Rev. Mod. Phys.* **57**, 827-863 (1985).
33. Elias H.G. *An Introduction to Polymer Science*. VCH Publishers, New York (1997).
34. Voinov O.V. Dynamic edge angles of wetting upon spreading of a drop over a solid surface. *J. Appl. Mech. Tech. Phys.* **40**, 86-92 (1999).
35. Daly D., Stevens R.F., Hutley M.C., & Davies N. The manufacture of microlenses by melting photoresist. *Meas. Sci. Technol.* **1**, 759 (1990).
36. Jay T.R. & Stern M.B. Preshaping photoresist for refractive microlens fabrication. *Opt. Eng.* **33**, 3552-3555 (1994).
37. Nerem, R.M., Alexander, R.W., *et al.* The Study of the influence of flow on vascular endothelial biology. *Am. J. Med. Sci.* **316** (3), 169-175 (1998).
38. Chien, S., Li, S., & Shyy, Y.J. Effects of mechanical forces on signal transduction and gene expression in endothelial cells. *Hypertension.* **31**, 162-169 (1998).
39. Li, Y.S., Haga, J.H., & Chien, S. Molecular basis of the effects of shear stress on vascular endothelial cells. *J. Biomech.* **38**, 1949-1971 (2005).
40. Lee, E.J., Vunjak-Novakovic, G., Wang, Y., & Niklason, L.E. A biocompatible endothelial cell delivery system for *in vitro* tissue engineering. *Cell Transplant.* **18**, 731-743 (2009).
41. Lee, E.J. & Niklason, L.E. A novel flow bioreactor for *in vitro* microvascularization. *Tissue Eng. Part C Methods.* **16**, 1191-1200 (2010).

42. Chau, L., Doran, M., & Cooper-White, J. A novel multishear microdevice for studying cell mechanics. *Lab Chip*. **9**, 1897-1902 (2009).
43. Meeson, A., Palmer, M., Calfon, M., & Lang, R. A relationship between apoptosis and flow during programmed capillary regression is revealed by vital analysis. *Development*. **122**, 3929-3938 (1996).
44. Van Royen, N.J., Piek, J., Schaper, W., Bode, C., & Buschmann, I. Arteriogenesis: mechanisms and modulation of collateral artery development. *J. Nucl. Cardiol.* **8**, 687-693 (2001).
45. Schaper, W. Therapeutic arteriogenesis has arrived. *Circulation*. **104** (17), 1994-1995 (2001).
46. Tarbell, J.M. Shear stress and the endothelial transport barrier. *Cardiovas. Res.* **87** (2), 320-330 (2010).
47. Potter, C.M., Lundberg, M.H., *et al.* Role of shear stress in endothelial cell morphology and expression of cyclooxygenase isoforms. *Arterioscler. Thromb. Vasc Biol.* **31**, 384-391 (2011).
48. Montesano, R. *In vitro* rapid organization of endothelial cells into capillary-like networks is promoted by collagen matrices. *J. Cell Biol.* **97**, 1648-1652 (1983).
49. Darland, D.C. & D'Amore, P.A. TGF beta is required for the formation of capillary-like structures in three-dimensional cocultures of 10T1/2 and endothelial cells. *Angiogenesis*. **4** (1), 11-20 (2001).
50. Lawley, T.J. & Kubota, Y. Induction to morphologic differentiation of endothelial cells in culture. *J. Invest. Dermatol.* **93**, 59S-61S (1989).
51. Kanzawa, S., Endo, H., & Shioya, N. Improved *in vitro* angiogenesis model by collagen density reduction and the use of type III collagen. *Ann. Plast. Surg.* **30**, 244-251 (1993).
52. Davis, G.E., Bayless, K.J., & Mavila, A. Molecular basis of endothelial cell morphogenesis in three-dimensional extracellular matrices. *Anat. Rec.* **268**, 252-275 (2002).
53. Velazquez, O.C., Snyder, R., Liu, Z., Fairman, R.M., & Herlyn, M. Fibroblast-dependent differentiation of human microvascular endothelial cells into capillary-like 3-dimensional networks. *FASEB J.* **16**, 1316-1318 (2002).
54. Donovan, D., Brown, N.J., & Bishop, E.T. Comparison of three *in vitro* human "angiogenesis" assays with capillaries formed *in vivo*. *Angiogenesis*. **4**, 113-121 (2001).
55. Tang, D.G. & Conti, C.J. Endothelial cell development, vasculogenesis, angiogenesis, and tumor neovascularization: an update. *Semin. Thromb. Hemost.* **30**, 109-117 (2004).
56. Takayama, S., McDonald, J.C., *et al.* Patterning cells and their environments using multiple laminar fluid flows in capillary networks. *Proc. Natl. Acad. Sci. U.S.A.* **96**, 5545-5548 (1999).
57. Cho, B.S., Schuster, T.G., *et al.* Passively driven integrated microfluidic system for separation of motile sperm. *Anal. Chem.* **75**, 1671-1675 (2003).
58. Parsa, H., Upadhyay, R., & Sia, S.K. Uncovering the behaviors of individual cells within a multicellular microvascular community. *Proc. Natl. Acad. Sci. U.S.A.* **108** (12), 5133-5138 (2011).